# NI Modular Instruments Python API Documentation

Release 1.0.1

**National Instruments** 

# Drivers

2 Installation 3 Contributing 4 Support / Feedback 5 Bugs / Feature Requests 6 Documentation 6.1 Additional Documentation 7 License 7.1 nidcpower module 7.1.1 Installation 7.1.2 Usage 7.1.3 API Reference 7.1.3.1 nidcpower.Session 7.1.3.2 Enums 7.1.3.3 Exceptions and Warnings 7.1.3.4 Examples	3 5 7
4 Support / Feedback  5 Bugs / Feature Requests  6 Documentation 6.1 Additional Documentation  7 License 7.1 nidcpower module 7.1.1 Installation 7.1.2 Usage 7.1.3 API Reference 7.1.3.1 nidcpower.Session 7.1.3.2 Enums 7.1.3.3 Exceptions and Warnings	
5 Bugs / Feature Requests           6 Documentation         6.1 Additional Documentation           7 License         7.1 nidcpower module           7.1.1 Installation         7.1.2 Usage           7.1.3 API Reference         7.1.3.1 nidcpower.Session           7.1.3.2 Enums         7.1.3.3 Exceptions and Warnings	7
6 Documentation 6.1 Additional Documentation	
6.1 Additional Documentation	9
7.1 nidcpower module 7.1.1 Installation 7.1.2 Usage 7.1.3 API Reference 7.1.3.1 nidcpower.Session 7.1.3.2 Enums 7.1.3.3 Exceptions and Warnings	<b>11</b>
7.2 nidmm module 7.2.1 Installation 7.2.2 Usage 7.2.3 API Reference 7.2.3.1 nidmm.Session 7.2.3.2 Enums 7.2.3.3 Exceptions and Warnings 7.2.3.4 Examples  7.3 nifgen module 7.3.1 Installation 7.3.2 Usage 7.3.3 API Reference 7.3.3.1 nifgen.Session 7.3.3.2 Enums	13 14 14 119 120 123 123 123 123 123 190 190 191 191

7.	1 niscope module       7.4.1 Installation         7.4.2 Usage       7.4.3 API Reference         7.4.3.1 niscope.Session       7.4.3.2 Enums         7.4.3.3 Exceptions and Warnings	287 290 290 290 293 293 367 373
7.:	5 niswitch module         7.5.1 Installation         7.5.2 Usage         7.5.3 API Reference         7.5.3.1 niswitch.Session         7.5.3.2 Enums         7.5.3.3 Exceptions and Warnings	374 378 378 378 379 379 417 423
7.	7.6.1 Installation 7.6.2 Usage 7.6.3 API Reference 7.6.3.1 nimodinst.Session 7.6.3.2 Exceptions and Warnings 7.6.3.3 Examples	428 428 428 428 428 433 434
7.	7.7.1 Installation 7.7.2 Usage 7.7.3 API Reference 7.7.3.1 nise.Session 7.7.3.2 Enums 7.7.3.3 Exceptions and Warnings	434 435 435 435 435 443 444 445
		447
Pytho	n Module Index	449

8

# CHAPTER 1

Intro

**nimi-python** is a collection of Python modules that provide an interface to the underlying NI driver. Currently, the following drivers are supported:

- NI-DCPower (Python module: nidcpower)
- NI-DMM (Python module: nidmm)
- NI-FGEN (Python module: nifgen)
- NI-SCOPE (Python module: niscope)
- NI-SWITCH (Python module: niswitch)
- NI-ModInst (Python module: nimodinst)
- NI Switch Executive (Python module: nise)

nimi-python supports all the Operating Systems supported by the underlying driver.

**nimi-python** supports Python 2.7, 3.4 and later using CPython or PyPy.

2 Chapter 1. Intro

# CHAPTER 2

Installation

Driver specific installation instructions can be found on Read The Docs:

- nidcpower
- nidmm
- nifgen
- niscope
- niswitch
- nimodinst
- nise

			$\cap$
$\bigcirc$ I	IΛ	דם	 າ ⊀
CH	1/4	ГΙ	7 U

	Contributing

We welcome contributions! You can clone the project repository, build it, and install it by following these instructions.

СН	AP	TFF	<b>4</b>
OI I	$\neg$ ı	1 -1	1 1

# Support / Feedback

The packages included in **nimi-python** package are supported by NI. For support, open a request through the NI support portal at ni.com.

NI Modular Instruments Python API Documentation, Release 1.0.1	

# CHAPTER 5

## Bugs / Feature Requests

To report a bug or submit a feature request specific to NI Modular Instruments Python bindings (nimi-python), please use the GitHub issues page.

Fill in the issue template as completely as possible and we will respond as soon as we can.

For hardware support or any other questions not specific to this GitHub project, please visit [NI Community Forums](https://forums.ni.com/).

NI Modular Instruments Python API Documentation, Release 1.0.1			

				$\sim$
$\cap$ L	<b>IAP</b>	TE	D	h
ОΓ	IAL		П	v

Documentation

Documentation is available here.

## **6.1 Additional Documentation**

Refer to your driver documentation for device-specific information and detailed API documentation.

# CHAPTER 7

License

**nimi-python** is licensed under an MIT-style license (see LICENSE). Other incorporated projects may be licensed under different licenses. All licenses allow for non-commercial and commercial use.

## 7.1 nidcpower module

#### 7.1.1 Installation

As a prerequisite to using the nidcpower module, you must install the NI-DCPower runtime on your system. Visit ni.com/downloads to download the driver runtime for your devices.

The nimi-python modules (i.e. for **NI-DCPower**) can be installed with pip:

```
$ python -m pip install nidcpower~=1.0.1
```

Or easy install from setuptools:

```
$ python -m easy_install nidcpower
```

#### **7.1.2 Usage**

The following is a basic example of using the **nidcpower** module to open a session to a Source Meter Unit and measure voltage and current.

```
import nidcpower
# Configure the session.

with nidcpower.Session(resource_name='PXI1S1ot2', channels='0') as session:
    session.measure_record_length = 20
    session.measure_record_length_is_finite = True
    session.measure_when = nidcpower.MeasureWhen.AUTOMATICALLY_AFTER_SOURCE_COMPLETE
```

(continues on next page)

(continued from previous page)

```
session.voltage_level = 5.0
   session.commit()
   print('Effective measurement rate: {0} S/s'.format(session.measure_record_delta_
\rightarrowtime / 1))
   samples\_acquired = 0
   print(' # Voltage
                         Current
                                    In Compliance')
   row_format = '{0:3d}: {1:8.6f} {2:8.6f} {3}'
   with session.initiate():
       while samples_acquired < 20:</pre>
          measurements = session.fetch_multiple(count=session.fetch_backlog)
          samples_acquired += len(measurements)
          for i in range(len(measurements)):
              print(row_format.format(i, measurements[i].voltage, measurements[i].
```

Additional examples for NI-DCPower are located in src/nidcpower/examples/ directory.

#### 7.1.3 API Reference

#### 7.1.3.1 nidcpower.Session

**class** nidepower.**Session**(*self*, *resource\_name*, *channels=None*, *reset=False*, *options={}*)

Creates and returns a new NI-DCPower session to the power supply or SMU specified in **resource name** to be used in all subsequent NI-DCPower method calls. With this method, you can optionally set the initial state of the following session properties:

- nidcpower.Session.simulate
- nidcpower.Session.driver setup

After calling this method, the session will be in the Uncommitted state. Refer to the Programming States topic for details about specific software states.

To place the device in a known start-up state when creating a new session, set **reset** to True. This action is equivalent to using the <code>nidcpower.Session.reset()</code> method immediately after initializing the session.

To open a session and leave the device in its existing configuration without passing through a transitional output state, set **reset** to False. Then configure the device as in the previous session, changing only the desired settings, and then call the nidcpower.Session.initiate() method.

#### **Related Topics:**

**Programming States** 

#### **Parameters**

- **resource\_name** (str) Specifies the **resourceName** assigned by Measurement & Automation Explorer (MAX), for example "PXI1Slot3" where "PXI1Slot3" is an instrument's **resourceName**. **resourceName** can also be a logical IVI name.
- **channels** (str) Specifies which output channel(s) to include in a new session. Specify multiple channels by using a channel list or a channel range. A channel list is a comma (,) separated sequence of channel names (for example, 0,2 specifies channels 0 and 2). A channel range is a lower bound channel followed by a hyphen (-) or colon (:) followed by an upper bound channel (for example, 0-2 specifies channels 0, 1, and 2). In the Running state, multiple output channel configurations are performed sequentially based on the order

specified in this parameter. If you do not specify any channels, by default all channels on the device are included in the session.

- **reset** (bool) Specifies whether to reset the device during the initialization procedure.
- **options** (*str*) Specifies the initial value of certain properties for the session. The syntax for **options** is a dictionary of properties with an assigned value. For example:

```
{ 'simulate': False }
```

You do not have to specify a value for all the properties. If you do not specify a value for a property, the default value is used.

Advanced Example: { 'simulate': True, 'driver\_setup': { 'Model': '<model number>', 'BoardType': '<type>' } }

Property	Default
range_check	True
query_instrument_status	False
cache	True
simulate	False
record_value_coersions	False
driver_setup	{}

#### **Properties**

Property	Datatype
aperture_time	float
aperture_time_units	ApertureTimeUnits
auto_zero	AutoZero
auxiliary_power_source_available	bool
channel_count	int
compliance_limit_symmetry	ComplianceLimitSymmetry
current_compensation_frequency	float
current_gain_bandwidth	float
current_level	float
current_level_autorange	bool
current_level_range	float
current_limit	float
current_limit_autorange	bool
current_limit_high	float
current_limit_low	float
current_limit_range	float
current_pole_zero_ratio	float
dc_noise_rejection	DCNoiseRejection
digital_edge_measure_trigger_input_terminal	str
digital_edge_pulse_trigger_input_terminal	str
digital_edge_sequence_advance_trigger_input_terminal	str
digital_edge_source_trigger_input_terminal	str
digital_edge_start_trigger_input_terminal	str
driver_setup	str
exported_measure_trigger_output_terminal	str
exported_pulse_trigger_output_terminal	str
exported_sequence_advance_trigger_output_terminal	str

Continued on next page

Table 1 – continued from previous page

Table 1 – continued from pre	
Property	Datatype
exported_source_trigger_output_terminal	str
exported_start_trigger_output_terminal	str
fetch_backlog	int
instrument_firmware_revision	str
instrument_manufacturer	str
instrument_model	str
interlock_input_open	bool
io_resource_descriptor	str
logical_name	str
measure_buffer_size	int
measure_complete_event_delay	float in seconds or datetime.timedelta
measure_complete_event_output_terminal	str
measure_complete_event_pulse_polarity	Polarity
measure_complete_event_pulse_width	float
measure_record_delta_time	float in seconds or datetime.timedelta
measure_record_length	int
measure_record_length_is_finite	bool
measure_trigger_type	TriggerType
measure_when	MeasureWhen
output_capacitance	OutputCapacitance
output_connected	bool
output_enabled	bool
output_function	OutputFunction
output_resistance	float
overranging_enabled	bool
ovp_enabled	bool
ovp_limit	float
power_line_frequency	float
power_source	PowerSource
power_source_in_use	PowerSourceInUse
pulse_bias_current_level	float
pulse_bias_current_limit	float
pulse_bias_current_limit_high	float
pulse_bias_current_limit_low	float
pulse_bias_delay	float
pulse_bias_voltage_level	float
pulse_bias_voltage_limit	float
pulse_bias_voltage_limit_high	float
pulse_bias_voltage_limit_low	float
pulse_complete_event_output_terminal	str
pulse_complete_event_pulse_polarity	Polarity
pulse_complete_event_pulse_width	float
pulse_current_level	float
pulse_current_level_range	float
pulse_current_limit	float
pulse_current_limit_high	float
pulse_current_limit_low	float
pulse_current_limit_range	float
pulse_off_time	float in seconds or datetime.timedelta

Continued on next page

Table 1 – continued from previous page

Property	Datatype
pulse_on_time	float in seconds or datetime.timedelta
pulse_trigger_type	TriggerType
pulse_voltage_level	float
pulse_voltage_level_range	float
pulse_voltage_limit	float
pulse_voltage_limit_high	float
pulse_voltage_limit_low	float
pulse_voltage_limit_range	float
query_instrument_status	bool
ready_for_pulse_trigger_event_output_terminal	str
ready_for_pulse_trigger_event_pulse_polarity	Polarity
ready_for_pulse_trigger_event_pulse_width	float
reset_average_before_measurement	bool
samples_to_average	int
self_calibration_persistence	SelfCalibrationPersistence
sense	Sense
sequence_advance_trigger_type	TriggerType
sequence_engine_done_event_output_terminal	str
sequence_engine_done_event_pulse_polarity	Polarity
sequence_engine_done_event_pulse_width	float
sequence_iteration_complete_event_output_terminal	str
sequence_iteration_complete_event_pulse_polarity	Polarity
sequence_iteration_complete_event_pulse_width	float
sequence_loop_count	int
sequence_loop_count_is_finite	bool
simulate	bool
source_complete_event_output_terminal	str
source_complete_event_pulse_polarity	Polarity
source_complete_event_pulse_width	float
source_delay	float in seconds or datetime.timedelta
source_mode	SourceMode
source_trigger_type	TriggerType
specific_driver_description	str
specific_driver_prefix	str
specific_driver_revision	
specific_driver_vendor	str
<u> </u>	str — Thi acconting
start_trigger_type	TriggerType
supported_instrument_models	str
transient_response	TransientResponse
voltage_compensation_frequency	float
voltage_gain_bandwidth	float
voltage_level	float
voltage_level_autorange	bool
voltage_level_range	float
voltage_limit	float
voltage_limit_autorange	bool
voltage_limit_high	float
voltage_limit_low	float
voltage_limit_range	float

Continued on next page

Table 1 – continued from previous page

Property	Datatype
voltage_pole_zero_ratio	float

#### **Public methods**

Method name
abort()
commit()
configure_aperture_time()
disable()
<pre>fetch_multiple()</pre>
<pre>get_channel_name()</pre>
<pre>get_ext_cal_last_date_and_time()</pre>
<pre>get_ext_cal_last_temp()</pre>
<pre>get_ext_cal_recommended_interval()</pre>
<pre>get_self_cal_last_date_and_time()</pre>
<pre>get_self_cal_last_temp()</pre>
lock()
measure()
measure_multiple()
query_in_compliance()
<pre>query_max_current_limit()</pre>
<pre>query_max_voltage_level()</pre>
<pre>query_min_current_limit()</pre>
query_output_state()
read_current_temperature()
reset()
reset_device()
reset_with_defaults()
self_cal()
self_test()
send_software_edge_trigger()
set_sequence()
unlock()
wait_for_event()

#### **Properties**

#### aperture\_time

#### nidcpower.Session.aperture\_time

Specifies the measurement aperture time for the channel configuration. Aperture time is specified in the units set by the <code>nidcpower.Session.aperture\_time\_units</code> property. for information about supported devices. Refer to the Aperture Time topic in the NI DC Power Supplies and SMUs Help for more information about how to configure your measurements and for information about valid values. Default Value: 0.01666666 seconds

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].aperture_time = var
var = session.channels[0,1].aperture_time
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement: Aperture Time
- C Attribute: NIDCPOWER\_ATTR\_APERTURE\_TIME

#### aperture\_time\_units

```
nidcpower.Session.aperture_time_units
```

Specifies the units of the *nidcpower.Session.aperture\_time* property for the channel configuration. for information about supported devices. Refer to the Aperture Time topic in the NI DC Power Supplies and SMUs Help for more information about how to configure your measurements and for information about valid values. Default Value: *SECONDS* 

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].aperture_time_units = var
var = session.channels[0,1].aperture_time_units
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ApertureTimeUnits
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement: Aperture Time Units
- C Attribute: NIDCPOWER ATTR APERTURE TIME UNITS

#### auto zero

#### nidcpower.Session.auto\_zero

Specifies the auto-zero method to use on the device. Refer to the NI PXI-4132 Measurement Configuration and Timing and Auto Zero topics for more information about how to configure your measurements. Default Value: The default value for the NI PXI-4132 is *ON*. The default value for all other devices is *OFF*, which is the only supported value for these devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].auto_zero = var
var = session.channels[0,1].auto_zero
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.AutoZero
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement: Auto Zero
- C Attribute: NIDCPOWER\_ATTR\_AUTO\_ZERO

#### auxiliary\_power\_source\_available

#### nidcpower.Session.auxiliary\_power\_source\_available

Indicates whether an auxiliary power source is connected to the device. A value of False may indicate that the auxiliary input fuse has blown. Refer to the Detecting Internal/Auxiliary Power topic in the NI DC Power Supplies and SMUs Help for more information about internal and auxiliary power. power source to generate power. Use the <code>nidcpower.Session.power\_source\_in\_use</code> property to retrieve this information.

**Note:** This property does not necessarily indicate if the device is using the auxiliary

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Advanced: Auxiliary Power Source Available
- C Attribute: NIDCPOWER\_ATTR\_AUXILIARY\_POWER\_SOURCE\_AVAILABLE

#### channel count

#### nidcpower.Session.channel\_count

Indicates the number of channels that NI-DCPower supports for the instrument that was chosen when the current session was opened. For channel-based properties, the IVI engine maintains a separate cache value for each channel.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Capabilities:Channel Count
- C Attribute: NIDCPOWER\_ATTR\_CHANNEL\_COUNT

#### compliance limit symmetry

#### nidcpower.Session.compliance\_limit\_symmetry

Specifies whether compliance limits for current generation and voltage generation for the device are applied symmetrically about 0 V and 0 A or asymmetrically with respect to 0 V and 0 A. When set to **Symmetric**, voltage limits and current limits are set using a single property with a positive value. The resulting range is bounded by this positive value and its opposite. When set to **Asymmetric**, you must separately set a limit high and a limit low using distinct properties. For asymmetric limits, the range bounded by the limit high and limit low must include zero. **Default Value:** Symmetric **Related Topics:** Compliance Ranges Changing Ranges Overranging

**Note:** Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].compliance_limit_symmetry = var
var = session.channels[0,1].compliance_limit_symmetry
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ComplianceLimitSymmetry
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Advanced:Compliance Limit Symmetry
- C Attribute: NIDCPOWER ATTR COMPLIANCE LIMIT SYMMETRY

#### current compensation frequency

nidcpower.Session.current\_compensation\_frequency

The frequency at which a pole-zero pair is added to the system when the channel is in Constant Current mode. for information about supported devices. Default Value: Determined by the value of the <code>NORMAL</code> setting of the <code>nidcpower.Session.transient\_response</code> property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_compensation_frequency = var
var = session.channels[0,1].current_compensation_frequency
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Custom Transient Response:Current:Compensation Frequency
- C Attribute: NIDCPOWER\_ATTR\_CURRENT\_COMPENSATION\_FREQUENCY

#### current gain bandwidth

#### nidcpower.Session.current\_gain\_bandwidth

The frequency at which the unloaded loop gain extrapolates to 0 dB in the absence of additional poles and zeroes. This property takes effect when the channel is in Constant Current mode. for information about supported devices. Default Value: Determined by the value of the <code>NORMAL</code> setting of the <code>nidcpower.Session.transient\_response</code> property.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_gain_bandwidth = var
var = session.channels[0,1].current_gain_bandwidth
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Custom Transient Response:Current:Gain Bandwidth
- C Attribute: NIDCPOWER ATTR CURRENT GAIN BANDWIDTH

#### current level

#### nidcpower.Session.current\_level

Specifies the current level, in amps, that the device attempts to generate on the specified channel(s). This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>DC\_CURRENT.nidcpower.Session.output\_enabled</code> property for more information about enabling the output channel. Valid Values: The valid values for this property are defined by the values to which the <code>nidcpower.Session.current\_level\_range</code> property is set.

**Note:** The channel must be enabled for the specified current level to take effect. Refer to the

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_level = var
var = session.channels[0,1].current_level
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Current:Current Level
- C Attribute: NIDCPOWER ATTR CURRENT LEVEL

#### current level autorange

```
nidcpower.Session.current_level_autorange
```

Specifies whether NI-DCPower automatically selects the current level range based on the desired current level for the specified channels. If you set this property to ON, NI-DCPower ignores any changes you make to the <code>nidcpower.Session.current\_level\_range</code> property. If you change the <code>nidcpower.Session.current\_level\_autorange</code> property from ON to OFF, NI-DCPower retains the last value the <code>nidcpower.Session.current\_level\_range</code> property was set to (or the default value if the property was never set) and uses that value as the current level range. Query the <code>nidcpower.Session.current\_level\_range</code> property by using the <code>nidcpower.Session.get\_attribute\_vi\_int32()</code> method for information about which range NI-DCPower automatically selects. The <code>nidcpower.Session.current\_level\_autorange</code> property is applicable only if the <code>nidcpower.Session.output function</code> property is set to DC CURRENT. Default Value: OFF

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_level_autorange = var
var = session.channels[0,1].current_level_autorange
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Current:Current Level Autorange
- C Attribute: NIDCPOWER\_ATTR\_CURRENT\_LEVEL\_AUTORANGE

#### current level range

```
nidcpower.Session.current_level_range
```

Specifies the current level range, in amps, for the specified channel(s). The range defines the valid value to which the current level can be set. Use the <code>nidcpower.Session.current\_level\_autorange</code> property to enable automatic selection of the current level range. The <code>nidcpower.Session.current\_level\_range</code> property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>DC\_CURRENT.nidcpower.Session.output\_enabled</code> property for more information about enabling the output channel. For valid ranges, refer to the Ranges topic for your device in the NI DC Power Supplies and SMUs Help.

**Note:** The channel must be enabled for the specified current level range to take effect. Refer to the

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_level_range = var
var = session.channels[0,1].current_level_range
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Source:DC Current:Current Level Range

#### • C Attribute: NIDCPOWER ATTR CURRENT LEVEL RANGE

#### current limit

```
nidcpower.Session.current limit
```

Specifies the current limit, in amps, that the output cannot exceed when generating the desired voltage level on the specified channel(s). This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>DC\_VOLTAGE</code> and the <code>nidcpower.Session.compliance\_limit\_symmetry</code> property is set to <code>NIDCPOWER\_VAL\_SYMMETRIC. nidcpower.Session.output\_enabled</code> property for more information about enabling the output channel. Valid Values: The valid values for this property are defined by the values to which <code>nidcpower.Session.current\_limit\_range</code> property is set.

Note: The channel must be enabled for the specified current limit to take effect. Refer to the

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_limit = var
var = session.channels[0,1].current_limit
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Voltage:Current Limit
- C Attribute: NIDCPOWER\_ATTR\_CURRENT\_LIMIT

#### current\_limit\_autorange

```
nidcpower.Session.current_limit_autorange
```

Specifies whether NI-DCPower automatically selects the current limit range based on the de-

sired current limit for the specified channel(s). If you set this property to <code>ON</code>, NI-DCPower ignores any changes you make to the <code>nidcpower.Session.current\_limit\_range</code> property. If you change this property from <code>ON</code> to <code>OFF</code>, NI-DCPower retains the last value the <code>nidcpower.Session.current\_limit\_range</code> property was set to (or the default value if the property was never set) and uses that value as the current limit range. Query the <code>nidcpower.Session.current\_limit\_range</code> property by using the <code>nidcpower.Session.get\_attribute\_vi\_int32()</code> method for information about which range NI-DCPower automatically selects. The <code>nidcpower.Session.current\_limit\_autorange</code> property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>DC\_VOLTAGE</code>. Default Value: <code>OFF</code>

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_limit_autorange = var
var = session.channels[0,1].current_limit_autorange
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Voltage:Current Limit Autorange
- C Attribute: NIDCPOWER\_ATTR\_CURRENT\_LIMIT\_AUTORANGE

#### current\_limit\_high

#### $\verb|nidcpower.Session.current_limit_high|\\$

Specifies the maximum current, in amps, that the output can produce when generating the desired voltage on the specified channel(s). This property is applicable only if the Compliance Limit Symmetry <p:py:meth: 'nidcpower.Session.ComplianceLimitSymmetry.html>' property is set to **Asymmetric** and the *Output Method <p:py:meth:'nidcpower.Session.OutputFunction.*html>' set to **DC** Voltage. Current property is You must also specify Limit Low <p:py:meth:'nidcpower.Session.CurrentLimitLow.html>'\_\_\_ to complete the asymmetric range. Valid Values: [1% of Current Limit Range <p:py:meth:'nidcpower.Session.CurrentLimitRange.html>'\_ Current Limit Range <p:py:meth: 'nidcpower.Session.CurrentLimitRange.html>'\_\_] The range bounded by the limit high and limit low must include zero. **Default Value:** Refer to Supported Properties by Device for the default value by device. Related Topics: Ranges Changing Ranges Overranging

Note: The limit may be extended beyond the selected limit range if the Overranging Enabled

<p:py:meth: 'nidcpower.Session.OverrangingEnabled.html>'\_\_\_ property is set to TRUE.

Note: One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_limit_high = var
var = session.channels[0,1].current_limit_high
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Voltage:Current Limit High
- C Attribute: NIDCPOWER\_ATTR\_CURRENT\_LIMIT\_HIGH

#### current limit low

```
nidcpower.Session.current_limit_low
```

Specifies the minimum current, in amps, that the output can produce when generating the desired voltage on the specified channel(s). This property is applicable only if the Compliance Limit Symmetry <p:py:meth:'nidcpower.Session.ComplianceLimitSymmetry.html>' property is set to **Asymmetric** and the *Output Method <p:py:meth:'nidcpower.Session.OutputFunction.*html>'\_\_ property is set to **DC** Voltage. You must also specify Current <p:py:meth:'nidcpower.Session.CurrentLimitHigh.html>'\_ Limit High to complete the asymmetric range. Valid Values: [-Current Range <p:py:meth:'nidcpower.Session.CurrentLimitRange.html>'\_\_\_, -1% of Current Limit Range <p:py:meth: 'nidcpower.Session.CurrentLimitRange.html>'\_\_] The range bounded by the limit high and limit low must include zero. Default Value: Refer to Supported Properties by Device for the default value by device. Related Topics: Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled* <*p:py:meth:'nidcpower.Session.OverrangingEnabled.*html>'\_\_ property is set to TRUE.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_limit_low = var
var = session.channels[0,1].current_limit_low
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Voltage:Current Limit Low
- C Attribute: NIDCPOWER\_ATTR\_CURRENT\_LIMIT\_LOW

#### current limit range

```
nidcpower.Session.current_limit_range
```

Specifies the current limit range, in amps, for the specified channel(s). The range defines the valid value to which the current limit can be set. Use the <code>nidcpower.Session.current\_limit\_autorange</code> property to enable automatic selection of the current limit range. The <code>nidcpower.Session.current\_limit\_range</code> property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>DC\_VOLTAGE.nidcpower.Session.output\_enabled</code> property for more information about enabling the output channel. For valid ranges, refer to the Ranges topic for your device in the NI DC Power Supplies and SMUs Help.

Note: The channel must be enabled for the specified current limit to take effect. Refer to the

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_limit_range = var
var = session.channels[0,1].current_limit_range
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Voltage:Current Limit Range
- C Attribute: NIDCPOWER\_ATTR\_CURRENT\_LIMIT\_RANGE

#### current pole zero ratio

```
nidcpower.Session.current_pole_zero_ratio
```

The ratio of the pole frequency to the zero frequency when the channel is in Constant Current mode. for information about supported devices. Default Value: Determined by the value of the <code>NORMAL</code> setting of the <code>nidcpower.Session.transient\_response</code> property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].current_pole_zero_ratio = var
var = session.channels[0,1].current_pole_zero_ratio
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source: Custom Transient Response: Current: Pole-Zero Ratio
- C Attribute: NIDCPOWER\_ATTR\_CURRENT\_POLE\_ZERO\_RATIO

#### dc noise rejection

nidcpower.Session.dc\_noise\_rejection

Determines the relative weighting of samples in a measurement. Refer to the NI PXIe-4140/4141 DC Noise Rejection, NI PXIe-4142/4143 DC Noise Rejection, or NI PXIe-4144/4145 DC Noise Rejection topic in the NI DC Power Supplies and SMUs Help for more information about noise rejection. for information about supported devices. Default Value: NORMAL

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.DCNoiseRejection
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement:Advanced:DC Noise Rejection
- C Attribute: NIDCPOWER\_ATTR\_DC\_NOISE\_REJECTION

#### digital\_edge\_measure\_trigger\_input\_terminal

#### nidcpower.Session.digital\_edge\_measure\_trigger\_input\_terminal

Specifies the input terminal for the Measure trigger. This property is used only when the <code>nidcpower.Session.measure\_trigger\_type</code> property is set to <code>DIGITAL\_EDGE</code>. for this property. You can specify any valid input terminal for this property. Valid terminals are listed in Measurement & Automation Explorer under the Device Routes tab. Input terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, <code>/Dev1/PXI\_Trig0</code>, or with the shortened terminal name, <code>PXI\_Trig0</code>. The input terminal can also be a terminal from another device. For example, you can set the input terminal on <code>Dev1</code> to be <code>/Dev2/SourceCompleteEvent</code>.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Measure Trigger:Digital Edge:Input Terminal
- C Attribute: NIDCPOWER ATTR DIGITAL EDGE MEASURE TRIGGER INPUT TERMINAL

#### digital\_edge\_pulse\_trigger\_input\_terminal

#### nidcpower.Session.digital\_edge\_pulse\_trigger\_input\_terminal

Specifies the input terminal for the Pulse trigger. This property is used only when the <code>nidcpower.Session.pulse\_trigger\_type</code> property is set to digital edge. You can specify any valid input terminal for this property. Valid terminals are listed in Measurement & Automation Explorer under the Device Routes tab. Input terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0. The input terminal can also be a terminal from another device. For example, you can set the input terminal on Dev1 to be /Dev2/SourceCompleteEvent.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Pulse Trigger:Digital Edge:Input Terminal
- C Attribute: NIDCPOWER\_ATTR\_DIGITAL\_EDGE\_PULSE\_TRIGGER\_INPUT\_TERMINAL

#### digital edge sequence advance trigger input terminal

nidcpower.Session.digital\_edge\_sequence\_advance\_trigger\_input\_terminal Specifies the input terminal for the Sequence Advance trigger. Use this property only when the nidcpower.Session.sequence\_advance\_trigger\_type property is set to DIGITAL\_EDGE. the NI DC Power Supplies and SMUs Help for information about supported devices. You can specify any valid input terminal for this property. Valid terminals are listed in Measurement & Automation Explorer under the Device Routes tab. Input terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0. The input terminal can also be a terminal from another device. For example, you can set the input terminal on Dev1 to be /Dev2/SourceCompleteEvent.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic in

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Sequence Advance Trigger:Digital Edge:Input Terminal
- C Attribute: NIDCPOWER ATTR DIGITAL EDGE SEQUENCE ADVANCE TRIGGER INPUT TERMINA

### digital\_edge\_source\_trigger\_input\_terminal

### nidcpower.Session.digital\_edge\_source\_trigger\_input\_terminal

Specifies the input terminal for the Source trigger. Use this property only when the <code>nidcpower.Session.source\_trigger\_type</code> property is set to <code>DIGITAL\_EDGE</code>. for information about supported devices. You can specify any valid input terminal for this property. Valid terminals are listed in Measurement & Automation Explorer under the Device Routes tab. Input terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, <code>/Dev1/PXI\_Trig0</code>, or with the shortened terminal name, <code>PXI\_Trig0</code>. The input terminal can also be a terminal from another device. For example, you can set the input terminal on <code>Dev1</code> to be <code>/Dev2/SourceCompleteEvent</code>.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Source Trigger:Digital Edge:Input Terminal
- $\bullet \ \ C \ Attribute: \ \textbf{NIDCPOWER\_ATTR\_DIGITAL\_EDGE\_SOURCE\_TRIGGER\_INPUT\_TERMINAL}$

### digital edge start trigger input terminal

nidcpower.Session.digital\_edge\_start\_trigger\_input\_terminal

Specifies the input terminal for the Start trigger. Use this property only when the nidcpower. Session.start\_trigger\_type property is set to DIGITAL\_EDGE. for information about supported devices. You can specify any valid input terminal for this property. Valid terminals are

listed in Measurement & Automation Explorer under the Device Routes tab. Input terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0. The input terminal can also be a terminal from another device. For example, you can set the input terminal on Dev1 to be /Dev2/SourceCompleteEvent.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Start Trigger:Digital Edge:Input Terminal
- C Attribute: NIDCPOWER ATTR DIGITAL EDGE START TRIGGER INPUT TERMINAL

### driver setup

nidcpower.Session.driver\_setup

Indicates the Driver Setup string that you specified when initializing the driver. Some cases exist where you must specify the instrument driver options at initialization time. An example of this case is specifying a particular device model from among a family of devices that the driver supports. This property is useful when simulating a device. You can specify the driver-specific options through the DriverSetup keyword in the optionsString parameter in the nidcpower.Session.\_\_init\_\_() method or through the IVI Configuration Utility. You can specify driver-specific options through the DriverSetup keyword in the optionsString parameter in the nidcpower.Session.\_\_init\_\_() method. If you do not specify a Driver Setup string, this property returns an empty string.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Advanced Session Information: Driver Setup
- C Attribute: NIDCPOWER\_ATTR\_DRIVER\_SETUP

# exported\_measure\_trigger\_output\_terminal

### nidcpower.Session.exported\_measure\_trigger\_output\_terminal

Specifies the output terminal for exporting the Measure trigger. Refer to the Device Routes tab in Measurement & Automation Explorer for a list of the terminals available on your device. for information about supported devices. Output terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI Trig0, or with the shortened terminal name, PXI Trig0.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Measure Trigger:Export Output Terminal
- C Attribute: NIDCPOWER ATTR EXPORTED MEASURE TRIGGER OUTPUT TERMINAL

### exported\_pulse\_trigger\_output\_terminal

# nidcpower.Session.exported\_pulse\_trigger\_output\_terminal

Specifies the output terminal for exporting the Pulse trigger. Refer to the Device Routes tab in Measurement & Automation Explorer for a list of the terminals available on your device. Output terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Triggers:Pulse Trigger:Export Output Terminal

• C Attribute: NIDCPOWER ATTR EXPORTED PULSE TRIGGER OUTPUT TERMINAL

# exported\_sequence\_advance\_trigger\_output\_terminal

### nidcpower.Session.exported\_sequence\_advance\_trigger\_output\_terminal

Specifies the output terminal for exporting the Sequence Advance trigger. Refer to the Device Routes tab in Measurement & Automation Explorer for a list of the terminals available on your device. for information about supported devices. Output terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Sequence Advance Trigger:Export Output Terminal
- C Attribute: NIDCPOWER ATTR EXPORTED SEQUENCE ADVANCE TRIGGER OUTPUT TERMINAL

# exported\_source\_trigger\_output\_terminal

#### nidcpower.Session.exported source trigger output terminal

Specifies the output terminal for exporting the Source trigger. Refer to the Device Routes tab in MAX for a list of the terminals available on your device. for information about supported devices. Output terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Triggers:Source Trigger:Export Output Terminal
- C Attribute: NIDCPOWER\_ATTR\_EXPORTED\_SOURCE\_TRIGGER\_OUTPUT\_TERMINAL

# exported\_start\_trigger\_output\_terminal

# nidcpower.Session.exported\_start\_trigger\_output\_terminal

Specifies the output terminal for exporting the Start trigger. Refer to the Device Routes tab in Measurement & Automation Explorer (MAX) for a list of the terminals available on your device. Output terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0. for information about supported devices.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Start Trigger:Export Output Terminal
- C Attribute: NIDCPOWER\_ATTR\_EXPORTED\_START\_TRIGGER\_OUTPUT\_TERMINAL

# fetch backlog

#### nidcpower.Session.fetch\_backlog

Returns the number of measurements acquired that have not been fetched yet.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Measurement:Fetch Backlog

#### • C Attribute: NIDCPOWER ATTR FETCH BACKLOG

# instrument firmware revision

nidcpower.Session.instrument firmware revision

Contains the firmware revision information for the device you are currently using.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Firmware Revision
- C Attribute: NIDCPOWER\_ATTR\_INSTRUMENT\_FIRMWARE\_REVISION

### instrument manufacturer

nidcpower.Session.instrument manufacturer

Contains the name of the manufacturer for the device you are currently using.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Manufacturer
- C Attribute: NIDCPOWER ATTR INSTRUMENT MANUFACTURER

### instrument model

nidcpower.Session.instrument\_model

Contains the model number or name of the device that you are currently using.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Model
- C Attribute: NIDCPOWER\_ATTR\_INSTRUMENT\_MODEL

### interlock input open

#### nidcpower.Session.interlock\_input\_open

Indicates whether the safety interlock circuit is open. Refer to the Safety Interlock topic in the NI DC Power Supplies and SMUs Help for more information about the safety interlock circuit. about supported devices.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Advanced:Interlock Input Open
- C Attribute: NIDCPOWER\_ATTR\_INTERLOCK\_INPUT\_OPEN

# io\_resource\_descriptor

### nidcpower.Session.io\_resource\_descriptor

Indicates the resource descriptor NI-DCPower uses to identify the physical device. If you initialize NI-DCPower with a logical name, this property contains the resource descriptor that corresponds to the entry in the IVI Configuration utility. If you initialize NI-DCPower with the resource descriptor, this property contains that value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

- LabVIEW Property: Inherent IVI Attributes:Advanced Session Information:Resource Descriptor
- C Attribute: NIDCPOWER ATTR IO RESOURCE DESCRIPTOR

# logical\_name

### nidcpower.Session.logical\_name

Contains the logical name you specified when opening the current IVI session. You can pass a logical name to the nidcpower.Session.\_\_init\_\_() method. The IVI Configuration utility must contain an entry for the logical name. The logical name entry refers to a method section in the IVI Configuration file. The method section specifies a physical device and initial user options.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Advanced Session Information: Logical Name
- C Attribute: NIDCPOWER\_ATTR\_LOGICAL\_NAME

#### measure buffer size

#### nidcpower.Session.measure buffer size

Specifies the number of samples that the active channel measurement buffer can hold. The default value is the maximum number of samples that a device is capable of recording in one second. for information about supported devices. Valid Values: 1000 to 2147483647 Default Value: Varies by device. Refer to Supported Properties by Device topic in the NI DC Power Supplies and SMUs Help for more information about default values.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Measurement:Advanced:Measure Buffer Size
- C Attribute: NIDCPOWER\_ATTR\_MEASURE\_BUFFER\_SIZE

#### measure complete event delay

#### nidcpower.Session.measure\_complete\_event\_delay

Specifies the amount of time to delay the generation of the Measure Complete event, in seconds. for information about supported devices. Valid Values: 0 to 167 seconds Default Value: The NI PXI-4132 and NI PXIe-4140/4141/4142/4143/4144/4145/4154 supports values from 0 seconds to 167 seconds.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Measure Complete Event:Event Delay
- C Attribute: NIDCPOWER\_ATTR\_MEASURE\_COMPLETE\_EVENT\_DELAY

# measure complete event output terminal

### nidcpower.Session.measure\_complete\_event\_output\_terminal

Specifies the output terminal for exporting the Measure Complete event. for information about supported devices. Output terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Measure Complete Event:Output Terminal
- C Attribute: NIDCPOWER ATTR MEASURE COMPLETE EVENT OUTPUT TERMINAL

### measure\_complete\_event\_pulse\_polarity

# nidcpower.Session.measure\_complete\_event\_pulse\_polarity

Specifies the behavior of the Measure Complete event. for information about supported devices. Default Value: HIGH

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.Polarity
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Measure Complete Event:Pulse:Polarity
- C Attribute: NIDCPOWER\_ATTR\_MEASURE\_COMPLETE\_EVENT\_PULSE\_POLARITY

### measure\_complete\_event\_pulse\_width

### nidcpower.Session.measure\_complete\_event\_pulse\_width

Specifies the width of the Measure Complete event, in seconds. The minimum event pulse width value for PXI devices is 150 ns, and the minimum event pulse width value for PXI Express devices is 250 ns. The maximum event pulse width value for all devices is 1.6 microseconds. for information about supported devices. Valid Values: 1.5e-7 to 1.6e-6 Default Value: The default value for PXI devices is 150 ns. The default value for PXI Express devices is 250 ns.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Measure Complete Event:Pulse:Width
- C Attribute: NIDCPOWER\_ATTR\_MEASURE\_COMPLETE\_EVENT\_PULSE\_WIDTH

### measure\_record\_delta\_time

### nidcpower.Session.measure\_record\_delta\_time

Queries the amount of time, in seconds, between between the start of two consecutive measurements in a measure record. Only query this property after the desired measurement settings are committed. for information about supported devices. two measurements and the rest would differ.

**Note:** This property is not available when Auto Zero is configured to Once because the amount of time between the first

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement:Measure Record Delta Time
- C Attribute: NIDCPOWER\_ATTR\_MEASURE\_RECORD\_DELTA\_TIME

### measure record length

# nidcpower.Session.measure\_record\_length

Specifies how many measurements compose a measure record. When this property is set to a value greater than 1, the <code>nidcpower.Session.measure\_when</code> property must be set to <code>AUTOMATICALLY\_AFTER\_SOURCE\_COMPLETE</code> or <code>ON\_MEASURE\_TRIGGER</code>. for information about supported devices. Valid Values: 1 to 16,777,216 Default Value: 1

**Note:** This property is not available in a session involving multiple channels.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement:Measure Record Length
- C Attribute: NIDCPOWER ATTR MEASURE RECORD LENGTH

### measure record length is finite

### nidcpower.Session.measure\_record\_length\_is\_finite

Specifies whether to take continuous measurements. Call the <code>nidcpower.Session.abort()</code> method to stop continuous measurements. When this property is set to False and the <code>nidcpower.Session.source\_mode</code> property is set to <code>SINGLE\_POINT</code>, the <code>nidcpower.Session.measure\_when</code> property must be set to <code>AUTOMATICALLY\_AFTER\_SOURCE\_COMPLETE</code> or <code>ON\_MEASURE\_TRIGGER</code>. When this property is set to False and the <code>nidcpower.Session.source\_mode</code> property is set to <code>SEQUENCE</code>, the <code>nidcpower.Session.measure\_when</code> property must be set to <code>ON\_MEASURE\_TRIGGER</code>. for information about supported devices. Default Value: True

**Note:** This property is not available in a session involving multiple channels.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement:Measure Record Length Is Finite
- C Attribute: NIDCPOWER\_ATTR\_MEASURE\_RECORD\_LENGTH\_IS\_FINITE

#### measure trigger type

nidcpower.Session.measure\_trigger\_type

Specifies the behavior of the Measure trigger. for information about supported devices. Default Value: <code>DIGITAL\_EDGE</code>

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerType
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Measure Trigger:Trigger Type
- C Attribute: NIDCPOWER\_ATTR\_MEASURE\_TRIGGER\_TYPE

### measure when

nidcpower.Session.measure\_when

Specifies when the measure unit should acquire measurements. Unless this property is configured to <code>ON\_MEASURE\_TRIGGER</code>, the <code>nidcpower.Session.measure\_trigger\_type</code> property is ignored. Refer to the Acquiring Measurements topic in the NI DC Power Supplies and SMUs Help for more information about how to configure your measurements. Default Value: If the <code>nidcpower.Session.source\_mode</code> property is set to <code>SINGLE\_POINT</code>, the default value is <code>ON\_DEMAND</code>. This value supports only the <code>nidcpower.Session.measure\_multiple()</code> method. If the <code>nidcpower.Session.source\_mode</code> property is set to <code>SEQUENCE</code>, the default value is <code>AUTOMATICALLY\_AFTER\_SOURCE\_COMPLETE</code>. This value supports only the <code>nidcpower.Session.fetch\_multiple()</code> method.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.MeasureWhen
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement:Advanced:Measure When
- C Attribute: NIDCPOWER ATTR MEASURE WHEN

#### output capacitance

```
nidcpower.Session.output_capacitance
```

Specifies whether to use a low or high capacitance on the output for the specified channel(s). for information about supported devices. Refer to the NI PXI-4130 Output Capacitance Selection topic in the NI DC Power Supplies and SMUs Help for more information about capacitance.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].output_capacitance = var
var = session.channels[0,1].output_capacitance
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.OutputCapacitance
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Advanced:Output Capacitance
- C Attribute: NIDCPOWER\_ATTR\_OUTPUT\_CAPACITANCE

# output\_connected

#### nidcpower.Session.output\_connected

Specifies whether the output relay is connected (closed) or disconnected (open). The <code>nidcpower.Session.output\_enabled</code> property does not change based on this property; they are independent of each other. about supported devices. Set this property to False to disconnect the output terminal from the output. to the output terminal might discharge unless the relay is disconnected. Excessive connecting and disconnecting of the output can cause premature wear on the relay. Default Value: True

**Note:** Only disconnect the output when disconnecting is necessary for your application. For example, a battery connected

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can

specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].output_connected = var
var = session.channels[0,1].output_connected
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Output Connected
- C Attribute: NIDCPOWER\_ATTR\_OUTPUT\_CONNECTED

### output enabled

```
nidcpower.Session.output enabled
```

Specifies whether the output is enabled (True) or disabled (False). Depending on the value you specify for the <code>nidcpower.Session.output\_function</code> property, you also must set the voltage level or current level in addition to enabling the output the <code>nidcpower.Session.initiate()</code> method. Refer to the Programming States topic in the NI DC Power Supplies and SMUs Help for more information about NI-DCPower programming states. Default Value: The default value is True if you use the <code>nidcpower.Session.\_\_init\_\_()</code> method to open the session. Otherwise the default value is False, including when you use a calibration session or the deprecated programming model.

**Note:** If the session is in the Committed or Uncommitted states, enabling the output does not take effect until you call

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].output_enabled = var
var = session.channels[0,1].output_enabled
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

• LabVIEW Property: Source:Output Enabled

• C Attribute: NIDCPOWER\_ATTR\_OUTPUT\_ENABLED

#### output function

nidcpower.Session.output\_function

Configures the method to generate on the specified channel(s). When DC VOLTAGE is selected, the device generates the desired voltage level on the output as long as the output current is below the current limit. You can use the following properties to configure the channel when DC VOLTAGE is selected: nidcpower.Session. voltage\_level nidcpower.Session.current\_limit nidcpower.Session. current\_limit\_high nidcpower.Session.current\_limit\_low nidcpower. Session.voltage\_level\_range nidcpower.Session.current\_limit\_range When DC\_CURRENT is selected, the device generates the desired current level on the output as long as the output voltage is below the voltage limit. You can use the following properties to configure the channel when DC\_CURRENT is selected: nidcpower.Session. current\_level nidcpower.Session.voltage\_limit nidcpower.Session. voltage\_limit\_high nidcpower.Session.voltage\_limit\_low nidcpower. Session.current\_level\_range nidcpower.Session.voltage\_limit\_range Default Value: DC\_VOLTAGE

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].output_function = var
var = session.channels[0,1].output_function
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.OutputFunction
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Source:Output Function

#### • C Attribute: NIDCPOWER ATTR OUTPUT FUNCTION

# output\_resistance

```
nidcpower.Session.output resistance
```

Specifies the output resistance that the device attempts to generate for the specified channel(s). This property is available only when you set the <code>nidcpower.Session.output\_function</code> property on a support device. Refer to a supported device's topic about output resistance for more information about selecting an output resistance. about supported devices. Default Value: 0.0

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic for information

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].output_resistance = var
var = session.channels[0,1].output_resistance
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Output Resistance
- C Attribute: NIDCPOWER\_ATTR\_OUTPUT\_RESISTANCE

### overranging\_enabled

### nidcpower.Session.overranging\_enabled

Specifies whether NI-DCPower allows setting the voltage level, current level, voltage limit and current limit outside the device specification limits. True means that overranging is enabled. Refer to the Ranges topic in the NI DC Power Supplies and SMUs Help for more information about overranging. Default Value: False

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Source:Advanced:Overranging Enabled
- $\bullet \ \ C \ Attribute: \ \textbf{NIDCPOWER\_ATTR\_OVERRANGING\_ENABLED}$

### ovp enabled

nidcpower.Session.ovp\_enabled

Enables (True) or disables (False) overvoltage protection (OVP). Refer to the Output Overvoltage Protection topic in the NI DC Power Supplies and SMUs Help for more information about overvoltage protection. for information about supported devices. Default Value: False

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Advanced:OVP Enabled
- C Attribute: NIDCPOWER\_ATTR\_OVP\_ENABLED

### ovp\_limit

nidcpower.Session.ovp\_limit

Determines the voltage limit, in volts, beyond which overvoltage protection (OVP) engages. for information about supported devices. Valid Values: 2 V to 210 V Default Value: 210 V

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

• LabVIEW Property: Source:Advanced:OVP Limit

• C Attribute: NIDCPOWER\_ATTR\_OVP\_LIMIT

### power line frequency

#### nidcpower.Session.power\_line\_frequency

Specifies the power line frequency for specified channel(s). NI-DCPower uses this value to select a timebase for setting the <code>nidcpower.Session.aperture\_time</code> property in power line cycles (PLCs). in the NI DC Power Supplies and SMUs Help for information about supported devices. Default Value: <code>NIDCPOWER\_VAL\_60\_HERTZ</code>

**Note:** This property is not supported by all devices. Refer to the Supported Properties by Device topic

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].power_line_frequency = var
var = session.channels[0,1].power_line_frequency
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Measurement:Power Line Frequency

### C Attribute: NIDCPOWER\_ATTR\_POWER\_LINE\_FREQUENCY

### power source

nidcpower.Session.power\_source

Specifies the power source to use. NI-DCPower switches the power source used by the device to the specified value. Default Value: AUTOMATIC is set to AUTOMATIC. However, if the session is in the Committed or Uncommitted state when you set this property, the power source selection only occurs after you call the nidcpower.Session.initiate() method.

Note: Automatic selection is not persistent and occurs only at the time this property

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.PowerSource
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Advanced:Power Source
- C Attribute: NIDCPOWER\_ATTR\_POWER\_SOURCE

### power source in use

nidcpower.Session.power\_source\_in\_use

Indicates whether the device is using the internal or auxiliary power source to generate power.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.PowerSourceInUse
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Advanced:Power Source In Use
- C Attribute: NIDCPOWER\_ATTR\_POWER\_SOURCE\_IN\_USE

# pulse\_bias\_current\_level

### nidcpower.Session.pulse\_bias\_current\_level

Specifies the pulse bias current level, in amps, that the device attempts to generate on the specified channel(s) during the off phase of a pulse. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_CURRENT</code>. Valid Values: The valid values for this property are defined by the values you specify for the <code>nidcpower.Session.pulse current level range</code> property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_bias_current_level = var
var = session.channels[0,1].pulse_bias_current_level
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Current:Pulse Bias Current Level
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_BIAS\_CURRENT\_LEVEL

#### pulse bias current limit

# $\verb|nidcpower.Session.pulse_bias_current_limit|\\$

Specifies the pulse bias current limit, in amps, that the output cannot exceed when generating the desired pulse bias voltage on the specified channel(s) during the off phase of a pulse. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_VOLTAGE</code>. Valid Values: The valid values for this property are defined by the values you specify for the <code>nidcpower.Session.pulse\_current\_limit\_range</code> property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_bias_current_limit = var
var = session.channels[0,1].pulse_bias_current_limit
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Voltage:Pulse Bias Current Limit
- C Attribute: NIDCPOWER ATTR PULSE BIAS CURRENT LIMIT

# pulse\_bias\_current\_limit\_high

# nidcpower.Session.pulse\_bias\_current\_limit\_high

Specifies the maximum current, in amps, that the output can produce when generating the desired pulse voltage on the specified channel(s) during the off phase This property is applicable only if the Compliance Limit Symmetry of a pulse. <p:py:meth:'nidcpower.Session.ComplianceLimitSymmetry.html>'\_\_\_ property is set to Asym**metric** and the Output Method <p:py:meth:'nidcpower.Session.OutputFunction.html>'\_\_ property is set to Pulse Voltage. You must also specify a Pulse Bias Current Limit Low <p:py:meth:'nidcpower.Session.PulseBiasCurrentLimitLow.html>'\_\_ to complete the asymmetric range. Valid Values: [1% of Pulse Current Limit Range <p:py:meth:'nidcpower.Session.PulseCurrentLimitRange.html>'\_\_\_, Pulse Current Limit Range <p:py:meth:'nidcpower.Session.PulseCurrentLimitRange.html>'\_\_] The range bounded by the limit high and limit low must include zero. **Default Value:** Refer to Supported Properties by Device for the default value by device. Related Topics: Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled* <p:py:meth:'nidcpower.Session.OverrangingEnabled.html>'\_\_ property is set to TRUE or if the *Output Method* <p:py:meth:'nidcpower.Session.OutputFunction.html>'\_\_ property is set to a pulsing method.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

Tip: This property can use repeated capabilities (channels). If set or get directly on the nid-

cpower.Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower.Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_bias_current_limit_high = var
var = session.channels[0,1].pulse_bias_current_limit_high
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Voltage:Pulse Bias Current Limit High
- C Attribute: NIDCPOWER ATTR PULSE BIAS CURRENT LIMIT HIGH

### pulse bias current limit low

nidcpower.Session.pulse\_bias\_current\_limit\_low

Specifies the minimum current, in amps, that the output can produce when generating the desired pulse voltage on the specified channel(s) during the off phase This property is applicable only if the Compliance Limit Symmetry of a pulse. <p:py:meth:'nidcpower.Session.ComplianceLimitSymmetry.html>'\_\_ property is set to Asym**metric** and the *Output Method <p:py:meth:'nidcpower.Session.OutputFunction.*html>'\_\_ property is set to Pulse Voltage. You must also specify a Pulse Bias Cur-<p:py:meth: 'nidcpower.Session.PulseBiasCurrentLimitHigh.html>'\_\_\_ rent Limit High Valid Values: to complete the asymmetric range. [-Pulse Current Limit Range <p:py:meth:'nidcpower.Session.PulseCurrentLimitRange.html>'\_\_, -1% of Pulse Current Limit Range <p:py:meth:'nidcpower.Session.PulseCurrentLimitRange.html>'\_\_] The range bounded by the limit high and limit low must include zero. **Default Value:** Refer to Supported Properties by Device for the default value by device. **Related Topics:** Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled* <p:py:meth:'nidcpower.Session.OverrangingEnabled.html>'\_\_ property is set to TRUE or if the *Output Method* <p:py:meth:'nidcpower.Session.OutputFunction.html>'\_\_ property is set to a pulsing method.

Note: One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can

specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_bias_current_limit_low = var
var = session.channels[0,1].pulse_bias_current_limit_low
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Voltage:Pulse Bias Current Limit Low
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_BIAS\_CURRENT\_LIMIT\_LOW

### pulse bias delay

```
nidcpower.Session.pulse_bias_delay
```

Determines when, in seconds, the device generates the Pulse Complete event after generating the off level of a pulse. Valid Values: 0 to 167 seconds Default Value: 16.67 milliseconds

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_bias_delay = var
var = session.channels[0,1].pulse_bias_delay
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Advanced:Pulse Bias Delay
- C Attribute: NIDCPOWER ATTR PULSE BIAS DELAY

# pulse\_bias\_voltage\_level

#### nidcpower.Session.pulse bias voltage level

Specifies the pulse bias voltage level, in volts, that the device attempts to generate on the specified channel(s) during the off phase of a pulse. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_VOLTAGE</code>. Valid Values: The valid values for this property are defined by the values you specify for the <code>nidcpower.Session.pulse\_voltage\_level\_range</code> property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_bias_voltage_level = var
var = session.channels[0,1].pulse_bias_voltage_level
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Voltage:Pulse Bias Voltage Level
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_BIAS\_VOLTAGE\_LEVEL

# pulse\_bias\_voltage\_limit

# $\verb|nidcpower.Session.pulse_bias_voltage_limit|\\$

Specifies the pulse voltage limit, in volts, that the output cannot exceed when generating the desired current on the specified channel(s) during the off phase of a pulse. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_CURRENT</code>. Valid Values: The valid values for this property are defined by the values you specify for the <code>nidcpower.Session.pulse\_voltage\_limit\_range</code> property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_bias_voltage_limit = var
var = session.channels[0,1].pulse_bias_voltage_limit
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Current:Pulse Bias Voltage Limit
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_BIAS\_VOLTAGE\_LIMIT

#### pulse bias voltage limit high

```
nidcpower.Session.pulse_bias_voltage_limit_high
```

Specifies the maximum voltage, in volts, that the output can produce when generating the desired pulse current on the specified channel(s) during the off phase This property is applicable only if the Compliance Limit Symmetry <p:py:meth:'nidcpower.Session.ComplianceLimitSymmetry.html>'\_\_\_ property is set to Asym**metric** and the *Output Method* <*p:py:meth:'nidcpower.Session.OutputFunction.*html>' property is set to **Pulse Current**. You must also specify a Pulse Bias Voltage <p:py:meth:'nidcpower.Session.PulseBiasVoltageLimitLow.html>'\_\_\_ Limit Low [1% of Pulse Voltage Limit Range plete the asymmetric range. Valid Values: <p:py:meth:'nidcpower.Session.PulseVoltageLimitRange.html>'\_\_, Pulse Voltage Limit Range <p:py:meth:'nidcpower.Session.PulseVoltageLimitRange.html>'\_\_] The range bounded by the limit high and limit low must include zero. Default Value: Refer to Supported Properties by Device for the default value by device. **Related Topics:** Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled* <p:py:meth:'nidcpower.Session.OverrangingEnabled.html>'\_\_ property is set to TRUE or if the *Output Method* <p:py:meth:'nidcpower.Session.OutputFunction.html>'\_\_ property is set to a pulsing method.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_bias_voltage_limit_high = var
var = session.channels[0,1].pulse_bias_voltage_limit_high
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Current:Pulse Bias Voltage Limit High
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_BIAS\_VOLTAGE\_LIMIT\_HIGH

#### pulse bias voltage limit low

### nidcpower.Session.pulse\_bias\_voltage\_limit\_low

Specifies the minimum voltage, in volts, that the output can produce when generating the desired pulse current on the specified channel(s) during the off phase of a pulse. This property is applicable only if the Compliance Limit Symmetry <p:py:meth:'nidcpower.Session.ComplianceLimitSymmetry.html>'\_\_ property is set to Asymmetric and the Output Method <p:py:meth:'nidcpower.Session.OutputFunction.html>'\_ property is set to Pulse Current. You must also specify a Pulse Bias Volt-Limit High <p:py:meth:'nidcpower.Session.PulseBiasVoltageLimitHigh.html>' to complete the asymmetric range. Valid Values: [-Pulse Voltage Limit Range <p:py:meth:'nidcpower.Session.PulseVoltageLimitRange.html>'\_\_, -1% of Pulse Voltage Limit Range <p:py:meth:'nidcpower.Session.PulseVoltageLimitRange.html>' ] The range bounded by the limit high and limit low must include zero. **Default Value:** Refer to Supported Properties by Device for the default value by device. **Related Topics:** Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled* <p:py:meth:'nidcpower.Session.OverrangingEnabled.html>'\_\_ property is set to TRUE or if the *Output Method* <p:py:meth:'nidcpower.Session.OutputFunction.html>'\_\_ property is set to a pulsing method.

Note: One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_bias_voltage_limit_low = var
var = session.channels[0,1].pulse_bias_voltage_limit_low
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Current:Pulse Bias Voltage Limit Low
- C Attribute: NIDCPOWER ATTR PULSE BIAS VOLTAGE LIMIT LOW

### pulse\_complete\_event\_output\_terminal

```
nidcpower.Session.pulse_complete_event_output_terminal
```

Specifies the output terminal for exporting the Pulse Complete event. Output terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0. Default Value:The default value for PXI Express devices is 250 ns.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Events:Pulse Complete Event:Output Terminal

### • C Attribute: NIDCPOWER\_ATTR\_PULSE\_COMPLETE\_EVENT\_OUTPUT\_TERMINAL

# pulse\_complete\_event\_pulse\_polarity

 $\verb|nidcpower.Session.pulse_complete_event_pulse_polarity|\\$ 

Specifies the behavior of the Pulse Complete event. Default Value: HIGH

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.Polarity
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Pulse Complete Event:Pulse:Polarity
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_COMPLETE\_EVENT\_PULSE\_POLARITY

# pulse\_complete\_event\_pulse\_width

nidcpower.Session.pulse\_complete\_event\_pulse\_width

Specifies the width of the Pulse Complete event, in seconds. The minimum event pulse width value for PXI Express devices is 250 ns. The maximum event pulse width value for PXI Express devices is 1.6 microseconds. Default Value: The default value for PXI Express devices is 250 ns.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Pulse Complete Event:Pulse:Width
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_COMPLETE\_EVENT\_PULSE\_WIDTH

#### pulse current level

```
nidcpower.Session.pulse_current_level
```

Specifies the pulse current level, in amps, that the device attempts to generate on the specified channel(s) during the on phase of a pulse. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_CURRENT</code>. Valid Values: The valid values for this property are defined by the values you specify for the <code>nidcpower.Session.pulse\_current\_level\_range</code> property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_current_level = var
var = session.channels[0,1].pulse_current_level
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Current:Pulse Current Level
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_CURRENT\_LEVEL

### pulse\_current\_level\_range

```
nidcpower.Session.pulse_current_level_range
```

Specifies the pulse current level range, in amps, for the specified channel(s). The range defines the valid values to which you can set the pulse current level and pulse bias current level. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_CURRENT</code>. For valid ranges, refer to the ranges topic for your device in the NI DC Power Supplies and SMUs Help.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_current_level_range = var
var = session.channels[0,1].pulse_current_level_range
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Current:Pulse Current Level Range
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_CURRENT\_LEVEL\_RANGE

#### pulse current limit

```
nidcpower.Session.pulse_current_limit
```

Specifies the pulse current limit, in amps, that the output cannot exceed when generating the desired pulse voltage on the specified channel(s) during the on phase of a pulse. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_VOLTAGE</code> and the <code>nidcpower.Session.compliance\_limit\_symmetry</code> property is set to <code>NIDCPOWER\_VAL\_SYMMETRIC</code>. Valid Values: The valid values for this property are defined by the values you specify for the <code>nidcpower.Session.pulse\_current\_limit\_range</code> property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can

specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_current_limit = var
var = session.channels[0,1].pulse_current_limit
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Voltage:Pulse Current Limit
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_CURRENT\_LIMIT

### pulse current limit high

#### nidcpower.Session.pulse current limit high

Specifies the maximum current, in amps, that the output can produce when generating the desired pulse voltage on the specified channel(s) during the on phase This property is applicable only if the Compliance Limit Symmetry of a pulse. <p:py:meth:'nidcpower.Session.ComplianceLimitSymmetry.html>'\_\_\_ property is set to Asym**metric** and the Output Method <p:py:meth:'nidcpower.Session.OutputFunction.html>'\_\_ set to **Pulse Voltage**. You must also specify a Pulse Current property is Low <p:py:meth: 'nidcpower.Session.PulseCurrentLimitLow.html>'\_\_\_ Limit plete the asymmetric range. Valid Values: [1% of Pulse Current Limit Range <p:py:meth:'nidcpower.Session.PulseCurrentLimitRange.html>'\_\_\_, Pulse Current Limit Range <p:py:meth:'nidcpower.Session.PulseCurrentLimitRange.html>'\_\_] The range bounded by the limit high and limit low must include zero. Default Value: Refer to Supported Properties by Device for the default value by device. **Related Topics:** Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled* <p:py:meth:'nidcpower.Session.OverrangingEnabled.html>'\_\_ property is set to TRUE or if the *Output Method* <p:py:meth:'nidcpower.Session.OutputFunction.html>'\_\_ property is set to a pulsing method.

Note: One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_current_limit_high = var
var = session.channels[0,1].pulse_current_limit_high
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Voltage:Pulse Current Limit High
- C Attribute: NIDCPOWER ATTR PULSE CURRENT LIMIT HIGH

### pulse current limit low

### nidcpower.Session.pulse\_current\_limit\_low

Specifies the minimum current, in amps, that the output can produce when generating the desired pulse voltage on the specified channel(s) during the on phase This property is applicable only if the Compliance Limit Symmetry <p:py:meth:'nidcpower.Session.ComplianceLimitSymmetry.html>'\_\_ property is set to Asym**metric** and the *Output Method* p:py:meth:'nidcpower.Session.OutputFunction.html' property is set to Pulse Voltage. You must also specify a Pulse Current Limit High <p:py:meth:'nidcpower.Session.PulseCurrentLimitHigh.html>'\_\_ asymmetric range. Valid Values: [-Pulse Current Limit plete Range <p:py:meth:'nidcpower.Session.PulseCurrentLimitRange.html>'\_\_, -1% of Pulse Current Limit Range <p:py:meth:'nidcpower.Session.PulseCurrentLimitRange.html>'\_\_] The range bounded by the limit high and limit low must include zero. **Default Value:** Refer to Supported Properties by Device for the default value by device. **Related Topics:** Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled* <p:py:meth:'nidcpower.Session.OverrangingEnabled.html>'\_\_ property is set to TRUE or if the Output Method <p:py:meth: 'nidcpower.Session.OutputFunction.html>' property is set to a pulsing method.

Note: One or more of the referenced methods are not in the Python API for this driver.

Tip: This property can use repeated capabilities (channels). If set or get directly on the nidcpower.Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session

repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_current_limit_low = var
var = session.channels[0,1].pulse_current_limit_low
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Voltage:Pulse Current Limit Low
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_CURRENT\_LIMIT\_LOW

### pulse\_current\_limit\_range

```
nidcpower.Session.pulse_current_limit_range
```

Specifies the pulse current limit range, in amps, for the specified channel(s). The range defines the valid values to which you can set the pulse current limit and pulse bias current limit. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_VOLTAGE</code>. For valid ranges, refer to the ranges topic for your device in the NI DC Power Supplies and SMUs Help.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_current_limit_range = var
var = session.channels[0,1].pulse_current_limit_range
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Voltage:Pulse Current Limit Range
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_CURRENT\_LIMIT\_RANGE

### pulse off time

#### nidcpower.Session.pulse off time

Determines the length, in seconds, of the off phase of a pulse. Valid Values: 10 microseconds to 167 seconds Default Value: 34 milliseconds

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_off_time = var
var = session.channels[0,1].pulse_off_time
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Advanced:Pulse Off Time
- C Attribute: NIDCPOWER ATTR PULSE OFF TIME

# pulse\_on\_time

### nidcpower.Session.pulse\_on\_time

Determines the length, in seconds, of the on phase of a pulse. Valid Values: 10 microseconds to 167 seconds Default Value: 34 milliseconds

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_on_time = var
var = session.channels[0,1].pulse_on_time
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	True
Resettable	No

Tip: This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Advanced:Pulse On Time
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_ON\_TIME

### pulse\_trigger\_type

```
nidcpower.Session.pulse_trigger_type
Specifies the behavior of the Pulse trigger. Default Value: NONE
```

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerType
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Pulse Trigger:Trigger Type
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_TRIGGER\_TYPE

#### pulse voltage level

#### nidcpower.Session.pulse\_voltage\_level

Specifies the pulse current limit, in amps, that the output cannot exceed when generating the desired pulse voltage on the specified channel(s) during the on phase of a pulse. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_VOLTAGE</code>. Valid Values: The valid values for this property are defined by the values you specify for the <code>nidcpower.Session.pulse\_current\_limit\_range</code> property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_voltage_level = var
var = session.channels[0,1].pulse_voltage_level
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Voltage:Pulse Voltage Level
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_VOLTAGE\_LEVEL

# pulse\_voltage\_level\_range

# $\verb|nidcpower.Session.pulse_voltage_level_range|\\$

Specifies the pulse voltage level range, in volts, for the specified channel(s). The range defines the valid values at which you can set the pulse voltage level and pulse bias voltage level. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_VOLTAGE</code>. For valid ranges, refer to the ranges topic for your device in the NI DC Power Supplies and SMUs Help.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_voltage_level_range = var
var = session.channels[0,1].pulse_voltage_level_range
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Voltage:Pulse Voltage Level Range
- C Attribute: NIDCPOWER ATTR PULSE VOLTAGE LEVEL RANGE

## pulse\_voltage\_limit

```
nidcpower.Session.pulse_voltage_limit
```

Specifies the pulse voltage limit, in volts, that the output cannot exceed when generating the desired pulse current on the specified channel(s) during the on phase of a pulse. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_CURRENT</code> and the <code>nidcpower.Session.compliance\_limit\_symmetry</code> property is set to <code>NIDCPOWER\_VAL\_SYMMETRIC</code>. Valid Values: The valid values for this property are defined by the values you specify for the <code>nidcpower.Session.pulse\_voltage\_limit\_range</code> property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_voltage_limit = var
var = session.channels[0,1].pulse_voltage_limit
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Current:Pulse Voltage Limit
- C Attribute: NIDCPOWER ATTR PULSE VOLTAGE LIMIT

# pulse\_voltage\_limit\_high

## nidcpower.Session.pulse\_voltage\_limit\_high

Specifies the maximum voltage, in volts, that the output can produce when generating the desired pulse current on the specified channel(s) during the on phase This property is applicable only if the Compliance Limit Symmetry <p:py:meth:'nidcpower.Session.ComplianceLimitSymmetry.html>'\_\_\_ property is set to Asym**metric** and the *Output Method* <*p:py:meth:'nidcpower.Session.OutputFunction.*html>'\_\_ property is set to Pulse Current. You must also specify a Pulse Voltage <p:py:meth:'nidcpower.Session.PulseVoltageLimitLow.html>' Limit Low plete the asymmetric range. Valid Values: [1% of Pulse Voltage Limit Range <p:py:meth:'nidcpower.Session.PulseVoltageLimitRange.html>'\_\_, Pulse Voltage Limit Range <p:py:meth:'nidcpower.Session.PulseVoltageLimitRange.html>'\_\_] The range bounded by the limit high and limit low must include zero. Default Value: Refer to Supported Properties by Device for the default value by device. **Related Topics:** Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled* <p:py:meth:'nidcpower.Session.OverrangingEnabled.html>'\_\_ property is set to TRUE or if the *Output Method* <p:py:meth:'nidcpower.Session.OutputFunction.html>'\_\_ property is set to a pulsing method.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_voltage_limit_high = var
var = session.channels[0,1].pulse_voltage_limit_high
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

- LabVIEW Property: Source:Pulse Current:Pulse Voltage Limit High
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_VOLTAGE\_LIMIT\_HIGH

## pulse voltage limit low

```
nidcpower.Session.pulse_voltage_limit_low
```

Specifies the minimum voltage, in volts, that the output can produce when generating the desired pulse current on the specified channel(s) during the on phase This property is applicable only if the Compliance Limit Symmetry of a pulse. <p:py:meth:'nidcpower.Session.ComplianceLimitSymmetry.html>'\_\_\_ property is set to Asym**metric** and the *Output Method <p:py:meth:'nidcpower.Session.OutputFunction.*html>'\_\_ property is set to **Pulse Current**. You must also specify a Pulse Limit High <p:py:meth:'nidcpower.Session.PulseVoltageLimitHigh.html>'\_\_ age complete the asymmetric range. Valid Values: [-Pulse Voltage Limit Range <p:py:meth:'nidcpower.Session.PulseVoltageLimitRange.html>'\_\_\_, -1% of Pulse Voltage Limit Range <p:py:meth:'nidcpower.Session.PulseVoltageLimitRange.html>' ] The range bounded by the limit high and limit low must include zero. **Default Value:** Refer to Supported Properties by Device for the default value by device. **Related Topics:** Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled* <p:py:meth:'nidcpower.Session.OverrangingEnabled.html>'\_\_ property is set to TRUE or if the *Output Method* <p:py:meth:'nidcpower.Session.OutputFunction.html>'\_\_ property is set to a pulsing method.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_voltage_limit_low = var
var = session.channels[0,1].pulse_voltage_limit_low
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

- LabVIEW Property: Source:Pulse Current:Pulse Voltage Limit Low
- C Attribute: NIDCPOWER\_ATTR\_PULSE\_VOLTAGE\_LIMIT\_LOW

### pulse voltage limit range

```
nidcpower.Session.pulse_voltage_limit_range
```

Specifies the pulse voltage limit range, in volts, for the specified channel(s). The range defines the valid values to which you can set the pulse voltage limit and pulse bias voltage limit. This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>PULSE\_CURRENT</code>. For valid ranges, refer to the ranges topic for your device in the NI DC Power Supplies and SMUs Help.

**Note:** The channel must be enabled for the specified current limit to take effect. Refer to the *nidcpower.Session.output\_enabled* property for more information about enabling the output channel.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].pulse_voltage_limit_range = var
var = session.channels[0,1].pulse_voltage_limit_range
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Pulse Current:Pulse Voltage Limit Range
- C Attribute: NIDCPOWER ATTR PULSE VOLTAGE LIMIT RANGE

## query instrument status

#### nidcpower.Session.query\_instrument\_status

Specifies whether NI-DCPower queries the device status after each operation. Querying the device status is useful for debugging. After you validate your program, you can set this property to False to disable status checking and maximize performance. NI-DCPower ignores status checking for particular properties regardless of the setting of this property. Use the nidcpower.Session.

init () method to override this value. Default Value: True

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: User Options: Query Instrument Status
- C Attribute: NIDCPOWER\_ATTR\_QUERY\_INSTRUMENT\_STATUS

### ready for pulse trigger event output terminal

## nidcpower.Session.ready\_for\_pulse\_trigger\_event\_output\_terminal

Specifies the output terminal for exporting the Ready For Pulse Trigger event. Output terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Ready For Pulse Trigger Event:Output Terminal
- C Attribute: NIDCPOWER ATTR READY FOR PULSE TRIGGER EVENT OUTPUT TERMINAL

## ready for pulse trigger event pulse polarity

nidcpower.Session.ready\_for\_pulse\_trigger\_event\_pulse\_polarity

Specifies the behavior of the Ready For Pulse Trigger event. Default Value: HIGH

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.Polarity
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Ready For Pulse Trigger Event:Pulse:Polarity
- C Attribute: NIDCPOWER\_ATTR\_READY\_FOR\_PULSE\_TRIGGER\_EVENT\_PULSE\_POLARITY

## ready for pulse trigger event pulse width

nidcpower.Session.ready\_for\_pulse\_trigger\_event\_pulse\_width

Specifies the width of the Ready For Pulse Trigger event, in seconds. The minimum event pulse width value for PXI Express devices is 250 ns. The maximum event pulse width value for all devices is 1.6 microseconds. Default Value: The default value for PXI Express devices is 250 ns

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information about supported devices.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Ready For Pulse Trigger Event:Pulse:Width
- C Attribute: NIDCPOWER ATTR READY FOR PULSE TRIGGER EVENT PULSE WIDTH

#### reset average before measurement

#### nidcpower.Session.reset\_average\_before\_measurement

Specifies whether the measurement returned from any measurement call starts with a new measurement call (True) or returns a measurement that has already begun or completed(False). for information about supported devices. When you set the <code>nidcpower.Session.samples\_to\_average</code> property in the Running state, the output channel measurements might move out of synchronization. While NI-DCPower automatically synchronizes measurements upon the initialization of a session, you can force a synchronization in the running state before you run the <code>nidcpower.Session.measure\_multiple()</code> method. To force a synchronization in the running state, set this property to True, and then run the <code>nidcpower.Session.measure\_multiple()</code> method name parameter. You can set the <code>nidcpower.Session.reset\_average\_before\_measurement</code> property to False after the <code>nidcpower.Session.measure\_multiple()</code> method completes. Default Value: True

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].reset_average_before_measurement = var
var = session.channels[0,1].reset_average_before_measurement
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement:Advanced:Reset Average Before Measurement
- C Attribute: NIDCPOWER\_ATTR\_RESET\_AVERAGE\_BEFORE\_MEASUREMENT

# samples\_to\_average

### nidcpower.Session.samples\_to\_average

Specifies the number of samples to average when you take a measurement. Increasing the number of samples to average decreases measurement noise but increases the time required to take a measurement. Refer to the NI PXI-4110, NI PXI-4130, NI PXI-4132, or NI PXIe-4154 Averaging topic for optional property settings to improve immunity to certain noise types, or refer to the

NI PXIe-4140/4141 DC Noise Rejection, NI PXIe-4142/4143 DC Noise Rejection, or NI PXIe-4144/4145 DC Noise Rejection topic for information about improving noise immunity for those devices. Default Value: NI PXI-4110 or NI PXI-4130—10 NI PXI-4132—1 NI PXIe-4112—1 NI PXIe-4113—1 NI PXIe-4140/4141—1 NI PXIe-4142/4143—1 NI PXIe-4144/4145—1 NI PXIe-4154—500

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].samples_to_average = var
var = session.channels[0,1].samples_to_average
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement:Samples To Average
- C Attribute: NIDCPOWER\_ATTR\_SAMPLES\_TO\_AVERAGE

### self\_calibration\_persistence

### nidcpower.Session.self\_calibration\_persistence

Specifies whether the values calculated during self-calibration should be written to hardware to be used until the next self-calibration or only used until the nidcpower.Session. reset\_device() method is called or the machine is powered down. This property affects the behavior of the nidcpower.Session.self\_cal() method. When set to KEEP\_IN\_MEMORY, the values calculated by the nidcpower.Session.self\_cal() method are used in the existing session, as well as in all further sessions until you call the nidcpower.Session.reset\_device() method or restart the machine. When you set this property to WRITE\_TO\_EEPROM, the values calculated by the nidcpower.Session.self\_cal() method are written to hardware and used in the existing session and in all subsequent sessions until another call to the nidcpower.Session.self\_cal() method is made. about supported devices. Default Value: KEEP\_IN\_MEMORY

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device for information

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.SelfCalibrationPersistence
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Advanced:Self-Calibration Persistence
- C Attribute: NIDCPOWER\_ATTR\_SELF\_CALIBRATION\_PERSISTENCE

#### sense

```
nidcpower.Session.sense
```

Selects either local or remote sensing of the output voltage for the specified channel(s). Refer to the Local and Remote Sense topic in the NI DC Power Supplies and SMUs Help for more information about sensing voltage on supported channels and about devices that support local and/or remote sensing. Default Value: The default value is LOCAL if the device supports local sense. Otherwise, the default and only supported value is REMOTE.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].sense = var
var = session.channels[0,1].sense
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.Sense
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Measurement:Sense
- C Attribute: NIDCPOWER ATTR SENSE

### sequence advance trigger type

```
\verb|nidcpower.Session.sequence_advance_trigger_type|\\
```

Specifies the behavior of the Sequence Advance trigger. for information about supported devices.

Default Value: NONE

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerType
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Sequence Advance Trigger:Trigger Type
- C Attribute: NIDCPOWER ATTR SEQUENCE ADVANCE TRIGGER TYPE

## sequence engine done event output terminal

## nidcpower.Session.sequence\_engine\_done\_event\_output\_terminal

Specifies the output terminal for exporting the Sequence Engine Done Complete event. for information about supported devices. Output terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Sequence Engine Done Event:Output Terminal
- C Attribute: NIDCPOWER\_ATTR\_SEQUENCE\_ENGINE\_DONE\_EVENT\_OUTPUT\_TERMINAL

## sequence engine done event pulse polarity

nidcpower.Session.sequence\_engine\_done\_event\_pulse\_polarity

Specifies the behavior of the Sequence Engine Done event. for information about supported devices.

Default Value: HIGH

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.Polarity
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Sequence Engine Done Event:Pulse:Polarity
- C Attribute: NIDCPOWER\_ATTR\_SEQUENCE\_ENGINE\_DONE\_EVENT\_PULSE\_POLARITY

## sequence engine done event pulse width

#### nidcpower.Session.sequence\_engine\_done\_event\_pulse\_width

Specifies the width of the Sequence Engine Done event, in seconds. The minimum event pulse width value for PXI devices is 150 ns, and the minimum event pulse width value for PXI Express devices is 250 ns. The maximum event pulse width value for all devices is 1.6 microseconds. for information about supported devices. Valid Values: 1.5e-7 to 1.6e-6 seconds Default Value: The default value for PXI devices is 150 ns. The default value for PXI Express devices is 250 ns.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Sequence Engine Done Event:Pulse:Width
- C Attribute: NIDCPOWER\_ATTR\_SEQUENCE\_ENGINE\_DONE\_EVENT\_PULSE\_WIDTH

## sequence iteration complete event output terminal

nidcpower.Session.sequence\_iteration\_complete\_event\_output\_terminal

Specifies the output terminal for exporting the Sequence Iteration Complete event. for information about supported devices. Output terminals can be specified in one of two ways. If the device is

named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Sequence Iteration Complete Event:Output Terminal
- C Attribute: NIDCPOWER\_ATTR\_SEQUENCE\_ITERATION\_COMPLETE\_EVENT\_OUTPUT\_TERMINAL

## sequence iteration complete event pulse polarity

nidcpower.Session.sequence\_iteration\_complete\_event\_pulse\_polarity
Specifies the behavior of the Sequence Iteration Complete event. for information about supported

devices. Default Value: HIGH

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.Polarity
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Sequence Iteration Complete Event:Pulse:Polarity
- C Attribute: NIDCPOWER\_ATTR\_SEQUENCE\_ITERATION\_COMPLETE\_EVENT\_PULSE\_POLARITY

### sequence iteration complete event pulse width

nidcpower.Session.sequence\_iteration\_complete\_event\_pulse\_width

Specifies the width of the Sequence Iteration Complete event, in seconds. The minimum event pulse width value for PXI devices is 150 ns, and the minimum event pulse width value for PXI Express devices is 250 ns. The maximum event pulse width value for all devices is 1.6 microseconds. the NI

DC Power Supplies and SMUs Help for information about supported devices. Valid Values: 1.5e-7 to 1.6e-6 seconds Default Value: The default value for PXI devices is 150 ns. The default value for PXI Express devices is 250 ns.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic in

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Sequence Iteration Complete Event:Pulse:Width
- C Attribute: NIDCPOWER\_ATTR\_SEQUENCE\_ITERATION\_COMPLETE\_EVENT\_PULSE\_WIDTH

## sequence loop count

nidcpower.Session.sequence\_loop\_count

Specifies the number of times a sequence is run after initiation. Refer to the Sequence Source Mode topic in the NI DC Power Supplies and SMUs Help for more information about the sequence loop count. for information about supported devices. When the <code>nidcpower.Session.sequence\_loop\_count\_is\_finite</code> property is set to False, the <code>nidcpower.Session.sequence\_loop\_count</code> property is ignored. Valid Range: 1 to 134217727 Default Value: 1

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Advanced:Sequence Loop Count
- C Attribute: NIDCPOWER\_ATTR\_SEQUENCE\_LOOP\_COUNT

## sequence loop count is finite

#### nidcpower.Session.sequence\_loop\_count\_is\_finite

Specifies whether a sequence should repeat indefinitely. Refer to the Sequence Source Mode topic in the NI DC Power Supplies and SMUs Help for more information about infinite sequencing. nidcpower.Session.sequence\_loop\_count\_is\_finite property is set to False, the nidcpower.Session.sequence\_loop\_count property is ignored. Default Value: True

**Note:** This property is not supported by all devices. When the

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Advanced:Sequence Loop Count Is Finite
- C Attribute: NIDCPOWER\_ATTR\_SEQUENCE\_LOOP\_COUNT\_IS\_FINITE

#### simulate

nidcpower.Session.simulate

Specifies whether to simulate NI-DCPower I/O operations. True specifies that operation is simulated. Default Value: False

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:User Options:Simulate
- C Attribute: NIDCPOWER\_ATTR\_SIMULATE

### source complete event output terminal

 $\verb|nidcpower.Session.source_complete_event_output\_terminal|\\$ 

Specifies the output terminal for exporting the Source Complete event. for information about sup-

ported devices. Output terminals can be specified in one of two ways. If the device is named Dev1 and your terminal is PXI\_Trig0, you can specify the terminal with the fully qualified terminal name, /Dev1/PXI\_Trig0, or with the shortened terminal name, PXI\_Trig0.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Source Complete Event:Output Terminal
- C Attribute: NIDCPOWER\_ATTR\_SOURCE\_COMPLETE\_EVENT\_OUTPUT\_TERMINAL

## source\_complete\_event\_pulse\_polarity

nidcpower.Session.source\_complete\_event\_pulse\_polarity

Specifies the behavior of the Source Complete event. for information about supported devices. Default Value: *HIGH* 

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.Polarity
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Source Complete Event:Pulse:Polarity
- C Attribute: NIDCPOWER\_ATTR\_SOURCE\_COMPLETE\_EVENT\_PULSE\_POLARITY

## source\_complete\_event\_pulse\_width

nidcpower.Session.source\_complete\_event\_pulse\_width

Specifies the width of the Source Complete event, in seconds. for information about supported devices. The minimum event pulse width value for PXI devices is 150 ns, and the minimum event

pulse width value for PXI Express devices is 250 ns. The maximum event pulse width value for all devices is 1.6 microseconds Valid Values: 1.5e-7 to 1.6e-6 seconds Default Value: The default value for PXI devices is 150 ns. The default value for PXI Express devices is 250 ns.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Source Complete Event:Pulse:Width
- C Attribute: NIDCPOWER\_ATTR\_SOURCE\_COMPLETE\_EVENT\_PULSE\_WIDTH

## source\_delay

nidcpower.Session.source\_delay

Determines when, in seconds, the device generates the Source Complete event, potentially starting a measurement if the <code>nidcpower.Session.measure\_when</code> property is set to <code>AUTOMATICALLY\_AFTER\_SOURCE\_COMPLETE</code>. Refer to the Single Point Source Mode and Sequence Source Mode topics for more information. Valid Values: 0 to 167 seconds Default Value: 0.01667 seconds

**Note:** Refer to Supported Properties by Device for information about supported devices.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].source_delay = var
var = session.channels[0,1].source_delay
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	True
Resettable	No

- LabVIEW Property: Source:Advanced:Source Delay
- C Attribute: NIDCPOWER\_ATTR\_SOURCE\_DELAY

## source\_mode

#### nidcpower.Session.source\_mode

Specifies whether to run a single output point or a sequence. Refer to the Single Point Source Mode and Sequence Source Mode topics in the NI DC Power Supplies and SMUs Help for more information about source modes. Default value:  $SINGLE\_POINT$ 

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.SourceMode
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Source Mode
- C Attribute: NIDCPOWER\_ATTR\_SOURCE\_MODE

## source\_trigger\_type

nidcpower.Session.source\_trigger\_type

Specifies the behavior of the Source trigger. for information about supported devices. Default Value: NONE

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerType
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Source Trigger:Trigger Type
- C Attribute: NIDCPOWER\_ATTR\_SOURCE\_TRIGGER\_TYPE

## specific driver description

nidcpower.Session.specific\_driver\_description

Contains a brief description of the specific driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Identification:Description
- C Attribute: NIDCPOWER\_ATTR\_SPECIFIC\_DRIVER\_DESCRIPTION

## specific driver prefix

nidcpower.Session.specific\_driver\_prefix

Contains the prefix for NI-DCPower. The name of each user-callable method in NI-DCPower begins with this prefix.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Identification:Driver Prefix
- C Attribute: NIDCPOWER\_ATTR\_SPECIFIC\_DRIVER\_PREFIX

## specific driver revision

nidcpower.Session.specific\_driver\_revision

Contains additional version information about NI-DCPower.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

- LabVIEW Property: Inherent IVI Attributes:Driver Identification:Revision
- C Attribute: NIDCPOWER\_ATTR\_SPECIFIC\_DRIVER\_REVISION

### specific driver vendor

nidcpower.Session.specific\_driver\_vendor

Contains the name of the vendor that supplies NI-DCPower.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Identification:Driver Vendor
- C Attribute: NIDCPOWER\_ATTR\_SPECIFIC\_DRIVER\_VENDOR

## start\_trigger\_type

nidcpower.Session.start\_trigger\_type

Specifies the behavior of the Start trigger. for information about supported devices. Default Value: NONE

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerType
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Triggers:Start Trigger:Trigger Type
- C Attribute: NIDCPOWER ATTR START TRIGGER TYPE

## supported instrument models

## nidcpower.Session.supported\_instrument\_models

Contains a comma-separated (,) list of supported NI-DCPower device models.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Capabilities:Supported Instrument Models
- C Attribute: NIDCPOWER\_ATTR\_SUPPORTED\_INSTRUMENT\_MODELS

## transient\_response

## nidcpower.Session.transient\_response

Specifies the transient response. Refer to the Transient Response topic in the NI DC Power Supplies and SMUs Help for more information about transient response. for information about supported devices. Default Value: NORMAL

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower.Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower.Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].transient_response = var
var = session.channels[0,1].transient_response
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TransientResponse
Permissions	read-write
Channel Based	True
Resettable	No

- LabVIEW Property: Source:Transient Response
- C Attribute: NIDCPOWER\_ATTR\_TRANSIENT\_RESPONSE

## voltage\_compensation\_frequency

nidcpower.Session.voltage\_compensation\_frequency

The frequency at which a pole-zero pair is added to the system when the channel is in Constant Voltage mode. for information about supported devices. Default value: Determined by the value of the NORMAL setting of the nidcpower. Session.transient\_response property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_compensation_frequency = var
var = session.channels[0,1].voltage_compensation_frequency
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Custom Transient Response:Voltage:Compensation Frequency
- C Attribute: NIDCPOWER\_ATTR\_VOLTAGE\_COMPENSATION\_FREQUENCY

#### voltage gain bandwidth

#### nidcpower.Session.voltage\_gain\_bandwidth

The frequency at which the unloaded loop gain extrapolates to 0 dB in the absence of additional poles and zeroes. This property takes effect when the channel is in Constant Voltage mode. for information about supported devices. Default Value: Determined by the value of the NORMAL setting of the nidcpower.Session.transient\_response property.

**Note:** This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_gain_bandwidth = var
var = session.channels[0,1].voltage_gain_bandwidth
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source: Custom Transient Response: Voltage: Gain Bandwidth
- C Attribute: NIDCPOWER\_ATTR\_VOLTAGE\_GAIN\_BANDWIDTH

# voltage\_level

## nidcpower.Session.voltage\_level

Specifies the voltage level, in volts, that the device attempts to generate on the specified channel(s). This property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>DC\_VOLTAGE.nidcpower.Session.output\_enabled</code> property for more information about enabling the output channel. Valid Values: The valid values for this property are defined by the values you specify for the <code>nidcpower.Session.voltage\_level\_range</code> property.

Note: The channel must be enabled for the specified voltage level to take effect. Refer to the

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_level = var
var = session.channels[0,1].voltage_level
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Voltage:Voltage Level
- C Attribute: NIDCPOWER\_ATTR\_VOLTAGE\_LEVEL

## voltage level autorange

### nidcpower.Session.voltage\_level\_autorange

Specifies whether NI-DCPower automatically selects the voltage level range based on the desired voltage level for the specified channel(s). If you set this property to ON, NI-DCPower ignores any changes you make to the <code>nidcpower.Session.voltage\_level\_range</code> property. If you change the <code>nidcpower.Session.voltage\_level\_autorange</code> property from ON to OFF, NI-DCPower retains the last value the <code>nidcpower.Session.voltage\_level\_range</code> property was set to (or the default value if the property was never set) and uses that value as the voltage level range. Query the <code>nidcpower.Session.voltage\_level\_range</code> property by using the <code>nidcpower.Session.get\_attribute\_vi\_int32()</code> method for information about which range NI-DCPower automatically selects. The <code>nidcpower.Session.voltage\_level\_autorange</code> property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>DC\_VOLTAGE</code>. Default Value: OFF

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_level_autorange = var
var = session.channels[0,1].voltage_level_autorange
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

- LabVIEW Property: Source:DC Voltage:Voltage Level Autorange
- C Attribute: NIDCPOWER ATTR VOLTAGE LEVEL AUTORANGE

### voltage level range

```
nidcpower.Session.voltage_level_range
```

Specifies the voltage level range, in volts, for the specified channel(s). The range defines the valid values to which the voltage level can be set. Use the <code>nidcpower.Session.voltage\_level\_autorange</code> property to enable automatic selection of the voltage level range. The <code>nidcpower.Session.voltage\_level\_range</code> property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>DC\_VOLTAGE.nidcpower.Session.output\_enabled</code> property for more information about enabling the output channel. For valid ranges, refer to the Ranges topic for your device in the NI DC Power Supplies and SMUs Help.

Note: The channel must be enabled for the specified voltage level range to take effect. Refer to the

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_level_range = var
var = session.channels[0,1].voltage_level_range
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Voltage:Voltage Level Range
- C Attribute: NIDCPOWER\_ATTR\_VOLTAGE\_LEVEL\_RANGE

## voltage\_limit

```
nidcpower.Session.voltage_limit
```

Specifies the voltage limit, in volts, that the output cannot exceed when generating the desired current level on the specified channels. This property is applicable

only if the nidcpower.Session.output\_function property is set to DC\_CURRENT and the nidcpower.Session.compliance\_limit\_symmetry property is set to NIDCPOWER\_VAL\_SYMMETRIC. nidcpower.Session.output\_enabled property for more information about enabling the output channel. Valid Values: The valid values for this property are defined by the values to which the nidcpower.Session.voltage\_limit\_range property is set.

**Note:** The channel must be enabled for the specified current level to take effect. Refer to the

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_limit = var
var = session.channels[0,1].voltage_limit
```

The following table lists the characteristics of this property.

Value
float
read-write
True
No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Current:Voltage Limit
- C Attribute: NIDCPOWER\_ATTR\_VOLTAGE\_LIMIT

## voltage limit autorange

#### nidcpower.Session.voltage\_limit\_autorange

Specifies whether NI-DCPower automatically selects the voltage limit range based on the desired voltage limit for the specified channel(s). If this property is set to *ON*, NI-DCPower ignores any changes you make to the <code>nidcpower.Session.voltage\_limit\_range</code> property. If you change the <code>nidcpower.Session.voltage\_limit\_autorange</code> property from *ON* to <code>OFF</code>, NI-DCPower retains the last value the <code>nidcpower.Session.voltage\_limit\_range</code> property was set to (or the default value if the property was never set) and uses that value as the voltage limit range. Query the <code>nidcpower.Session.voltage\_limit\_range</code> property by using the <code>nidcpower.Session.get\_attribute\_vi\_int32</code>() method to find out which range NI-DCPower automatically selects. The <code>nidcpower.Session</code>.

voltage\_limit\_autorange property is applicable only if the nidcpower.Session.output\_function property is set to DC\_CURRENT. Default Value: OFF

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_limit_autorange = var
var = session.channels[0,1].voltage_limit_autorange
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Current:Voltage Limit Autorange
- C Attribute: NIDCPOWER\_ATTR\_VOLTAGE\_LIMIT\_AUTORANGE

## voltage\_limit\_high

## nidcpower.Session.voltage\_limit\_high

Specifies the maximum voltage, in volts, that the output can produce when generating the desired current on the specified channel(s). This property is applicable only if the Compliance Limit Symmetry <p:py:meth: 'nidcpower.Session.ComplianceLimitSymmetry.html>'\_\_\_ property is set to **Asymmetric** and the *Output Method <p:py:meth:'nidcpower.Session.OutputFunction.*html>'\_\_ also property is to **DC** Current. You must specify Voltage Low <p:py:meth:'nidcpower.Session.VoltageLimitLow.html>' Limit comto plete the asymmetric range. Valid Values: [1% of Voltage Limit Range high and limit low must include zero. Default Value: Refer to Supported Properties by Device for the default value by device. **Related Topics:** Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled <p:py:meth:'nidcpower.Session.OverrangingEnabled.*html>'\_\_ property is set to TRUE.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can

specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_limit_high = var
var = session.channels[0,1].voltage_limit_high
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Current:Voltage Limit High
- C Attribute: NIDCPOWER\_ATTR\_VOLTAGE\_LIMIT\_HIGH

## voltage limit low

```
nidcpower.Session.voltage limit low
```

Specifies the minimum voltage, in volts, that the output can produce when generating the desired current on the specified channel(s). This property is applicable only if the *Compliance Limit Symmetry <p:py:meth: 'nidcpower.Session.ComplianceLimitSymmetry.*html>'\_\_ property is set to **Asymmetric** and the *Output Method <p:py:meth: 'nidcpower.Session.OutputFunction.*html>'\_\_ property is set to **DC Current**. You must also specify a *Voltage Limit High <p:py:meth: 'nidcpower.Session.VoltageLimitHigh.*html>'\_\_ to complete the asymmetric range. **Valid Values:** [-Voltage Limit Range <p:py:meth: 'nidcpower.Session.VoltageLimitRange.html>'\_\_, -1% of *Voltage Limit Range <p:py:meth: 'nidcpower.Session.VoltageLimitRange.*html>'\_\_] The range bounded by the limit high and limit low must include zero. **Default Value:** Refer to Supported Properties by Device for the default value by device. **Related Topics:** Ranges Changing Ranges Overranging

**Note:** The limit may be extended beyond the selected limit range if the *Overranging Enabled* <p:py:meth:'nidcpower.Session.OverrangingEnabled.html>'\_\_ property is set to TRUE.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_limit_low = var
var = session.channels[0,1].voltage_limit_low
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Current:Voltage Limit Low
- C Attribute: NIDCPOWER\_ATTR\_VOLTAGE\_LIMIT\_LOW

# voltage\_limit\_range

```
nidcpower.Session.voltage_limit_range
```

Specifies the voltage limit range, in volts, for the specified channel(s). The range defines the valid values to which the voltage limit can be set. Use the <code>nidcpower.Session.voltage\_limit\_autorange</code> property to enable automatic selection of the voltage limit range. The <code>nidcpower.Session.voltage\_limit\_range</code> property is applicable only if the <code>nidcpower.Session.output\_function</code> property is set to <code>DC\_CURRENT.nidcpower.Session.output\_enabled</code> property for more information about enabling the output channel. For valid ranges, refer to the Ranges topic for your device in the NI DC Power Supplies and SMUs Help.

**Note:** The channel must be enabled for the specified voltage limit range to take effect. Refer to the

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_limit_range = var
var = session.channels[0,1].voltage_limit_range
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:DC Current:Voltage Limit Range
- C Attribute: NIDCPOWER\_ATTR\_VOLTAGE\_LIMIT\_RANGE

#### voltage pole zero ratio

```
nidcpower.Session.voltage pole zero ratio
```

The ratio of the pole frequency to the zero frequency when the channel is in Constant Voltage mode. for information about supported devices. Default value: Determined by the value of the *NORMAL* setting of the *nidcpower.Session.transient\_response* property.

Note: This property is not supported by all devices. Refer to Supported Properties by Device topic

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the nid-cpower. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].voltage_pole_zero_ratio = var
var = session.channels[0,1].voltage_pole_zero_ratio
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Source:Custom Transient Response:Voltage:Pole-Zero Ratio
- C Attribute: NIDCPOWER ATTR VOLTAGE POLE ZERO RATIO

#### **Methods**

#### abort

```
nidcpower.Session.abort()
```

Transitions the NI-DCPower session from the Running state to the Committed state. If a sequence is running, it is stopped. Any configuration methods called after this method are not applied until the <code>nidcpower.Session.initiate()</code> method is called. If power output is enabled when you call the <code>nidcpower.Session.abort()</code> method, the output channels remain in their current state and continue providing power.

Use the nidcpower.Session.ConfigureOutputEnabled() method to disable power output on a per channel basis. Use the nidcpower.Session.reset() method to disable output on all channels.

Refer to the Programming States topic in the *NI DC Power Supplies and SMUs Help* for information about the specific NI-DCPower software states.

#### **Related Topics:**

**Programming States** 

**Note:** One or more of the referenced methods are not in the Python API for this driver.

#### commit

```
nidcpower.Session.commit()
```

Applies previously configured settings to the device. Calling this method moves the NI-DCPower session from the Uncommitted state into the Committed state. After calling this method, modifying any property reverts the NI-DCPower session to the Uncommitted state. Use the nidcpower. Session.initiate() method to transition to the Running state. Refer to the Programming States topic in the NI DC Power Supplies and SMUs Help for details about the specific NI-DCPower software states.

## **Related Topics:**

**Programming States** 

# configure\_aperture\_time

nidcpower.Session.configure\_aperture\_time (aperture\_time,

units=nidcpower.ApertureTimeUnits.SECONDS)

Configures the aperture time on the specified channel(s).

The supported values depend on the **units**. Refer to the *Aperture Time* topic for your device in the *NI DC Power Supplies and SMUs Help* for more information. In general, devices support discrete **aperture Time** values, and if you configure **aperture Time** to some unsupported value, NI-DCPower coerces it up to the next supported value.

Refer to the *Measurement Configuration and Timing* or *DC Noise Rejection* topic for your device in the *NI DC Power Supplies and SMUs Help* for more information about how to configure your measurements.

#### **Related Topics:**

Aperture Time

**Note:** This method is not supported on all devices. Refer to Supported Methods by Device for more information about supported devices.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].configure_aperture_time(aperture_time, _ 

→units=nidcpower.ApertureTimeUnits.SECONDS)
```

#### **Parameters**

- aperture\_time (float) Specifies the aperture time. Refer to the *Aperture Time* topic for your device in the *NI DC Power Supplies and SMUs Help* for more information.
- units (nidcpower.ApertureTimeUnits) Specifies the units for apertureTime. Defined Values:

SECONDS (1028)	Specifies seconds.
POWER_LINE_CYCLES (1029)	Specifies Power Line Cycles.

#### disable

```
nidcpower.Session.disable()
```

This method performs the same actions as the <code>nidcpower.Session.reset()</code> method, except that this method also immediately sets the <code>nidcpower.Session.output\_enabled</code> property to False.

This method opens the output relay on devices that have an output relay.

### fetch multiple

nidcpower.Session.fetch\_multiple (count, timeout=datetime.timedelta(seconds=1.0))
Returns a list of named tuples (Measurement) that were previously taken and are stored in the NI-DCPower buffer. This method should not be used when the nidcpower.Session.

measure\_when property is set to ON\_DEMAND. You must first call nidcpower.Session.
initiate() before calling this method.

Fields in Measurement:

- voltage (float)
- current (float)
- in\_compliance (bool)

**Note:** This method is not supported on all devices. Refer to Supported Methods by Device for more information about supported devices.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].fetch_multiple(count, timeout=datetime.

→timedelta(seconds=1.0))
```

#### **Parameters**

- **count** (*int*) Specifies the number of measurements to fetch.
- **timeout** (*float in seconds or datetime.timedelta*) Specifies the maximum time allowed for this method to complete. If the method does not complete within this time interval, NI-DCPower returns an error.

**Note:** When setting the timeout interval, ensure you take into account any triggers so that the timeout interval is long enough for your application.

## Return type list of Measurement

#### Returns

List of named tuples with fields:

- voltage (float)
- current (float)
- in\_compliance (bool)

## get channel name

```
nidcpower.Session.get_channel_name(index)
```

Retrieves the output **channelName** that corresponds to the requested **index**. Use the *nidcpower*. Session.channel\_count property to determine the upper bound of valid values for **index**.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].get_channel_name(index)
```

**Parameters index** (*int*) – Specifies which output channel name to return. The index values begin at 1.

## get\_ext\_cal\_last\_date\_and\_time

```
nidcpower.Session.get_ext_cal_last_date_and_time()
```

Returns the date and time of the last successful calibration.

Return type datetime.datetime

**Returns** Indicates date and time of the last calibration.

### get ext cal last temp

```
nidcpower.Session.get_ext_cal_last_temp()
```

Returns the onboard **temperature** of the device, in degrees Celsius, during the last successful external calibration.

#### Return type float

**Returns** Returns the onboard **temperature** of the device, in degrees Celsius, during the last successful external calibration.

### get ext cal recommended interval

```
nidcpower.Session.get_ext_cal_recommended_interval()
```

Returns the recommended maximum interval, in months, between external calibrations.

Return type datetime.timedelta

**Returns** Specifies the recommended maximum interval, in **months**, between external calibrations.

## get\_self\_cal\_last\_date\_and\_time

```
nidcpower.Session.get self cal last date and time()
```

Returns the date and time of the oldest successful self-calibration from among the channels in the session.

**Note:** This method is not supported on all devices.

Return type datetime.datetime

**Returns** Returns the date and time the device was last calibrated.

#### get self cal last temp

```
nidcpower.Session.get_self_cal_last_temp()
```

Returns the onboard temperature of the device, in degrees Celsius, during the oldest successful self-calibration from among the channels in the session.

For example, if you have a session using channels 1 and 2, and you perform a self-calibration on channel 1 with a device temperature of 25 degrees Celsius at 2:00, and a self-calibration was performed on channel 2 at 27 degrees Celsius at 3:00 on the same day, this method returns 25 for the **temperature** parameter.

**Note:** This method is not supported on all devices. Refer to Supported Methods by Device for more information about supported devices.

# Return type float

**Returns** Returns the onboard **temperature** of the device, in degrees Celsius, during the oldest successful calibration.

#### lock

```
nidcpower.Session.lock()
```

Obtains a multithread lock on the device session. Before doing so, the software waits until all other execution threads release their locks on the device session.

Other threads may have obtained a lock on this session for the following reasons:

- The application called the *nidcpower.Session.lock()* method.
- A call to NI-DCPower locked the session.
- After a call to the <code>nidcpower.Session.lock()</code> method returns successfully, no other threads can access the device session until you call the <code>nidcpower.Session.unlock()</code> method or exit out of the with block when using lock context manager.
- Use the nidcpower.Session.lock() method and the nidcpower.Session.unlock() method around a sequence of calls to instrument driver methods if you require that the device retain its settings through the end of the sequence.

You can safely make nested calls to the <code>nidcpower.Session.lock()</code> method within the same thread. To completely unlock the session, you must balance each call to the <code>nidcpower.Session.lock()</code> method with a call to the <code>nidcpower.Session.unlock()</code> method.

One method for ensuring there are the same number of unlock method calls as there is lock calls is to use lock as a context manager

```
with nidcpower.Session('dev1') as session:
    with session.lock():
        # Calls to session within a single lock context
```

The first with block ensures the session is closed regardless of any exceptions raised

The second with block ensures that unlock is called regardless of any exceptions raised

**Return type** context manager

**Returns** When used in a *with* statement, *nidcpower.Session.lock()* acts as a context manager and unlock will be called when the *with* block is exited

### measure

```
nidcpower.Session.measure(measurement_type)
```

Returns the measured value of either the voltage or current on the specified output channel. Each call to this method blocks other method calls until the hardware returns the **measurement**. To measure multiple output channels, use the <code>nidcpower.Session.measure\_multiple()</code> method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower.Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower.Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].measure(measurement_type)
```

**Parameters** measurement\_type (nidcpower.MeasurementTypes) - Specifies whether a voltage or current value is measured. **Defined Values**:

VOLTAGE (1)	The device measures voltage.
CURRENT (0)	The device measures current.

## Return type float

**Returns** Returns the value of the measurement, either in volts for voltage or amps for

## measure\_multiple

```
nidcpower.Session.measure_multiple()
```

Returns a list of named tuples (Measurement) containing the measured voltage and current values on the specified output channel(s). Each call to this method blocks other method calls until the measurements are returned from the device. The order of the measurements returned in the array corresponds to the order on the specified output channel(s).

Fields in Measurement:

- voltage (float)
- current (float)
- in\_compliance (bool) Always None

**Note:** This method is not supported on all devices. Refer to Supported Methods by Device for more information about supported devices.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].measure_multiple()
```

## Return type list of Measurement

#### Returns

List of named tuples with fields:

- voltage (float)
- current (float)
- in\_compliance (bool) Always None

## query in compliance

```
nidcpower.Session.query_in_compliance()
```

Queries the specified output device to determine if it is operating at the compliance limit.

The compliance limit is the current limit when the output method is set to DC\_VOLTAGE. If the output is operating at the compliance limit, the output reaches the current limit before the desired

voltage level. Refer to the nidcpower.Session.ConfigureOutputFunction() method and the nidcpower.Session.ConfigureCurrentLimit() method for more information about output method and current limit, respectively.

The compliance limit is the voltage limit when the output method is set to *DC\_CURRENT*. If the output is operating at the compliance limit, the output reaches the voltage limit before the desired current level. Refer to the nidcpower.Session.ConfigureOutputFunction() method and the nidcpower.Session.ConfigureVoltageLimit() method for more information about output method and voltage limit, respectively.

## **Related Topics:**

Compliance

Note: One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].query_in_compliance()
```

### Return type bool

**Returns** Returns whether the device output channel is in compliance.

#### query max current limit

nidcpower.Session.query\_max\_current\_limit(voltage\_level)

Queries the maximum current limit on an output channel if the output channel is set to the specified **voltageLevel**.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].query_max_current_limit(voltage_level)
```

**Parameters voltage\_level** (float) – Specifies the voltage level to use when calculating the **maxCurrentLimit**.

Return type float

**Returns** Returns the maximum current limit that can be set with the specified **voltageLevel**.

#### query max voltage level

nidcpower.Session.query\_max\_voltage\_level(current\_limit)

Queries the maximum voltage level on an output channel if the output channel is set to the specified **currentLimit**.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

session.channels[0,1].query\_max\_voltage\_level(current\_limit)

**Parameters** current\_limit (float) - Specifies the current limit to use when calculating the maxVoltageLevel.

Return type float

**Returns** Returns the maximum voltage level that can be set on an output channel with the specified **currentLimit**.

### query min current limit

nidcpower.Session.query\_min\_current\_limit(voltage\_level)

Queries the minimum current limit on an output channel if the output channel is set to the specified **voltageLevel**.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

session.channels[0,1].query\_min\_current\_limit(voltage\_level)

**Parameters voltage\_level** (float) – Specifies the voltage level to use when calculating the **minCurrentLimit**.

Return type float

**Returns** Returns the minimum current limit that can be set on an output channel with the specified **voltageLevel**.

### query\_output\_state

nidcpower.Session.query\_output\_state(output\_state)

Queries the specified output channel to determine if the output channel is currently in the state specified by **outputState**.

# **Related Topics:**

Compliance

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].query_output_state(output_state)
```

**Parameters output\_state** (nidcpower.OutputStates) - Specifies the output state of the output channel that is being queried. **Defined Values**:

VOLTAGE	(0)	The device maintains a constant voltage by adjusting the current.
CURRENT	(1)	The device maintains a constant current by adjusting the voltage.

### Return type bool

Returns Returns whether the device output channel is in the specified output state.

#### read current temperature

```
nidcpower.Session.read_current_temperature()
```

Returns the current onboard temperature, in degrees Celsius, of the device.

## Return type float

**Returns** Returns the onboard **temperature**, in degrees Celsius, of the device.

#### reset

```
nidcpower.Session.reset()
```

Resets the device to a known state. This method disables power generation, resets session properties to their default values, commits the session properties, and leaves the session in the Uncommitted state. Refer to the Programming States topic for more information about NI-DCPower software states.

### reset device

```
nidcpower.Session.reset_device()
```

Resets the device to a known state. The method disables power generation, resets session properties to their default values, clears errors such as overtemperature and unexpected loss of auxiliary power, commits the session properties, and leaves the session in the Uncommitted state. This method also performs a hard reset on the device and driver software. This method has the same functionality as using reset in Measurement & Automation Explorer. Refer to the Programming States topic for more information about NI-DCPower software states.

This will also open the output relay on devices that have an output relay.

#### reset with defaults

```
nidcpower.Session.reset_with_defaults()
```

Resets the device to a known state. This method disables power generation, resets session properties to their default values, commits the session properties, and leaves the session in the Running state. In addition to exhibiting the behavior of the <code>nidcpower.Session.reset()</code> method, this method can assign user-defined default values for configurable properties from the IVI configuration.

### self\_cal

```
nidcpower.Session.self_cal()
```

Performs a self-calibration upon the specified channel(s).

This method disables the output, performs several internal calculations, and updates calibration values. The updated calibration values are written to the device hardware if the <code>nidcpower.Session.self\_calibration\_persistence</code> property is set to <code>WRITE\_TO\_EEPROM</code>. Refer to the <code>nidcpower.Session.self\_calibration\_persistence</code> property topic for more information about the settings for this property.

When calling  $nidepower.Session.self\_cal()$  with the PXIe-4162/4163, specify all channels of your PXIe-4162/4163 with the channelName input. You cannot self-calibrate a subset of PXIe-4162/4163 channels.

Refer to the Self-Calibration topic for more information about this method.

### **Related Topics:**

Self-Calibration

**Note:** This method is not supported on all devices. Refer to Supported Methods by Device for more information about supported devices.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].self_cal()
```

### self\_test

```
nidcpower.Session.self_test()
```

Performs the device self-test routine and returns the test result(s). Calling this method implicitly calls the <code>nidcpower.Session.reset()</code> method.

When calling nidcpower.Session.self\_test() with the PXIe-4162/4163, specify all channels of your PXIe-4162/4163 with the channels input of nidcpower.Session. \_\_init\_\_(). You cannot self test a subset of PXIe-4162/4163 channels.

Raises SelfTestError on self test failure. Properties on exception object:

• code - failure code from driver

· message - status message from driver

Self-Test Code	Description
0	Self test passed.
1	Self test failed.

## send\_software\_edge\_trigger

nidcpower.Session.send\_software\_edge\_trigger(trigger)

Asserts the specified trigger. This method can override an external edge trigger.

### **Related Topics:**

**Triggers** 

**Note:** This method is not supported on all devices. Refer to Supported Methods by Device for more information about supported devices.

**Parameters trigger** (nidcpower.SendSoftwareEdgeTriggerType) – Specifies which trigger to assert. **Defined Values:** 

NIDCPOWER_VAL_START_TRIGGER (1034)	Asserts the Start trigger.	
NIDCPOWER_VAL_SOURCE_TRIGGER (1035)	Asserts the Source trigger.	
NIDCPOWER_VAL_MEASURE_TRIGGER	Asserts the Measure trigger.	
(1036)		
NIDCPOWER_VAL_SEQUENCE_ADVANCE_TRI	GASBerts the Sequence Ad-	
(1037)	vance trigger.	
NIDCPOWER_VAL_PULSE_TRIGGER (1053	Asserts the Pulse trigger.	

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

### set\_sequence

nidcpower.Session.set\_sequence(values, source\_delays)

Configures a series of voltage or current outputs and corresponding source delays. The source mode must be set to Sequence for this method to take effect.

Refer to the Configuring the Source Unit topic in the NI DC Power Supplies and SMUs Help for more information about how to configure your device.

Use this method in the Uncommitted or Committed programming states. Refer to the Programming States topic in the *NI DC Power Supplies and SMUs Help* for more information about NI-DCPower programming states.

**Note:** This method is not supported on all devices. Refer to Supported Methods by Device for more information about supported devices.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nid-cpower. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nidcpower. Session repeated capabilities container, and calling this method on the result.:

session.channels[0,1].set\_sequence(values, source\_delays)

#### **Parameters**

- **values** (*list of float*) Specifies the series of voltage levels or current levels, depending on the configured output method. **Valid values**: The valid values for this parameter are defined by the voltage level range or current level range.
- **source\_delays** (*list of float*) Specifies the source delay that follows the configuration of each value in the sequence. **Valid Values**: The valid values are between 0 and 167 seconds.

#### unlock

nidcpower.Session.unlock()

Releases a lock that you acquired on an device session using nidcpower.Session.lock(). Refer to nidcpower.Session.unlock() for additional information on session locks.

### wait for event

nidcpower.Session.wait\_for\_event (event\_id, timeout=datetime.timedelta(seconds=10.0)) Waits until the device has generated the specified event.

The session monitors whether each type of event has occurred at least once since the last time this method or the nidcpower.Session.initiate() method were called. If an event has only been generated once and you call this method successively, the method times out. Individual events must be generated between separate calls of this method.

**Note:** Refer to Supported Methods by Device for more information about supported devices.

### **Parameters**

• event\_id (nidcpower.Event) - Specifies which event to wait for. Defined Values:

NIDCPOWER_VAL_SOURCE_COMPLETE_EVE	NW aits for the Source Com-	
(1030) plete event.		
NIDCPOWER_VAL_MEASURE_COMPLETE_EV	EWaits for the Measure Com-	
(1031)	plete event.	
NIDCPOWER_VAL_SEQUENCE_ITERATION_	Weirs for the Sequence Itera-	
(1032)	tion Complete event.	
NIDCPOWER_VAL_SEQUENCE_ENGINE_DON	SEQUENCE_ENGINE_DONE Waive Not the Sequence En-	
(1033)	gine Done event.	
NIDCPOWER_VAL_PULSE_COMPLETE_EVEN	TWaits for the Pulse Complete	
(1051)	event.	
NIDCPOWER_VAL_READY_FOR_PULSE_TR	CWairs_fov the Ready for Pulse	
(1052)	Trigger event.	

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

• **timeout** (*float in seconds or datetime.timedelta*) – Specifies the maximum time allowed for this method to complete, in seconds. If the method does not complete within this time interval, NI-DCPower returns an error.

**Note:** When setting the timeout interval, ensure you take into account any triggers so that the timeout interval is long enough for your application.

# **Properties**

Property	Datatype
nidcpower.Session.aperture_time	float
nidcpower.Session.aperture_time_units	ApertureTimeUnits
nidcpower.Session.auto_zero	AutoZero
nidcpower.Session.auxiliary_power_source_available	bool
nidcpower.Session.channel_count	int
nidcpower.Session.compliance_limit_symmetry	ComplianceLimitSy
nidcpower.Session.current_compensation_frequency	float
nidcpower.Session.current_gain_bandwidth	float
nidcpower.Session.current_level	float
nidcpower.Session.current_level_autorange	bool
nidcpower.Session.current_level_range	float
nidcpower.Session.current_limit	float
nidcpower.Session.current_limit_autorange	bool
nidcpower.Session.current_limit_high	float
nidcpower.Session.current_limit_low	float
nidcpower.Session.current_limit_range	float
nidcpower.Session.current_pole_zero_ratio	float
nidcpower.Session.dc_noise_rejection	DCNoiseRejection
nidcpower.Session.digital_edge_measure_trigger_input_terminal	str
nidcpower.Session.digital_edge_pulse_trigger_input_terminal	str

Continued of

Table 2 – continued from previous page

Table 2 – continued from previous page		
Property	Datatype	
nidcpower.Session.digital_edge_sequence_advance_trigger_input_terminal	str	
nidcpower.Session.digital_edge_source_trigger_input_terminal	str	
nidcpower.Session.digital_edge_start_trigger_input_terminal	str	
nidcpower.Session.driver_setup	str	
nidcpower.Session.exported_measure_trigger_output_terminal	str	
nidcpower.Session.exported_pulse_trigger_output_terminal	str	
nidcpower.Session.exported_sequence_advance_trigger_output_terminal	str	
nidcpower.Session.exported_source_trigger_output_terminal	str	
nidcpower.Session.exported_start_trigger_output_terminal	str	
nidcpower.Session.fetch_backlog	int	
nidcpower.Session.instrument_firmware_revision	str	
nidcpower.Session.instrument_manufacturer	str	
nidcpower.Session.instrument_model	str	
nidcpower.Session.interlock_input_open	bool	
nidcpower.Session.io_resource_descriptor	str	
nidcpower.Session.logical_name	str	
nidcpower.Session.measure_buffer_size	int	
nidcpower.Session.measure_complete_event_delay	float in seconds or datetir	
nidcpower.Session.measure_complete_event_output_terminal	str	
nidcpower.Session.measure_complete_event_pulse_polarity	Polarity	
nidcpower.Session.measure_complete_event_pulse_width	float	
nidcpower.Session.measure_record_delta_time	float in seconds or dateting	
nidcpower.Session.measure_record_length	int	
nidcpower.Session.measure_record_length_is_finite	bool	
nidcpower.Session.measure_trigger_type	TriggerType	
nidcpower.Session.measure_when	MeasureWhen	
nidcpower.Session.output_capacitance	OutputCapacitance	
nidcpower.Session.output_connected	bool	
nidcpower.Session.output_enabled	bool	
nidcpower.Session.output_function	OutputFunction	
nidcpower.Session.output_resistance	float	
nidcpower.Session.overranging_enabled	bool	
nidcpower.Session.ovp_enabled	bool	
nidcpower.Session.ovp_limit	float	
nidcpower.Session.power_line_frequency	float	
nidcpower.Session.power_source	PowerSource	
nidcpower.Session.power_source_in_use	PowerSourceInUse	
nidcpower.Session.pulse_bias_current_level	float	
nidcpower.Session.pulse_bias_current_limit	float	
nidcpower.Session.pulse_bias_current_limit_high	float	
nidcpower.Session.pulse_bias_current_limit_low	float	
nidcpower.Session.pulse_bias_delay	float	
nidcpower.Session.pulse_bias_voltage_level	float	
nidcpower.Session.pulse_bias_voltage_limit	float	
nidcpower.Session.pulse_bias_voltage_limit_high	float	
nidcpower.Session.pulse_bias_voltage_limit_low	float	
nidcpower.Session.pulse_complete_event_output_terminal	str	
nidcpower.Session.pulse_complete_event_pulse_polarity	Polarity	
nidcpower.Session.pulse_complete_event_pulse_width	float	
	<del></del>	

Continued c

Table 2 – continued from previous page

Property Property	Datatype
nidcpower.Session.pulse_current_level	float
	float
nidcpower.Session.pulse_current_level_range	
nidcpower.Session.pulse_current_limit	float float
nidcpower.Session.pulse_current_limit_high	
nidcpower.Session.pulse_current_limit_low	float
nidcpower.Session.pulse_current_limit_range	float float in seconds or datetin
nidcpower.Session.pulse_off_time	
nidcpower.Session.pulse_on_time	float in seconds or datetin
nidcpower.Session.pulse_trigger_type	TriggerType
nidcpower.Session.pulse_voltage_level	float
nidcpower.Session.pulse_voltage_level_range	float
nidcpower.Session.pulse_voltage_limit	float
nidcpower.Session.pulse_voltage_limit_high	float
nidcpower.Session.pulse_voltage_limit_low	float
nidcpower.Session.pulse_voltage_limit_range	float
nidcpower.Session.query_instrument_status	bool
nidcpower.Session.ready_for_pulse_trigger_event_output_terminal	str
nidcpower.Session.ready_for_pulse_trigger_event_pulse_polarity	Polarity
nidcpower.Session.ready_for_pulse_trigger_event_pulse_width	float
nidcpower.Session.reset_average_before_measurement	bool
nidcpower.Session.samples_to_average	int
nidcpower.Session.self_calibration_persistence	SelfCalibrationPe
nidcpower.Session.sense	Sense
nidcpower.Session.sequence_advance_trigger_type	TriggerType
nidcpower.Session.sequence_engine_done_event_output_terminal	str
nidcpower.Session.sequence_engine_done_event_pulse_polarity	Polarity
nidcpower.Session.sequence_engine_done_event_pulse_width	float
nidcpower.Session.sequence_iteration_complete_event_output_terminal	str
nidcpower.Session.sequence_iteration_complete_event_pulse_polarity	Polarity
nidcpower.Session.sequence_iteration_complete_event_pulse_width	float
nidcpower.Session.sequence_loop_count	int
nidcpower.Session.sequence_loop_count_is_finite	bool
nidcpower.Session.simulate	bool
nidcpower.Session.source_complete_event_output_terminal	str
nidcpower.Session.source_complete_event_pulse_polarity	Polarity
nidcpower.Session.source_complete_event_pulse_width	float
nidcpower.Session.source_delay	float in seconds or datetin
nidcpower.Session.source_mode	SourceMode
nidcpower.Session.source_trigger_type	TriggerType
nidcpower.Session.specific_driver_description	str
nidcpower.Session.specific_driver_prefix	str
nidcpower.Session.specific_driver_revision	str
nidcpower.Session.specific_driver_vendor	str
nidcpower.Session.start_trigger_type	TriggerType
nidcpower.Session.supported_instrument_models	str
nidcpower.Session.transient_response	TransientResponse
nidcpower.Session.voltage_compensation_frequency	float
nidcpower.Session.voltage_gain_bandwidth	float
nidcpower.Session.voltage_level	float
	Continued

Continued of

Table 2 – continued from previous page

Property	Datatype
nidcpower.Session.voltage_level_autorange	bool
nidcpower.Session.voltage_level_range	float
nidcpower.Session.voltage_limit	float
nidcpower.Session.voltage_limit_autorange	bool
nidcpower.Session.voltage_limit_high	float
nidcpower.Session.voltage_limit_low	float
nidcpower.Session.voltage_limit_range	float
nidcpower.Session.voltage_pole_zero_ratio	float

# Methods

Method name
nidcpower.Session.abort()
nidcpower.Session.commit()
nidcpower.Session.configure_aperture_time()
nidcpower.Session.disable()
nidcpower.Session.fetch_multiple()
nidcpower.Session.get_channel_name()
<pre>nidcpower.Session.get_ext_cal_last_date_and_time()</pre>
nidcpower.Session.get_ext_cal_last_temp()
<pre>nidcpower.Session.get_ext_cal_recommended_interval()</pre>
<pre>nidcpower.Session.get_self_cal_last_date_and_time()</pre>
<pre>nidcpower.Session.get_self_cal_last_temp()</pre>
nidcpower.Session.lock()
nidcpower.Session.measure()
nidcpower.Session.measure_multiple()
<pre>nidcpower.Session.query_in_compliance()</pre>
<pre>nidcpower.Session.query_max_current_limit()</pre>
<pre>nidcpower.Session.query_max_voltage_level()</pre>
<pre>nidcpower.Session.query_min_current_limit()</pre>
<pre>nidcpower.Session.query_output_state()</pre>
<pre>nidcpower.Session.read_current_temperature()</pre>
nidcpower.Session.reset()
nidcpower.Session.reset_device()
<pre>nidcpower.Session.reset_with_defaults()</pre>
nidcpower.Session.self_cal()
nidcpower.Session.self_test()
<pre>nidcpower.Session.send_software_edge_trigger()</pre>
nidcpower.Session.set_sequence()
nidcpower.Session.unlock()
nidcpower.Session.wait_for_event()

# 7.1.3.2 Enums

Enums used in NI-DCPower

# **ApertureTimeUnits**

```
class nidcpower.ApertureTimeUnits
```

#### **SECONDS**

Specifies aperture time in seconds.

#### POWER LINE CYCLES

Specifies aperture time in power line cycles (PLCs).

### **AutoZero**

```
class nidcpower.AutoZero
```

OFF

Disables auto zero.

ON

Makes zero conversions for every measurement.

ONCE

Makes zero conversions following the first measurement after initiating the device. The device uses these zero conversions for the preceding measurement and future measurements until the device is reinitiated.

## ComplianceLimitSymmetry

```
class nidcpower.ComplianceLimitSymmetry
```

## SYMMETRIC

Compliance limits are specified symmetrically about 0.

#### ASYMMETRIC

Compliance limits can be specified asymmetrically with respect to 0.

# **DCNoiseRejection**

```
class nidcpower.DCNoiseRejection
```

### SECOND ORDER

Second-order rejection of DC noise.

### NORMAL

Normal rejection of DC noise.

# **Event**

```
class nidcpower.Event
```

SOURCE\_COMPLETE

```
MEASURE_COMPLETE

SEQUENCE_ITERATION_COMPLETE

SEQUENCE_ENGINE_DONE

PULSE_COMPLETE

READY_FOR_PULSE_TRIGGER
```

## MeasureWhen

class nidcpower.MeasureWhen

#### AUTOMATICALLY AFTER SOURCE COMPLETE

Acquires a measurement after each Source Complete event completes.

### ON DEMAND

Acquires a measurement when the nidcpower.Session.measure() method or nidcpower. Session.measure\_multiple() method is called.

#### ON MEASURE TRIGGER

Acquires a measurement when a Measure trigger is received.

# MeasurementTypes

```
class nidcpower.MeasurementTypes
```

#### CURRENT

The device measures current.

#### **VOLTAGE**

The device measures voltage.

## **OutputCapacitance**

```
class nidcpower.OutputCapacitance
```

LOW

Output Capacitance is low.

HIGH

Output Capacitance is high.

## **OutputFunction**

class nidcpower.OutputFunction

#### DC VOLTAGE

Sets the output method to DC voltage.

### DC\_CURRENT

Sets the output method to DC current.

#### PULSE VOLTAGE

Sets the output method to pulse voltage.

#### PULSE CURRENT

Sets the output method to pulse current.

## **OutputStates**

## class nidcpower.OutputStates

#### **VOLTAGE**

The device maintains a constant voltage by adjusting the current

#### CURRENT

The device maintains a constant current by adjusting the voltage.

# **Polarity**

### class nidcpower.Polarity

#### HIGH

A high pulse occurs when the event is generated. The exported signal is low level both before and after the event is generated.

#### LOW

A low pulse occurs when the event is generated. The exported signal is high level both before and after the event is generated.

## **PowerSource**

## class nidcpower.PowerSource

# INTERNAL

Uses the PXI chassis power source.

### AUXILIARY

Uses the auxiliary power source connected to the device.

#### AUTOMATIC

Uses the auxiliary power source if it is available; otherwise uses the PXI chassis power source.

## **PowerSourceInUse**

### class nidcpower.PowerSourceInUse

## INTERNAL

Uses the PXI chassis power source.

#### **AUXILIARY**

Uses the auxiliary power source connected to the device. Only the NI PXI-4110, NI PXIe-4112, NI PXIe-4113, and NI PXI-4130 support this value. This is the only supported value for the NI PXIe-4112 and NI PXIe-4113.

#### **SelfCalibrationPersistence**

```
class nidcpower.SelfCalibrationPersistence
```

### KEEP\_IN\_MEMORY

Keep new self calibration values in memory only.

### WRITE\_TO\_EEPROM

Write new self calibration values to hardware.

## SendSoftwareEdgeTriggerType

```
class nidcpower.SendSoftwareEdgeTriggerType
```

START

SOURCE

**MEASURE** 

SEQUENCE\_ADVANCE

PULSE

### Sense

```
class nidcpower.Sense
```

# LOCAL

Local sensing is selected.

#### REMOTE

Remote sensing is selected.

#### SourceMode

```
class nidcpower.SourceMode
```

#### SINGLE POINT

The source unit applies a single source configuration.

### **SEQUENCE**

The source unit applies a list of voltage or current configurations sequentially.

## **TransientResponse**

### class nidcpower. Transient Response

#### NORMAL

The output responds to changes in load at a normal speed.

#### FAST

The output responds to changes in load quickly.

#### SLOW

The output responds to changes in load slowly.

#### CUSTOM

The output responds to changes in load based on specified values.

## **TriggerType**

```
class nidcpower.TriggerType
```

#### NONE

No trigger is configured.

### DIGITAL EDGE

The data operation starts when a digital edge is detected.

#### SOFTWARE EDGE

The data operation starts when a software trigger occurs.

## 7.1.3.3 Exceptions and Warnings

### **DriverError**

```
exception nidepower.DriverError
An error originating from the NI-DCPower driver
```

## UnsupportedConfigurationError

```
exception nidepower.UnsupportedConfigurationError
An error due to using this module in an usupported platform.
```

#### **DriverNotInstalledError**

### exception nidcpower.DriverNotInstalledError

An error due to using this module without the driver runtime installed.

## InvalidRepeatedCapabilityError

### exception nidcpower.InvalidRepeatedCapabilityError

An error due to an invalid character in a repeated capability

### SelfTestError

```
exception nidepower.SelfTestError
An error due to a failed self-test
```

## **DriverWarning**

```
exception nidcpower.DriverWarning
A warning originating from the NI-DCPower driver
```

#### **7.1.3.4 Examples**

### nidcpower measure record.py

Listing 1: (nidcpower\_measure\_record.py)

```
#!/usr/bin/python
2
   import argparse
   import nidcpower
   import sys
   def example(resource_name, channels, options, voltage, length):
8
       with nidcpower.Session(resource_name=resource_name, channels=channels,_
9
   →options=options) as session:
10
           # Configure the session.
11
           session.measure_record_length = length
12
           session.measure_record_length_is_finite = True
13
           session.measure_when = nidcpower.MeasureWhen.AUTOMATICALLY_AFTER_SOURCE_
14
   →COMPLETE
           session.voltage_level = voltage
15
           session.commit()
17
           print('Effective measurement rate: {0} S/s'.format(session.measure_record_
18
   \rightarrow delta_time / 1))
19
           samples\_acquired = 0
20
           print(' # Voltage
                                                In Compliance')
                                   Current
21
           row_format = '{0:3d}:
                                              {2:8.6f} {3}'
                                   {1:8.6f}
22
           with session.initiate():
23
               while samples_acquired < length:</pre>
24
                    measurements = session.fetch_multiple(count=session.fetch_backlog)
25
                    samples_acquired += len(measurements)
26
                    for i in range(len(measurements)):
27
                        print(row_format.format(i, measurements[i].voltage,_
   →measurements[i].current, measurements[i].in_compliance))
29
30
   def _main(argsv):
31
       parser = argparse.ArgumentParser(description='Outputs the specified voltage, then_
32
   →takes the specified number of voltage and current readings.', formatter_
   →class=argparse.ArgumentDefaultsHelpFormatter)
```

(continues on next page)

(continued from previous page)

```
parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource...
33
   →name of a National Instruments SMU')
       parser.add_argument('-c', '--channels', default='0', help='Channel(s) to use')
34
       parser.add_argument('-1', '--length', default='20', type=int, help='Measure_
35
   →record length')
       parser.add_argument('-v', '--voltage', default=5.0, type=float, help='Voltage_
36
   →level (V)')
       parser.add_argument('-op', '--option-string', default='', type=str, help='Option.
37
   ⇔string')
       args = parser.parse_args(argsv)
38
       example(args.resource_name, args.channels, args.option_string, args.voltage, args.
39
   →length)
41
   def main():
42.
       _main(sys.argv[1:])
43
44
45
   def test_example():
46
       options = {'simulate': True, 'driver_setup': {'Model': '4162', 'BoardType': 'PXIe
47
   example ('PXI1Slot2', '0', options, 5.0, 20)
48
40
50
   def test_main():
       cmd_line = ['--option-string', 'Simulate=1, DriverSetup=Model:4162; BoardType:PXIe
   _main(cmd_line)
53
54
55
   if __name__ == '__main__':
56
57
       main()
58
59
```

### nidcpower source delay measure.py

Listing 2: (nidcpower\_source\_delay\_measure.py)

```
#!/usr/bin/python
2
   import argparse
   import datetime
   import nidcpower
   import sys
   def print_fetched_measurements(measurements):
9
       print('
                          Voltage : {:f} V'.format(measurements[0].voltage))
10
                            Current: {:f} A'.format(measurements[0].current))
       print('
11
                     In compliance: {0}'.format(measurements[0].in_compliance))
       print('
12
13
14
   def example(resource_name, channels, options, voltage1, voltage2, delay):
```

(continues on next page)

(continued from previous page)

```
timeout = datetime.timedelta(seconds=(delay + 1.0))
16
17
       with nidcpower.Session(resource_name=resource_name, channels=channels,__
18
   →options=options) as session:
19
            # Configure the session.
20
           session.source_mode = nidcpower.SourceMode.SINGLE_POINT
21
           session.output_function = nidcpower.OutputFunction.DC_VOLTAGE
22
           session.current_limit = .06
23
           session.voltage_level_range = 5.0
24
           session.current_limit_range = .06
25
           session.source_delay = datetime.timedelta(seconds=delay)
           session.measure_when = nidcpower.MeasureWhen.AUTOMATICALLY_AFTER_SOURCE_
   COMPLETE
           session.voltage_level = voltage1
28
29
           with session.initiate():
30
               print('Voltage 1:')
31
                print_fetched_measurements(session.fetch_multiple(count=1,_
32
   →timeout=timeout))
                session.voltage_level = voltage2 # on-the-fly set
33
               print('Voltage 2:')
34
               print_fetched_measurements(session.fetch_multiple(count=1,_
35
   →timeout=timeout))
                session.output_enabled = False
37
38
   def main(argsv):
39
       parser = argparse.ArgumentParser(description='Outputs voltage 1, waits for source...)
40
   →delay, and then takes a measurement. Then orepeat with voltage 2.', formatter_
   →class=argparse.ArgumentDefaultsHelpFormatter)
41
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource_
   →name of a National Instruments SMU')
       parser.add_argument('-c', '--channels', default='0', help='Channel(s) to use')
42
       parser.add_argument('-v1', '--voltage1', default=1.0, type=float, help='Voltage...
43
   \rightarrow level 1 (V)')
       parser.add_argument('-v2', '--voltage2', default=2.0, type=float, help='Voltage_
44
   \rightarrowlevel 2 (V)')
       parser.add_argument('-d', '--delay', default=0.05, type=float, help='Source delay_
   \hookrightarrow (s) ')
       parser.add_argument('-op', '--option-string', default='', type=str, help='Option.
46
   →string')
47
       args = parser.parse_args(argsv)
       example(args.resource_name, args.channels, args.option_string, args.voltage1,...
   ⇒args.voltage2, args.delay)
49
50
   def main():
51
52
       _main(sys.argv[1:])
53
54
   def test_main():
       cmd_line = ['--option-string', 'Simulate=1, DriverSetup=Model:4162; BoardType:PXIe
56
   _main(cmd_line)
57
58
```

(continues on next page)

(continued from previous page)

```
def test_example():
    options = {'simulate': True, 'driver_setup': {'Model': '4162', 'BoardType': 'PXIe
    o', },
    example('PXI1Slot2', '0', options, 1.0, 2.0, 0.05)

if __name__ == '__main__':
    main()
```

# 7.2 nidmm module

#### 7.2.1 Installation

As a prerequisite to using the nidmm module, you must install the NI-DMM runtime on your system. Visit ni.com/downloads to download the driver runtime for your devices.

The nimi-python modules (i.e. for **NI-DMM**) can be installed with pip:

```
$ python -m pip install nidmm~=1.0.1
```

Or easy\_install from setuptools:

```
$ python -m easy_install nidmm
```

## **7.2.2 Usage**

The following is a basic example of using the **nidmm** module to open a session to a DMM and perform a 5.5 digits of resolution voltage measurement in the 10 V range.

```
import nidmm
with nidmm.Session("Dev1") as session:
    session.configureMeasurementDigits(nidmm.Function.DC_VOLTS, 10, 5.5)
    print("Measurement: " + str(session.read()))
```

Additional examples for NI-DMM are located in src/nidmm/examples/ directory.

# 7.2.3 API Reference

#### 7.2.3.1 nidmm.Session

class nidmm.Session(self, resource\_name, id\_query=False, reset\_device=False, options={})
 This method completes the following tasks:

• Creates a new IVI instrument driver session and, optionally, sets the initial state of the following session properties: nidmm.Session.range\_check, nidmm.Session.QUERY\_INSTR\_STATUS, nidmm.Session.cache, nidmm.Session.simulate, nidmm.Session.record\_coercions.

- Opens a session to the device you specify for the Resource\_Name parameter. If the ID\_Query parameter
  is set to True, this method queries the instrument ID and checks that it is valid for this instrument driver.
- If the **Reset\_Device** parameter is set to True, this method resets the instrument to a known state. Sends initialization commands to set the instrument to the state necessary for the operation of the instrument driver.
- Returns a ViSession handle that you use to identify the instrument in all subsequent instrument driver method calls.

**Note:** One or more of the referenced properties are not in the Python API for this driver.

#### **Parameters**

• resource\_name (str) -

**Caution:** All IVI names for the **Resource\_Name**, such as logical names or virtual names, are case-sensitive. If you use logical names, driver session names, or virtual names in your program, you must make sure that the name you use matches the name in the IVI Configuration Store file exactly, without any variations in the case of the characters in the name.

Contains the **resource\_name** of the device to initialize. The **resource\_name** is assigned in Measurement & Automation Explorer (MAX). Refer to Related Documentation for the *NI Digital Multimeters Getting Started Guide* for more information about configuring and testing the DMM in MAX.

Valid Syntax:

- NI-DAQmx name
- DAQ::NI-DAQmx name[::INSTR]
- DAQ::Traditional NI-DAQ device number[::INSTR]
- IVI logical name
- id\_query (bool) Verifies that the device you initialize is one that the driver supports. NI-DMM automatically performs this query, so setting this parameter is not necessary. Defined Values:

True (default)	1	Perform ID Query
False	0	Skip ID Query

• **reset\_device** (bool) – Specifies whether to reset the instrument during the initialization procedure. Defined Values:

True (default)	1	Reset Device
False	0	Don't Reset

• **options** (str) – Specifies the initial value of certain properties for the session. The syntax for **options** is a dictionary of properties with an assigned value. For example:

```
{ 'simulate': False }
```

You do not have to specify a value for all the properties. If you do not specify a value for a property, the default value is used.

Advanced Example: { 'simulate': True, 'driver\_setup': { 'Model': '<model number>', 'BoardType': '<type>' } }

Property	Default
range_check	True
query_instrument_status	False
cache	True
simulate	False
record_value_coersions	False
driver_setup	{}

# **Properties**

Property	Datatype
ac_max_freq	float
ac_min_freq	float
adc_calibration	ADCCalibration
aperture_time	float
aperture_time_units	ApertureTimeUnits
auto_range_value	float
auto_zero	AutoZero
buffer_size	int
cable_comp_type	CableCompensationType
channel_count	int
current_source	float
dc_bias	bool
dc_noise_rejection	DCNoiseRejection
driver_setup	str
freq_voltage_auto_range	float
freq_voltage_range	float
function	Function
input_resistance	float
<pre>instrument_firmware_revision</pre>	str
instrument_manufacturer	str
instrument_model	str
instrument_product_id	int
io_resource_descriptor	str
lc_calculation_model	LCCalculationModel
lc_number_meas_to_average	int
logical_name	str
meas_complete_dest	<i>MeasurementCompleteDest</i>
number_of_averages	int
offset_comp_ohms	bool
open_cable_comp_conductance	float
open_cable_comp_susceptance	float
operation_mode	OperationMode
powerline_freq	float

Continued on next page

Table 3 – continued from previous page

Property	Datatype
range	float
resolution_absolute	float
resolution_digits	float
sample_count	int
sample_interval	float in seconds or datetime.timedelta
sample_trigger	SampleTrigger
serial_number	str
settle_time	float in seconds or datetime.timedelta
short_cable_comp_reactance	float
short_cable_comp_resistance	float
simulate	bool
specific_driver_description	str
specific_driver_major_version	int
specific_driver_minor_version	int
specific_driver_revision	str
specific_driver_vendor	str
supported_instrument_models	str
temp_rtd_a	float
temp_rtd_b	float
temp_rtd_c	float
temp_rtd_res	float
temp_rtd_type	RTDType
temp_tc_fixed_ref_junc	float
temp_tc_ref_junc_type	ThermocoupleReferenceJunctionType
temp_tc_type	ThermocoupleType
temp_thermistor_a	float
temp_thermistor_b	float
temp_thermistor_c	float
temp_thermistor_type	ThermistorType
temp_transducer_type	TransducerType
trigger_count	int
trigger_delay	float in seconds or datetime.timedelta
trigger_source	TriggerSource
waveform_coupling	WaveformCoupling
waveform_points	int
waveform_rate	float

# **Public methods**

Method name	
abort()	
<pre>configure_measurement_absolute()</pre>	
<pre>configure_measurement_digits()</pre>	
configure_multi_point()	
configure_rtd_custom()	
<pre>configure_rtd_type()</pre>	
<pre>configure_thermistor_custom()</pre>	
configure_thermocouple()	
configure_trigger()	

Continued on next page

Table 4 – continued from previous page

Method name
configure_waveform_acquisition()
disable()
fetch()
fetch_multi_point()
<pre>fetch_waveform()</pre>
<pre>fetch_waveform_into()</pre>
<pre>get_cal_date_and_time()</pre>
<pre>get_dev_temp()</pre>
<pre>get_ext_cal_recommended_interval()</pre>
<pre>get_last_cal_temp()</pre>
<pre>get_self_cal_supported()</pre>
lock()
<pre>perform_open_cable_comp()</pre>
<pre>perform_short_cable_comp()</pre>
read()
read_multi_point()
read_status()
read_waveform()
reset()
reset_with_defaults()
self_cal()
self_test()
send_software_trigger()
unlock()

## **Properties**

### ac\_max\_freq

### nidmm.Session.ac\_max\_freq

Specifies the maximum frequency component of the input signal for AC measurements. This property is used only for error checking and verifies that the value of this parameter is less than the maximum frequency of the device. This property affects the DMM only when you set the nidmm. Session.method property to AC measurements. The valid range is 1 Hz-300 kHz for the NI 4070/4071/4072, 10 Hz-100 kHz for the NI 4065, and 20 Hz-25 kHz for the NI 4050 and NI 4060.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Max Frequency
- C Attribute: NIDMM\_ATTR\_AC\_MAX\_FREQ

## ac\_min\_freq

#### nidmm.Session.ac\_min\_freq

Specifies the minimum frequency component of the input signal for AC measurements. This property affects the DMM only when you set the nidmm. Session.method property to AC measurements. The valid range is 1 Hz-300 kHz for the NI 4070/4071/4072, 10 Hz-100 kHz for the NI 4065, and 20 Hz-25 kHz for the NI 4050 and NI 4060.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Min Frequency
- C Attribute: NIDMM\_ATTR\_AC\_MIN\_FREQ

### adc calibration

#### nidmm.Session.adc\_calibration

For the NI 4070/4071/4072 only, specifies the ADC calibration mode.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ADCCalibration
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration: Measurement Options: ADC Calibration
- C Attribute: NIDMM\_ATTR\_ADC\_CALIBRATION

# aperture\_time

## nidmm.Session.aperture\_time

Specifies the measurement aperture time for the current configuration. Aperture time is specified in units set by <code>nidmm.Session.aperture\_time\_units</code>. To override the default aperture, set this property to the desired aperture time after calling <code>nidmm.Session.ConfigureMeasurement()</code>. To return to the default, set this property to <code>NIDMM\_VAL\_APERTURE\_TIME\_AUTO(-1)</code>. On the NI 4070/4071/4072, the minimum aperture

time is 8.89 usec, and the maximum aperture time is 149 sec. Any number of powerline cycles (PLCs) within the minimum and maximum ranges is allowed on the NI 4070/4071/4072. On the NI 4065 the minimum aperture time is 333  $\mu$ s, and the maximum aperture time is 78.2 s. If setting the number of averages directly, the total measurement time is aperture time X the number of averages, which must be less than 72.8 s. The aperture times allowed are 333  $\mu$ s, 667  $\mu$ s, or multiples of 1.11 ms-for example 1.11 ms, 2.22 ms, 3.33 ms, and so on. If you set an aperture time other than 333  $\mu$ s, 667  $\mu$ s, or multiples of 1.11 ms, the value will be coerced up to the next supported aperture time. On the NI 4060, when the powerline frequency is 60 Hz, the PLCs allowed are 1 PLC, 6 PLC, 12 PLC, and 120 PLC. When the powerline frequency is 50 Hz, the PLCs allowed are 1 PLC, 5 PLC, 10 PLC, and 100 PLC.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Advanced:Aperture Time
- C Attribute: NIDMM\_ATTR\_APERTURE\_TIME

## aperture\_time\_units

nidmm.Session.aperture\_time\_units

Specifies the units of aperture time for the current configuration. The NI 4060 does not support an aperture time set in seconds.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ApertureTimeUnits
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Advanced:Aperture Time Units
- C Attribute: NIDMM\_ATTR\_APERTURE\_TIME\_UNITS

### auto\_range\_value

nidmm.Session.auto\_range\_value

Specifies the value of the range. If auto ranging, shows the actual value of the active range. The value of this property is set during a read operation.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration: Auto Range Value
- C Attribute: NIDMM\_ATTR\_AUTO\_RANGE\_VALUE

## auto\_zero

nidmm.Session.auto zero

Specifies the AutoZero mode. The NI 4050 is not supported.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.AutoZero
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Auto Zero
- C Attribute: NIDMM\_ATTR\_AUTO\_ZERO

## buffer size

nidmm.Session.buffer\_size

Size in samples of the internal data buffer. Maximum is 134,217,727 (OX7FFFFF) samples. When set to NIDMM\_VAL\_BUFFER\_SIZE\_AUTO (-1), NI-DMM chooses the buffer size.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Multi Point Acquisition: Advanced: Buffer Size
- C Attribute: NIDMM\_ATTR\_BUFFER\_SIZE

### cable\_comp\_type

#### nidmm.Session.cable\_comp\_type

For the NI 4072 only, the type of cable compensation that is applied to the current capacitance or inductance measurement for the current range. Changing the method or the range through this property or through <code>nidmm.Session.configure\_measurement\_digits()</code> resets the value of this property to the default value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.CableCompensationType
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Capacitance and Inductance:Cable Compensation Type
- C Attribute: NIDMM ATTR CABLE COMP TYPE

### channel count

#### nidmm.Session.channel\_count

Indicates the number of channels that the specific instrument driver supports. For each property for which the IVI\_VAL\_MULTI\_CHANNEL flag property is set, the IVI engine maintains a separate cache value for each channel.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Capabilities:Channel Count
- C Attribute: NIDMM\_ATTR\_CHANNEL\_COUNT

## current\_source

nidmm.Session.current\_source

Specifies the current source provided during diode measurements. The NI 4050 and NI 4060 are not supported.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Current Source
- C Attribute: NIDMM\_ATTR\_CURRENT\_SOURCE

## dc\_bias

nidmm.Session.dc bias

For the NI 4072 only, controls the available DC bias for capacitance measurements.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Capacitance and Inductance:Advanced:DC Bias
- C Attribute: NIDMM\_ATTR\_DC\_BIAS

### dc\_noise\_rejection

### nidmm.Session.dc\_noise\_rejection

Specifies the DC noise rejection mode. The NI 4050 and NI 4060 are not supported.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.DCNoiseRejection
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration: Measurement Options: DC Noise Rejection
- C Attribute: NIDMM\_ATTR\_DC\_NOISE\_REJECTION

## driver\_setup

### nidmm.Session.driver\_setup

This property indicates the Driver Setup string that the user specified when initializing the driver. Some cases exist where the end-user must specify instrument driver options at initialization time. An example of this is specifying a particular instrument model from among a family of instruments that the driver supports. This is useful when using simulation. The end-user can specify driver-specific options through the DriverSetup keyword in the optionsString parameter to the niDMM Init With Options.vi. If the user does not specify a Driver Setup string, this property returns an empty string.

The following table lists the characteristics of this property.

Value
str
read only
False
No

Tip: This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: User Options: Driver Setup
- C Attribute: NIDMM\_ATTR\_DRIVER\_SETUP

### freq voltage auto range

#### nidmm.Session.freq\_voltage\_auto\_range

For the NI 4070/4071/4072 only, specifies the value of the frequency voltage range. If Auto Ranging, shows the actual value of the active frequency voltage range. If not Auto Ranging, the value of this property is the same as that of nidmm.Session.freq\_voltage\_range.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Frequency Voltage Auto Range Value
- C Attribute: NIDMM\_ATTR\_FREQ\_VOLTAGE\_AUTO\_RANGE

# freq\_voltage\_range

#### nidmm.Session.freq\_voltage\_range

Specifies the maximum amplitude of the input signal for frequency measurements.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Frequency Voltage Range
- C Attribute: NIDMM\_ATTR\_FREQ\_VOLTAGE\_RANGE

#### function

## nidmm.Session.function

Specifies the measurement method. Refer to the nidmm.Session.method topic in the NI Digital Multimeters Help for device-specific information. If you are setting this property directly, you must also set the nidmm.Session.operation\_mode property, which controls whether the DMM takes standard single or multipoint measurements, or acquires a waveform. If you are

programming properties directly, you must set the <code>nidmm.Session.operation\_mode</code> property before setting other configuration properties. If the <code>nidmm.Session.operation\_mode</code> property is set to <code>WAVEFORM</code>, the only valid method types are <code>WAVEFORM\_VOLTAGE</code> and <code>WAVEFORM\_CURRENT</code>. Set the <code>nidmm.Session.operation\_mode</code> property to <code>IVIDMM</code> to set all other method values.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.Function
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Configuration:Function

• C Attribute: NIDMM\_ATTR\_FUNCTION

## input\_resistance

nidmm.Session.input\_resistance

Specifies the input resistance of the instrument. The NI 4050 and NI 4060 are not supported.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Configuration:Measurement Options:Input Resistance

• C Attribute: NIDMM\_ATTR\_INPUT\_RESISTANCE

# instrument\_firmware\_revision

nidmm.Session.instrument\_firmware\_revision

A string containing the instrument firmware revision number.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Instrument Firmware Revision
- C Attribute: NIDMM\_ATTR\_INSTRUMENT\_FIRMWARE\_REVISION

## instrument\_manufacturer

nidmm.Session.instrument\_manufacturer

A string containing the manufacturer of the instrument.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Instrument Manufacturer
- C Attribute: NIDMM\_ATTR\_INSTRUMENT\_MANUFACTURER

## instrument\_model

nidmm.Session.instrument\_model

A string containing the instrument model.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Instrument Model
- C Attribute: NIDMM\_ATTR\_INSTRUMENT\_MODEL

### instrument\_product\_id

nidmm.Session.instrument\_product\_id

The PCI product ID.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Instrument Product ID
- C Attribute: NIDMM\_ATTR\_INSTRUMENT\_PRODUCT\_ID

### io resource descriptor

nidmm.Session.io resource descriptor

A string containing the resource descriptor of the instrument.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Advanced Session Information: I/O Resource Descriptor
- C Attribute: NIDMM\_ATTR\_IO\_RESOURCE\_DESCRIPTOR

## lc\_calculation\_model

nidmm.Session.lc\_calculation\_model

For the NI 4072 only, specifies the type of algorithm that the measurement processing uses for capacitance and inductance measurements.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.LCCalculationModel
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Capacitance and Inductance:Advanced:Calculation Model
- C Attribute: NIDMM\_ATTR\_LC\_CALCULATION\_MODEL

# lc\_number\_meas\_to\_average

nidmm.Session.lc\_number\_meas\_to\_average

For the NI 4072 only, specifies the number of LC measurements that are averaged to produce one reading.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Capacitance and Inductance:Number of LC Measurements To Average
- C Attribute: NIDMM\_ATTR\_LC\_NUMBER\_MEAS\_TO\_AVERAGE

## logical\_name

nidmm.Session.logical\_name

A string containing the logical name of the instrument.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Advanced Session Information: Logical Name
- C Attribute: NIDMM\_ATTR\_LOGICAL\_NAME

## meas\_complete\_dest

## nidmm.Session.meas\_complete\_dest

Specifies the destination of the measurement complete (MC) signal. The NI 4050 is not supported. To determine which values are supported by each device, refer to the LabWindows/CVI Trigger Routing section in the NI Digital Multimeters Help.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.MeasurementCompleteDest
Permissions	read-write
Channel Based	False
Resettable	No

Tip: This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Trigger:Measurement Complete Dest
- C Attribute: NIDMM\_ATTR\_MEAS\_COMPLETE\_DEST

### number of averages

## nidmm.Session.number\_of\_averages

Specifies the number of averages to perform in a measurement. For the NI 4070/4071/4072, applies only when the aperture time is not set to AUTO and Auto Zero is ON. The default is 1. The NI 4050 and NI 4060 are not supported.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Advanced:Number Of Averages
- C Attribute: NIDMM\_ATTR\_NUMBER\_OF\_AVERAGES

### offset comp ohms

#### nidmm.Session.offset\_comp\_ohms

For the NI 4070/4071/4072 only, enables or disables offset compensated ohms.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration: Measurement Options: Offset Compensated Ohms
- C Attribute: NIDMM\_ATTR\_OFFSET\_COMP\_OHMS

### open cable comp conductance

### nidmm.Session.open\_cable\_comp\_conductance

For the NI 4072 only, specifies the active part (conductance) of the open cable compensation. The valid range is any real number greater than 0. The default value (-1.0) indicates that compensation has not taken place. Changing the method or the range through this property or through <code>nidmm.Session.configure\_measurement\_digits()</code> resets the value of this property to the default value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Capacitance and Inductance:Open Cable Compensation Values:Conductance
- C Attribute: NIDMM\_ATTR\_OPEN\_CABLE\_COMP\_CONDUCTANCE

### open cable comp susceptance

# nidmm.Session.open\_cable\_comp\_susceptance

For the NI 4072 only, specifies the reactive part (susceptance) of the open cable compensation. The valid range is any real number greater than 0. The default value (-1.0) indicates that compensation has not taken place. Changing the method or the range through this property or through

nidmm.Session.configure\_measurement\_digits() resets the value of this property to the default value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Capacitance and Inductance:Open Cable Compensation Values:Susceptance
- C Attribute: NIDMM ATTR OPEN CABLE COMP SUSCEPTANCE

## operation mode

#### nidmm.Session.operation\_mode

Specifies how the NI 4065 and NI 4070/4071/4072 acquire data. When you call <code>nidmm.Session.configure\_measurement\_digits()</code>, NI-DMM sets this property to <code>IVIDMM</code>. When you call <code>nidmm.Session.configure\_waveform\_acquisition()</code>, NI-DMM sets this property to <code>WAVEFORM</code>. If you are programming properties directly, you must set this property before setting other configuration properties.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.OperationMode
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Advanced:Operation Mode
- C Attribute: NIDMM\_ATTR\_OPERATION\_MODE

# powerline\_freq

# nidmm.Session.powerline\_freq

Specifies the powerline frequency. The NI 4050 and NI 4060 use this value to select an aperture time to reject powerline noise by selecting the appropriate internal sample clock and filter. The NI 4065 and NI 4070/4071/4072 use this value to select a timebase for setting the <code>nidmm.Session.aperture\_time</code> property in powerline cycles (PLCs). After configuring powerline frequency, set the <code>nidmm.Session.aperture\_time\_units</code> property to PLCs. When setting the <code>nidmm.Session.aperture\_time</code> property, select the number of PLCs for the powerline frequency.

For example, if powerline frequency = 50 Hz (or 20 ms) and aperture time in PLCs = 5, then aperture time in Seconds = 20 ms \* 5 PLCs = 100 ms. Similarly, if powerline frequency = 60 Hz (or 16.667 ms) and aperture time in PLCs = 6, then aperture time in Seconds = 16.667 ms \* 6 PLCs = 100 ms.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Powerline Frequency
- C Attribute: NIDMM ATTR POWERLINE FREQ

#### range

nidmm.Session.range

Specifies the measurement range. Use positive values to represent the absolute value of the maximum expected measurement. The value is in units appropriate for the current value of the nidmm.Session.method property. For example, if nidmm.Session.method is set to NIDMM\_VAL\_VOLTS, the units are volts. The NI 4050 and NI 4060 only support Auto Range when the trigger and sample trigger is set to IMMEDIATE. NIDMM\_VAL\_AUTO\_RANGE\_ON -1.0 NI-DMM performs an Auto Range before acquiring the measurement. NIDMM\_VAL\_AUTO\_RANGE\_OFF -2.0 NI-DMM sets the Range to the current nidmm.Session.auto\_range\_value and uses this range for all subsequent measurements until the measurement configuration is changed. NIDMM\_VAL\_AUTO\_RANGE\_ONCE -3.0 NI-DMM performs an Auto Range before acquiring the next measurement. The nidmm.Session.auto\_range\_value is stored and used for all subsequent measurements until the measurement configuration is changed.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Configuration:Range

• C Attribute: NIDMM\_ATTR\_RANGE

## resolution\_absolute

#### nidmm.Session.resolution absolute

Specifies the measurement resolution in absolute units. Setting this property to higher values increases the measurement accuracy. Setting this property to lower values increases the measurement speed. NI-DMM ignores this property for capacitance and inductance measurements on the NI 4072. To achieve better resolution for such measurements, use the <code>nidmm.Session.lc\_number\_meas\_to\_average</code> property.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Configuration: Absolute Resolution

• C Attribute: NIDMM\_ATTR\_RESOLUTION\_ABSOLUTE

## resolution\_digits

### nidmm.Session.resolution\_digits

Specifies the measurement resolution in digits. Setting this property to higher values increases the measurement accuracy. Setting this property to lower values increases the measurement speed. NI-DMM ignores this property for capacitance and inductance measurements on the NI 4072. To achieve better resolution for such measurements, use the <code>nidmm.Session.lc\_number\_meas\_to\_average</code> property.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Configuration:Digits Resolution

• C Attribute: NIDMM\_ATTR\_RESOLUTION\_DIGITS

## sample count

nidmm.Session.sample\_count

Specifies the number of measurements the DMM takes each time it receives a trigger in a multiple point acquisition.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Multi Point Acquisition:Sample Count
- C Attribute: NIDMM\_ATTR\_SAMPLE\_COUNT

## sample\_interval

nidmm.Session.sample\_interval

Specifies the amount of time in seconds the DMM waits between measurement cycles. This property only applies when the <code>nidmm.Session.sample\_trigger</code> property is set to INTERVAL. On the NI 4060, the value for this property is used as the settling time. When this property is set to 0, the NI 4060 does not settle between measurement cycles. The onboard timing resolution is 1 µs on the NI 4060. The NI 4065 and NI 4070/4071/4072 use the value specified in this property as additional delay. On the NI 4065 and NI 4070/4071/4072, the onboard timing resolution is 34.72 ns and the valid range is 0-149 s. Only positive values are valid when setting the sample interval. The NI 4050 is not supported.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Multi Point Acquisition:Sample Interval
- C Attribute: NIDMM\_ATTR\_SAMPLE\_INTERVAL

## sample\_trigger

nidmm.Session.sample\_trigger

Specifies the sample trigger source. To determine which values are supported by each device, refer

to the LabWindows/CVI Trigger Routing section in the NI Digital Multimeters Help.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.SampleTrigger
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Multi Point Acquisition:Sample Trigger
- C Attribute: NIDMM\_ATTR\_SAMPLE\_TRIGGER

# serial\_number

nidmm.Session.serial\_number

A string containing the serial number of the instrument. This property corresponds to the serial number label that is attached to most products.

The following table lists the characteristics of this property.

Value
str
read only
False
No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Instrument Serial Number
- C Attribute: NIDMM\_ATTR\_SERIAL\_NUMBER

## settle\_time

nidmm.Session.settle\_time

Specifies the settling time in seconds. To override the default settling time, set this property. To return to the default, set this property to <code>NIDMM\_VAL\_SETTLE\_TIME\_AUTO</code> (-1). The NI 4050 and NI 4060 are not supported.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

• LabVIEW Property: Configuration:Advanced:Settle Time

• C Attribute: NIDMM\_ATTR\_SETTLE\_TIME

## short cable comp reactance

#### nidmm.Session.short\_cable\_comp\_reactance

For the NI 4072 only, represents the reactive part (reactance) of the short cable compensation. The valid range is any real number greater than 0. The default value (-1) indicates that compensation has not taken place. Changing the method or the range through this property or through <code>nidmm.Session.configure\_measurement\_digits()</code> resets the value of this property to the default value.

The following table lists the characteristics of this property.

Value
float
read-write
False
No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Capacitance and Inductance:Short Cable Compensation Values:Reactance
- C Attribute: NIDMM\_ATTR\_SHORT\_CABLE\_COMP\_REACTANCE

# short\_cable\_comp\_resistance

# $\verb| nidmm.Session.short_cable_comp_resistance| \\$

For the NI 4072 only, represents the active part (resistance) of the short cable compensation. The valid range is any real number greater than 0. The default value (-1) indicates that compensation has not taken place. Changing the method or the range through this property or through <code>nidmm.Session.configure\_measurement\_digits()</code> resets the value of this property to the default value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Configuration:Measurement Options:Capacitance and Inductance:Short Cable Compensation Values:Resistance
- C Attribute: NIDMM\_ATTR\_SHORT\_CABLE\_COMP\_RESISTANCE

#### simulate

#### nidmm.Session.simulate

Specifies whether or not to simulate instrument driver I/O operations. If simulation is enabled, instrument driver methods perform range checking and call IVI Get and Set methods, but they do not perform instrument I/O. For output parameters that represent instrument data, the instrument driver methods return calculated values. The default value is False (0). Use the nidmm.Session. \_\_init\_\_() method to override this setting. Simulate can only be set within the InitWithOptions method. The property value cannot be changed outside of the method.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: User Options: Simulate
- C Attribute: NIDMM ATTR SIMULATE

### specific driver description

 $\verb|nidmm.Session.specific_driver_description| \\$ 

A string containing a description of the specific driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

- LabVIEW Property: Inherent IVI Attributes: Specific Driver Identification: Specific Driver Description
- C Attribute: NIDMM ATTR SPECIFIC DRIVER DESCRIPTION

## specific\_driver\_major\_version

 $\verb| nidmm.Session.specific_driver_major_version| \\$ 

Returns the major version number of this instrument driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Version Info: Specific Driver Major Version
- C Attribute: NIDMM\_ATTR\_SPECIFIC\_DRIVER\_MAJOR\_VERSION

# specific\_driver\_minor\_version

nidmm.Session.specific\_driver\_minor\_version

The minor version number of this instrument driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Version Info: Specific Driver Minor Version
- C Attribute: NIDMM\_ATTR\_SPECIFIC\_DRIVER\_MINOR\_VERSION

## specific driver revision

nidmm.Session.specific\_driver\_revision

A string that contains additional version information about this specific instrument driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Version Info: Specific Driver Revision
- C Attribute: NIDMM\_ATTR\_SPECIFIC\_DRIVER\_REVISION

## specific driver vendor

nidmm.Session.specific\_driver\_vendor

A string containing the vendor of the specific driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Specific Driver Identification: Specific Driver Vendor
- C Attribute: NIDMM\_ATTR\_SPECIFIC\_DRIVER\_VENDOR

### supported\_instrument\_models

nidmm.Session.supported\_instrument\_models

A string containing the instrument models supported by the specific driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

- LabVIEW Property: Inherent IVI Attributes:Specific Driver Capabilities:Supported Instrument Models
- C Attribute: NIDMM\_ATTR\_SUPPORTED\_INSTRUMENT\_MODELS

## temp\_rtd\_a

nidmm.Session.temp\_rtd\_a

Specifies the Callendar-Van Dusen A coefficient for RTD scaling when the RTD Type property is set to Custom. The default value is 3.9083e-3 (Pt3851).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Temperature:Resistance Temperature Detector:RTD A
- C Attribute: NIDMM\_ATTR\_TEMP\_RTD\_A

# temp\_rtd\_b

 $\verb|nidmm.Session.temp_rtd_b|\\$ 

Specifies the Callendar-Van Dusen B coefficient for RTD scaling when the RTD Type property is set to Custom. The default value is -5.775e-7(Pt3851).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Configuration:Measurement Options:Temperature:Resistance Temperature Detector:RTD B
- C Attribute: NIDMM\_ATTR\_TEMP\_RTD\_B

## temp\_rtd\_c

nidmm.Session.temp\_rtd\_c

Specifies the Callendar-Van Dusen C coefficient for RTD scaling when the RTD Type property is set to Custom. The default value is -4.183e-12(Pt3851).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Temperature:Resistance Temperature Detector:RTD C
- C Attribute: NIDMM\_ATTR\_TEMP\_RTD\_C

### temp rtd res

nidmm.Session.temp\_rtd\_res

Specifies the RTD resistance at 0 degrees Celsius. This applies to all supported RTDs, including custom RTDs. The default value is 100 (?).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Temperature:Resistance Temperature Detector:RTD Resistance
- C Attribute: NIDMM\_ATTR\_TEMP\_RTD\_RES

## temp rtd type

#### nidmm.Session.temp\_rtd\_type

Specifies the type of RTD used to measure temperature. The default value is PT3851. Refer to the nidmm. Session.temp\_rtd\_type topic in the NI Digital Multimeters Help for additional information about defined values.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.RTDType
Permissions	read-write
Channel Based	False
Resettable	No

Tip: This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Temperature:Resistance Temperature Detector:RTD Type
- C Attribute: NIDMM\_ATTR\_TEMP\_RTD\_TYPE

# temp\_tc\_fixed\_ref\_junc

## nidmm.Session.temp\_tc\_fixed\_ref\_junc

Specifies the reference junction temperature when a fixed reference junction is used to take a thermocouple measurement. The default value is 25.0 (°C).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Configuration:Measurement tions:Temperature:Thermocouple:Fixed Reference Junction

• C Attribute: NIDMM\_ATTR\_TEMP\_TC\_FIXED\_REF\_JUNC

## temp\_tc\_ref\_junc\_type

# nidmm.Session.temp\_tc\_ref\_junc\_type

Specifies the type of reference junction to be used in the reference junction compensation of a thermocouple. The only supported value, NIDMM VAL TEMP REF JUNC FIXED, is fixed.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ThermocoupleReferenceJunctionType
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Configuration:Measurement tions:Temperature:Thermocouple:Reference Junction Type

• C Attribute: NIDMM\_ATTR\_TEMP\_TC\_REF\_JUNC\_TYPE

## temp\_tc\_type

nidmm.Session.temp\_tc\_type

Specifies the type of thermocouple used to measure the temperature. The default value is J.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ThermocoupleType
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Configuration:Measurement Options:Temperature:Thermocouple:Thermocouple Type

• C Attribute: NIDMM\_ATTR\_TEMP\_TC\_TYPE

## temp thermistor a

nidmm.Session.temp\_thermistor\_a

Specifies the Steinhart-Hart A coefficient for thermistor scaling when the Thermistor Type property is set to Custom. The default value is 0.0010295 (44006).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Configuration:Measurement Options:Temperature:Thermistor:Thermistor A
- C Attribute: NIDMM\_ATTR\_TEMP\_THERMISTOR\_A

## temp\_thermistor\_b

nidmm.Session.temp\_thermistor\_b

Specifies the Steinhart-Hart B coefficient for thermistor scaling when the Thermistor Type proerty is set to Custom. The default value is 0.0002391 (44006).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration:Measurement Options:Temperature:Thermistor:Thermistor B
- C Attribute: NIDMM\_ATTR\_TEMP\_THERMISTOR\_B

## temp\_thermistor\_c

nidmm.Session.temp\_thermistor\_c

Specifies the Steinhart-Hart C coefficient for thermistor scaling when the Thermistor Type property is set to Custom. The default value is 1.568e-7 (44006).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

LabVIEW Property: Configuration:Measurement tions:Temperature:Thermistor:Thermistor C

• C Attribute: NIDMM\_ATTR\_TEMP\_THERMISTOR\_C

## temp thermistor type

#### nidmm.Session.temp\_thermistor\_type

Specifies the type of thermistor used to measure the temperature. The default value is *THERMISTOR\_44006*. Refer to the *nidmm.Session.temp\_thermistor\_type* topic in the NI Digital Multimeters Help for additional information about defined values.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ThermistorType
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Configuration:Measurement tions:Temperature:Thermistor:Thermistor Type

• C Attribute: NIDMM\_ATTR\_TEMP\_THERMISTOR\_TYPE

### temp transducer type

nidmm.Session.temp\_transducer\_type

Specifies the type of device used to measure the temperature. The default value is NIDMM\_VAL\_4\_THERMOCOUPLE.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TransducerType
Permissions	read-write
Channel Based	False
Resettable	No

Tip: This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Configuration: Measurement Options: Temperature: Transducer Type
- C Attribute: NIDMM ATTR TEMP TRANSDUCER TYPE

## trigger\_count

#### nidmm.Session.trigger count

Specifies the number of triggers the DMM receives before returning to the Idle state. This property can be set to any positive ViInt32 value for the NI 4065 and NI 4070/4071/4072. The NI 4050 and NI 4060 support this property being set to 1. Refer to the Multiple Point Acquisitions section of the NI Digital Multimeters Help for more information.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Multi Point Acquisition: Trigger Count
- C Attribute: NIDMM\_ATTR\_TRIGGER\_COUNT

## trigger\_delay

### nidmm.Session.trigger\_delay

Specifies the time (in seconds) that the DMM waits after it has received a trigger before taking a measurement. The default value is AUTO DELAY (-1), which means that the DMM waits an appropriate settling time before taking the measurement. (-1) signifies that AUTO DELAY is on, and (-2) signifies that AUTO DELAY is off. The NI 4065 and NI 4070/4071/4072 use the value specified in this property as additional settling time. For the The NI 4065 and NI 4070/4071/4072, the valid range for Trigger Delay is AUTO DELAY (-1) or 0.0-149.0 seconds and the onboard timing resolution is 34.72 ns. On the NI 4060, if this property is set to 0, the DMM does not settle before taking the measurement. On the NI 4060, the valid range for AUTO DELAY (-1) is 0.0-12.0 seconds and the onboard timing resolution is 100 ms. When using the NI 4050, this property must be set to AUTO DELAY (-1). Use positive values to set the trigger delay in seconds. Valid Range: NIDMM\_VAL\_AUTO\_DELAY (-1.0), 0.0-12.0 seconds (NI 4060 only) Default Value: NIDMM\_VAL\_AUTO\_DELAY

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

• LabVIEW Property: Trigger:Trigger Delay

• C Attribute: NIDMM\_ATTR\_TRIGGER\_DELAY

## trigger\_source

#### nidmm.Session.trigger\_source

Specifies the trigger source. When nidmm.Session.\_initiate() is called, the DMM waits for the trigger specified with this property. After it receives the trigger, the DMM waits the length of time specified with the nidmm.Session.trigger\_delay property. The DMM then takes a measurement. This property is not supported on the NI 4050. To determine which values are supported by each device, refer to the LabWindows/CVI Trigger Routing section in the NI Digital Multimeters Help.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerSource
Permissions	read-write
Channel Based	False
Resettable	No

Tip: This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Trigger:Trigger Source

• C Attribute: NIDMM\_ATTR\_TRIGGER\_SOURCE

## waveform\_coupling

# $\verb| nidmm.Session.waveform_coupling| \\$

For the NI 4070/4071/4072 only, specifies the coupling during a waveform acquisition.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.WaveformCoupling
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Waveform Acquisition: Waveform Coupling
- C Attribute: NIDMM\_ATTR\_WAVEFORM\_COUPLING

## waveform points

## nidmm.Session.waveform\_points

For the NI 4070/4071/4072 only, specifies the number of points to acquire in a waveform acquisition.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Waveform Acquisition: Waveform Points
- C Attribute: NIDMM\_ATTR\_WAVEFORM\_POINTS

## waveform\_rate

### nidmm.Session.waveform\_rate

For the NI 4070/4071/4072 only, specifies the rate of the waveform acquisition in Samples per second (S/s). The valid Range is 10.0-1,800,000 S/s. Values are coerced to the closest integer divisor of 1,800,000. The default value is 1,800,000.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Waveform Acquisition: Waveform Rate
- C Attribute: NIDMM\_ATTR\_WAVEFORM\_RATE

## **Methods**

#### abort

```
nidmm.Session.abort()
```

Aborts a previously initiated measurement and returns the DMM to the Idle state.

## configure measurement absolute

Configures the common properties of the measurement. These properties include nidmm.Session.method, nidmm.Session.range, and nidmm.Session.resolution absolute.

#### **Parameters**

- measurement\_function (nidmm.Function) Specifies the measurement\_function used to acquire the measurement. The driver sets nidmm. Session.method to this value.
- range (float) Specifies the range for the method specified in the Measurement\_Function parameter. When frequency is specified in the Measurement\_Function parameter, you must supply the minimum frequency expected in the range parameter. For example, you must type in 100 Hz if you are measuring 101 Hz or higher. For all other methods, you must supply a range that exceeds the value that you are measuring. For example, you must type in 10 V if you are measuring 9 V. range values are coerced up to the closest input range. Refer to the Devices Overview for a list of valid ranges. The driver sets nidmm.Session.range to this value. The default is 0.02 V.

NIDMM_VA	L_AU	ONIRIAMM point forms an Auto Range before acquiring the
	1.0	measurement.
NIDMM_VA	L <u> </u>	ONIRANIM sets the Range to the current nidmm. Session.
	2.0	<pre>auto_range_value and uses this range for all subse-</pre>
		quent measurements until the measurement configuration is
		changed.
NIDMM_VA	LAU	ONIRAMIMI_quefforms an Auto Range before acquir-
	3.0	ing the measurement. The nidmm.Session.
		<pre>auto_range_value is stored and used for all sub-</pre>
		sequent measurements until the measurement configuration
		is changed.

**Note:** The NI 4050, NI 4060, and NI 4065 only support Auto Range when the trigger and sample trigger are set to IMMEDIATE.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

• resolution\_absolute (float) - Specifies the absolute resolution for the measurement. NI-DMM sets nidmm.Session.resolution absolute to

this value. This parameter is ignored when the Range parameter is set to NIDMM\_VAL\_AUTO\_RANGE\_ON (-1.0) or NIDMM\_VAL\_AUTO\_RANGE\_ONCE (-3.0). The default is 0.001~V.

**Note:** NI-DMM ignores this parameter for capacitance and inductance measurements on the NI 4072. To achieve better resolution for such measurements, use the <code>nidmm.Session.lc\_number\_meas\_to\_average</code> property.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

## configure measurement digits

nidmm.Session.configure\_measurement\_digits (measurement\_function, range, resolution digits)

Configures the common properties of the measurement. These properties include nidmm.Session.method, nidmm.Session.range, and nidmm.Session.resolution\_digits.

#### **Parameters**

- measurement\_function (nidmm.Function) Specifies the measurement\_function used to acquire the measurement. The driver sets nidmm. Session.method to this value.
- range (float) Specifies the range for the method specified in the Measure-surement\_Function parameter. When frequency is specified in the Measurement\_Function parameter, you must supply the minimum frequency expected in the range parameter. For example, you must type in 100 Hz if you are measuring 101 Hz or higher. For all other methods, you must supply a range that exceeds the value that you are measuring. For example, you must type in 10 V if you are measuring 9 V. range values are coerced up to the closest input range. Refer to the Devices Overview for a list of valid ranges. The driver sets nidmm.Session.range to this value. The default is 0.02 V.

NIDMM_VA	L_AU:	ONIRAMM_pointforms an Auto Range before acquiring the
	1.0	measurement.
NIDMM_VA	LAU:	NIRAMM sets the Range to the current nidmm. Session.
	2.0	auto_range_value and uses this range for all subse-
		quent measurements until the measurement configuration is
		changed.
NIDMM_VA	L_AU:	ONIRAMM_quectorms an Auto Range before acquir-
	3.0	ing the measurement. The nidmm.Session.
		auto_range_value is stored and used for all sub-
		sequent measurements until the measurement configuration
		is changed.

**Note:** The NI 4050, NI 4060, and NI 4065 only support Auto Range when the trigger and sample trigger are set to IMMEDIATE.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

• **resolution\_digits** (*float*) – Specifies the resolution of the measurement in digits. The driver sets the Devices Overview for a list of valid ranges. The driver sets *nidmm.Session.resolution\_digits* property to this value. This parameter is ignored when the **Range** parameter is set to NIDMM\_VAL\_AUTO\_RANGE\_ON (-1.0) or NIDMM\_VAL\_AUTO\_RANGE\_ONCE (-3.0). The default is 5½.

**Note:** NI-DMM ignores this parameter for capacitance and inductance measurements on the NI 4072. To achieve better resolution for such measurements, use the <code>nidmm.Session.lc\_number\_meas\_to\_average</code> property.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

## configure multi point

nidmm.Session.configure\_multi\_point (trigger\_count, sample\_count, sample\_trigger=nidmm.SampleTrigger.IMMEDIATE, sample\_interval=datetime.timedelta(seconds=-1))

Configures the properties for multipoint measurements. These properties include nidmm. Session.trigger\_count, nidmm.Session.sample\_count, nidmm.Session.sample\_trigger, and nidmm.Session.sample\_interval.

For continuous acquisitions, set nidmm. Session.trigger\_count or nidmm. Session. sample\_count to zero. For more information, refer to Multiple Point Acquisitions, Triggering, and Using Switches.

#### **Parameters**

- trigger\_count (int) Sets the number of triggers you want the DMM to receive before returning to the Idle state. The driver sets nidmm.Session. trigger\_count to this value. The default value is 1.
- **sample\_count** (*int*) Sets the number of measurements the DMM makes in each measurement sequence initiated by a trigger. The driver sets *nidmm*. Session.sample\_count to this value. The default value is 1.
- sample\_trigger (nidmm.SampleTrigger) Specifies the sample\_trigger source you want to use. The driver sets nidmm.Session.sample\_trigger to this value. The default is Immediate.

**Note:** To determine which values are supported by each device, refer to the Lab-Windows/CVI Trigger Routing section.

• sample\_interval (float in seconds or datetime.timedelta)
- Sets the amount of time in seconds the DMM waits between measurement cycles.
The driver sets nidmm.Session.sample\_interval to this value. Specify a

sample interval to add settling time between measurement cycles or to decrease the measurement rate. **sample\_interval** only applies when the **Sample\_Trigger** is set to INTERVAL.

On the NI 4060, the **sample\_interval** value is used as the settling time. When sample interval is set to 0, the DMM does not settle between measurement cycles. The NI 4065 and NI 4070/4071/4072 use the value specified in **sample\_interval** as additional delay. The default value (-1) ensures that the DMM settles for a recommended time. This is the same as using an Immediate trigger.

**Note:** This property is not used on the NI 4080/4081/4082 and the NI 4050.

## configure\_rtd\_custom

nidmm.Session.configure\_rtd\_custom(rtd\_a, rtd\_b, rtd\_c) Configures the A, B, and C parameters for a custom RTD.

#### **Parameters**

- rtd\_a (float) Specifies the Callendar-Van Dusen A coefficient for RTD scaling when RTD Type parameter is set to Custom in the nidmm.Session.configure\_rtd\_type() method. The default is 3.9083e-3 (Pt3851)
- rtd\_b (float) Specifies the Callendar-Van Dusen B coefficient for RTD scaling when RTD Type parameter is set to Custom in the nidmm.Session.configure\_rtd\_type() method. The default is -5.775e-7 (Pt3851).
- rtd\_c (float) Specifies the Callendar-Van Dusen C coefficient for RTD scaling when RTD Type parameter is set to Custom in the nidmm.Session.configure\_rtd\_type() method. The default is -4.183e-12 (Pt3851).

# configure\_rtd\_type

nidmm.Session.configure\_rtd\_type (rtd\_type, rtd\_resistance)
Configures the RTD Type and RTD Resistance parameters for an RTD.

#### **Parameters**

• **rtd\_type** (nidmm.RTDType) – Specifies the type of RTD used to measure the temperature resistance. NI-DMM uses this value to set the RTD Type property. The default is PT3851.

Enum	Standards	Ma- te- rial	TCR (α)	Typ-ical $R_0$ $(\Omega)$	Notes	
Callendar-						
Van						
Dusen						
Coeffi-						
cient						
PT3851	IEC-751	Plat-	.0038	5100	A = 3.9083	Most
	DIN 43760	inum		Ω	$\times 10^{-3}$ B =	com-
	BS 1904			1000	−5.775×10:sup:−7	mon
	ASTM-E1137			Ω	C =	RTDs
	EN-60751				-4.183×10:sup:− <i>12</i>	
PT3750	Low-cost ven-	Plat-	.0037	<b>50</b> 000	A = 3.81	Low-
	dor compliant	inum		Ω	$\times 10^{-3}$ B =	cost
	RTD*				−6.02×10:sup:−7	RTD
					C =	
					-6.0×10:sup:− <i>12</i>	
PT3916	JISC 1604	Plat-	.0039	1600	A = 3.9739	Used in
		inum		Ω	$\times 10^{-3}$ B =	primar-
					−5.870×10:sup:−7	ily in
					$C = -4.4 \times 10^{-12}$	Japan
PT3920	US Industrial	Plat-	.0039	2000	A = 3.9787	Low-
	Standard D-	inum		Ω	$\times$ 10 <sup>-3</sup> B =	cost
	100 American				−5.8686×10:sup:−7	RTD
					C = -4.167	
					×10 <sup>-12</sup>	
PT3911	US Indus-	Plat-	.0039	1100	A = 3.9692	Low-
	trial Standard	inum		Ω	$\times 10^{-3}$ B =	cost
	American				−5.8495×10:sup:−7	RTD
					C = -4.233	
					×10 <sup>-12</sup>	
PT3928	ITS-90	Plat-	.0039	2800	A = 3.9888	The
		inum		Ω	$\times$ 10 <sup>-3</sup> B =	defini-
					-5.915×10:sup:-7	tion of
					$C = -3.85 \times 10^{-12}$	temper-
						ature
*No						
stan-						
dard.						
Check						
the						
TCR.						

• rtd\_resistance (float) – Specifies the RTD resistance in ohms at 0 °C. NI-DMM uses this value to set the RTD Resistance property. The default is  $100 \ (\Omega)$ .

# configure\_thermistor\_custom

nidmm.Session.configure\_thermistor\_custom(thermistor\_a, thermistor\_b, thermistor\_c)

Configures the A, B, and C parameters for a custom thermistor.

#### **Parameters**

• thermistor\_a (float) - Specifies the Steinhart-Hart A coefficient for thermistor scaling when Thermistor Type is set to Custom in the nidmm. Session. ConfigureThermistorType() method. The default is 1.0295e-3 (44006).

**Note:** One or more of the referenced methods are not in the Python API for this driver.

• thermistor\_b (float) - Specifies the Steinhart-Hart B coefficient for thermistor scaling when Thermistor Type is set to Custom in the nidmm.Session. ConfigureThermistorType() method. The default is 2.391e-4 (44006).

**Note:** One or more of the referenced methods are not in the Python API for this driver.

• thermistor\_c (float) - Specifies the Steinhart-Hart C coefficient for thermistor scaling when Thermistor Type is set to Custom in the nidmm. Session. ConfigureThermistorType() method. The default is 1.568e-7 (44006).

**Note:** One or more of the referenced methods are not in the Python API for this driver.

#### configure thermocouple

nidmm.Session.configure\_thermocouple (thermocouple\_type, reference\_junction\_type=nidmm.ThermocoupleReferenceJunctionType.FIXED Configures the thermocouple type and reference junction type for a chosen thermocouple.

#### **Parameters**

• thermocouple\_type (nidmm. Thermocouple Type) - Specifies the type of thermocouple used to measure the temperature. NI-DMM uses this value to set the Thermocouple Type property. The default is J.

B	Thermocouple type B
E	Thermocouple type E
J	Thermocouple type J
K	Thermocouple type K
N	Thermocouple type N
R	Thermocouple type R
S	Thermocouple type S
T	Thermocouple type T

• reference\_junction\_type(nidmm.ThermocoupleReferenceJunctionType) - Specifies the type of reference junction to be used in the reference junction compensation of a thermocouple measurement. NI-DMM uses this value to set the Reference Junction Type property. The only supported value is NIDMM\_VAL\_TEMP\_REF\_JUNC\_FIXED.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

## configure trigger

nidmm.Session.configure\_trigger(trigger\_source, trigger\_delay=datetime.timedelta(seconds=1))

Configures the DMM **Trigger\_Source** and **Trigger\_Delay**. Refer to Triggering and Using Switches for more information.

#### **Parameters**

• trigger\_source (nidmm.TriggerSource) - Specifies the trigger\_source that initiates the acquisition. The driver sets nidmm.Session. trigger\_source to this value. Software configures the DMM to wait until nidmm.Session.send\_software\_trigger() is called before triggering the DMM.

**Note:** To determine which values are supported by each device, refer to the Lab-Windows/CVI Trigger Routing section.

• **trigger\_delay** (float in seconds or datetime.timedelta) — Specifies the time that the DMM waits after it has received a trigger before taking a measurement. The driver sets the nidmm.Session.trigger\_delay property to this value. By default, **trigger\_delay** is NIDMM\_VAL\_AUTO\_DELAY (-1), which means the DMM waits an appropriate settling time before taking the measurement. On the NI 4060, if you set **trigger\_delay** to 0, the DMM does not settle before taking the measurement. The NI 4065 and NI 4070/4071/4072 use the value specified in **trigger\_delay** as additional settling time.

**Note:** When using the NI 4050, **Trigger\_Delay** must be set to NIDMM\_VAL\_AUTO\_DELAY (-1).

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

### configure\_waveform\_acquisition

nidmm.Session.configure\_waveform\_acquisition (measurement\_function, range, rate, waveform\_points)

Configures the DMM for waveform acquisitions. This feature is supported on the NI 4080/4081/4082 and the NI 4070/4071/4072.

#### **Parameters**

• measurement\_function (nidmm.Function) - Specifies the measurement\_function used in a waveform acquisition. The driver sets nidmm. Session.method to this value.

WAVEFORM_VOLTAGE (default)	1003	Voltage Waveform
WAVEFORM_CURRENT	1004	Current Waveform

• range (float) - Specifies the expected maximum amplitude of the input signal and sets the range for the Measurement\_Function. NI-DMM sets nidmm. Session.range to this value. range values are coerced up to the closest input range. The default is 10.0.

For valid ranges refer to the topics in Devices.

Auto-ranging is not supported during waveform acquisitions.

• rate (float) - Specifies the rate of the acquisition in samples per second. NI-DMM sets nidmm.Session.waveform\_rate to this value.

The valid **Range** is 10.0–1,800,000 S/s. **rate** values are coerced to the closest integer divisor of 1,800,000. The default value is 1,800,000.

• waveform\_points (int) - Specifies the number of points to acquire before the waveform acquisition completes. NI-DMM sets nidmm.Session. waveform\_points to this value.

To calculate the maximum and minimum number of waveform points that you can acquire in one acquisition, refer to the Waveform Acquisition Measurement Cycle.

The default value is 500.

## disable

```
nidmm.Session.disable()
```

Places the instrument in a quiescent state where it has minimal or no impact on the system to which it is connected. If a measurement is in progress when this method is called, the measurement is aborted.

#### fetch

```
nidmm.Session.fetch(maximum_time=datetime.timedelta(milliseconds=-1))
```

Returns the value from a previously initiated measurement. You must call nidmm. Session. \_initiate() before calling this method.

Parameters maximum\_time (float in seconds or datetime.

timedelta) – Specifies the **maximum\_time** allowed for this method to complete in milliseconds. If the method does not complete within this time interval, the method returns the NIDMM\_ERROR\_MAX\_TIME\_EXCEEDED error code. This may happen if an external trigger has not been received, or if the specified timeout is not long enough for the acquisition to complete.

The valid range is 0–86400000. The default value is  $\tt NIDMM\_VAL\_TIME\_LIMIT\_AUTO$  (-1). The DMM calculates the timeout automatically.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

# Return type float

**Returns** The measured value returned from the DMM.

## fetch multi point

nidmm.Session.fetch\_multi\_point (array\_size, maximum\_time=datetime.timedelta(milliseconds=-1))

Returns an array of values from a previously initiated multipoint measurement. The number of measurements the DMM makes is determined by the values you specify for the **Trigger\_Count** and **Sample\_Count** parameters of <code>nidmm.Session.configure\_multi\_point()</code>. You must first call <code>nidmm.Session.\_initiate()</code> to initiate a measurement before calling this method.

#### **Parameters**

• array\_size (int) - Specifies the number of measurements to acquire. The maximum number of measurements for a finite acquisition is the (**Trigger Count** x **Sample Count**) parameters in nidmm. Session.configure\_multi\_point().

For continuous acquisitions, up to 100,000 points can be returned at once. The number of measurements can be a subset. The valid range is any positive ViInt32. The default value is 1.

• maximum\_time (float in seconds or datetime.timedelta) — Specifies the maximum\_time allowed for this method to complete in milliseconds. If the method does not complete within this time interval, the method returns the NIDMM\_ERROR\_MAX\_TIME\_EXCEEDED error code. This may happen if an external trigger has not been received, or if the specified timeout is not long enough for the acquisition to complete.

The valid range is 0-86400000. The default value is NIDMM\_VAL\_TIME\_LIMIT\_AUTO (-1). The DMM calculates the timeout automatically.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

### Return type

```
tuple (reading_array, actual_number_of_points)
WHERE
reading_array (array.array("d")):
    An array of measurement values.
```

**Note:** The size of the **Reading\_Array** must be at least the size that you specify for the **Array\_Size** parameter.

actual\_number\_of\_points (int):

Indicates the number of measured values actually retrieved from the DMM.

## fetch\_waveform

nidmm.Session.fetch\_waveform(array\_size, maximum\_time=datetime.timedelta(milliseconds=-

For the NI 4080/4081/4082 and the NI 4070/4071/4072, returns an array of values from a previously initiated waveform acquisition. You must call nidmm.Session.\_initiate() before calling this method.

#### **Parameters**

- array\_size (int) Specifies the number of waveform points to return. You specify the total number of points that the DMM acquires in the Waveform Points parameter of nidmm.Session.configure\_waveform\_acquisition(). The default value is 1.
- maximum\_time (float in seconds or datetime.timedelta) Specifies the maximum\_time allowed for this method to complete in milliseconds. If the method does not complete within this time interval, the method returns the NIDMM\_ERROR\_MAX\_TIME\_EXCEEDED error code. This may happen if an external trigger has not been received, or if the specified timeout is not long enough for the acquisition to complete.

The valid range is 0–86400000. The default value is  $\mbox{NIDMM\_VAL\_TIME\_LIMIT\_AUTO}$  (-1). The DMM calculates the timeout automatically.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

### Return type

```
tuple (waveform array, actual number of points)
```

#### **WHERE**

waveform\_array (array.array("d")):

**Waveform Array** is an array of measurement values stored in waveform data type.

actual\_number\_of\_points (int):

Indicates the number of measured values actually retrieved from the DMM.

### fetch waveform into

```
nidmm.Session.fetch_waveform_into(array_size, maximum_time=datetime.timedelta(milliseconds=-1))
```

For the NI 4080/4081/4082 and the NI 4070/4071/4072, returns an array of values from a previously initiated waveform acquisition. You must call nidmm.Session.\_initiate() before calling this method.

## **Parameters**

- waveform\_array (numpy.array(dtype=numpy.float64)) Waveform Array is an array of measurement values stored in waveform data type.
- maximum\_time (float in seconds or datetime.timedelta) Specifies the maximum\_time allowed for this method to complete in milliseconds. If the method does not complete within this time interval, the method returns the NIDMM\_ERROR\_MAX\_TIME\_EXCEEDED error code. This may happen if an external trigger has not been received, or if the specified timeout is not long enough for the acquisition to complete.

The valid range is 0–86400000. The default value is  $\tt NIDMM\_VAL\_TIME\_LIMIT\_AUTO$  (-1). The DMM calculates the timeout automatically.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

## Return type

tuple (waveform\_array, actual\_number\_of\_points)

WHERE

waveform\_array (numpy.array(dtype=numpy.float64)):

Waveform Array is an array of measurement values stored in waveform data type.

actual\_number\_of\_points (int):

Indicates the number of measured values actually retrieved from the DMM.

### get\_cal\_date\_and\_time

nidmm.Session.get\_cal\_date\_and\_time(cal\_type)

Returns the date and time of the last calibration performed.

**Note:** The NI 4050 and NI 4060 are not supported.

**Parameters** cal\_type (int) – Specifies the type of calibration performed (external or self-calibration).

NIDMM_VAL_INTERNAL_AREA (default)	0	Self-Calibration
NIDMM_VAL_EXTERNAL_AREA	1	External Calibration

**Note:** The NI 4065 does not support self-calibration.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

Return type datetime.datetime

**Returns** Indicates date and time of the last calibration.

## get\_dev\_temp

nidmm.Session.get\_dev\_temp(options="")
Returns the current Temperature of the device.

Note: The NI 4050 and NI 4060 are not supported.

**Parameters options** (str) – Reserved.

Return type float

**Returns** Returns the current **temperature** of the device.

# get\_ext\_cal\_recommended\_interval

```
nidmm.Session.get_ext_cal_recommended_interval()
```

Returns the recommended interval between external recalibration in **Months**.

**Note:** The NI 4050 and NI 4060 are not supported.

Return type datetime.timedelta

**Returns** Returns the recommended number of **months** between external calibrations.

## get\_last\_cal\_temp

```
nidmm.Session.get_last_cal_temp(cal_type)
```

Returns the **Temperature** during the last calibration procedure.

**Note:** The NI 4050 and NI 4060 are not supported.

**Parameters** cal\_type (int) – Specifies the type of calibration performed (external or self-calibration).

NIDMM_VAL_INTERNAL_AREA (default)	0	Self-Calibration
NIDMM_VAL_EXTERNAL_AREA	1	External Calibration

**Note:** The NI 4065 does not support self-calibration.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

## Return type float

**Returns** Returns the **temperature** during the last calibration.

## get\_self\_cal\_supported

```
nidmm.Session.get_self_cal_supported()
```

Returns a Boolean value that expresses whether or not the DMM that you are using can perform self-calibration.

### Return type bool

#### Returns

Returns whether Self Cal is supported for the device specified by the given session.

		The DMM that you are using can perform self-calibration.
False	0	The DMM that you are using cannot perform self-calibration.

#### lock

```
nidmm.Session.lock()
```

Obtains a multithread lock on the device session. Before doing so, the software waits until all other execution threads release their locks on the device session.

Other threads may have obtained a lock on this session for the following reasons:

- The application called the *nidmm*. Session.lock() method.
- A call to NI-DMM locked the session.
- After a call to the nidmm. Session.lock() method returns successfully, no other threads can access
  the device session until you call the nidmm. Session.unlock() method or exit out of the with block
  when using lock context manager.
- Use the nidmm.Session.lock() method and the nidmm.Session.unlock() method around a sequence of calls to instrument driver methods if you require that the device retain its settings through the end of the sequence.

You can safely make nested calls to the <code>nidmm.Session.lock()</code> method within the same thread. To completely unlock the session, you must balance each call to the <code>nidmm.Session.lock()</code> method with a call to the <code>nidmm.Session.unlock()</code> method.

One method for ensuring there are the same number of unlock method calls as there is lock calls is to use lock as a context manager

```
with nidmm.Session('dev1') as session:
    with session.lock():
        # Calls to session within a single lock context
```

The first with block ensures the session is closed regardless of any exceptions raised

The second with block ensures that unlock is called regardless of any exceptions raised

Return type context manager

**Returns** When used in a *with* statement, *nidmm.Session.lock()* acts as a context manager and unlock will be called when the *with* block is exited

## perform\_open\_cable\_comp

```
nidmm.Session.perform_open_cable_comp()
```

For the NI 4082 and NI 4072 only, performs the open cable compensation measurements for the current capacitance/inductance range, and returns open cable compensation **Conductance** and **Susceptance** values. You can use the return values of this method as inputs to nidmm.Session. ConfigureOpenCableCompValues().

This method returns an error if the value of the nidmm. Session.method property is not set to CAPACITANCE (1005) or INDUCTANCE (1006).

Note: One or more of the referenced methods are not in the Python API for this driver.

## Return type

```
tuple (conductance, susceptance)

WHERE

conductance (float):

conductance is the measured value of open cable compensation conductance.

susceptance (float):

susceptance is the measured value of open cable compensation susceptance.
```

## perform\_short\_cable\_comp

```
nidmm.Session.perform_short_cable_comp()
```

Performs the short cable compensation measurements for the current capacitance/inductance range, and returns short cable compensation **Resistance** and **Reactance** values. You can use the return values of this method as inputs to nidmm.Session.ConfigureShortCableCompValues().

This method returns an error if the value of the nidmm. Session.method property is not set to CAPACITANCE (1005) or INDUCTANCE (1006).

**Note:** One or more of the referenced methods are not in the Python API for this driver.

# Return type

```
tuple (resistance, reactance)
WHERE
resistance (float):
    resistance is the measured value of short cable compensation resistance.
reactance (float):
    reactance is the measured value of short cable compensation reactance.
```

#### read

nidmm. Session. **read** (maximum\_time=datetime.timedelta(milliseconds=-1))
Acquires a single measurement and returns the measured value.

#### Parameters maximum\_time

(float in seconds or datetime.

timedelta) – Specifies the **maximum\_time** allowed for this method to complete in milliseconds. If the method does not complete within this time interval, the method returns the NIDMM\_ERROR\_MAX\_TIME\_EXCEEDED error code. This may happen if an external trigger has not been received, or if the specified timeout is not long enough for the acquisition to complete.

The valid range is 0–86400000. The default value is NIDMM\_VAL\_TIME\_LIMIT\_AUTO (-1). The DMM calculates the timeout automatically.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

## Return type float

**Returns** The measured value returned from the DMM.

## read multi point

nidmm.Session.read\_multi\_point (array\_size, maximum\_time=datetime.timedelta(milliseconds=1))

Acquires multiple measurements and returns an array of measured values. The number of measurements the DMM makes is determined by the values you specify for the **Trigger\_Count** and **Sample\_Count** parameters in nidmm.Session.configure\_multi\_point().

### **Parameters**

• array\_size (int) - Specifies the number of measurements to acquire. The maximum number of measurements for a finite acquisition is the (**Trigger Count** x **Sample Count**) parameters in nidmm.Session.configure\_multi\_point().

For continuous acquisitions, up to 100,000 points can be returned at once. The number of measurements can be a subset. The valid range is any positive ViInt32. The default value is 1.

• maximum\_time (float in seconds or datetime.timedelta) – Specifies the maximum\_time allowed for this method to complete in milliseconds. If the method does not complete within this time interval, the method returns the NIDMM\_ERROR\_MAX\_TIME\_EXCEEDED error code. This may happen if an external trigger has not been received, or if the specified timeout is not long enough for the acquisition to complete.

The valid range is 0–86400000. The default value is  $\tt NIDMM\_VAL\_TIME\_LIMIT\_AUTO$  (-1). The DMM calculates the timeout automatically.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

#### Return type

```
tuple (reading_array, actual_number_of_points)
```

WHERE

reading\_array (array.array("d")):

An array of measurement values.

**Note:** The size of the **Reading\_Array** must be at least the size that you specify for the **Array\_Size** parameter.

```
actual_number_of_points (int):
```

Indicates the number of measured values actually retrieved from the DMM.

### read\_status

```
nidmm.Session.read_status()
```

Returns measurement backlog and acquisition status. Use this method to determine how many measurements are available before calling nidmm.Session.fetch(), nidmm.Session.fetch\_multi\_point(), or nidmm.Session.fetch\_waveform().

**Note:** The NI 4050 is not supported.

## Return type

tuple (acquisition\_backlog, acquisition\_status)

**WHERE** 

acquisition\_backlog (int):

The number of measurements available to be read. If the backlog continues to increase, data is eventually overwritten, resulting in an error.

**Note:** On the NI 4060, the **Backlog** does not increase when autoranging. On the NI 4065, the **Backlog** does not increase when Range is set to AUTO RANGE ON (-1), or before the first point is fetched when Range is set to AUTO RANGE ONCE (-3). These behaviors are due to the autorange model of the devices.

acquisition\_status (nidmm.AcquisitionStatus):

Indicates status of the acquisition. The following table shows the acquisition states:

0	Running
1	Finished with backlog
2	Finished with no backlog
3	Paused
4	No acquisition in progress

## read waveform

nidmm.Session.read\_waveform(array\_size, maximum\_time=datetime.timedelta(milliseconds=-1))

For the NI 4080/4081/4082 and the NI 4070/4071/4072, acquires a waveform and returns data as an array of values or as a waveform data type. The number of elements in the **Waveform\_Array** is determined by the values you specify for the **Waveform\_Points** parameter in nidmm.Session.configure\_waveform\_acquisition().

#### **Parameters**

- array\_size (int) Specifies the number of waveform points to return. You specify the total number of points that the DMM acquires in the Waveform Points parameter of nidmm.Session.configure\_waveform\_acquisition(). The default value is 1.
- maximum\_time (float in seconds or datetime.timedelta) Specifies the maximum\_time allowed for this method to complete in milliseconds. If the method does not complete within this time interval, the method returns the NIDMM\_ERROR\_MAX\_TIME\_EXCEEDED error code. This may happen if an external trigger has not been received, or if the specified timeout is not long enough for the acquisition to complete.

The valid range is 0–86400000. The default value is  $\mbox{NIDMM\_VAL\_TIME\_LIMIT\_AUTO}$  (-1). The DMM calculates the timeout automatically.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

#### Return type

```
tuple (waveform_array, actual_number_of_points)
```

#### WHERE

waveform\_array (array.array("d")):

An array of measurement values.

**Note:** The size of the **Waveform\_Array** must be at least the size that you specify for the **Array\_Size** parameter.

```
actual_number_of_points (int):
```

Indicates the number of measured values actually retrieved from the DMM.

### reset

```
nidmm.Session.reset()
```

Resets the instrument to a known state and sends initialization commands to the instrument. The initialization commands set instrument settings to the state necessary for the operation of the instrument driver.

#### reset with defaults

```
nidmm.Session.reset_with_defaults()
```

Resets the instrument to a known state and sends initialization commands to the DMM. The initialization commands set the DMM settings to the state necessary for the operation of NI-DMM. All user-defined default values associated with a logical name are applied after setting the DMM.

### self\_cal

```
nidmm.Session.self cal()
```

For the NI 4080/4081/4082 and the NI 4070/4071/4072, executes the self-calibration routine to maintain measurement accuracy.

**Note:** This method calls *nidmm.Session.reset()*, and any configurations previous to the call will be lost. All properties will be set to their default values after the call returns.

## self\_test

```
nidmm.Session.self_test()
```

Performs a self-test on the DMM to ensure that the DMM is functioning properly. Self-test does not calibrate the DMM. Zero indicates success.

On the NI 4080/4082 and NI 4070/4072, the error code 1013 indicates that you should check the fuse and replace it, if necessary.

Raises SelfTestError on self test failure. Properties on exception object:

- code failure code from driver
- · message status message from driver

**Note:** Self-test does not check the fuse on the NI 4065, NI 4071, and NI 4081. Hence, even if the fuse is blown on the device, self-test does not return error code 1013.

**Note:** This method calls *nidmm.Session.reset()*, and any configurations previous to the call will be lost. All properties will be set to their default values after the call returns.

## send software trigger

```
nidmm.Session.send_software_trigger()
```

Sends a command to trigger the DMM. Call this method if you have configured either the <code>nidmm.Session.trigger\_source</code> or <code>nidmm.Session.sample\_trigger</code> properties. If the <code>nidmm.Session.trigger\_source</code> and/or <code>nidmm.Session.sample\_trigger</code> properties are set to <code>NIDMM\_VAL\_EXTERNAL</code> or <code>NIDMM\_VAL\_TTLn</code>, you can use this method to override the trigger source that you configured and trigger the device. The NI 4050 and NI 4060 are not supported.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

## unlock

nidmm.Session.unlock()

Releases a lock that you acquired on an device session using nidmm. Session.lock(). Refer to nidmm. Session.unlock() for additional information on session locks.

# **Properties**

Property	Datatype
nidmm.Session.ac_max_freq	float
nidmm.Session.ac_min_freq	float
nidmm.Session.adc_calibration	ADCCalibration
nidmm.Session.aperture_time	float
nidmm.Session.aperture_time_units	ApertureTimeUnits
nidmm.Session.auto_range_value	float
nidmm.Session.auto_zero	AutoZero
nidmm.Session.buffer_size	int
nidmm.Session.cable_comp_type	CableCompensationType
nidmm.Session.channel_count	int
nidmm.Session.current_source	float
nidmm.Session.dc_bias	bool
nidmm.Session.dc_noise_rejection	DCNoiseRejection
nidmm.Session.driver_setup	str
nidmm.Session.freq_voltage_auto_range	float
nidmm.Session.freq_voltage_range	float
nidmm.Session.function	Function
nidmm.Session.input_resistance	float
nidmm.Session.instrument_firmware_revision	str
nidmm.Session.instrument_manufacturer	str
nidmm.Session.instrument_model	str
nidmm.Session.instrument_product_id	int
nidmm.Session.io_resource_descriptor	str
nidmm.Session.lc_calculation_model	LCCalculationModel
nidmm.Session.lc_number_meas_to_average	int
nidmm.Session.logical_name	str
nidmm.Session.meas_complete_dest	MeasurementCompleteDest
nidmm.Session.number_of_averages	int
nidmm.Session.offset_comp_ohms	bool
nidmm.Session.open_cable_comp_conductance	float
nidmm.Session.open_cable_comp_susceptance	float
nidmm.Session.operation_mode	OperationMode
nidmm.Session.powerline_freq	float
nidmm.Session.range	float
nidmm.Session.resolution_absolute	float

Continued on next page

7.2. nidmm module 177

Table 5 – continued from previous page

Property Property	Datatype	
nidmm.Session.resolution_digits	float	
nidmm.Session.sample_count	int	
nidmm.Session.sample_interval	float in seconds or datetime.timedelta	
nidmm.Session.sample_trigger	SampleTrigger	
nidmm.Session.serial_number	str	
nidmm.Session.settle_time	float in seconds or datetime.timedelta	
nidmm.Session.short_cable_comp_reactance	float	
nidmm.Session.short_cable_comp_resistance	float	
nidmm.Session.simulate	bool	
nidmm.Session.specific_driver_description	str	
nidmm.Session.specific_driver_major_version	int	
nidmm.Session.specific_driver_minor_version	int	
nidmm.Session.specific_driver_revision	str	
nidmm.Session.specific_driver_vendor	str	
nidmm.Session.supported_instrument_models	str	
nidmm.Session.temp_rtd_a	float	
nidmm.Session.temp_rtd_b	float	
nidmm.Session.temp_rtd_c	float	
nidmm.Session.temp_rtd_res	float	
nidmm.Session.temp_rtd_type	RTDType	
nidmm.Session.temp_tc_fixed_ref_junc	float	
<pre>nidmm.Session.temp_tc_ref_junc_type</pre>	ThermocoupleReferenceJunctionType	
nidmm.Session.temp_tc_type	ThermocoupleType	
nidmm.Session.temp_thermistor_a	float	
nidmm.Session.temp_thermistor_b	float	
nidmm.Session.temp_thermistor_c	float	
nidmm.Session.temp_thermistor_type	ThermistorType	
nidmm.Session.temp_transducer_type	TransducerType	
nidmm.Session.trigger_count	int	
nidmm.Session.trigger_delay	float in seconds or datetime.timedelta	
nidmm.Session.trigger_source	TriggerSource	
nidmm.Session.waveform_coupling	WaveformCoupling	
nidmm.Session.waveform_points	int	
nidmm.Session.waveform_rate	float	

# Methods

Method name
nidmm.Session.abort()
<pre>nidmm.Session.configure_measurement_absolute()</pre>
<pre>nidmm.Session.configure_measurement_digits()</pre>
<pre>nidmm.Session.configure_multi_point()</pre>
nidmm.Session.configure_rtd_custom()
<pre>nidmm.Session.configure_rtd_type()</pre>
nidmm.Session.configure_thermistor_custom()
nidmm.Session.configure_thermocouple()
nidmm.Session.configure_trigger()
<pre>nidmm.Session.configure_waveform_acquisition()</pre>

Continued on next page

Table 6 – continued from previous page

Method name
nidmm.Session.disable()
nidmm.Session.fetch()
nidmm.Session.fetch_multi_point()
nidmm.Session.fetch_waveform()
nidmm.Session.fetch_waveform_into()
nidmm.Session.get_cal_date_and_time()
nidmm.Session.get_dev_temp()
<pre>nidmm.Session.get_ext_cal_recommended_interval()</pre>
nidmm.Session.get_last_cal_temp()
nidmm.Session.get_self_cal_supported()
nidmm.Session.lock()
nidmm.Session.perform_open_cable_comp()
nidmm.Session.perform_short_cable_comp()
nidmm.Session.read()
nidmm.Session.read_multi_point()
nidmm.Session.read_status()
nidmm.Session.read_waveform()
nidmm.Session.reset()
<pre>nidmm.Session.reset_with_defaults()</pre>
nidmm.Session.self_cal()
nidmm.Session.self_test()
<pre>nidmm.Session.send_software_trigger()</pre>
nidmm.Session.unlock()

## 7.2.3.2 Enums

Enums used in NI-DMM

## **ADCCalibration**

```
class nidmm.ADCCalibration
```

## AUTO

The DMM enables or disables ADC calibration for you.

## OFF

The DMM does not compensate for changes to the gain.

ON

The DMM measures an internal reference to calculate the correct gain for the measurement.

## **AcquisitionStatus**

## class nidmm.AcquisitionStatus

## RUNNING

Running

## FINISHED\_WITH\_BACKLOG

Finished with Backlog

7.2. nidmm module 179

#### FINISHED WITH NO BACKLOG

Finished with no Backlog

#### **PAUSED**

Paused

## NO ACQUISITION IN PROGRESS

No acquisition in progress

## **ApertureTimeUnits**

```
class nidmm.ApertureTimeUnits
```

#### **SECONDS**

Seconds

## POWER\_LINE\_CYCLES

Powerline Cycles

## **AutoZero**

class nidmm.AutoZero

#### AUTO

The drivers chooses the AutoZero setting based on the configured method and resolution.

## OFF

Disables AutoZero.

ON

The DMM internally disconnects the input signal following each measurement and takes a zero reading. It then subtracts the zero reading from the preceding reading.

#### ONCE

The DMM internally disconnects the input signal for the first measurement and takes a zero reading. It then subtracts the zero reading from the first reading and the following readings.

## CableCompensationType

## class nidmm.CableCompensationType

#### NONE

No Cable Compensation

#### OPEN

Open Cable Compensation

#### SHORT

**Short Cable Compensation** 

## OPEN\_AND\_SHORT

Open and Short Cable Compensation

## **DCNoiseRejection**

## class nidmm.DCNoiseRejection

#### AUTO

The driver chooses the DC noise rejection setting based on the configured method and resolution.

#### NORMAL

NI-DMM weighs all samples equally.

## SECOND\_ORDER

NI-DMM weighs the samples taken in the middle of the aperture time more than samples taken at the beginning and the end of the measurement using a triangular weighing method.

## HIGH ORDER

NI-DMM weighs the samples taken in the middle of the aperture time more than samples taken at the beginning and the end of the measurement using a bell-curve weighing method.

## **Function**

#### class nidmm.Function

DC VOLTS

DC Voltage

AC\_VOLTS

AC Voltage

DC CURRENT

DC Current

AC\_CURRENT

AC Current

TWO\_WIRE\_RES

2-Wire Resistance

FOUR\_WIRE\_RES

4-Wire Resistance

**FREQ** 

Frequency

PERIOD

Period

#### **TEMPERATURE**

NI 4065, and NI 4070/4071/4072 supported.

## AC\_VOLTS\_DC\_COUPLED

AC Voltage with DC Coupling

DIODE

Diode

## WAVEFORM\_VOLTAGE

Waveform voltage

7.2. nidmm module 181

#### WAVEFORM CURRENT

Waveform current

#### CAPACITANCE

Capacitance

## INDUCTANCE

Inductance

## **LCCalculationModel**

## class nidmm.LCCalculationModel

#### AUTO

NI-DMM chooses the algorithm based on method and range

#### SERIES

NI-DMM uses the series impedance model to calculate capacitance and inductance

#### PARALLEL

NI-DMM uses the parallel admittance model to calculate capacitance and inductance

## MeasurementCompleteDest

#### class nidmm.MeasurementCompleteDest

## NONE

No Trigger

#### EXTERNAL

AUX I/O Connector

## PXI\_TRIG0

PXI Trigger Line 0

## PXI\_TRIG1

PXI Trigger Line 1

## PXI TRIG2

PXI Trigger Line 2

## PXI TRIG3

PXI Trigger Line 3

## PXI\_TRIG4

PXI Trigger Line 4

## PXI\_TRIG5

PXI Trigger Line 5

## PXI\_TRIG6

PXI Trigger Line 6

## PXI\_TRIG7

PXI Trigger Line 7

## LBR TRIGO

Internal Trigger Line of a PXI/SCXI Combination Chassis

## **OperationMode**

## class nidmm.OperationMode

## **IVIDMM**

IviDmm Mode

## WAVEFORM

Waveform acquisition mode

## **RTDType**

## class nidmm.RTDType

#### CUSTOM

Performs Callendar-Van Dusen RTD scaling with the user-specified A, B, and C coefficients.

## **PT3750**

Performs scaling for a Pt 3750 RTD.

#### PT3851

Performs scaling for a Pt 3851 RTD.

## PT3911

Performs scaling for a Pt 3911 RTD.

#### PT3916

Performs scaling for a Pt 3916 RTD.

#### PT3920

Performs scaling for a Pt 3920 RTD.

#### PT3928

Performs scaling for a Pt 3928 RTD.

## **SampleTrigger**

## class nidmm.SampleTrigger

#### IMMEDIATE

No Trigger

# EXTERNAL

AUX I/O Connector Trigger Line 0

## SOFTWARE\_TRIG

Software Trigger

## INTERVAL

Interval Trigger

#### PXI TRIGO

PXI Trigger Line 0

## PXI\_TRIG1

PXI Trigger Line 1

7.2. nidmm module 183

```
PXI TRIG2
    PXI Trigger Line 2
PXI TRIG3
    PXI Trigger Line 3
PXI TRIG4
    PXI Trigger Line 4
PXI TRIG5
    PXI Trigger Line 5
PXI_TRIG6
    PXI Trigger Line 6
PXI_TRIG7
    PXI Trigger Line 7
PXI_STAR
    PXI Star Trigger Line
AUX TRIG1
    AUX I/0 Connector Trigger Line 1
LBR_TRIG1
    Internal Trigger Line of a PXI/SCXI Combination Chassis
```

## **ThermistorType**

```
class nidmm.ThermistorType
```

#### CUSTOM

Custom

 ${\tt THERMISTOR\_44004}$ 

44004

THERMISTOR\_44006

44006

THERMISTOR 44007

44007

## **ThermocoupleReferenceJunctionType**

```
class nidmm.ThermocoupleReferenceJunctionType
```

#### FIXED

Thermocouple reference juction is fixed at the user-specified temperature.

# ThermocoupleType

class nidmm.ThermocoupleType

```
В
    Thermocouple type B
Е
    Thermocouple type E
J
    Thermocouple type J
K
    Thermocouple type K
N
    Thermocouple type N
R
    Thermocouple type R
s
    Thermocouple type S
т
    Thermocouple type T
```

# TransducerType

class nidmm.TransducerType

## THERMOCOUPLE

Thermocouple

## THERMISTOR

Thermistor

## TWO\_WIRE\_RTD

2-wire RTD

## FOUR\_WIRE\_RTD

4-wire RTD

## **TriggerSource**

class nidmm.TriggerSource

## IMMEDIATE

No Trigger

## EXTERNAL

AUX I/O Connector Trigger Line 0

## SOFTWARE\_TRIG

Software Trigger

## PXI\_TRIG0

PXI Trigger Line 0

## PXI\_TRIG1

PXI Trigger Line 1

7.2. nidmm module 185

```
PXI TRIG2
```

PXI Trigger Line 2

PXI TRIG3

PXI Trigger Line 3

PXI TRIG4

PXI Trigger Line 4

PXI TRIG5

PXI Trigger Line 5

PXI\_TRIG6

PXI Trigger Line 6

PXI\_TRIG7

PXI Trigger Line 7

PXI\_STAR

PXI Star Trigger Line

AUX TRIG1

AUX I/O Connector Trigger Line 1

LBR TRIG1

Internal Trigger Line of a PXI/SCXI Combination Chassis

## WaveformCoupling

class nidmm.WaveformCoupling

AC

AC Coupled

DC

DC Coupled

## 7.2.3.3 Exceptions and Warnings

## **DriverError**

```
exception nidmm.DriverError
```

An error originating from the NI-DMM driver

## UnsupportedConfigurationError

## ${\tt exception} \ {\tt nidmm.UnsupportedConfigurationError}$

An error due to using this module in an usupported platform.

## DriverNotInstalledError

## exception nidmm.DriverNotInstalledError

An error due to using this module without the driver runtime installed.

## InvalidRepeatedCapabilityError

# **exception** nidmm. **InvalidRepeatedCapabilityError**An error due to an invalid character in a repeated capability

## SelfTestError

```
exception nidmm.SelfTestError
An error due to a failed self-test
```

## **DriverWarning**

```
exception nidmm.DriverWarning
A warning originating from the NI-DMM driver
```

## 7.2.3.4 Examples

## nidmm\_fetch\_waveform.py

Listing 3: (nidmm\_fetch\_waveform.py)

```
#!/usr/bin/python
2
   import argparse
   import nidmm
   import sys
   import time
   def example(resource_name, options, function, range, points, rate):
       with nidmm.Session(resource_name=resource_name, options=options) as session:
10
           session.configure_waveform_acquisition(measurement_function=nidmm.
11
   →Function[function], range=range, rate=rate, waveform_points=points)
           with session.initiate():
12
               while True:
13
                    time.sleep(0.1)
14
                    backlog, acquisition_state = session.read_status()
15
                    if acquisition_state == nidmm.AcquisitionStatus.FINISHED_WITH_NO_
16
   →BACKLOG:
                        break
17
                    measurements = session.fetch_waveform(array_size=backlog)
18
                    print (measurements)
19
20
21
   def _main(argsv):
22
       parser = argparse.ArgumentParser(description='Performs a waveform acquisition_
23
   →using the NI-DMM API.', formatter_class=argparse.ArgumentDefaultsHelpFormatter)
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource_
24
   →name of a National Instruments Digital Multimeter.')
       parser.add_argument('-f', '--function', default='WAVEFORM_VOLTAGE', choices=nidmm.
   →Function.__members__.keys(), type=str.upper, help='Measurement function.')
       parser.add_argument('-r', '--range', default=10, type=float, help='Measurement...
26
   →range.')
                                                                               (continues on next page)
```

7.2. nidmm module 187

(continued from previous page)

```
parser.add_argument('-p', '--points', default=10, type=int, help='Specifies the...
27
   →number of points to acquire before the waveform acquisition completes.')
       parser.add_argument('-s', '--rate', default=1000, type=int, help='Specifies the_
28
   →rate of the acquisition in samples per second.')
       parser.add_argument('-op', '--option-string', default='', type=str, help='Option_
29
   ⇔string')
       args = parser.parse_args(argsv)
30
       example(args.resource_name, args.option_string, args.function, args.range, args.
31
   →points, args.rate)
32
33
   def main():
       _main(sys.argv[1:])
36
37
   def test_example():
38
       options = {'simulate': True, 'driver_setup': {'Model': '4082', 'BoardType': 'PXIe
39
   example ('PXI1Slot2', options, 'WAVEFORM_VOLTAGE', 10, 10, 1000)
40
41
42
   def test main():
43
       cmd_line = ['--option-string', 'Simulate=1, DriverSetup=Model:4082; BoardType:PXIe
44
   _main(cmd_line)
45
47
   if __name__ == '__main__':
48
       main()
49
50
51
```

#### nidmm measurement.pv

Listing 4: (nidmm\_measurement.py)

```
#!/usr/bin/python
2
   import argparse
3
   import nidmm
   import sys
   def example(resource_name, option_string, function, range, digits):
       with nidmm.Session(resource_name=resource_name, options=option_string) as session:
           session.configure_measurement_digits(measurement_function=nidmm.
10
   →Function[function], range=range, resolution_digits=digits)
           print (session.read())
11
12
13
   def _main(argsv):
14
       supported_functions = list(nidmm.Function.__members__.keys())
15
       parser = argparse.ArgumentParser(description='Performs a single measurement using,
   → the NI-DMM API.', formatter_class=argparse.ArgumentDefaultsHelpFormatter)
```

(continues on next page)

(continued from previous page)

```
parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource.
17
   →name of a National Instruments Digital Multimeter.')
       parser.add_argument('-f', '--function', default=supported_functions[0],_
18
   →choices=supported_functions, type=str.upper, help='Measurement function.')
       parser.add_argument('-r', '--range', default=10, type=float, help='Measurement_
19
   →range.')
       parser.add_argument('-d', '--digits', default=6.5, type=float, help='Digits of...
20
   ⇒resolution for the measurement.')
       parser.add_argument('-op', '--option-string', default='', type=str, help='Option_
21
   ⇔string')
       args = parser.parse_args(argsv)
22
       example(args.resource_name, args.option_string, args.function, args.range, args.
   →digits)
24
25
   def main():
26
       _main(sys.argv[1:])
27
28
29
   def test_example():
30
       options = {'simulate': True, 'driver_setup': {'Model': '4082', 'BoardType': 'PXIe
31
   example('PXI1Slot2', options, 'DC_VOLTS', 10, 6.5)
32
33
   def test_main():
      cmd_line = ['--option-string', 'Simulate=1, DriverSetup=Model:4082; BoardType:PXIe
36
   37
       _main(cmd_line)
38
39
   if __name__ == '__main__':
40
       main()
41
42
43
```

## nidmm multi point measurement.py

Listing 5: (nidmm\_multi\_point\_measurement.py)

```
#!/usr/bin/python
2
   import argparse
   import nidmm
   import sys
   def example(resource_name, options, function, range, digits, samples, triggers):
       with nidmm.Session(resource_name=resource_name, options=options) as session:
           session.configure_measurement_digits(measurement_function=nidmm.
10
   →Function[function], range=range, resolution_digits=digits)
           session.configure_multi_point(trigger_count=triggers, sample_count=samples)
11
           measurements = session.read_multi_point(array_size=samples)
12
           print('Measurements: ', measurements)
```

(continues on next page)

7.2. nidmm module 189

(continued from previous page)

```
14
15
   def _main(argsv):
16
       parser = argparse.ArgumentParser(description='Performs a multipoint measurement_
   →using the NI-DMM API.', formatter_class=argparse.ArgumentDefaultsHelpFormatter)
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource_
18
   →name of a National Instruments Digital Multimeter.')
       parser.add_argument('-f', '--function', default='DC_VOLTS', choices=nidmm.
19
   {\color{red} \hookrightarrow} \texttt{Function.} \underline{\quad} \texttt{members} \underline{\quad} . \texttt{keys(), type=str.upper, help='Measurement function.')}
       parser.add_argument('-r', '--range', default=10, type=float, help='Measurement_
20
   →range.')
       parser.add_argument('-d', '--digits', default=6.5, type=float, help='Digits of...
21
   ⇒resolution for the measurement.')
       parser.add_argument('-s', '--samples', default=10, type=int, help='The number of...
22
   →measurements the DMM makes.')
       parser.add_argument('-t', '--triggers', default=1, type=int, help='Sets the_
23
   →number of triggers you want the DMM to receive before returning to the Idle state.')
       parser.add_argument('-op', '--option-string', default='', type=str, help='Option_
    ⇔string')
       args = parser.parse_args(argsv)
25
       example(args.resource_name, args.option_string, args.function, args.range, args.
26
   →digits, args.samples, args.triggers)
27
28
   def main():
       _main(sys.argv[1:])
31
32
33
   def test_example():
       options = {'simulate': True, 'driver_setup': {'Model': '4082', 'BoardType': 'PXIe
34
   → ', }, }
       example('PXI1Slot2', options, 'DC_VOLTS', 10, 6.5, 10, 1)
35
36
37
   def test main():
38
       cmd_line = ['--option-string', 'Simulate=1, DriverSetup=Model:4082; BoardType:PXIe
   _main(cmd_line)
42
   if name == ' main ':
43
44
       main()
45
46
```

# 7.3 nifgen module

## 7.3.1 Installation

As a prerequisite to using the nifgen module, you must install the NI-FGEN runtime on your system. Visit ni.com/downloads to download the driver runtime for your devices.

The nimi-python modules (i.e. for **NI-FGEN**) can be installed with pip:

```
$ python -m pip install nifgen~=1.0.1
```

## Or easy\_install from setuptools:

```
$ python -m easy_install nifgen
```

## **7.3.2 Usage**

The following is a basic example of using the **nifgen** module to open a session to a Function Generator and generate a sine wave for 5 seconds.

Additional examples for NI-FGEN are located in src/nifgen/examples/ directory.

## 7.3.3 API Reference

## 7.3.3.1 nifgen.Session

class nifgen. Session (self, resource\_name, channel\_name=None, reset\_device=False, options={})

Creates and returns a new NI-FGEN session to the specified channel of a waveform generator that is used in all subsequent NI-FGEN method calls.

#### **Parameters**

• resource\_name (str) -

**Caution:** Traditional NI-DAQ and NI-DAQmx device names are not case-sensitive. However, all IVI names, such as logical names, are case-sensitive. If you use logical names, driver session names, or virtual names in your program, you must ensure that the name you use matches the name in the IVI Configuration Store file exactly, without any variations in the case of the characters.

Specifies the resource name of the device to initialize.

For Traditional NI-DAQ devices, the syntax is DAQ::n, where n is the device number assigned by MAX, as shown in Example 1.

For NI-DAQmx devices, the syntax is just the device name specified in MAX, as shown in Example 2. Typical default names for NI-DAQmx devices in MAX are Dev1 or PXI1Slot1. You can rename an NI-DAQmx device by right-clicking on the name in MAX and entering a new name.

An alternate syntax for NI-DAQmx devices consists of DAQ::*NI-DAQmx device name*, as shown in Example 3. This naming convention allows for the use of an NI-DAQmx device in an application that was originally designed for a Traditional NI-DAQ device. For example, if the application expects DAQ::1, you can rename the NI-DAQmx device to 1 in MAX and pass in DAQ::1 for the resource name, as shown in Example 4.

If you use the DAQ::*n* syntax and an NI-DAQmx device name already exists with that same name, the NI-DAQmx device is matched first.

You can also pass in the name of an IVI logical name or an IVI virtual name configured with the IVI Configuration utility, as shown in Example 5. A logical name identifies a particular virtual instrument. A virtual name identifies a specific device and specifies the initial settings for the session.

Ex- ample #	Device Type	Syntax	Variable
1	Traditional NI-DAQ device	DAQ::1	(1 = device number)
2	NI-DAQmx device	myDAQmxDevice	(myDAQmxDevice = device name)
3	NI-DAQmx device	DAQ::myDAQmxD	ev(inneyDAQmxDevice = de- vice name)
4	NI-DAQmx device	DAQ::2	(2 = device name)
5	IVI logical name or IVI virtual name	myLogicalName	(myLogicalName = name)

• **channel\_name** (str) – Specifies the channel that this VI uses.

Default Value: "0"

• **reset\_device** (bool) – Specifies whether you want to reset the device during the initialization procedure. True specifies that the device is reset and performs the same method as the nifgen.Session.Reset() method.

## \*\*Defined Values\*\*

Default Value: False

True	Reset device
False	Do not reset device

• **options** (*str*) – Specifies the initial value of certain properties for the session. The syntax for **options** is a dictionary of properties with an assigned value. For example:

```
{ 'simulate': False }
```

You do not have to specify a value for all the properties. If you do not specify a value for a property, the default value is used.

Advanced Example: { 'simulate': True, 'driver\_setup': { 'Model': '<model number>', 'BoardType': '<type>' } }

Property	Default
range_check	True
query_instrument_status	False
cache	True
simulate	False
record_value_coersions	False
driver_setup	{}

# **Properties**

Property	Datatype	
all_marker_events_latched_status	int	
all_marker_events_live_status	int	
analog_data_mask	int	
analog_filter_enabled	bool	
analog_path	AnalogPath	
analog_static_value	int	
arb_gain	float	
arb_marker_position	int	
arb_offset	float	
arb_repeat_count	int	
arb_sample_rate	float	
arb_sequence_handle	int	
arb_waveform_handle	int	
aux_power_enabled	bool	
bus_type	BusType	
channel_count	int	
channel_delay	float	
clock_mode	ClockMode	
common_mode_offset	float	
data_marker_events_count	int	
data_marker_event_data_bit_number	int	
data_marker_event_level_polarity	DataMarkerEventLevelPolarity	
data_marker_event_output_terminal	str	
data_transfer_block_size	int	
data_transfer_maximum_bandwidth	float	
data_transfer_maximum_in_flight_reads	int	
data_transfer_preferred_packet_size	int	
digital_data_mask	int	
digital_edge_script_trigger_edge	ScriptTriggerDigitalEdgeEdge	
digital_edge_script_trigger_source	str	
digital_edge_start_trigger_edge	StartTriggerDigitalEdgeEdge	
digital_edge_start_trigger_source	str	
digital_filter_enabled	bool	
digital_filter_interpolation_factor	float	
digital_gain	float	
digital_pattern_enabled	bool	
digital_static_value	int	
done_event_output_terminal	str	
driver_setup	str Continued on next page	

Continued on next page

Table 7 – continued from previous page

Property Table / – continued from previous	Datatype
exported_onboard_reference_clock_output_terminal	* *
exported_reference_clock_output_terminal	str
exported_sample_clock_divisor	int
exported_sample_clock_output_terminal	str
exported_sample_clock_timebase_divisor	int
exported_sample_clock_timebase_output_terminal	str
exported_script_trigger_output_terminal	str
exported_start_trigger_output_terminal	str
external_clock_delay_binary_value	int
external_sample_clock_multiplier	float
file_transfer_block_size	int
filter_correction_frequency	float
flatness_correction_enabled	bool
fpga_bitfile_path	str
freq_list_duration_quantum	float
freq_list_handle	int
func_amplitude	float
func_buffer_size	int
func_dc_offset	float
func_duty_cycle_high	float
func_frequency	float
func_max_buffer_size	int
func_start_phase	float
func_waveform	Waveform
idle_behavior	IdleBehavior
idle_value	int
instrument_firmware_revision	str
instrument_manufacturer	str
instrument_model	str
io_resource_descriptor	str
load_impedance	float
logical_name	str
major_version	int
marker_events_count	int
marker_event_output_terminal	str
max_freq_list_duration	float
max_freq_list_length	int
max_loop_count	int
max_num_freq_lists	int
max_num_sequences	int
max_num_waveforms	int
max_sequence_length	int
max_waveform_size	int
memory_size	int
memory_size minor_version	int int
minor_version	int
minor_version min_freq_list_duration	int float

Continued on next page

Table 7 – continued from previous page

Property	Datatype	
module_revision	str	
output_enabled	bool	
output_impedance	float	
output_mode	OutputMode	
ready_for_start_event_output_terminal	str	
reference_clock_source	ReferenceClockSource	
ref_clock_frequency	float	
sample_clock_source	SampleClockSource	
sample_clock_timebase_rate	float	
sample_clock_timebase_source	SampleClockTimebaseSource	
script_to_generate	str	
script_triggers_count	int	
script_trigger_type	ScriptTriggerType	
serial_number	str	
simulate	bool	
specific_driver_description	str	
specific_driver_revision	str	
specific_driver_vendor	str	
started_event_output_terminal	str	
start_trigger_type	StartTriggerType	
streaming_space_available_in_waveform	int	
streaming_waveform_handle	int	
streaming_waveform_name	str	
streaming_write_timeout	float in seconds or datetime.timedelta	
supported_instrument_models	str	
terminal_configuration	TerminalConfiguration	
trigger_mode	TriggerMode	
wait_behavior	WaitBehavior	
wait_value	int	
waveform_quantum	int	

## **Public methods**

Method name
abort()
allocate_named_waveform()
allocate_waveform()
clear_arb_memory()
clear_arb_sequence()
clear_freq_list()
clear_user_standard_waveform()
commit()
configure_arb_sequence()
configure_arb_waveform()
<pre>configure_custom_fir_filter_coefficients()</pre>
configure_freq_list()
<pre>configure_standard_waveform()</pre>
create_advanced_arb_sequence()
create_arb_sequence()

Continued on next page

Table 8 – continued from previous page

Method name
create_freq_list()
create_waveform_from_file_f64()
create_waveform_from_file_i16()
create_waveform_numpy()
define_user_standard_waveform()
delete_script()
delete_waveform()
disable()
<pre>get_ext_cal_last_date_and_time()</pre>
<pre>get_ext_cal_last_temp()</pre>
<pre>get_ext_cal_recommended_interval()</pre>
<pre>get_hardware_state()</pre>
<pre>get_self_cal_last_date_and_time()</pre>
<pre>get_self_cal_last_temp()</pre>
<pre>get_self_cal_supported()</pre>
is_done()
lock()
query_arb_seq_capabilities()
<pre>query_arb_wfm_capabilities()</pre>
<pre>query_freq_list_capabilities()</pre>
read_current_temperature()
reset()
reset_device()
reset_with_defaults()
self_cal()
self_test()
send_software_edge_trigger()
set_next_write_position()
unlock()
wait_until_done()
write_script()
write_waveform()

## **Properties**

## all\_marker\_events\_latched\_status

# $\verb|nifgen.Session.all_marker_events_latched_status|\\$

Returns a bit field of the latched status of all Marker Events. Write 0 to this property to clear the latched status of all Marker Events.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Marker:Advanced:All Marker Events Latched Status
- C Attribute: NIFGEN\_ATTR\_ALL\_MARKER\_EVENTS\_LATCHED\_STATUS

# all\_marker\_events\_live\_status

## nifgen.Session.all\_marker\_events\_live\_status

Returns a bit field of the live status of all Marker Events.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Marker:Advanced:All Marker Events Live Status
- C Attribute: NIFGEN\_ATTR\_ALL\_MARKER\_EVENTS\_LIVE\_STATUS

## analog\_data\_mask

## nifgen.Session.analog\_data\_mask

Specifies the mask to apply to the analog output. The masked data is replaced with the data in nifgen.Session.analog\_static\_value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Data Mask:Analog Data Mask
- C Attribute: NIFGEN\_ATTR\_ANALOG\_DATA\_MASK

## analog filter enabled

#### nifgen.Session.analog\_filter\_enabled

Controls whether the signal generator applies to an analog filter to the output signal. This property is valid in arbitrary waveform, arbitrary sequence, and script modes. This property can also be used in standard method and frequency list modes for user-defined waveforms.

The following table lists the characteristics of this property.

Value
bool
read-write
read write
False
No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Filters:Analog Filter Enabled
- C Attribute: NIFGEN\_ATTR\_ANALOG\_FILTER\_ENABLED

## analog\_path

## nifgen.Session.analog\_path

Specifies the analog signal path that should be used. The main path allows you to configure gain, offset, analog filter status, output impedance, and output enable. The main path has two amplifier options, high- and low-gain. The direct path presents a much smaller gain range, and you cannot adjust offset or the filter status. The direct path also provides a smaller output range but also lower distortion. NI-FGEN normally chooses the amplifier based on the user-specified gain.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.AnalogPath
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Output: Analog Path

• C Attribute: NIFGEN\_ATTR\_ANALOG\_PATH

## analog static value

```
nifgen.Session.analog_static_value
```

Specifies the static value that replaces data masked by nifgen. Session. analog\_data\_mask.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Data Mask:Analog Static Value
- C Attribute: NIFGEN\_ATTR\_ANALOG\_STATIC\_VALUE

## arb gain

## nifgen.Session.arb\_gain

Specifies the factor by which the signal generator scales the arbitrary waveform data. When you create arbitrary waveforms, you must first normalize the data points to the range -1.0 to +1.0. Use this property to scale the arbitrary waveform to other ranges. For example, when you set this property to 2.0, the output signal ranges from -2.0 V to +2.0 V. Use this property when nifgen.Session.output\_mode is set to ARB or SEQ.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Arbitrary Waveform: Gain

• C Attribute: NIFGEN\_ATTR\_ARB\_GAIN

## arb marker position

## nifgen.Session.arb\_marker\_position

Specifies the position for a marker to be asserted in the arbitrary waveform. This property defaults to -1 when no marker position is specified. Use this property when <code>nifgen.Session.output\_mode</code> is set to <code>ARB</code>. Use <code>nifgen.Session.ExportSignal()</code> to export the marker signal.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This property can use repeated capabilities (markers). If set or get directly on the nifgen. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling set/get value on the result.:

```
session.markers[0,1].arb_marker_position = var
var = session.markers[0,1].arb_marker_position
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

Tip: This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Arbitrary Waveform Mode: Marker Position
- C Attribute: NIFGEN\_ATTR\_ARB\_MARKER\_POSITION

## arb offset

#### nifgen.Session.arb\_offset

Specifies the value that the signal generator adds to the arbitrary waveform data. When you create arbitrary waveforms, you must first normalize the data points to the range -1.0 to +1.0. Use this property to shift the arbitrary waveform range. For example, when you set this property to 1.0, the output signal ranges from 2.0 V to 0.0 V. Use this property when <code>nifgen.Session.output\_mode</code> is set to <code>ARB</code> or <code>SEQ</code>. Units: Volts

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Arbitrary Waveform:Offset

C Attribute: NIFGEN\_ATTR\_ARB\_OFFSET

#### arb repeat count

#### nifgen.Session.arb\_repeat\_count

Specifies number of times to repeat the arbitrary waveform when the triggerMode parameter of nifgen.Session.ConfigureTriggerMode() is set to SINGLE or STEPPED. This property is ignored if the triggerMode parameter is set to CONTINUOUS or BURST. Use this property when nifgen.Session.output\_mode is set to ARB. When used during streaming, this property specifies the number of times to repeat the streaming waveform (the onboard memory allocated for streaming). For more information about streaming, refer to the Streaming topic.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Arbitrary Waveform Mode: Repeat Count
- C Attribute: NIFGEN\_ATTR\_ARB\_REPEAT\_COUNT

## arb sample rate

## nifgen.Session.arb\_sample\_rate

Specifies the rate at which the signal generator outputs the points in arbitrary waveforms. Use this property when <code>nifgen.Session.output\_mode</code> is set to <code>ARB</code> or <code>SEQ</code>. Units: Samples/s

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Sample Clock:Rate
- C Attribute: NIFGEN\_ATTR\_ARB\_SAMPLE\_RATE

#### arb sequence handle

## nifgen.Session.arb\_sequence\_handle

This channel-based property identifies which sequence the signal generator produces. You can

create multiple sequences using nifgen. Session.create\_arb\_sequence(). nifgen. Session.create\_arb\_sequence() returns a handle that you can use to identify the particular sequence. To configure the signal generator to produce a particular sequence, set this property to the sequence handle. Use this property only when nifgen.Session.output\_mode is set to SEQ.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Arbitrary Sequence Mode: Arbitrary Sequence Handle
- C Attribute: NIFGEN\_ATTR\_ARB\_SEQUENCE\_HANDLE

## arb waveform handle

#### nifgen.Session.arb\_waveform\_handle

Selects which arbitrary waveform the signal generator produces. You can create multiple arbitrary waveforms using one of the following niFgen Create Waveform methods: nifgen.Session.create\_waveform() nifgen.Session.create\_waveform\_from\_file\_i16() nifgen.Session.create\_waveform\_from\_file\_i16() nifgen.Session.create\_waveform\_from\_file\_i64() nifgen.Session.CreateWaveformFromFileHWS() These methods return a handle that you can use to identify the particular waveform. To configure the signal generator to produce a particular waveform, set this property to the waveform handle. Use this property only when nifgen.Session.output\_mode is set to ARB.

Note: One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Arbitrary Waveform Mode: Arbitrary Waveform Handle
- C Attribute: NIFGEN\_ATTR\_ARB\_WAVEFORM\_HANDLE

## aux\_power\_enabled

## nifgen.Session.aux\_power\_enabled

Controls the specified auxiliary power pin. Setting this property to TRUE energizes the auxiliary power when the session is committed. When this property is FALSE, the power pin of the connector outputs no power.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Advanced:AUX Power Enabled
- C Attribute: NIFGEN\_ATTR\_AUX\_POWER\_ENABLED

## bus\_type

## nifgen.Session.bus\_type

The bus type of the signal generator.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.BusType
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Bus Type
- C Attribute: NIFGEN\_ATTR\_BUS\_TYPE

## channel count

## nifgen.Session.channel\_count

Returns the number of channels that the specific instrument driver supports. For each property for which IVI\_VAL\_MULTI\_CHANNEL is set, the IVI Engine maintains a separate cache value for each channel.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Inherent IVI Attributes:Driver Capabilities:Channel Count
- C Attribute: NIFGEN\_ATTR\_CHANNEL\_COUNT

## channel\_delay

## nifgen.Session.channel\_delay

Specifies, in seconds, the delay to apply to the analog output of the channel specified by the channel string. You can use the channel delay to configure the timing relationship between channels on a multichannel device. Values for this property can be zero or positive. A value of zero indicates that the channels are aligned. A positive value delays the analog output by the specified number of seconds.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output: Channel Delay
- C Attribute: NIFGEN\_ATTR\_CHANNEL\_DELAY

## clock\_mode

## nifgen.Session.clock\_mode

Controls which clock mode is used for the signal generator. For signal generators that support it, this property allows switching the sample clock to High-Resolution mode. When in Divide-Down mode, the sample rate can only be set to certain frequences, based on dividing down the update clock. However, in High-Resolution mode, the sample rate may be set to any value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ClockMode
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Clocks:Sample Clock:Mode

• C Attribute: NIFGEN\_ATTR\_CLOCK\_MODE

#### common mode offset

## nifgen.Session.common\_mode\_offset

Specifies, in volts, the value the signal generator adds to or subtracts from the arbitrary waveform data. This property applies only when you set the <code>nifgen.Session.terminal\_configuration</code> property to <code>DIFFERENTIAL</code>. Common mode offset is applied to the signals generated at each differential output terminal.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

LabVIEW Property: Output:Common Mode Offset

• C Attribute: NIFGEN\_ATTR\_COMMON\_MODE\_OFFSET

## data\_marker\_events\_count

## nifgen.Session.data\_marker\_events\_count

Returns the number of Data Marker Events supported by the device.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Data Marker Events Count
- C Attribute: NIFGEN\_ATTR\_DATA\_MARKER\_EVENTS\_COUNT

## data marker event data bit number

```
nifgen.Session.data_marker_event_data_bit_number
Specifies the bit number to assign to the Data Marker Event.
```

**Tip:** This property can use repeated capabilities (markers). If set or get directly on the nifgen. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling set/get value on the result.:

```
session.markers[0,1].data_marker_event_data_bit_number = var
var = session.markers[0,1].data_marker_event_data_bit_number
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Data Marker:Data Bit Number
- C Attribute: NIFGEN\_ATTR\_DATA\_MARKER\_EVENT\_DATA\_BIT\_NUMBER

## data marker event level polarity

```
nifgen.Session.data_marker_event_level_polarity
Specifies the output polarity of the Data marker event.
```

**Tip:** This property can use repeated capabilities (markers). If set or get directly on the nifgen. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling set/get value on the result.:

```
session.markers[0,1].data_marker_event_level_polarity = var
var = session.markers[0,1].data_marker_event_level_polarity
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.DataMarkerEventLevelPolarity
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Data Marker:Level:Active Level
- C Attribute: NIFGEN\_ATTR\_DATA\_MARKER\_EVENT\_LEVEL\_POLARITY

## data marker event output terminal

nifgen.Session.data\_marker\_event\_output\_terminal

Specifies the destination terminal for the Data Marker Event.

**Tip:** This property can use repeated capabilities (markers). If set or get directly on the nifgen. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling set/get value on the result.:

```
session.markers[0,1].data_marker_event_output_terminal = var
var = session.markers[0,1].data_marker_event_output_terminal
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Data Marker:Output Terminal
- C Attribute: NIFGEN\_ATTR\_DATA\_MARKER\_EVENT\_OUTPUT\_TERMINAL

## data\_transfer\_block\_size

nifgen.Session.data\_transfer\_block\_size

The number of samples at a time to download to onboard memory. Useful when the total data to be transferred to onboard memory is large.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform:Data Transfer:Data Transfer Block Size
- C Attribute: NIFGEN\_ATTR\_DATA\_TRANSFER\_BLOCK\_SIZE

#### data transfer maximum bandwidth

#### nifgen.Session.data\_transfer\_maximum\_bandwidth

Specifies the maximum amount of bus bandwidth (in bytes per second) to use for data transfers. The signal generator limits data transfer speeds on the PCIe bus to the value you specify for this property. Set this property to optimize bus bandwidth usage for multi-device streaming applications by preventing the signal generator from consuming all of the available bandwidth on a PCI express link when waveforms are being written to the onboard memory of the device.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform:Data Transfer:Maximum Bandwidth
- C Attribute: NIFGEN\_ATTR\_DATA\_TRANSFER\_MAXIMUM\_BANDWIDTH

#### data transfer maximum in flight reads

#### nifgen.Session.data transfer maximum in flight reads

Specifies the maximum number of concurrent PCI Express read requests the signal generator can issue. When transferring data from computer memory to device onboard memory across the PCI Express bus, the signal generator can issue multiple memory reads at the same time. In general, the larger the number of read requests, the more efficiently the device uses the bus because the multiple read requests keep the data flowing, even in a PCI Express topology that has high latency due to PCI Express switches in the data path. Most NI devices can issue a large number of read requests (typically 8 or 16). By default, this property is set to the highest value the signal generator supports. If other devices in your system cannot tolerate long data latencies, it may be helpful to decrease the number of in-flight read requests the NI signal generator issues. This helps to reduce the amount of data the signal generator reads at one time.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform:Data Transfer:Advanced:Maximum In-Flight Read Requests
- C Attribute: NIFGEN\_ATTR\_DATA\_TRANSFER\_MAXIMUM\_IN\_FLIGHT\_READS

## data\_transfer\_preferred\_packet\_size

## nifgen.Session.data\_transfer\_preferred\_packet\_size

Specifies the preferred size of the data field in a PCI Express read request packet. In general, the larger the packet size, the more efficiently the device uses the bus. By default, NI signal generators use the largest packet size allowed by the system. However, due to different system implementations, some systems may perform better with smaller packet sizes. Recommended values for this property are powers of two between 64 and 512. In some cases, the signal generator generates packets smaller than the preferred size you set with this property. You cannot change this property while the device is generating a waveform. If you want to change the device configuration, call the <code>nifgen.Session.abort()</code> method or wait for the generation to complete.

#### Note: :

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Data Transfer: Advanced: Preferred Packet Size
- C Attribute: NIFGEN\_ATTR\_DATA\_TRANSFER\_PREFERRED\_PACKET\_SIZE

## digital\_data\_mask

## nifgen.Session.digital\_data\_mask

Specifies the mask to apply to the output on the digital connector. The masked data is replaced with the data in nifgen. Session.digital\_static\_value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Data Mask:Digital Data Mask
- C Attribute: NIFGEN\_ATTR\_DIGITAL\_DATA\_MASK

## digital\_edge\_script\_trigger\_edge

## nifgen.Session.digital\_edge\_script\_trigger\_edge

Specifies the active edge for the Script trigger. This property is used when nifgen. Session. script\_trigger\_type is set to Digital Edge.

**Tip:** This property can use repeated capabilities (script\_triggers). If set or get directly on the nifgen. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling set/get value on the result.:

```
session.script_triggers[0,1].digital_edge_script_trigger_edge = var
var = session.script_triggers[0,1].digital_edge_script_trigger_edge
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ScriptTriggerDigitalEdgeEdge
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Script:Digital Edge:Edge
- $\bullet \ \ C \ Attribute: \\ \textbf{NIFGEN\_ATTR\_DIGITAL\_EDGE\_SCRIPT\_TRIGGER\_EDGE}$

## digital\_edge\_script\_trigger\_source

```
nifgen.Session.digital_edge_script_trigger_source
```

Specifies the source terminal for the Script trigger. This property is used when nifgen.  $Session.script\_trigger\_type$  is set to Digital Edge.

**Tip:** This property can use repeated capabilities (script\_triggers). If set or get directly on the nifgen. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling set/get value on the result.:

```
session.script_triggers[0,1].digital_edge_script_trigger_source = var
var = session.script_triggers[0,1].digital_edge_script_trigger_source
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Script:Digital Edge:Source
- C Attribute: NIFGEN\_ATTR\_DIGITAL\_EDGE\_SCRIPT\_TRIGGER\_SOURCE

# digital\_edge\_start\_trigger\_edge

```
nifgen.Session.digital_edge_start_trigger_edge
```

Specifies the active edge for the Start trigger. This property is used only when nifgen. Session. start\_trigger\_type is set to Digital Edge.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.StartTriggerDigitalEdgeEdge
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Start:Digital Edge:Edge
- C Attribute: NIFGEN\_ATTR\_DIGITAL\_EDGE\_START\_TRIGGER\_EDGE

## digital\_edge\_start\_trigger\_source

```
nifgen.Session.digital_edge_start_trigger_source
```

Specifies the source terminal for the Start trigger. This property is used only when nifgen. Session.start\_trigger\_type is set to Digital Edge.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Start:Digital Edge:Source
- C Attribute: NIFGEN\_ATTR\_DIGITAL\_EDGE\_START\_TRIGGER\_SOURCE

## digital\_filter\_enabled

## nifgen.Session.digital\_filter\_enabled

Controls whether the signal generator applies a digital filter to the output signal. This property is valid in arbitrary waveform, arbitrary sequence, and script modes. This property can also be used in standard method and frequency list modes for user-defined waveforms.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Filters:Digital Filter Enabled
- C Attribute: NIFGEN\_ATTR\_DIGITAL\_FILTER\_ENABLED

## digital filter interpolation factor

# $\verb|nifgen.Session.digital_filter_interpolation_factor|\\$

This property only affects the device when nifgen. Session.digital\_filter\_enabled is set to True. If you do not set this property directly, NI-FGEN automatically selects the maximum interpolation factor allowed for the current sample rate. Valid values are 2, 4, and 8.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Output: Filters: Digital Filter Interpolation Factor
- C Attribute: NIFGEN ATTR DIGITAL FILTER INTERPOLATION FACTOR

# digital\_gain

# nifgen.Session.digital\_gain

Specifies a factor by which the signal generator digitally multiplies generated data before converting it to an analog signal in the DAC. For a digital gain greater than 1.0, the product of digital gain times the generated data must be inside the range plus or minus 1.0 (assuming floating point data). If the product exceeds these limits, the signal generator clips the output signal, and an error results. Some signal generators support both digital gain and an analog gain (analog gain is specified with the nifgen.Session.func\_amplitude property or the nifgen.Session.arb\_gain property). Digital gain can be changed during generation without the glitches that may occur when changing analog gains, due to relay switching. However, the DAC output resolution is a method of analog gain, so only analog gain makes full use of the resolution of the DAC.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Digital Gain
- C Attribute: NIFGEN\_ATTR\_DIGITAL\_GAIN

# digital\_pattern\_enabled

#### nifgen.Session.digital\_pattern\_enabled

Controls whether the signal generator generates a digital pattern of the output signal.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Output:Advanced:Digital Pattern Enabled

## • C Attribute: NIFGEN\_ATTR\_DIGITAL\_PATTERN\_ENABLED

# digital\_static\_value

nifgen.Session.digital\_static\_value

Specifies the static value that replaces data masked by nifgen. Session. digital\_data\_mask.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Data Mask:Digital Static Value
- C Attribute: NIFGEN\_ATTR\_DIGITAL\_STATIC\_VALUE

#### done event output terminal

nifgen.Session.done\_event\_output\_terminal

Specifies the destination terminal for the Done Event.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Done:Output Terminal
- C Attribute: NIFGEN\_ATTR\_DONE\_EVENT\_OUTPUT\_TERMINAL

### driver setup

nifgen.Session.driver\_setup

Specifies the driver setup portion of the option string that was passed into the nifgen. Session. InitWithOptions () method.

Note: One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	0
Resettable	0

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NIFGEN\_ATTR\_DRIVER\_SETUP

# exported\_onboard\_reference\_clock\_output\_terminal

nifgen.Session.exported\_onboard\_reference\_clock\_output\_terminal Specifies the terminal to which to export the Onboard Reference Clock.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Reference Clock:Onboard Reference Clock:Export Output Terminal
- C Attribute: NIFGEN\_ATTR\_EXPORTED\_ONBOARD\_REFERENCE\_CLOCK\_OUTPUT\_TERMINAL

# exported\_reference\_clock\_output\_terminal

nifgen.Session.exported\_reference\_clock\_output\_terminal

Specifies the terminal to which to export the Reference Clock.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

- LabVIEW Property: Clocks:Reference Clock:Export Output Terminal
- C Attribute: NIFGEN\_ATTR\_EXPORTED\_REFERENCE\_CLOCK\_OUTPUT\_TERMINAL

## exported\_sample\_clock\_divisor

# nifgen.Session.exported\_sample\_clock\_divisor

Specifies the factor by which to divide the Sample clock, also known as the Update clock, before it is exported. To export the Sample clock, use the nifgen.Session.ExportSignal() method or the nifgen.Session.exported\_sample\_clock\_output\_terminal property.

Note: One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Sample Clock:Exported Sample Clock Divisor
- C Attribute: NIFGEN\_ATTR\_EXPORTED\_SAMPLE\_CLOCK\_DIVISOR

# exported\_sample\_clock\_output\_terminal

nifgen.Session.exported\_sample\_clock\_output\_terminal

Specifies the terminal to which to export the Sample Clock.

The following table lists the characteristics of this property.

Value
str
read-write
False
Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Sample Clock:Export Output Terminal
- C Attribute: NIFGEN\_ATTR\_EXPORTED\_SAMPLE\_CLOCK\_OUTPUT\_TERMINAL

## exported sample clock timebase divisor

## nifgen.Session.exported\_sample\_clock\_timebase\_divisor

Specifies the factor by which to divide the sample clock timebase (board clock) before it is exported. To export the Sample clock timebase, use the nifgen.Session.ExportSignal() method or the nifgen.Session.exported\_sample\_clock\_timebase\_output\_terminal property.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Sample Clock Timebase:Exported Sample Clock Timebase
   Divisor
- C Attribute: NIFGEN ATTR EXPORTED SAMPLE CLOCK TIMEBASE DIVISOR

### exported\_sample\_clock\_timebase\_output\_terminal

## nifgen.Session.exported\_sample\_clock\_timebase\_output\_terminal

Specifies the terminal to which to export the Sample clock timebase. If you specify a divisor with the <code>nifgen.Session.exported\_sample\_clock\_timebase\_divisor</code> property, the Sample clock exported with the <code>nifgen.Session.exported\_sample\_clock\_timebase\_output\_terminal</code> property is the value of the Sample clock timebase after it is divided-down. For a list of the terminals available on your device, refer to the Device Routes tab in MAX. To change the device configuration, call <code>nifgen.Session.abort()</code> or wait for the generation to complete.

**Note:** The signal generator must not be in the Generating state when you change this property.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Sample Clock Timebase:Export Output Terminal
- C Attribute: NIFGEN\_ATTR\_EXPORTED\_SAMPLE\_CLOCK\_TIMEBASE\_OUTPUT\_TERMINAL

### exported script trigger output terminal

#### nifgen.Session.exported\_script\_trigger\_output\_terminal

Specifies the output terminal for the exported Script trigger. Setting this property to an empty string means that when you commit the session, the signal is removed from that terminal and, if possible, the terminal is tristated.

**Tip:** This property can use repeated capabilities (script\_triggers). If set or get directly on the nifgen. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling set/get value on the result.:

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Script:Output Terminal
- C Attribute: NIFGEN\_ATTR\_EXPORTED\_SCRIPT\_TRIGGER\_OUTPUT\_TERMINAL

# exported\_start\_trigger\_output\_terminal

# $\verb|nifgen.Session.exported_start_trigger_output_terminal|\\$

Specifies the destination terminal for exporting the Start trigger.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

- LabVIEW Property: Triggers:Start:Output Terminal
- C Attribute: NIFGEN\_ATTR\_EXPORTED\_START\_TRIGGER\_OUTPUT\_TERMINAL

# external\_clock\_delay\_binary\_value

nifgen.Session.external\_clock\_delay\_binary\_value

Binary value of the external clock delay.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Advanced:External Clock Delay Binary Value
- C Attribute: NIFGEN\_ATTR\_EXTERNAL\_CLOCK\_DELAY\_BINARY\_VALUE

# external\_sample\_clock\_multiplier

### nifgen.Session.external\_sample\_clock\_multiplier

Specifies a multiplication factor to use to obtain a desired sample rate from an external Sample clock. The resulting sample rate is equal to this factor multiplied by the external Sample clock rate. You can use this property to generate samples at a rate higher than your external clock rate. When using this property, you do not need to explicitly set the external clock rate.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Advanced:External Sample Clock Multiplier
- C Attribute: NIFGEN\_ATTR\_EXTERNAL\_SAMPLE\_CLOCK\_MULTIPLIER

## file transfer block size

### nifgen.Session.file\_transfer\_block\_size

The number of samples at a time to read from the file and download to onboard memory. Used in conjunction with the Create From File and Write From File methods.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform:Data Transfer:File Transfer Block Size
- C Attribute: NIFGEN\_ATTR\_FILE\_TRANSFER\_BLOCK\_SIZE

## filter\_correction\_frequency

#### nifgen.Session.filter correction frequency

Controls the filter correction frequency of the analog filter. This property corrects for the ripples in the analog filter frequency response at the frequency specified. For standard waveform output, the filter correction frequency should be set to be the same as the frequency of the standard waveform. To have no filter correction, set this property to 0 Hz.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:5401/5411/5431:Filter Correction Frequency
- C Attribute: NIFGEN\_ATTR\_FILTER\_CORRECTION\_FREQUENCY

# flatness\_correction\_enabled

# nifgen.Session.flatness\_correction\_enabled

When True, the signal generator applies a flatness correction factor to the generated sine wave in order to ensure the same output power level at all frequencies. This property should be set to False when performing Flatness Calibration.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	Yes

- LabVIEW Property: Output:Filters:Flatness Correction Enabled
- C Attribute: NIFGEN\_ATTR\_FLATNESS\_CORRECTION\_ENABLED

### fpga bitfile path

### nifgen.Session.fpga\_bitfile\_path

Gets the absolute file path to the bitfile loaded on the FPGA.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:FPGA Bitfile Path
- C Attribute: NIFGEN\_ATTR\_FPGA\_BITFILE\_PATH

# freq\_list\_duration\_quantum

## nifgen.Session.freq\_list\_duration\_quantum

Returns the quantum of which all durations must be a multiple in a frequency list.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Standard Function:Frequency List Mode:Frequency List Duration Quantum

## • C Attribute: NIFGEN\_ATTR\_FREQ\_LIST\_DURATION\_QUANTUM

# freq\_list\_handle

#### nifgen.Session.freq list handle

Sets which frequency list the signal generator produces. Create a frequency list using <code>nifgen.Session.create\_freq\_list().nifgen.Session.create\_freq\_list()</code> returns a handle that you can use to identify the list.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function:Frequency List Mode:Frequency List Handle
- C Attribute: NIFGEN\_ATTR\_FREQ\_LIST\_HANDLE

# func\_amplitude

#### nifgen.Session.func amplitude

Controls the amplitude of the standard waveform that the signal generator produces. This value is the amplitude at the output terminal. For example, to produce a waveform ranging from -5.00 V to +5.00 V, set the amplitude to 10.00 V. set the Waveform parameter to DC. Units: Vpk-pk

**Note:** This parameter does not affect signal generator behavior when you

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function: Amplitude
- C Attribute: NIFGEN\_ATTR\_FUNC\_AMPLITUDE

## func buffer size

#### nifgen.Session.func\_buffer\_size

This property contains the number of samples used in the standard method waveform buffer. This property is only valid on devices that implement standard method mode in software, and is read-only for all other devices. implementation of Standard Method Mode on your device.

**Note:** Refer to the Standard Method Mode topic for more information on the

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function:Standard Function Mode:Buffer Size
- C Attribute: NIFGEN ATTR FUNC BUFFER SIZE

# func\_dc\_offset

## nifgen.Session.func\_dc\_offset

Controls the DC offset of the standard waveform that the signal generator produces. This value is the offset at the output terminal. The value is the offset from ground to the center of the waveform that you specify with the Waveform parameter. For example, to configure a waveform with an amplitude of 10.00 V to range from 0.00 V to +10.00 V, set DC Offset to 5.00 V. Units: volts

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function:DC Offset
- C Attribute: NIFGEN\_ATTR\_FUNC\_DC\_OFFSET

# func\_duty\_cycle\_high

#### nifgen.Session.func\_duty\_cycle\_high

Controls the duty cycle of the square wave the signal generator produces. Specify this property as a percentage of the time the square wave is high in a cycle. set the Waveform parameter to *SQUARE*. Units: Percentage of time the waveform is high

**Note:** This parameter only affects signal generator behavior when you

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function: Duty Cycle High
- C Attribute: NIFGEN ATTR FUNC DUTY CYCLE HIGH

# func\_frequency

# nifgen.Session.func\_frequency

Controls the frequency of the standard waveform that the signal generator produces. Units: hertz (1) This parameter does not affect signal generator behavior when you set the Waveform parameter of the nifgen.Session.configure\_standard\_waveform() method to DC. (2) For SINE, the range is between 0 MHz and 16 MHz, but the range is between 0 MHz and 1 MHz for all other waveforms.

#### Note: :

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function:Standard Function Mode:Frequency
- C Attribute: NIFGEN\_ATTR\_FUNC\_FREQUENCY

# func\_max\_buffer\_size

#### nifgen.Session.func\_max\_buffer\_size

This property sets the maximum number of samples that can be used in the standard method waveform buffer. Increasing this value may increase the quality of the waveform. This property is only valid on devices that implement standard method mode in software, and is read-only for all other devices. implementation of Standard Method Mode on your device.

**Note:** Refer to the Standard Method Mode topic for more information on the

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function:Standard Function Mode:Maximum Buffer Size
- C Attribute: NIFGEN ATTR FUNC MAX BUFFER SIZE

# func\_start\_phase

## nifgen.Session.func\_start\_phase

Controls horizontal offset of the standard waveform the signal generator produces. Specify this property in degrees of one waveform cycle. A start phase of 180 degrees means output generation begins halfway through the waveform. A start phase of 360 degrees offsets the output by an entire waveform cycle, which is identical to a start phase of 0 degrees. set the Waveform parameter to DC. Units: Degrees of one cycle

**Note:** This parameter does not affect signal generator behavior when you

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function:Start Phase
- C Attribute: NIFGEN\_ATTR\_FUNC\_START\_PHASE

## func waveform

#### nifgen.Session.func\_waveform

This channel-based property specifies which standard waveform the signal generator produces. Use this property only when <code>nifgen.Session.output\_mode</code> is set to <code>FUNC.SINE</code> - Sinusoid waveform <code>SQUARE</code> - Square waveform <code>TRIANGLE</code> - Triangle waveform <code>RAMP\_UP</code> - Positive ramp waveform <code>RAMP\_DOWN</code> - Negative ramp waveform <code>DC</code> - Constant voltage <code>NOISE</code> - White noise <code>USER</code> - User-defined waveform as defined with <code>nifgen.Session.define\_user\_standard\_waveform()</code>

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.Waveform
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function: Waveform
- C Attribute: NIFGEN\_ATTR\_FUNC\_WAVEFORM

# idle behavior

# nifgen.Session.idle\_behavior

Specifies the behavior of the output during the Idle state. The output can be configured to hold the last generated voltage before entering the Idle state or jump to the Idle Value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.IdleBehavior
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Advanced:Idle Behavior
- C Attribute: NIFGEN\_ATTR\_IDLE\_BEHAVIOR

### idle value

### nifgen.Session.idle\_value

Specifies the value to generate in the Idle state. The Idle Behavior must be configured to jump to this value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Output:Advanced:Idle Value

• C Attribute: NIFGEN\_ATTR\_IDLE\_VALUE

## instrument firmware revision

# nifgen.Session.instrument\_firmware\_revision

A string that contains the firmware revision information for the device that you are currently using.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Inherent IVI Attributes:Instrument Identification:Firmware Revision
- C Attribute: NIFGEN\_ATTR\_INSTRUMENT\_FIRMWARE\_REVISION

# instrument\_manufacturer

### nifgen.Session.instrument\_manufacturer

A string that contains the name of the device manufacturer you are currently using.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Inherent IVI Attributes:Instrument Identification:Manufacturer
- C Attribute: NIFGEN\_ATTR\_INSTRUMENT\_MANUFACTURER

# instrument\_model

### nifgen.Session.instrument\_model

A string that contains the model number or name of the device that you are currently using.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Inherent IVI Attributes:Instrument Identification:Model
- C Attribute: NIFGEN\_ATTR\_INSTRUMENT\_MODEL

# io\_resource\_descriptor

# nifgen.Session.io\_resource\_descriptor

Indicates the resource descriptor that NI-FGEN uses to identify the physical device. If you initialize NI-FGEN with a logical name, this property contains the resource descriptor that corresponds to the entry in the IVI Configuration Utility. If you initialize NI-FGEN with the resource descriptor, this property contains that value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Inherent IVI Attributes:Advanced Session Information:Resource Descriptor
- C Attribute: NIFGEN ATTR IO RESOURCE DESCRIPTOR

# load\_impedance

#### nifgen.Session.load\_impedance

This channel-based property specifies the load impedance connected to the analog output of the channel. If you set this property to NIFGEN\_VAL\_MATCHED\_LOAD\_IMPEDANCE (-1.0), NIFGEN assumes that the load impedance matches the output impedance. NI-FGEN compensates to give the desired peak-to-peak voltage amplitude or arbitrary gain (relative to 1 V).

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Output:Load Impedance

• C Attribute: NIFGEN ATTR LOAD IMPEDANCE

# logical\_name

## nifgen.Session.logical\_name

A string containing the logical name that you specified when opening the current IVI session. You may pass a logical name to nifgen.Session.init() or nifgen.Session. InitWithOptions(). The IVI Configuration Utility must contain an entry for the logical name. The logical name entry refers to a virtual instrument section in the IVI Configuration file. The virtual instrument section specifies a physical device and initial user options.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Instrument:Inherent IVI Attributes:Advanced Session Information:Logical Name

## • C Attribute: NIFGEN\_ATTR\_LOGICAL\_NAME

# major version

nifgen.Session.major\_version

Returns the major version number of NI-FGEN.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Obsolete:Major Version
- C Attribute: NIFGEN\_ATTR\_MAJOR\_VERSION

# marker\_events\_count

nifgen.Session.marker\_events\_count

Returns the number of markers supported by the device. Use this property when nifgen. Session.output mode is set to SCRIPT.

The following table lists the characteristics of this property.

Value
int
read only
False
No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Marker Events Count
- C Attribute: NIFGEN\_ATTR\_MARKER\_EVENTS\_COUNT

### marker event output terminal

nifgen.Session.marker\_event\_output\_terminal Specifies the destination terminal for the Marker Event.

**Tip:** This property can use repeated capabilities (markers). If set or get directly on the nifgen. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling set/get value on the result.:

```
session.markers[0,1].marker_event_output_terminal = var
var = session.markers[0,1].marker_event_output_terminal
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Marker:Output Terminal
- C Attribute: NIFGEN\_ATTR\_MARKER\_EVENT\_OUTPUT\_TERMINAL

# max\_freq\_list\_duration

```
nifgen.Session.max_freq_list_duration
```

Returns the maximum duration of any one step in the frequency list.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function:Frequency List Mode:Maximum Frequency List Duration
- C Attribute: NIFGEN\_ATTR\_MAX\_FREQ\_LIST\_DURATION

# max\_freq\_list\_length

```
nifgen.Session.max_freq_list_length
```

Returns the maximum number of steps that can be in a frequency list.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

- LabVIEW Property: Standard Function:Frequency List Mode:Maximum Frequency List Length
- C Attribute: NIFGEN\_ATTR\_MAX\_FREQ\_LIST\_LENGTH

# max\_loop\_count

# nifgen.Session.max\_loop\_count

Returns the maximum number of times that the signal generator can repeat a waveform in a sequence. Typically, this value is constant for the signal generator.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Arbitrary Sequence Mode: Max Loop Count
- C Attribute: NIFGEN\_ATTR\_MAX\_LOOP\_COUNT

# max\_num\_freq\_lists

# $\verb|nifgen.Session.max_num_freq_lists|\\$

Returns the maximum number of frequency lists the signal generator allows.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function:Frequency List Mode:Maximum Number Of Frequency Lists
- C Attribute: NIFGEN\_ATTR\_MAX\_NUM\_FREQ\_LISTS

#### max\_num\_sequences

### nifgen.Session.max\_num\_sequences

Returns the maximum number of arbitrary sequences that the signal generator allows. Typically, this value is constant for the signal generator.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Arbitrary Sequence Mode: Max Number of Sequences
- C Attribute: NIFGEN\_ATTR\_MAX\_NUM\_SEQUENCES

### max\_num\_waveforms

### nifgen.Session.max\_num\_waveforms

Returns the maximum number of arbitrary waveforms that the signal generator allows. Typically, this value is constant for the signal generator.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Capabilities: Max Number of Waveforms
- C Attribute: NIFGEN\_ATTR\_MAX\_NUM\_WAVEFORMS

## max sequence length

#### nifgen.Session.max\_sequence\_length

Returns the maximum number of arbitrary waveforms that the signal generator allows in a sequence. Typically, this value is constant for the signal generator.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Arbitrary Sequence Mode: Max Sequence Length
- C Attribute: NIFGEN\_ATTR\_MAX\_SEQUENCE\_LENGTH

# max waveform size

### nifgen.Session.max\_waveform\_size

Returns the size, in samples, of the largest waveform that can be created. This property reflects the space currently available, taking into account previously allocated waveforms and instructions.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Capabilities: Max Waveform Size
- C Attribute: NIFGEN\_ATTR\_MAX\_WAVEFORM\_SIZE

# memory\_size

# nifgen.Session.memory\_size

The total amount of memory, in bytes, on the signal generator.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

• LabVIEW Property: Instrument:Memory Size

• C Attribute: NIFGEN\_ATTR\_MEMORY\_SIZE

# minor\_version

### nifgen.Session.minor\_version

Returns the minor version number of NI-FGEN.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Instrument:Obsolete:Minor Version

• C Attribute: NIFGEN\_ATTR\_MINOR\_VERSION

# min\_freq\_list\_duration

## nifgen.Session.min\_freq\_list\_duration

Returns the minimum number of steps that can be in a frequency list.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Standard Function:Frequency List Mode:Minimum Frequency List Duration

## • C Attribute: NIFGEN\_ATTR\_MIN\_FREQ\_LIST\_DURATION

### min freq list length

#### nifgen.Session.min freg list length

Returns the minimum number of frequency lists that the signal generator allows.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Standard Function:Frequency List Mode:Minimum Frequency List Length
- C Attribute: NIFGEN\_ATTR\_MIN\_FREQ\_LIST\_LENGTH

#### min sequence length

### nifgen.Session.min\_sequence\_length

Returns the minimum number of arbitrary waveforms that the signal generator allows in a sequence. Typically, this value is constant for the signal generator.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Arbitrary Sequence Mode: Min Sequence Length
- C Attribute: NIFGEN\_ATTR\_MIN\_SEQUENCE\_LENGTH

## min\_waveform\_size

# nifgen.Session.min\_waveform\_size

Returns the minimum number of points that the signal generator allows in an arbitrary waveform. Typically, this value is constant for the signal generator.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Capabilities: Min Waveform Size
- C Attribute: NIFGEN\_ATTR\_MIN\_WAVEFORM\_SIZE

## module revision

# nifgen.Session.module\_revision

A string that contains the module revision for the device that you are currently using.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Inherent IVI Attributes:Instrument Identification:Module Revision
- C Attribute: NIFGEN\_ATTR\_MODULE\_REVISION

# output\_enabled

### nifgen.Session.output\_enabled

This channel-based property specifies whether the signal that the signal generator produces appears at the output connector.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	Yes

• LabVIEW Property: Output:Output Enabled

• C Attribute: NIFGEN ATTR OUTPUT ENABLED

# output\_impedance

### nifgen.Session.output\_impedance

This channel-based property specifies the signal generator output impedance at the output connector. NI signal sources modules have an output impedance of 50 ohms and an optional 75 ohms on select modules. If the load impedance matches the output impedance, then the voltage at the signal output connector is at the needed level. The voltage at the signal output connector varies with load output impedance, up to doubling the voltage for a high-impedance load.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Output:Output Impedance

• C Attribute: NIFGEN ATTR OUTPUT IMPEDANCE

### output mode

## nifgen.Session.output\_mode

Sets which output mode the signal generator will use. The value you specify determines which methods and properties you use to configure the waveform the signal generator produces.

**Note:** The signal generator must not be in the Generating state when you change this property. To change the device configuration, call <code>nifgen.Session.abort()</code> or wait for the generation to complete.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.OutputMode
Permissions	read-write
Channel Based	False
Resettable	No

• LabVIEW Property: Output:Output Mode

• C Attribute: NIFGEN ATTR OUTPUT MODE

# ready\_for\_start\_event\_output\_terminal

# nifgen.Session.ready\_for\_start\_event\_output\_terminal

Specifies the destination terminal for the Ready for Start Event.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Ready For Start:Output Terminal
- C Attribute: NIFGEN\_ATTR\_READY\_FOR\_START\_EVENT\_OUTPUT\_TERMINAL

# reference\_clock\_source

### nifgen.Session.reference\_clock\_source

Specifies the reference clock source used by the signal generator. The signal generator derives the frequencies and sample rates that it uses to generate waveforms from the source you specify. For example, when you set this property to ClkIn, the signal generator uses the signal it receives at the CLK IN front panel connector as the Reference clock. To change the device configuration, call <code>nifgen.Session.abort()</code> or wait for the generation to complete.

**Note:** The signal generator must not be in the Generating state when you change this property.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ReferenceClockSource
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Clocks:Reference Clock:Source

#### • C Attribute: NIFGEN ATTR REFERENCE CLOCK SOURCE

# ref\_clock\_frequency

#### nifgen.Session.ref clock frequency

Sets the frequency of the signal generator reference clock. The signal generator uses the reference clock to derive frequencies and sample rates when generating output.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Reference Clock:Frequency
- C Attribute: NIFGEN\_ATTR\_REF\_CLOCK\_FREQUENCY

#### sample clock source

### nifgen.Session.sample\_clock\_source

Specifies the Sample clock source. If you specify a divisor with the nifgen.Session.exported\_sample\_clock\_divisor property, the Sample clock exported with the nifgen.Session.exported\_sample\_clock\_output\_terminal property is the value of the Sample clock after it is divided-down. For a list of the terminals available on your device, refer to the Device Routes tab in MAX. To change the device configuration, call nifgen.Session.abort () or wait for the generation to complete.

**Note:** The signal generator must not be in the Generating state when you change this property.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.SampleClockSource
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Sample Clock:Source
- C Attribute: NIFGEN\_ATTR\_SAMPLE\_CLOCK\_SOURCE

## sample clock timebase rate

#### nifgen.Session.sample\_clock\_timebase\_rate

Specifies the Sample clock timebase rate. This property applies only to external Sample clock timebases. To change the device configuration, call <code>nifgen.Session.abort()</code> or wait for the generation to complete.

**Note:** The signal generator must not be in the Generating state when you change this property.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Sample Clock Timebase:Rate
- C Attribute: NIFGEN ATTR SAMPLE CLOCK TIMEBASE RATE

# sample clock timebase source

# nifgen.Session.sample\_clock\_timebase\_source

Specifies the Sample Clock Timebase source. To change the device configuration, call the nifgen. Session.abort () method or wait for the generation to complete.

**Note:** The signal generator must not be in the Generating state when you change this property.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.SampleClockTimebaseSource
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocks:Sample Clock Timebase:Source
- C Attribute: NIFGEN\_ATTR\_SAMPLE\_CLOCK\_TIMEBASE\_SOURCE

## script to generate

### nifgen.Session.script\_to\_generate

Specifies which script the generator produces. To configure the generator to run a particular script, set this property to the name of the script. Use nifgen.Session.write\_script() to create multiple scripts. Use this property when nifgen.Session.output\_mode is set to SCRIPT.

**Note:** The signal generator must not be in the Generating state when you change this property. To change the device configuration, call <code>nifgen.Session.abort()</code> or wait for the generation to complete.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform:Script Mode:Script to Generate
- C Attribute: NIFGEN\_ATTR\_SCRIPT\_TO\_GENERATE

### script\_triggers\_count

## nifgen.Session.script\_triggers\_count

Specifies the number of Script triggers supported by the device. Use this property when nifgen. Session.output\_mode is set to SCRIPT.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Script Triggers Count
- C Attribute: NIFGEN\_ATTR\_SCRIPT\_TRIGGERS\_COUNT

#### script trigger type

```
nifgen.Session.script_trigger_type
```

Specifies the Script trigger type. Depending upon the value of this property, additional properties may need to be configured to fully configure the trigger.

**Tip:** This property can use repeated capabilities (script\_triggers). If set or get directly on the nifgen. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling set/get value on the result.:

```
session.script_triggers[0,1].script_trigger_type = var
var = session.script_triggers[0,1].script_trigger_type
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ScriptTriggerType
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Script:Trigger Type
- C Attribute: NIFGEN\_ATTR\_SCRIPT\_TRIGGER\_TYPE

### serial number

 $\verb|nifgen.Session.serial_number|\\$ 

The signal generator's serial number.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Serial Number
- C Attribute: NIFGEN\_ATTR\_SERIAL\_NUMBER

#### simulate

#### nifgen.Session.simulate

Specifies whether to simulate NI-FGEN I/O operations. If simulation is enabled, NI-FGEN methods perform range checking and call Ivi\_GetAttribute and Ivi\_SetAttribute, but they do not perform device I/O. For output parameters that represent device data, NI-FGEN methods return calculated values. Default Value: False Use nifgen.Session.InitWithOptions() to override default value.

Note: One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Inherent IVI Attributes:User Options:Simulate
- C Attribute: NIFGEN\_ATTR\_SIMULATE

### specific\_driver\_description

 $\verb|nifgen.Session.specific_driver_description|\\$ 

Returns a brief description of NI-FGEN.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- $\bullet \ \ Lab VIEW \ Property: \textbf{Instrument:Inherent IVI Attributes:Driver Identification:Description}$
- C Attribute: NIFGEN\_ATTR\_SPECIFIC\_DRIVER\_DESCRIPTION

### specific driver revision

nifgen.Session.specific\_driver\_revision

A string that contains additional version information about NI-FGEN.

The following table lists the characteristics of this property.

Characteristic	Value
Characteristic	value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Inherent IVI Attributes:Driver Identification:Revision
- C Attribute: NIFGEN\_ATTR\_SPECIFIC\_DRIVER\_REVISION

# specific\_driver\_vendor

# nifgen.Session.specific\_driver\_vendor

A string that contains the name of the vendor that supplies NI-FGEN.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Inherent IVI Attributes:Driver Identification:Driver Vendor
- C Attribute: NIFGEN\_ATTR\_SPECIFIC\_DRIVER\_VENDOR

# started\_event\_output\_terminal

### nifgen.Session.started\_event\_output\_terminal

Specifies the destination terminal for the Started Event.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Events:Started:Output Terminal
- C Attribute: NIFGEN ATTR STARTED EVENT OUTPUT TERMINAL

#### start\_trigger\_type

#### nifgen.Session.start trigger type

Specifies whether you want the Start trigger to be a Digital Edge, or Software trigger. You can also choose None as the value for this property.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.StartTriggerType
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggers:Start:Trigger Type
- C Attribute: NIFGEN\_ATTR\_START\_TRIGGER\_TYPE

### streaming\_space\_available\_in\_waveform

### nifgen.Session.streaming\_space\_available\_in\_waveform

Indicates the space available (in samples) in the streaming waveform for writing new data. During generation, this available space may be in multiple locations with, for example, part of the available space at the end of the streaming waveform and the rest at the beginning. In this situation, writing a block of waveform data the size of the total space available in the streaming waveform causes NI-FGEN to return an error, as NI-FGEN will not wrap the data from the end of the waveform to the beginning and cannot write data past the end of the waveform buffer. To avoid writing data past the end of the waveform, write new data to the waveform in a fixed size that is an integer divisor of the total size of the streaming waveform. Used in conjunction with the nifgen. Session.streaming\_waveform\_handle or nifgen. Session.streaming\_waveform\_name properties.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Arbitrary Waveform:Data Transfer:Streaming:Space Available in Streaming Waveform

### C Attribute: NIFGEN\_ATTR\_STREAMING\_SPACE\_AVAILABLE\_IN\_WAVEFORM

# streaming waveform handle

#### nifgen.Session.streaming waveform handle

Specifies the waveform handle of the waveform used to continuously stream data during generation. This property defaults to -1 when no streaming waveform is specified. Used in conjunction with nifgen.Session.streaming\_space\_available\_in\_waveform.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: **Arbitrary Waveform:Data Transfer:Streaming:Streaming Waveform Handle**
- C Attribute: NIFGEN\_ATTR\_STREAMING\_WAVEFORM\_HANDLE

## streaming waveform name

# nifgen.Session.streaming\_waveform\_name

Specifies the name of the waveform used to continuously stream data during generation. This property defaults to // when no streaming waveform is specified. Use in conjunction with nifgen. Session.streaming\_space\_available\_in\_waveform.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: **Arbitrary Waveform:Data Transfer:Streaming:Streaming Waveform**Name
- C Attribute: NIFGEN\_ATTR\_STREAMING\_WAVEFORM\_NAME

## streaming write timeout

#### nifgen.Session.streaming\_write\_timeout

Specifies the maximum amount of time allowed to complete a streaming write operation.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform:Data Transfer:Streaming:Streaming Write Timeout
- C Attribute: NIFGEN\_ATTR\_STREAMING\_WRITE\_TIMEOUT

## supported\_instrument\_models

#### nifgen.Session.supported instrument models

Returns a model code of the device. For NI-FGEN versions that support more than one device, this property contains a comma-separated list of supported device models.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Instrument:Inherent IVI Attributes:Driver Capabilities:Supported Instrument Models
- $\bullet \ \ C \ Attribute: \ \textbf{NIFGEN\_ATTR\_SUPPORTED\_INSTRUMENT\_MODELS}$

# terminal\_configuration

# nifgen.Session.terminal\_configuration

Specifies whether gain and offset values will be analyzed based on single-ended or differential operation.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TerminalConfiguration
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Output:Terminal Configuration

• C Attribute: NIFGEN\_ATTR\_TERMINAL\_CONFIGURATION

# trigger mode

nifgen.Session.trigger\_mode

Controls the trigger mode.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerMode
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Triggers:Trigger Mode

• C Attribute: NIFGEN\_ATTR\_TRIGGER\_MODE

## wait behavior

nifgen.Session.wait\_behavior

Specifies the behavior of the output while waiting for a script trigger or during a wait instruction. The output can be configured to hold the last generated voltage before waiting or jump to the Wait Value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.WaitBehavior
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Advanced:Wait Behavior
- C Attribute: NIFGEN\_ATTR\_WAIT\_BEHAVIOR

### wait value

#### nifgen.Session.wait value

Specifies the value to generate while waiting. The Wait Behavior must be configured to jump to this value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Output:Advanced:Wait Value
- C Attribute: NIFGEN\_ATTR\_WAIT\_VALUE

### waveform\_quantum

# nifgen.Session.waveform\_quantum

The size of each arbitrary waveform must be a multiple of a quantum value. This property returns the quantum value that the signal generator allows. For example, when this property returns a value of 8, all waveform sizes must be a multiple of 8. Typically, this value is constant for the signal generator.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Arbitrary Waveform: Capabilities: Waveform Quantum
- C Attribute: NIFGEN\_ATTR\_WAVEFORM\_QUANTUM

#### **Methods**

#### abort

```
nifgen.Session.abort()
```

Aborts any previously initiated signal generation. Call the nifgen.Session.initiate() method to cause the signal generator to produce a signal again.

# allocate named waveform

```
nifgen.Session.allocate_named_waveform(waveform_name, waveform_size)
```

Specifies the size of a named waveform up front so that it can be allocated in onboard memory before loading the associated data. Data can then be loaded in smaller blocks with the niFgen Write (Binary16) Waveform methods.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels [0,1].allocate\_named\_waveform (waveform\_name, waveform\_ \\ \hookrightarrow size)
```

#### **Parameters**

- waveform\_name (str) Specifies the name to associate with the allocated waveform.
- waveform\_size (int) Specifies the size of the waveform to allocate in samples.

Default Value: "4096"

# allocate waveform

```
nifgen.Session.allocate_waveform(waveform_size)
```

Specifies the size of a waveform so that it can be allocated in onboard memory before loading the associated data. Data can then be loaded in smaller blocks with the Write Binary 16 Waveform methods.

**Note:** The signal generator must not be in the Generating state when you call this method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].allocate_waveform(waveform_size)
```

**Parameters waveform\_size** (int) - Specifies, in samples, the size of the waveform to allocate.

## Return type int

**Returns** The handle that identifies the new waveform. This handle is used later when referring to this waveform.

# clear\_arb\_memory

```
nifgen.Session.clear_arb_memory()
```

Removes all previously created arbitrary waveforms, sequences, and scripts from the signal generator memory and invalidates all waveform handles, sequence handles, and waveform names.

**Note:** The signal generator must not be in the Generating state when you call this method.

# clear\_arb\_sequence

```
nifgen.Session.clear_arb_sequence (sequence_handle)
```

Removes a previously created arbitrary sequence from the signal generator memory and invalidates the sequence handle.

Note: The signal generator must not be in the Generating state when you call this method.

Parameters sequence\_handle (int) - Specifies the handle of the arbitrary sequence that you want the signal generator to remove. You can create an arbitrary sequence using the nifgen.Session.create\_arb\_sequence() or nifgen.Session.create\_advanced\_arb\_sequence() method. These methods return a handle that you use to identify the sequence.

#### **Defined Value:**

Default Value: None

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

### clear freq list

```
nifgen.Session.clear_freq_list (frequency_list_handle)
```

Removes a previously created frequency list from the signal generator memory and invalidates the frequency list handle.

**Note:** The signal generator must not be in the Generating state when you call this method.

Parameters frequency\_list\_handle (int) - Specifies the handle of the frequency list you want the signal generator to remove. You create multiple frequency lists using nifgen.Session.create\_freq\_list(). nifgen.Session.create\_freq\_list() returns a handle that you use to identify each list. Specify a value of -1 to clear all frequency lists.

#### **Defined Value**

NIFGEN\_VAL\_ALL\_FLISTS—Remove all frequency lists from the signal generator.

**Default Value**: None

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

# clear user standard waveform

```
nifgen.Session.clear_user_standard_waveform()

Clears the user-defined waveform created by the nifgen.Session.

define user standard waveform() method.
```

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].clear_user_standard_waveform()
```

#### commit

```
nifgen.Session.commit()
```

Causes a transition to the Committed state. This method verifies property values, reserves the device, and commits the property values to the device. If the property values are all valid, NI-FGEN sets the device hardware configuration to match the session configuration. This method does not support the NI 5401/5404/5411/5431 signal generators.

In the Committed state, you can load waveforms, scripts, and sequences into memory. If any properties are changed, NI-FGEN implicitly transitions back to the Idle state, where you can program all session properties before applying them to the device. This method has no effect if the device is already in the Committed or Generating state and returns a successful status value.

Calling this VI before the niFgen Initiate Generation VI is optional but has the following benefits:

- Routes are committed, so signals are exported or imported.
- Any Reference Clock and external clock circuits are phase-locked.

• A subsequent nifgen.Session.initiate() method can run faster because the device is already configured.

# configure arb sequence

nifgen.Session.configure\_arb\_sequence(sequence\_handle, gain, offset)

Configures the signal generator properties that affect arbitrary sequence generation. Sets the nifgen.Session.arb\_sequence\_handle, nifgen.Session.arb\_gain, and nifgen.Session.arb\_offset properties.

**Note:** The signal generator must not be in the Generating state when you call this method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

#### **Parameters**

• **sequence\_handle** (*int*) – Specifies the handle of the arbitrary sequence that you want the signal generator to produce. NI-FGEN sets the *nifgen.Session.arb\_sequence\_handle* property to this value. You can create an arbitrary sequence using the *nifgen.Session.create\_arb\_sequence()* or *nifgen.Session.create\_advanced\_arb\_sequence()* method. These methods return a handle that you use to identify the sequence.

Default Value: None

• **gain** (float) – Specifies the factor by which the signal generator scales the arbitrary waveforms in the sequence. When you create an arbitrary waveform, you must first normalize the data points to a range of –1.00 to +1.00. You can use this parameter to scale the waveform to other ranges. The gain is applied before the offset is added.

For example, to configure the output signal to range from -2.00 to +2.00 V, set **gain** to 2.00.

Units: unitless

Default Value: None

• **offset** (float) – Specifies the value the signal generator adds to the arbitrary waveform data. When you create arbitrary waveforms, you must first normalize the data points to a range of –1.00 to +1.00 V. You can use this parameter to shift the range of the arbitrary waveform. NI-FGEN sets the nifgen. Session. arb\_offset property to this value.

For example, to configure the output signal to range from 0.00 to 2.00 V instead of -1.00 to 1.00 V, set the offset to 1.00.

Units: volts

**Default Value**: None

#### configure arb waveform

nifgen.Session.configure\_arb\_waveform(waveform\_handle, gain, offset)

Configures the properties of the signal generator that affect arbitrary waveform generation.

Sets the nifgen.Session.arb\_waveform\_handle, nifgen.Session.arb\_gain, and nifgen.Session.arb offset properties.

Note: The signal generator must not be in the Generating state when you call this method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].configure_arb_waveform(waveform_handle, gain, \_ \hookrightarrow offset)
```

#### **Parameters**

- waveform\_handle (int) Specifies the handle of the arbitrary waveform you want the signal generator to produce. NI-FGEN sets the nifgen.Session. arb\_waveform\_handle property to this value. You can create an arbitrary waveform using one of the following niFgen Create Waveform methods:
  - nifgen.Session.create waveform()
  - nifgen.Session.create\_waveform()
  - nifgen.Session.create\_waveform\_from\_file\_i16()
  - nifgen.Session.create\_waveform\_from\_file\_f64()
  - nifgen.Session.CreateWaveformFromFileHWS()

These methods return a handle that you use to identify the waveform.

Default Value: None

**Note:** One or more of the referenced methods are not in the Python API for this driver.

• **gain** (float) – Specifies the factor by which the signal generator scales the arbitrary waveforms in the sequence. When you create an arbitrary waveform, you must first normalize the data points to a range of –1.00 to +1.00. You can use this parameter to scale the waveform to other ranges. The gain is applied before the offset is added.

For example, to configure the output signal to range from -2.00 to +2.00 V, set **gain** to 2.00.

Units: unitless

Default Value: None

• **offset** (float) – Specifies the value the signal generator adds to the arbitrary waveform data. When you create arbitrary waveforms, you must first normalize the data points to a range of –1.00 to +1.00 V. You can use this parameter to shift the range of the arbitrary waveform. NI-FGEN sets the nifgen. Session. arb\_offset property to this value.

For example, to configure the output signal to range from 0.00 to 2.00 V instead of -1.00 to 1.00 V, set the offset to 1.00.

Units: volts

Default Value: None

# configure\_custom\_fir\_filter\_coefficients

nifgen.Session.configure\_custom\_fir\_filter\_coefficients (coefficients\_array)

Sets the FIR filter coefficients used by the onboard signal processing block. The values are coerced to the closest settings achievable by the signal generator.

Refer to the *FIR Filter* topic for your device in the *NI Signal Generators Help* for more information about FIR filter coefficients. This method is supported only for the NI 5441.

Note: The signal generator must not be in the Generating state when you call this method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].configure_custom_fir_filter_

→coefficients(coefficients_array)
```

**Parameters coefficients\_array** (list of float)—Specifies the array of data the onboard signal processor uses for the FIR filter coefficients. For the NI 5441, provide a symmetric array of 95 coefficients to this parameter. The array must have at least as many elements as the value that you specify in the **numberOfCoefficients** parameter in this method. The coefficients should range between –1.00 and +1.00.

## configure freq list

```
nifgen.Session.configure_freq_list (frequency_list_handle, dc\_offset=0.0, start\_phase=0.0) amplitude,
```

Configures the properties of the signal generator that affect frequency list generation (the nifgen.Session.freq\_list\_handle, nifgen.Session.func\_amplitude, nifgen.Session.func\_dc\_offset, and nifgen.Session.func\_start\_phase properties).

**Note:** The signal generator must not be in the Generating state when you call this method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].configure_freq_list(frequency_list_handle, _ → amplitude, dc_offset=0.0, start_phase=0.0)
```

#### **Parameters**

- **frequency\_list\_handle** (*int*) Specifies the handle of the frequency list that you want the signal generator to produce. NI-FGEN sets the *nifgen*. Session.freq\_list\_handle property to this value. You can create a frequency list using the *nifgen*.Session.create\_freq\_list() method, which returns a handle that you use to identify the list. **Default Value**: None
- amplitude (float) Specifies the amplitude of the standard waveform that you want the signal generator to produce. This value is the amplitude at the output terminal. NI-FGEN sets the nifgen.Session.func\_amplitude property to this value.

For example, to produce a waveform ranging from -5.00 V to +5.00 V, set the amplitude to 10.00 V.

Units: peak-to-peak voltage

**Default Value**: None

Note: This parameter does not affect signal generator behavior when you set the waveform parameter of the nifgen. Session. configure\_standard\_waveform() method to DC.

• dc\_offset (float) – Specifies the DC offset of the standard waveform that you want the signal generator to produce. The value is the offset from ground to the center of the waveform you specify with the waveform parameter, observed at the output terminal. For example, to configure a waveform with an amplitude of 10.00 V to range from 0.00 V to +10.00 V, set the dcOffset to 5.00 V. NI-FGEN sets the nifgen.Session.func\_dc\_offset property to this value.

Units: volts

**Default Value**: None

• **start\_phase** (float) – Specifies the horizontal offset of the standard waveform you want the signal generator to produce. Specify this property in degrees of one waveform cycle. NI-FGEN sets the nifgen.Session.func\_start\_phase property to this value. A start phase of 180 degrees means output generation begins halfway through the waveform. A start phase of 360 degrees offsets the output by an entire waveform cycle, which is identical to a start phase of 0 degrees.

Units: degrees of one cycleDefault Value: None degrees

**Note:** This parameter does not affect signal generator behavior when you set the **waveform** parameter to DC.

# configure\_standard\_waveform

nifgen.Session.configure\_standard\_waveform(waveform, amplitude, frequency,  $dc\_offset=0.0, start\_phase=0.0$ )

Configures the following properties of the signal generator that affect standard waveform generation:

- nifgen.Session.func\_waveform
- nifgen.Session.func\_amplitude
- nifgen.Session.func\_dc\_offset
- nifgen.Session.func\_frequency
- nifgen.Session.func\_start\_phase

**Note:** You must call the nifgen.Session.ConfigureOutputMode() method with the **outputMode** parameter set to *FUNC* before calling this method.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].configure_standard_waveform(waveform, amplitude, \_ \rightarrow frequency, dc_offset=0.0, start_phase=0.0)
```

#### **Parameters**

• waveform (nifgen. Waveform) - Specifies the standard waveform that you want the signal generator to produce. NI-FGEN sets the nifgen. Session. func\_waveform property to this value.

\*\*Defined Values\*\*

**Default Value**: SINE

SINE	Specifies that the signal generator produces a sinusoid waveform.	
SQUAR	ESpecifies that the signal generator produces a square waveform.	
TRIAN	Specifies that the signal generator produces a triangle waveform.	
RAMP_	USpecifies that the signal generator produces a positive ramp waveform.	
RAMP_	DSpēcifies that the signal generator produces a negative ramp waveform.	
DC	Specifies that the signal generator produces a constant voltage.	
NOISE	Specifies that the signal generator produces white noise.	
USER	Specifies that the signal generator produces a user-defined	
	waveform as defined with the nifgen.Session.	
	define_user_standard_waveform() method.	

• amplitude (float) - Specifies the amplitude of the standard waveform that you want the signal generator to produce. This value is the amplitude at the output terminal. NI-FGEN sets the nifqen. Session. func\_amplitude property to this value.

For example, to produce a waveform ranging from -5.00 V to +5.00 V, set the amplitude to 10.00 V.

Units: peak-to-peak voltage

**Default Value**: None

Note: This parameter does not affect signal generator behavior when you set the waveform parameter of the nifgen. Session. configure\_standard\_waveform() method to DC.

## • frequency (float) -

Specifies the frequency of the standard waveform that you want the signal generator to produce. NI-FGEN sets the

nifgen.Session.func\_frequency property to this value.

Units: hertz

Default Value: None

Note: This parameter does not affect signal generator behavior when you set the waveform parameter of the nifqen. Session. configure\_standard\_waveform() method to DC.

• dc\_offset (float) - Specifies the DC offset of the standard waveform that you want the signal generator to produce. The value is the offset from ground to the center of the waveform you specify with the **waveform** parameter, observed at the output terminal. For example, to configure a waveform with an amplitude of 10.00 V to range from 0.00 V to +10.00 V, set the **dcOffset** to 5.00 V. NI-FGEN sets the nifgen.Session.func\_dc\_offset property to this value.

Units: volts

Default Value: None

• start\_phase (float) - Specifies the horizontal offset of the standard waveform that you want the signal generator to produce. Specify this parameter

in degrees of one waveform cycle. NI-FGEN sets the *nifgen.Session.* func\_start\_phase property to this value. A start phase of 180 degrees means output generation begins halfway through the waveform. A start phase of 360 degrees offsets the output by an entire waveform cycle, which is identical to a start phase of 0 degrees.

Units: degrees of one cycle

**Default Value**: 0.00

**Note:** This parameter does not affect signal generator behavior when you set the **waveform** parameter to DC.

#### create advanced arb sequence

```
nifgen.Session.create_advanced_arb_sequence (waveform_handles_array, loop_counts_array, sample_counts_array=None, marker location array=None)
```

Creates an arbitrary sequence from an array of waveform handles and an array of corresponding loop counts. This method returns a handle that identifies the sequence. You pass this handle to the nifgen. Session.configure\_arb\_sequence() method to specify what arbitrary sequence you want the signal generator to produce.

The nifgen.Session.create\_advanced\_arb\_sequence() method extends on the nifgen.Session.create\_arb\_sequence() method by adding the ability to set the number of samples in each sequence step and to set marker locations.

An arbitrary sequence consists of multiple waveforms. For each waveform, you specify the number of times the signal generator produces the waveform before proceeding to the next waveform. The number of times to repeat a specific waveform is called the loop count.

**Note:** The signal generator must not be in the Generating state when you call this method. You must call the nifgen.Session.ConfigureOutputMode() method to set the **outputMode** parameter to SEQ before calling this method.

#### **Parameters**

waveform\_handles\_array (list of int) – Specifies the array of waveform handles from which you want to create a new arbitrary sequence. The array must have at least as many elements as the value that you specify in sequenceLength. Each waveformHandlesArray element has a corresponding loop-CountsArray element that indicates how many times that waveform is repeated. You obtain waveform handles when you create arbitrary waveforms with the nifgen.Session.allocate\_waveform() method or one of the following niFgen CreateWaveform methods:

```
nifgen.Session.create_waveform()
nifgen.Session.create_waveform()
nifgen.Session.create_waveform_from_file_i16()
nifgen.Session.create_waveform_from_file_f64()
```

- nifgen.Session.CreateWaveformFromFileHWS()

**Default Value**: None

• loop\_counts\_array (list of int) - Specifies the array of loop counts you want to use to create a new arbitrary sequence. The array must have at least as many elements as the value that you specify in the sequenceLength parameter. Each loop-CountsArray element corresponds to a waveformHandlesArray element and indicates how many times to repeat that waveform. Each element of the loopCountsArray must be less than or equal to the maximum number of loop counts that the signal generator allows. You can obtain the maximum loop count from maximum-LoopCount in the nifgen. Session.query\_arb\_seq\_capabilities() method.

Default Value: None

• **sample\_counts\_array** (*list of int*) – Specifies the array of sample counts that you want to use to create a new arbitrary sequence. The array must have at least as many elements as the value you specify in the **sequenceLength** parameter. Each **sampleCountsArray** element corresponds to a **waveformHandlesArray** element and indicates the subset, in samples, of the given waveform to generate. Each element of the **sampleCountsArray** must be larger than the minimum waveform size, a multiple of the waveform quantum and no larger than the number of samples in the corresponding waveform. You can obtain these values by calling the <code>nifgen.Session.query\_arb\_wfm\_capabilities()</code> method.

**Default Value:** None

• marker\_location\_array (list of int) - Specifies the array of marker locations to where you want a marker to be generated in the sequence. The array must have at least as many elements as the value you specify in the sequenceLength parameter. Each markerLocationArray element corresponds to a waveformHandlesArray element and indicates where in the waveform a marker is to generate. The marker location must be less than the size of the waveform the marker is in. The markers are coerced to the nearest marker quantum and the coerced values are returned in the coercedMarkersArray parameter.

If you do not want a marker generated for a particular sequence stage, set this parameter to NIFGEN\_VAL\_NO\_MARKER.

**Defined Value**: NIFGEN\_VAL\_NO\_MARKER

Default Value: None

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

# Return type

tuple (coerced\_markers\_array, sequence\_handle)

WHERE

coerced\_markers\_array (list of int):

Returns an array of all given markers that are coerced (rounded) to the nearest marker quantum. Not all devices coerce markers.

Default Value: None

```
sequence handle (int):
```

Returns the handle that identifies the new arbitrary sequence. You can pass this handle to <code>nifgen.Session.configure\_arb\_sequence()</code> to generate the arbitrary sequence.

#### create arb sequence

Creates an arbitrary sequence from an array of waveform handles and an array of corresponding loop counts. This method returns a handle that identifies the sequence. You pass this handle to the nifgen. Session.configure\_arb\_sequence() method to specify what arbitrary sequence you want the signal generator to produce.

An arbitrary sequence consists of multiple waveforms. For each waveform, you can specify the number of times that the signal generator produces the waveform before proceeding to the next waveform. The number of times to repeat a specific waveform is called the loop count.

**Note:** You must call the nifgen.Session.ConfigureOutputMode() method to set the **outputMode** parameter to SEQ before calling this method.

#### **Parameters**

waveform\_handles\_array (list of int) - Specifies the array of waveform handles from which you want to create a new arbitrary sequence. The array must have at least as many elements as the value that you specify in sequenceLength. Each waveformHandlesArray element has a corresponding loop-CountsArray element that indicates how many times that waveform is repeated. You obtain waveform handles when you create arbitrary waveforms with the nifgen.Session.allocate\_waveform() method or one of the following niFgen CreateWaveform methods:

```
    nifgen.Session.create_waveform()
    nifgen.Session.create_waveform()
    nifgen.Session.create_waveform_from_file_i16()
    nifgen.Session.create_waveform_from_file_f64()
    nifgen.Session.CreateWaveformFromFileHWS()
```

#### **Default Value**: None

• loop\_counts\_array (list of int) - Specifies the array of loop counts you want to use to create a new arbitrary sequence. The array must have at least as many elements as the value that you specify in the sequenceLength parameter. Each loop-CountsArray element corresponds to a waveformHandlesArray element and indicates how many times to repeat that waveform. Each element of the loopCountsArray must be less than or equal to the maximum number of loop counts that the signal generator allows. You can obtain the maximum loop count from maximum-LoopCount in the nifgen. Session.query\_arb\_seq\_capabilities() method.

Default Value: None

Return type int

**Returns** Returns the handle that identifies the new arbitrary sequence. You can pass this handle to nifgen. Session.configure\_arb\_sequence() to generate the arbitrary sequence.

# create\_freq\_list

nifgen.Session.create\_freq\_list (waveform, frequency\_array, duration\_array)

Creates a frequency list from an array of frequencies (**frequencyArray**) and an array of durations (**durationArray**). The two arrays should have the same number of elements, and this value must also be the size of the **frequencyListLength**. The method returns a handle that identifies the frequency list (the **frequencyListHandle**). You can pass this handle to <code>nifgen.Session.configure\_freq\_list()</code> to specify what frequency list you want the signal generator to produce.

A frequency list consists of a list of frequencies and durations. The signal generator generates each frequency for the given amount of time and then proceeds to the next frequency. When the end of the list is reached, the signal generator starts over at the beginning of the list.

**Note:** The signal generator must not be in the Generating state when you call this method.

#### **Parameters**

• waveform (nifgen. Waveform) - Specifies the standard waveform that you want the signal generator to produce. NI-FGEN sets the nifgen. Session. func\_waveform property to this value.

# \*\*Defined Values\*\*

**Default Value:** SINE

SINE	Specifies that the signal generator produces a sinusoid waveform.		
SQUAR	ESpecifies that the signal generator produces a square waveform.		
TRIAN	Specifies that the signal generator produces a triangle waveform.		
RAMP_	USpecifies that the signal generator produces a positive ramp waveform.		
RAMP_	RAMP_DSpecifies that the signal generator produces a negative ramp waveform.		
DC	Specifies that the signal generator produces a constant voltage.		
NOISE	NOISE Specifies that the signal generator produces white noise.		
USER	Specifies that the signal generator produces a user-defined		
	waveform as defined with the nifgen.Session.		
	define_user_standard_waveform() method.		

• **frequency\_array** (list of float) – Specifies the array of frequencies to form the frequency list. The array must have at least as many elements as the value you specify in **frequencyListLength**. Each **frequencyArray** element has a corresponding **durationArray** element that indicates how long that frequency is repeated.

Units: hertz

**Default Value**: None

• duration\_array(list of float) - Specifies the array of durations to form the frequency list. The array must have at least as many elements as the value that

you specify in **frequencyListLength**. Each **durationArray** element has a corresponding **frequencyArray** element and indicates how long in seconds to generate the corresponding frequency.

Units: seconds

**Default Value**: None

# Return type int

**Returns** Returns the handle that identifies the new frequency list. You can pass this handle to nifgen. Session.configure\_freq\_list() to generate the arbitrary sequence.

### create waveform from file f64

```
nifgen.Session.create_waveform_from_file_f64 (file_name, byte_order)
```

This method takes the floating point double (F64) data from the specified file and creates an onboard waveform for use in Arbitrary Waveform or Arbitrary Sequence output mode. The **waveformHandle** returned by this method can later be used for setting the active waveform, changing the data in the waveform, building sequences of waveforms, or deleting the waveform when it is no longer needed.

Note: The F64 data must be between -1.0 and +1.0 V. Use the nifgen. Session. digital\_gain property to generate different voltage outputs.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].create_waveform_from_file_f64(file_name, byte_
order)
```

## **Parameters**

- **file\_name** (str) The full path and name of the file where the waveform data resides
- byte\_order (nifgen.ByteOrder) Specifies the byte order of the data in the file.

\*\*Defined Values\*\*

\*\*Default Value:\*\* LITTLE

LITTLE Endian Data—The least significant bit is stored at the lowest address, followed by the other bits, in order of increasing significance.

BIG Big Endian Data—The most significant bit is stored at the lowest address, followed by the other bits, in order of decreasing significance.

**Note:** Data written by most applications in Windows (including LabWindows<sup>TM</sup>/CVI<sup>TM</sup>) is in Little Endian format. Data written to a file from LabVIEW is in Big Endian format by default on all platforms. Big Endian and Little Endian refer to the way data is stored in memory, which can differ on different processors.

# Return type int

**Returns** The handle that identifies the new waveform. This handle is used later when referring to this waveform.

### create\_waveform\_from\_file\_i16

```
nifgen.Session.create_waveform_from_file_i16 (file_name, byte_order)
```

Takes the binary 16-bit signed integer (I16) data from the specified file and creates an onboard waveform for use in Arbitrary Waveform or Arbitrary Sequence output mode. The **waveformHandle** returned by this method can later be used for setting the active waveform, changing the data in the waveform, building sequences of waveforms, or deleting the waveform when it is no longer needed.

**Note:** The I16 data (values between -32768 and +32767) is assumed to represent -1 to +1 V. Use the nifgen.Session.digital\_gain property to generate different voltage outputs.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

#### **Parameters**

- file\_name (str) The full path and name of the file where the waveform data resides.
- byte\_order (nifgen.ByteOrder) Specifies the byte order of the data in the file.

\*\*Defined Values\*\*

\*\*Default Value:\*\* LITTLE

LITTLELittle Endian Data—The least significant bit is stored at the lowest ad-		
	dress, followed by the other bits, in order of increasing significance.	
BIG	Big Endian Data—The most significant bit is stored at the lowest address,	
	followed by the other bits, in order of decreasing significance.	

**Note:** Data written by most applications in Windows (including LabWindows<sup>TM</sup>/CVI<sup>TM</sup>) is in Little Endian format. Data written to a file from LabVIEW is in Big Endian format by default on all platforms. Big Endian and Little Endian refer to the way data is stored in memory, which can differ on different processors.

# Return type int

**Returns** The handle that identifies the new waveform. This handle is used later when referring to this waveform.

### create\_waveform\_numpy

nifgen.Session.create\_waveform\_numpy(waveform\_data\_array)

Creates an onboard waveform for use in Arbitrary Waveform output mode or Arbitrary Sequence output mode.

**Note:** You must set nifgen. Session.output\_mode to ARB or SEQ before calling this method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

 ${\tt session.channels[0,1].create\_waveform(waveform\_data\_array)}$ 

**Parameters waveform\_data\_array** (list of float) – Array of data for the new arbitrary waveform. This may be an iterable of float, or for best performance a numpy.ndarray of dtype int16 or float64.

# Return type int

**Returns** The handle that identifies the new waveform. This handle is used in other methods when referring to this waveform.

# define\_user\_standard\_waveform

nifgen.Session.define\_user\_standard\_waveform (waveform\_data\_array)

Defines a user waveform for use in either Standard Method or Frequency List output mode.

To select the waveform, set the waveform parameter to <code>USER</code> with either the <code>nifgen.Session.configure\_standard\_waveform()</code> or the <code>nifgen.Session.configure\_list()</code> method.

The waveform data must be scaled between -1.0 and 1.0. Use the **amplitude** parameter in the  $nifgen.Session.configure\_standard\_waveform()$  method to generate different output voltages.

**Note:** You must call the nifgen.Session.ConfigureOutputMode() method to set the **outputMode** parameter to *FUNC* or *FREQ\_LIST* before calling this method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

session.channels[0,1].define\_user\_standard\_waveform(waveform\_data\_array)

**Parameters waveform\_data\_array** (list of float) – Specifies the array of data you want to use for the new arbitrary waveform. The array must have at least as many elements as the value that you specify in **waveformSize**.

You must normalize the data points in the array to be between -1.00 and +1.00.

Default Value: None

# delete\_script

nifgen.Session.delete\_script (script\_name)

Deletes the specified script from onboard memory.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

session.channels[0,1].delete\_script(script\_name)

**Parameters** script\_name (str) – Specifies the name of the script you want to delete. The script name appears in the text of the script following the script keyword.

# delete waveform

nifgen.Session.delete\_waveform(waveform\_name\_or\_handle)

Removes a previously created arbitrary waveform from the signal generator memory.

**Note:** The signal generator must not be in the Generating state when you call this method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].delete_waveform(waveform_name_or_handle)
```

Parameters waveform\_name\_or\_handle (str or int) - The name (str) or handle (int) of an arbitrary waveform previously allocated with nifgen.Session.allocate\_named\_waveform(), nifgen.Session.allocate\_waveform() or nifgen.Session.create\_waveform().

#### disable

```
nifgen.Session.disable()
```

Places the instrument in a quiescent state where it has minimal or no impact on the system to which it is connected. The analog output and all exported signals are disabled.

# get ext cal last date and time

```
nifgen.Session.get_ext_cal_last_date_and_time()
```

Returns the date and time of the last successful external calibration. The time returned is 24-hour (military) local time; for example, if the device was calibrated at 2:30 PM, this method returns 14 for the **hour** parameter and 30 for the **minute** parameter.

Return type datetime.datetime

**Returns** Indicates date and time of the last calibration.

#### get ext cal last temp

```
nifgen.Session.get_ext_cal_last_temp()
```

Returns the temperature at the last successful external calibration. The temperature is returned in degrees Celsius.

Return type float

**Returns** Specifies the temperature at the last successful calibration in degrees Celsius.

## get ext cal recommended interval

```
nifgen.Session.get_ext_cal_recommended_interval()
```

Returns the recommended interval between external calibrations in months.

Return type datetime.timedelta

Returns Specifies the recommended interval between external calibrations in months.

# get\_hardware\_state

```
nifgen.Session.get_hardware_state()
```

Returns the current hardware state of the device and, if the device is in the hardware error state, the current hardware error.

Note: Hardware states do not necessarily correspond to NI-FGEN states.

Return type nifgen. Hardware State

#### Returns

Returns the hardware state of the signal generator.

### **Defined Values**

IDLE	The device is in the Idle state.
WAITING_FOR_START_TRIGGER	The device is waiting for Start Trigger.
RUNNING	The device is in the Running state.
DONE	The generation has completed success-
	fully.
HARDWARE_ERROR	There is a hardware error.

# get self cal last date and time

```
nifgen.Session.get_self_cal_last_date_and_time()
```

Returns the date and time of the last successful self-calibration.

Return type datetime.datetime

Returns Returns the date and time the device was last calibrated.

# get\_self\_cal\_last\_temp

```
nifgen.Session.get_self_cal_last_temp()
```

Returns the temperature at the last successful self-calibration. The temperature is returned in degrees Celsius.

Return type float

Returns Specifies the temperature at the last successful calibration in degrees Celsius.

# get\_self\_cal\_supported

```
nifgen.Session.get_self_cal_supported()
```

Returns whether the device supports self-calibration.

Return type bool

**Returns** 

Returns whether the device supports self-calibration.

\*\*Defined Values\*\*

True	Self–calibration is supported.
False	Self–calibration is not supported.

# is done

```
nifgen.Session.is_done()
```

Determines whether the current generation is complete. This method sets the **done** parameter to True if the session is in the Idle or Committed states.

**Note:** NI-FGEN only reports the **done** parameter as True after the current generation is complete in Single trigger mode.

# Return type bool

#### Returns

Returns information about the completion of waveform generation.

#### **Defined Values**

	Generation is complete.
False	Generation is not complete.

#### lock

```
nifgen.Session.lock()
```

Obtains a multithread lock on the device session. Before doing so, the software waits until all other execution threads release their locks on the device session.

Other threads may have obtained a lock on this session for the following reasons:

- The application called the nifgen.Session.lock() method.
- A call to NI-FGEN locked the session.
- After a call to the nifgen. Session. lock() method returns successfully, no other threads can access the device session until you call the nifgen. Session.unlock() method or exit out of the with block when using lock context manager.
- Use the nifgen. Session. lock () method and the nifgen. Session. unlock () method around a sequence of calls to instrument driver methods if you require that the device retain its settings through the end of the sequence.

You can safely make nested calls to the <code>nifgen.Session.lock()</code> method within the same thread. To completely unlock the session, you must balance each call to the <code>nifgen.Session.lock()</code> method with a call to the <code>nifgen.Session.unlock()</code> method.

One method for ensuring there are the same number of unlock method calls as there is lock calls is to use lock as a context manager

```
with nifgen.Session('dev1') as session:
    with session.lock():
     # Calls to session within a single lock context
```

The first with block ensures the session is closed regardless of any exceptions raised

The second with block ensures that unlock is called regardless of any exceptions raised

Return type context manager

**Returns** When used in a *with* statement, *nifgen.Session.lock()* acts as a context manager and unlock will be called when the *with* block is exited

# query\_arb\_seq\_capabilities

# nifgen.Session.query\_arb\_seq\_capabilities()

Returns the properties of the signal generator that are related to creating arbitrary sequences (the nifgen.Session.max\_num\_sequences, nifgen.Session.max\_sequence\_length, nifgen.Session.max\_sequence\_length, and nifgen.Session.max\_loop\_count properties).

### Return type

tuple (maximum\_number\_of\_sequences, minimum\_sequence\_length, maximum\_sequence\_length, maximum\_loop\_count)

#### WHERE

maximum number of sequences (int):

Returns the maximum number of arbitrary waveform sequences that the signal generator allows. NI-FGEN obtains this value from the nifgen.Session.max\_num\_sequences property.

minimum\_sequence\_length (int):

Returns the minimum number of arbitrary waveforms the signal generator allows in a sequence. NI-FGEN obtains this value from the nifgen. Session. min\_sequence\_length property.

maximum\_sequence\_length (int):

Returns the maximum number of arbitrary waveforms the signal generator allows in a sequence. NI-FGEN obtains this value from the nifgen.Session. max\_sequence\_length property.

maximum\_loop\_count (int):

Returns the maximum number of times the signal generator can repeat an arbitrary waveform in a sequence. NI-FGEN obtains this value from the *nifgen*. Session.max\_loop\_count property.

## query arb wfm capabilities

### nifgen.Session.query\_arb\_wfm\_capabilities()

Returns the properties of the signal generator that are related to creating arbitrary waveforms. These properties are the maximum number of waveforms, waveform quantum, minimum waveform size, and maximum waveform size.

**Note:** If you do not want to obtain the waveform quantum, pass a value of VI\_NULL for this parameter.

# Return type

tuple (maximum\_number\_of\_waveforms, waveform\_quantum, minimum\_waveform\_size, maximum\_waveform\_size)

#### WHERE

maximum\_number\_of\_waveforms (int):

Returns the maximum number of arbitrary waveforms that the signal generator allows. NI-FGEN obtains this value from the nifgen.Session.

max num waveforms property.

waveform\_quantum (int):

The size (number of points) of each waveform must be a multiple of a constant quantum value. This parameter obtains the quantum value that the signal generator uses. NI-FGEN returns this value from the nifgen.Session. waveform\_quantum property.

For example, when this property returns a value of 8, all waveform sizes must be a multiple of 8.

minimum\_waveform\_size (int):

Returns the minimum number of points that the signal generator allows in a waveform. NI-FGEN obtains this value from the nifgen.Session.min\_waveform\_size property.

maximum\_waveform\_size (int):

Returns the maximum number of points that the signal generator allows in a waveform. NI-FGEN obtains this value from the nifgen.Session. max\_waveform\_size property.

### query freq list capabilities

```
nifgen.Session.query_freq_list_capabilities()
```

Returns the properties of the signal generator that are related to creating frequency lists. These properties are nifgen.Session.max\_num\_freq\_lists, nifgen.Session.min\_freq\_list\_length, nifgen.Session.min\_freq\_list\_length, nifgen.Session.min\_freq\_list\_duration, nifgen.Session.max\_freq\_list\_duration, and nifgen.Session.freq\_list\_duration\_quantum.

# Return type

tuple (maximum\_number\_of\_freq\_lists, minimum\_frequency\_list\_length, maximum\_frequency\_list\_length, minimum\_frequency\_list\_duration, maximum frequency list duration, frequency list duration quantum)

#### WHERE

maximum\_number\_of\_freq\_lists (int):

Returns the maximum number of frequency lists that the signal generator allows. NI-FGEN obtains this value from the nifgen.Session.

max\_num\_freq\_lists property.

minimum\_frequency\_list\_length (int):

Returns the minimum number of steps that the signal generator allows in a frequency list. NI-FGEN obtains this value from the nifgen.Session.min\_freq\_list\_length property.

maximum\_frequency\_list\_length (int):

Returns the maximum number of steps that the signal generator allows in a frequency list. NI-FGEN obtains this value from the *nifgen.Session.max\_freq\_list\_length* property.

minimum\_frequency\_list\_duration (float):

Returns the minimum duration that the signal generator allows in a step of a frequency list. NI-FGEN obtains this value from the nifgen. Session. min\_freq\_list\_duration property.

maximum\_frequency\_list\_duration (float):

Returns the maximum duration that the signal generator allows in a step of a frequency list. NI-FGEN obtains this value from the nifgen.Session.

max\_freq\_list\_duration property.

frequency\_list\_duration\_quantum (float):

Returns the quantum of which all durations must be a multiple in a frequency list. NI-FGEN obtains this value from the nifgen.Session. freq\_list\_duration\_quantum property.

# read\_current\_temperature

```
nifgen.Session.read_current_temperature()
```

Reads the current onboard temperature of the device. The temperature is returned in degrees Celsius.

Return type float

**Returns** Returns the current temperature read from onboard temperature sensors, in degrees Celsius.

## reset

```
nifgen.Session.reset()
```

Resets the instrument to a known state. This method aborts the generation, clears all routes, and resets session properties to the default values. This method does not, however, commit the session properties or configure the device hardware to its default state.

**Note:** For the NI 5401/5404/5411/5431, this method exhibits the same behavior as the *nifgen*. *Session.reset\_device()* method.

### reset\_device

```
nifgen.Session.reset device()
```

Performs a hard reset on the device. Generation is stopped, all routes are released, external bidirectional terminals are tristated, FPGAs are reset, hardware is configured to its default state, and all session properties are reset to their default states.

# reset\_with\_defaults

```
nifgen.Session.reset_with_defaults()
```

Resets the instrument and reapplies initial user-specified settings from the logical name that was

used to initialize the session. If the session was created without a logical name, this method is equivalent to the nifgen.Session.reset() method.

# self\_cal

```
nifgen.Session.self_cal()
```

Performs a full internal self-calibration on the device. If the calibration is successful, new calibration data and constants are stored in the onboard EEPROM.

### self test

```
nifgen.Session.self_test()
```

Runs the instrument self-test routine and returns the test result(s).

Raises SelfTestError on self test failure. Properties on exception object:

- · code failure code from driver
- message status message from driver

Self-Test Code	Description
0	Passed self-test
1	Self-test failed

**Note:** When used on some signal generators, the device is reset after the *nifgen.Session.*  $self\_test()$  method runs. If you use the *nifgen.Session.self\_test()* method, your device may not be in its previously configured state after the method runs.

### send software edge trigger

```
nifgen.Session.send_software_edge_trigger()
```

Sends a command to trigger the signal generator. This VI can act as an override for an external edge trigger.

If called directly on the session, this will send a software start trigger.

```
..code:: python
session.send_software_edge_trigger()
```

If called using the script trigger repeated capability container, this will send a software trigger to the specified script trigger

```
..code:: python
    session.script_triggers[1].send_software_edge_trigger()
..note:
```

# set\_next\_write\_position

Sets the position in the waveform at which the next waveform data is written. This method allows you to write to arbitrary locations within the waveform. These settings apply only to the next write to the waveform specified by the waveformHandle parameter. Subsequent writes to that waveform begin where the last write left off, unless this method is called again. The waveformHandle passed in must have been created by a call to the <code>nifgen.Session.allocate\_waveform()</code> method or one of the following <code>nifgen.Session.create\_waveform()</code> method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].set_next_write_position(waveform_name_or_handle, → relative_to, offset)
```

#### **Parameters**

- waveform\_name\_or\_handle (str or int) The name (str) or handle (int) of an arbitrary waveform previously allocated with nifgen.

  Session.allocate\_named\_waveform(), nifgen.Session.

  allocate\_waveform() or nifgen.Session.create\_waveform().
- relative\_to (nifgen.RelativeTo) Specifies the reference position in the waveform. This position and offset together determine where to start loading data into the waveform.

#### \*\*Defined Values\*\*

START (0)	Use the start of the waveform as the reference position.		
CURRENT	Use the current position within the waveform as the reference		
(1)	position.		

• **offset** (*int*) – Specifies the offset from **relativeTo** at which to start loading the data into the waveform.

#### unlock

```
nifgen.Session.unlock()
```

Releases a lock that you acquired on an device session using nifgen.Session.lock(). Refer to nifgen.Session.unlock() for additional information on session locks.

#### wait until done

```
nifgen.Session.wait_until_done (max_time=datetime.timedelta(seconds=10.0)) Waits until the device is done generating or until the maximum time has expired.
```

Parameters max\_time (float in seconds or datetime.timedelta) - Specifies the timeout value in milliseconds.

# write script

```
nifgen.Session.write_script(script)
```

Writes a string containing one or more scripts that govern the generation of waveforms.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].write_script(script)
```

**Parameters** script (str) – Contains the text of the script you want to use for your generation operation. Refer to scripting Instructions for more information about writing scripts.

# write\_waveform

```
nifgen.Session.write_waveform(waveform_name_or_handle, data)
Writes data to the waveform in onboard memory.
```

By default, subsequent calls to this method continue writing data from the position of the last sample written. You can set the write position and offset by calling the <code>nifgen.Session.set\_next\_write\_position()</code> <code>nifgen.Session.set\_next\_write\_position()</code> method.

**Tip:** This method requires repeated capabilities (channels). If called directly on the nifgen. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an nifgen. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].write_waveform(waveform_name_or_handle, data)
```

#### **Parameters**

- waveform\_name\_or\_handle (str or int) The name (str) or handle (int) of an arbitrary waveform previously allocated with nifgen.

  Session.allocate\_named\_waveform(), nifgen.Session.

  allocate\_waveform() or nifgen.Session.create\_waveform().
- data (list of float) Array of data to load into the waveform. This may be an iterable of float, or for best performance a numpy.ndarray of dtype int16 or float64.

## **Properties**

Property	Datatype
nifgen.Session.all_marker_events_latched_status	int
nifgen.Session.all_marker_events_live_status	int
nifgen.Session.analog_data_mask	int
nifgen.Session.analog_filter_enabled	bool
nifgen.Session.analog_path	AnalogPath
nifgen.Session.analog_static_value	int
nifgen.Session.arb_gain	float
nifgen.Session.arb_marker_position	int
nifgen.Session.arb_offset	float
nifgen.Session.arb_repeat_count	int
nifgen.Session.arb_sample_rate	float
nifgen.Session.arb_sequence_handle	int
nifgen.Session.arb_waveform_handle	int
nifgen.Session.aux_power_enabled	bool
nifgen.Session.bus_type	BusType
nifgen.Session.channel_count	int
nifgen.Session.channel_delay	float
nifgen.Session.clock_mode	ClockMode
nifgen.Session.common_mode_offset	float
nifgen.Session.data_marker_events_count	int
nifgen.Session.data_marker_event_data_bit_number	int
nifgen.Session.data_marker_event_level_polarity	DataMarkerEventLevelPola
nifgen.Session.data_marker_event_output_terminal	str
nifgen.Session.data_transfer_block_size	int
nifgen.Session.data_transfer_maximum_bandwidth	float
nifgen.Session.data_transfer_maximum_in_flight_reads	int
nifgen.Session.data_transfer_preferred_packet_size	int
nifgen.Session.digital_data_mask	int
nifgen.Session.digital_edge_script_trigger_edge	ScriptTriggerDigitalEdge
nifgen.Session.digital_edge_script_trigger_source	str
nifgen.Session.digital_edge_start_trigger_edge	StartTriggerDigitalEdgeE
nifgen.Session.digital_edge_start_trigger_source	str
nifgen.Session.digital_filter_enabled nifgen.Session.digital_filter_interpolation_factor	bool float
nifgen.Session.digital_gain	float
nifgen.Session.digital_pattern_enabled	bool
nifgen.Session.digital_static_value	int
nifgen.Session.done_event_output_terminal	str
nifgen.Session.driver_setup	str
nifgen.Session.exported_onboard_reference_clock_output_terminal	str
nifgen.Session.exported_onBoard_reference_clock_output_terminal	str
nifgen.Session.exported_reference_clock_divisor	int
nifgen.Session.exported_sample_clock_output_terminal	str
nifgen.Session.exported_sample_clock_timebase_divisor	int
nifgen.Session.exported_sample_clock_timebase_output_terminal	str
nifgen.Session.exported_sample_clock_timebase_output_terminal	str
nifgen.Session.exported_start_trigger_output_terminal	str
nifgen.Session.exported_start_trigger_output_terminar	int
nifgen.Session.external_sample_clock_multiplier	float
nifgen.Session.file_transfer_block_size	int
1114C11.00001011.1110_014110101_01001_0140	****

Continued on next

Table 9 – continued from previous page

Property	Datatype
nifgen.Session.filter_correction_frequency	float
nifgen.Session.flatness_correction_enabled	bool
nifgen.Session.fpga_bitfile_path	str
nifgen.Session.freq_list_duration_quantum	float
nifgen.Session.freq_list_handle	int
nifgen.Session.func_amplitude	float
nifgen.Session.func_ampfitude nifgen.Session.func_buffer_size	int
nifgen.Session.func_dc_offset	float
nifgen.Session.func_duty_cycle_high	float
nifgen.Session.func_frequency	float
nifgen.Session.func_max_buffer_size	int
nifgen.Session.func_max_burrer_size nifgen.Session.func_start_phase	float
nifgen.Session.func_start_phase	Waveform
nifgen.Session.idle_behavior	IdleBehavior
nifgen.Session.idle_value	int
nifgen.Session.instrument_firmware_revision	str
nifgen.Session.instrument_manufacturer	str
nifgen.Session.instrument_model	str
nifgen.Session.io_resource_descriptor	str
nifgen.Session.load_impedance	float
nifgen.Session.logical_name	str
nifgen.Session.major_version	int
nifgen.Session.marker_events_count	int
nifgen.Session.marker_event_output_terminal	str
nifgen.Session.max_freq_list_duration	float
nifgen.Session.max_freq_list_length	int
nifgen.Session.max_loop_count	int
nifgen.Session.max_num_freq_lists	int
nifgen.Session.max_num_sequences	int
nifgen.Session.max_num_waveforms	int
nifgen.Session.max_sequence_length	int
nifgen.Session.max_waveform_size	int
nifgen.Session.memory_size	int
nifgen.Session.minor_version	int
nifgen.Session.min_freq_list_duration	float
nifgen.Session.min_freq_list_length	int
nifgen.Session.min_sequence_length	int
nifgen.Session.min_waveform_size	int
nifgen.Session.module_revision	str
nifgen.Session.output_enabled	bool
nifgen.Session.output_impedance	float
nifgen.Session.output_mode	OutputMode
nifgen.Session.ready_for_start_event_output_terminal	str
nifgen.Session.reference_clock_source	ReferenceClockSource
nifgen.Session.ref_clock_frequency	float
nifgen.Session.sample_clock_source	SampleClockSource
nifgen.Session.sample_clock_timebase_rate	float
nifgen.Session.sample_clock_timebase_source	SampleClockTimebaseSourc
nifgen.Session.script_to_generate	str
miligon, coopion, collips_co_generate	501

Continued on next

Table 9 – continued from previous page

Property	Datatype
nifgen.Session.script_triggers_count	int
nifgen.Session.script_trigger_type	ScriptTriggerType
nifgen.Session.serial_number	str
nifgen.Session.simulate	bool
nifgen.Session.specific_driver_description	str
nifgen.Session.specific_driver_revision	str
nifgen.Session.specific_driver_vendor	str
nifgen.Session.started_event_output_terminal	str
nifgen.Session.start_trigger_type	StartTriggerType
nifgen.Session.streaming_space_available_in_waveform	int
nifgen.Session.streaming_waveform_handle	int
nifgen.Session.streaming_waveform_name	str
nifgen.Session.streaming_write_timeout	float in seconds or datetime.timedelt
nifgen.Session.supported_instrument_models	str
nifgen.Session.terminal_configuration	TerminalConfiguration
nifgen.Session.trigger_mode	TriggerMode
nifgen.Session.wait_behavior	WaitBehavior
niigen.bebbion.waie_benavior	Warebellavior
nifgen.Session.wait_senavior	int

# Methods

Method name	
nifgen.Session.abort()	
nifgen.Session.allocate_named_waveform()	
nifgen.Session.allocate_waveform()	
nifgen.Session.clear_arb_memory()	
nifgen.Session.clear_arb_sequence()	
nifgen.Session.clear_freq_list()	
nifgen.Session.clear_user_standard_waveform()	
nifgen.Session.commit()	
nifgen.Session.configure_arb_sequence()	
nifgen.Session.configure_arb_waveform()	
nifgen.Session.configure_custom_fir_filter_coefficient	s()
nifgen.Session.configure_freq_list()	
nifgen.Session.configure_standard_waveform()	
nifgen.Session.create_advanced_arb_sequence()	
nifgen.Session.create_arb_sequence()	
nifgen.Session.create_freq_list()	
nifgen.Session.create_waveform_from_file_f64()	
nifgen.Session.create_waveform_from_file_i16()	
nifgen.Session.create_waveform_numpy()	
nifgen.Session.define_user_standard_waveform()	
nifgen.Session.delete_script()	
nifgen.Session.delete_waveform()	
nifgen.Session.disable()	
nifgen.Session.get_ext_cal_last_date_and_time()	
nifgen.Session.get_ext_cal_last_temp()	

Continued on next page

Table 10 – continued from previous page

Method name
nifgen.Session.get_ext_cal_recommended_interval()
nifgen.Session.get_hardware_state()
nifgen.Session.get_self_cal_last_date_and_time()
nifgen.Session.get_self_cal_last_temp()
nifgen.Session.get_self_cal_supported()
nifgen.Session.is_done()
nifgen.Session.lock()
nifgen.Session.query_arb_seq_capabilities()
nifgen.Session.query_arb_wfm_capabilities()
nifgen.Session.query_freq_list_capabilities()
nifgen.Session.read_current_temperature()
nifgen.Session.reset()
nifgen.Session.reset_device()
nifgen.Session.reset_with_defaults()
nifgen.Session.self_cal()
nifgen.Session.self_test()
nifgen.Session.send_software_edge_trigger()
nifgen.Session.set_next_write_position()
nifgen.Session.unlock()
nifgen.Session.wait_until_done()
nifgen.Session.write_script()
nifgen.Session.write_waveform()

### 7.3.3.2 Enums

Enums used in NI-FGEN

# **AnalogPath**

# class nifgen.AnalogPath

#### MAIN

Specifies use of the main path. NI-FGEN chooses the amplifier based on the user-specified gain.

## DIRECT

Specifies use of the direct path.

# FIXED\_LOW\_GAIN

Specifies use of the low-gain amplifier in the main path, no matter what value the user specifies for gain. This setting limits the output range.

# FIXED\_HIGH\_GAIN

Specifies use of the high-gain amplifier in the main path.

# **BusType**

### class nifgen.BusType

# INVALID

Indicates an invalid bus type.

ΑT

Indicates the signal generator is the AT bus type.

PCI

Indicates the signal generator is the PCI bus type.

PXI

Indicates the signal generator is the PXI bus type.

VXI

Indicates the signal generator is the VXI bus type.

**PCMCIA** 

Indicates the signal generator is the PCI-CMA bus type.

PXIE

Indicates the signal generator is the PXI Express bus type.

# **ByteOrder**

```
class nifgen.ByteOrder
```

LITTLE

BIG

# ClockMode

```
class nifgen.ClockMode
```

#### HIGH RESOLUTION

High resolution sampling—Sample rate is generated by a high–resolution clock source.

#### DIVIDE DOWN

Divide down sampling—Sample rates are generated by dividing the source frequency.

# AUTOMATIC

Automatic Selection—NI-FGEN selects between the divide-down and high-resolution clocking modes.

# **DataMarkerEventLevelPolarity**

# class nifgen.DataMarkerEventLevelPolarity

HIGH

When the operation is ready to start, the Ready for Start event level is high.

LOW

When the operation is ready to start, the Ready for Start event level is low.

### **HardwareState**

```
class nifgen.HardwareState
```

IDLE

WAITING\_FOR\_START\_TRIGGER

RUNNING

DONE

HARDWARE ERROR

### **IdleBehavior**

#### class nifgen. IdleBehavior

#### HOLD LAST

While in an Idle or Wait state, the output signal remains at the last voltage generated prior to entering the state

#### JUMP TO

While in an Idle or Wait state, the output signal remains at the value configured in the Idle or Wait value property.

# OutputMode

### class nifgen.OutputMode

#### FUNC

Standard Method mode—Generates standard method waveforms such as sine, square, triangle, and so on.

#### ARB

Arbitrary waveform mode—Generates waveforms from user-created/provided waveform arrays of numeric data.

#### SEQ

Arbitrary sequence mode — Generates downloaded waveforms in an order your specify.

# FREQ\_LIST

Frequency List mode—Generates a standard method using a list of frequencies you define.

#### SCRIPT

Script mode—Allows you to use scripting to link and loop multiple waveforms in complex combinations.

#### ReferenceClockSource

# class nifgen.ReferenceClockSource

#### CLOCK IN

Specifies that the CLK IN input signal from the front panel connector is used as the Reference Clock source.

# NONE

Specifies that a Reference Clock is not used.

### ONBOARD REFERENCE CLOCK

Specifies that the onboard Reference Clock is used as the Reference Clock source.

#### PXI CLOCK

Specifies the PXI Clock is used as the Reference Clock source.

#### RTSI 7

Specifies that the RTSI line 7 is used as the Reference Clock source.

### RelativeTo

```
class nifgen.RelativeTo
```

START

CURRENT

# SampleClockSource

### class nifgen.SampleClockSource

#### CLOCK IN

Specifies that the signal at the CLK IN front panel connector is used as the Sample Clock source.

#### DDC CLOCK IN

Specifies that the Sample Clock from DDC connector is used as the Sample Clock source.

#### ONBOARD CLOCK

Specifies that the onboard clock is used as the Sample Clock source.

## PXI\_STAR\_LINE

Specifies that the PXI\_STAR trigger line is used as the Sample Clock source.

# PXI\_TRIGGER\_LINE\_ORTSI\_0

Specifies that the PXI or RTSI line 0 is used as the Sample Clock source.

### PXI\_TRIGGER\_LINE\_1RTSI\_1

Specifies that the PXI or RTSI line 1 is used as the Sample Clock source.

# PXI\_TRIGGER\_LINE\_2RTSI\_2

Specifies that the PXI or RTSI line 2 is used as the Sample Clock source.

# PXI\_TRIGGER\_LINE\_3RTSI\_3

Specifies that the PXI or RTSI line 3 is used as the Sample Clock source.

#### PXI TRIGGER LINE 4RTSI 4

Specifies that the PXI or RTSI line 4 is used as the Sample Clock source.

# PXI\_TRIGGER\_LINE\_5RTSI\_5

Specifies that the PXI or RTSI line 5 is used as the Sample Clock source.

#### PXI\_TRIGGER\_LINE\_6RTSI\_6

Specifies that the PXI or RTSI line 6 is used as the Sample Clock source.

#### PXI TRIGGER LINE 7RTSI 7

Specifies that the PXI or RTSI line 7 is used as the Sample Clock source.

# SampleClockTimebaseSource

# class nifgen.SampleClockTimebaseSource

### CLOCK\_IN

Specifies that the external signal on the CLK IN front panel connector is used as the source.

#### ONBOARD CLOCK

Specifies that the onboard Sample Clock timebase is used as the source.

# ScriptTriggerDigitalEdgeEdge

```
class nifgen.ScriptTriggerDigitalEdgeEdge
```

#### RISING

Rising Edge

### **FALLING**

Falling Edge

# ScriptTriggerType

```
class nifgen.ScriptTriggerType
```

#### TRIG NONE

No trigger is configured. Signal generation starts immediately.

## DIGITAL\_EDGE

Trigger is asserted when a digital edge is detected.

#### DIGITAL LEVEL

Trigger is asserted when a digital level is detected.

### SOFTWARE\_EDGE

Trigger is asserted when a software edge is detected.

# StartTriggerDigitalEdgeEdge

```
class nifgen.StartTriggerDigitalEdgeEdge
```

#### RISING

Rising Edge

# **FALLING**

Falling Edge

# StartTriggerType

```
class nifgen.StartTriggerType
```

TRIG NONE

None

DIGITAL EDGE

Digital Edge

SOFTWARE EDGE

Software Edge

P2P ENDPOINT FULLNESS

P2P Endpoint Fullness

#### **TerminalConfiguration**

class nifgen. TerminalConfiguration

SINGLE ENDED

Single-ended operation

**DIFFERENTIAL** 

Differential operation

## **TriggerMode**

class nifgen.TriggerMode

# SINGLE

Single Trigger Mode - The waveform you describe in the sequence list is generated only once by going through the entire staging list. Only one trigger is required to start the waveform generation. You can use Single trigger mode with the output mode in any mode. After a trigger is received, the waveform generation starts from the first stage and continues through to the last stage. Then, the last stage generates repeatedly until you stop the waveform generation.

#### CONTINUOUS

Continuous Trigger Mode - The waveform you describe in the staging list generates infinitely by repeatedly cycling through the staging list. After a trigger is received, the waveform generation starts from the first stage and continues through to the last stage. After the last stage completes, the waveform generation loops back to the start of the first stage and continues until it is stopped. Only one trigger is required to start the waveform generation.

#### STEPPED

Stepped Trigger Mode - After a start trigger is received, the waveform described by the first stage generates. Then, the device waits for the next trigger signal. On the next trigger, the waveform described by the second stage generates, and so on. After the staging list completes, the waveform generation returns to the first stage and continues in a cyclic fashion. After any stage has generated completely, the first eight samples of the next stage are repeated continuously until the next trigger is received. trigger mode.

**Note:** In Frequency List mode, Stepped trigger mode is the same as Burst

# BURST

Burst Trigger Mode - After a start trigger is received, the waveform described by the first stage generates until another trigger is received. At the next trigger, the buffer of the previous stage completes, and then the waveform described by the second stage generates. After the staging list completes, the waveform

7.3. nifgen module 285

generation returns to the first stage and continues in a cyclic fashion. In Frequency List mode, the duration instruction is ignored, and the trigger switches the frequency to the next frequency in the list. trigger mode.

Note: In Frequency List mode, Stepped trigger mode is the same as Burst

#### WaitBehavior

```
class nifgen.WaitBehavior
```

#### HOLD LAST

While in an Idle or Wait state, the output signal remains at the last voltage generated prior to entering the state.

#### JUMP\_TO

While in an Idle or Wait state, the output signal remains at the value configured in the Idle or Wait value property.

#### Waveform

```
class nifgen. Waveform
```

#### SINE

Sinusoid waveform

# SQUARE

Square waveform

#### TRIANGLE

Triange waveform

#### RAMP UP

Positive ramp waveform

## RAMP DOWN

Negative ramp waveform

DC

Constant voltage

#### NOISE

White noise

#### USER

User-defined waveform as defined by the nifgen.Session. define\_user\_standard\_waveform() method.

# 7.3.3.3 Exceptions and Warnings

#### **DriverError**

# exception nifgen.DriverError

An error originating from the NI-FGEN driver

# UnsupportedConfigurationError

```
{\tt exception} \ {\tt nifgen. Unsupported Configuration Error}
```

An error due to using this module in an usupported platform.

#### **DriverNotInstalledError**

```
exception nifgen.DriverNotInstalledError
```

An error due to using this module without the driver runtime installed.

# InvalidRepeatedCapabilityError

```
exception nifgen.InvalidRepeatedCapabilityError
```

An error due to an invalid character in a repeated capability

#### SelfTestError

```
exception nifgen.SelfTestError
```

An error due to a failed self-test

# **DriverWarning**

```
exception nifgen.DriverWarning
```

A warning originating from the NI-FGEN driver

## **7.3.3.4 Examples**

# nifgen\_arb\_waveform.py

#### Listing 6: (nifgen\_arb\_waveform.py)

```
#!/usr/bin/python
   import argparse
   import math
   import nifgen
   import sys
   import time
   def create_waveform_data(number_of_samples):
10
       waveform_data = []
11
       angle_per_sample = (2 * math.pi) / number_of_samples
12
       for i in range(number_of_samples):
13
           waveform_data.append(math.sin(i * angle_per_sample) * math.sin(i * angle_per_
   \rightarrowsample * 20))
       return waveform_data
15
16
```

(continues on next page)

7.3. nifgen module 287

(continued from previous page)

```
def example (resource name, options, samples, gain, offset, gen_time):
18
       waveform_data = create_waveform_data(samples)
19
       with nifgen.Session(resource_name=resource_name, options=options) as session:
20
           session.output_mode = nifgen.OutputMode.ARB
21
           waveform = session.create_waveform(waveform_data_array=waveform_data)
22
            session.configure_arb_waveform(waveform_handle=waveform, gain=gain,_
23
   \hookrightarrowoffset=offset)
           with session.initiate():
24
                time.sleep(gen_time)
25
26
27
   def _main(argsv):
28
       parser = argparse.ArgumentParser(description='Continuously generates an arbitrary,...
   →waveform.', formatter_class=argparse.ArgumentDefaultsHelpFormatter)
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource...
30
   →name of a National Instruments Arbitrary Waveform Generator')
       parser.add_argument('-s', '--samples', default=100000, type=int, help='Number of_
31
   →samples')
       parser.add_argument('-g', '--gain', default=1.0, type=float, help='Gain')
32
       parser.add_argument('-o', '--offset', default=0.0, type=float, help='DC offset (V)
33
       parser.add_argument('-t', '--time', default=5.0, type=float, help='Generation...
34
   →time (s)')
       parser.add_argument('-op', '--option-string', default='', type=str, help='Option_
35
   ⇔string')
       args = parser.parse_args(argsv)
       example(args.resource_name, args.option_string, args.samples, args.gain, args.
37
   →offset, args.time)
38
39
   def main():
40
41
       _main(sys.argv[1:])
42
43
   def test example():
44
       options = {'simulate': True, 'driver_setup': {'Model': '5433 (2CH)', 'BoardType':
45
   → 'PXIe', }, }
       example('PXI1Slot2', options, 100000, 1.0, 0.0, 5.0)
47
48
   def test main():
49
       cmd_line = ['--option-string', 'Simulate=1, DriverSetup=Model:5433 (2CH);
50
   →BoardType:PXIe', ]
       _main(cmd_line)
51
52
53
   if __name__ == '__main__':
54
       main()
55
56
57
```

## nifgen\_standard\_function.py

Listing 7: (nifgen\_standard\_function.py)

```
#!/usr/bin/python
2
   import argparse
   import nifgen
   import svs
   import time
   def example (resource_name, options, waveform, frequency, amplitude, offset, phase,...
   ⇒gen time):
       with nifgen. Session (resource name=resource name, options=options) as session:
10
           session.output_mode = nifgen.OutputMode.FUNC
11
           session.configure_standard_waveform(waveform=nifgen.Waveform[waveform],...
12
   →amplitude-amplitude, frequency-frequency, dc_offset-offset, start_phase=phase)
           with session.initiate():
                time.sleep(gen_time)
14
15
16
   def _main(argsv):
17
       supported_waveforms = list(nifgen.Waveform.__members__.keys())[:-1] # no support_
18
   → for user-defined waveforms in example
       parser = argparse.ArgumentParser(description='Generates the standard function.',...
19
   →formatter_class=argparse.ArgumentDefaultsHelpFormatter)
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource...
20
   →name of a National Instruments Function Generator')
       parser.add_argument('-w', '--waveform', default=supported_waveforms[0],...
21
   →choices=supported_waveforms, type=str.upper, help='Standard waveform')
22
       parser.add_argument('-f', '--frequency', default=1000, type=float, help=
   →'Frequency (Hz)')
       parser.add_argument('-a', '--amplitude', default=1.0, type=float, help='Amplitude_
23
   \hookrightarrow (Vpk-pk) ')
       parser.add_argument('-o', '--offset', default=0.0, type=float, help='DC offset (V)
24
       parser.add_argument('-p', '--phase', default=0.0, type=float, help='Start phase,
   \hookrightarrow (deg) ')
       parser.add_argument('-t', '--time', default=5.0, type=float, help='Generation_
26
   \rightarrowtime (s)')
       parser.add_argument('-op', '--option-string', default='', type=str, help='Option_
27
   ⇒string')
28
       args = parser.parse_args(argsv)
       example(args.resource_name, args.option_string, args.waveform, args.frequency,_
   →args.amplitude, args.offset, args.phase, args.time)
30
31
   def main():
32
       _main(sys.argv[1:])
33
34
   def test_example():
36
       options = {'simulate': True, 'driver_setup': {'Model': '5433 (2CH)', 'BoardType':
37
   → 'PXIe', }, }
       example('PXI1Slot2', options, 'SINE', 1000, 1.0, 0.0, 0.0, 5.0)
38
39
   def test_main():
```

(continues on next page)

(continued from previous page)

# 7.4 niscope module

#### 7.4.1 Installation

As a prerequisite to using the niscope module, you must install the NI-SCOPE runtime on your system. Visit ni.com/downloads to download the driver runtime for your devices.

The nimi-python modules (i.e. for NI-SCOPE) can be installed with pip:

```
$ python -m pip install niscope~=1.0.1
```

Or easy\_install from setuptools:

```
$ python -m easy_install niscope
```

# 7.4.2 **Usage**

The following is a basic example of using the **niscope** module to open a session to a High Speed Digitizer and capture a single record of 1000 points.

```
import niscope
with niscope. Session ("Dev1") as session:
    session.channels[0].configure_vertical(range=1.0, coupling=niscope.
→VerticalCoupling.AC)
   session.channels[1].configure_vertical(range=10.0, coupling=niscope.
→VerticalCoupling.DC)
   session.configure_horizontal_timing(min_sample_rate=50000000, min_num_pts=1000,_
→ref_position=50.0, num_records=5, enforce_realtime=True)
   with session.initiate():
       waveforms = session.channels[0,1].fetch(num_records=5)
    for wfm in waveforms:
       print('Channel {0}, record {1} samples acquired: {2:,}\n'.format(wfm.channel,_
→wfm.record, len(wfm.samples)))
    # Find all channel 1 records (Note channel name is always a sting even if,
→integers used in channel[])
   chan1 = [wfm for wfm in waveforms if wfm.channel == '0']
    # Find all record number 3
    rec3 = [wfm for wfm in waveforms if wfm.record == 3]
```

The waveform returned from fetch is a flat list of Python objects

#### • Attributes:

- relative\_initial\_x (float) the time (in seconds) from the trigger to the first sample in the fetched waveform
- absolute\_initial\_x (float) timestamp (in seconds) of the first fetched sample. This timestamp is comparable between records and acquisitions; devices that do not support this parameter use 0 for this output.
- x\_increment (float) the time between points in the acquired waveform in seconds
- channel (str) channel name this waveform was asquire from
- record (int) record number of this waveform
- gain (float) the gain factor of the given channel; useful for scaling binary data with the following formula:

```
voltage = binarydata * gainfactor + offset
```

offset (float) the offset factor of the given channel; useful for scaling binary data with the following formula:

```
voltage = binarydata * gainfactor + offset
```

- samples (array of float) floating point array of samples. Length will be of the actual samples acquired
- Such that all record 0 waveforms are first. For example, with a channel list of 0,1, you would have the following
  index values:
  - index 0 = record 0, channel 0
  - index 1 = record 0, channel 1
  - index 2 = record 1, channel 0
  - index 3 = record 1, channel 1
  - etc.

If you need more performance or need to work with SciPy, you can use the *fetch\_into()* method instead of *fetch()*. This method takes an already allocated numpy array and puts the acquired samples in it. Data types supported:

- numpy.float64
- numpy.int8
- numpy.in16
- numpy.int32

The waveform\_infos returned from fetch\_into is a 1D list of Python objects

- Attributes:
  - relative\_initial\_x (float) the time (in seconds) from the trigger to the first sample in the fetched waveform
  - absolute\_initial\_x (float) timestamp (in seconds) of the first fetched sample. This timestamp is comparable between records and acquisitions; devices that do not support this parameter use 0 for this output.
  - x\_increment (float) the time between points in the acquired waveform in seconds
  - channel (str) channel name this waveform was asquire from
  - record (int) record number of this waveform
  - gain (float) the gain factor of the given channel; useful for scaling binary data with the following formula:

$$voltage = binarydata * gainfactor + offset$$

offset (float) the offset factor of the given channel; useful for scaling binary data with the following formula:

$$voltage = binarydata * gainfactor + offset$$

 samples (numpy array of datatype used) floating point array of samples. Length will be of the actual samples acquired

```
Note: Python 3 only
```

- Such that all record 0 waveforms are first. For example, with a channel list of 0,1, you would have the following index values:
  - index 0 = record 0, channel 0
  - index 1 = record 0, channel 1
  - index 2 = record 1, channel 0
  - index 3 = record 1, channel 1
  - etc.

**Note:** When using Python 2, the waveform\_infos objects do not include the waveform for that record. Instead, samples are in the waveform passed into the function using the following layout:

- index 0 = record 0, channel 0
- index x = record 0, channel 1
- index 2x = record 1, channel 0
- index 3x = record 1, channel 1
- etc.
- Where x = the record length

Additional examples for NI-SCOPE are located in src/niscope/examples/ directory.

# 7.4.3 API Reference

# 7.4.3.1 niscope.Session

**class** niscope.**Session** (*self*, *resource\_name*, *id\_query=False*, *reset\_device=False*, *options={}*) Performs the following initialization actions:

- Creates a new IVI instrument driver and optionally sets the initial state of the following session properties: Range Check, Cache, Simulate, Record Value Coercions
- Opens a session to the specified device using the interface and address you specify for the resourceName
- Resets the digitizer to a known state if **resetDevice** is set to True
- Queries the instrument ID and verifies that it is valid for this instrument driver if the IDQuery is set to True
- Returns an instrument handle that you use to identify the instrument in all subsequent instrument driver method calls

#### **Parameters**

• resource\_name (str) -

**Caution:** Traditional NI-DAQ and NI-DAQmx device names are not case-sensitive. However, all IVI names, such as logical names, are case-sensitive. If you use logical names, driver session names, or virtual names in your program, you must make sure that the name you use matches the name in the IVI Configuration Store file exactly, without any variations in the case of the characters.

Specifies the resource name of the device to initialize

For Traditional NI-DAQ devices, the syntax is DAQ::*n*, where *n* is the device number assigned by MAX, as shown in Example 1.

For NI-DAQmx devices, the syntax is just the device name specified in MAX, as shown in Example 2. Typical default names for NI-DAQmx devices in MAX are Dev1 or PXI1Slot1. You can rename an NI-DAQmx device by right-clicking on the name in MAX and entering a new name.

An alternate syntax for NI-DAQmx devices consists of DAQ::NI-DAQmx device name, as shown in Example 3. This naming convention allows for the use of an NI-DAQmx device in an application that was originally designed for a Traditional NI-DAQ device. For example, if the application expects DAQ::1, you can rename the NI-DAQmx device to 1 in MAX and pass in DAQ::1 for the resource name, as shown in Example 4.

If you use the DAQ::*n* syntax and an NI-DAQmx device name already exists with that same name, the NI-DAQmx device is matched first.

You can also pass in the name of an IVI logical name or an IVI virtual name configured with the IVI Configuration utility, as shown in Example 5. A logical name identifies a particular virtual instrument. A virtual name identifies a specific device and specifies the initial settings for the session.

Exam-	Device Type	Syntax
ple		
1	Traditional NI-DAQ device	DAQ::1 (1 = device number)
2	NI-DAQmx device	myDAQmxDevice (myDAQmxDevice = de-
		vice name)
3	NI-DAQmx device	DAQ::myDAQmxDevice (myDAQmxDevice
		= device name)
4	NI-DAQmx device	DAQ::2 (2 = device name)
5	IVI logical name or IVI vir-	myLogicalName (myLogicalName = name)
	tual name	

•  $id_query(bool)$  – Specify whether to perform an ID query.

When you set this parameter to True, NI-SCOPE verifies that the device you initialize is a type that it supports.

When you set this parameter to False, the method initializes the device without performing an ID query.

# **Defined Values**

True—Perform ID query False—Skip ID query

Default Value: True

• **reset\_device** (bool) – Specify whether to reset the device during the initialization process.

Default Value: True

**Defined Values** 

True (1)—Reset device

False (0)—Do not reset device

**Note:** For the NI 5112, repeatedly resetting the device may cause excessive wear on the electromechanical relays. Refer to NI 5112 Electromechanical Relays for recommended programming practices.

• **options** (str) – Specifies the initial value of certain properties for the session. The syntax for **options** is a dictionary of properties with an assigned value. For example:

```
{ 'simulate': False }
```

You do not have to specify a value for all the properties. If you do not specify a value for a property, the default value is used.

Advanced Example: { 'simulate': True, 'driver\_setup': { 'Model': '<model number>', 'BoardType': '<type>' } }

Property	Default
range_check	True
query_instrument_status	False
cache	True
simulate	False
record_value_coersions	False
driver_setup	{}

# **Properties**

Property	Datatype
absolute_sample_clock_offset	float in seconds or datetime.timedelta
accessory_gain	float
accessory_offset	float
acquisition_start_time	float in seconds or datetime.timedelta
acquisition_type	AcquisitionType
acq_arm_source	str
adv_trig_src	str
allow_more_records_than_memory	bool
arm_ref_trig_src	str
backlog	float
bandpass_filter_enabled	bool
binary_sample_width	int
channel_count	int
channel_enabled	bool
channel_terminal_configuration	TerminalConfiguration
data_transfer_block_size	int
data_transfer_maximum_bandwidth	float
data_transfer_preferred_packet_size	int
device_temperature	float
driver_setup	str
enable_dc_restore	bool
enable_time_interleaved_sampling	bool
end_of_acquisition_event_output_terminal	str
end_of_record_event_output_terminal	str
<pre>end_of_record_to_advance_trigger_holdoff</pre>	float in seconds or datetime.timedelta
equalization_filter_enabled	bool
equalization_num_coefficients	int
exported_advance_trigger_output_terminal	str
exported_ref_trigger_output_terminal	str
exported_start_trigger_output_terminal	str
flex_fir_antialias_filter_type	FlexFIRAntialiasFilterType
fpga_bitfile_path	str
high_pass_filter_frequency	float
horz_enforce_realtime	bool
horz_min_num_pts	int
horz_num_records	int
horz_record_length	int
horz_record_ref_position	float
horz_sample_rate	float

Continued on next page

Table 11 – continued from previous page

Property Table 11 – continued from p	Datatype
horz_time_per_record	float in seconds or datetime.timedelta
input_clock_source	str
input_impedance	float
instrument_firmware_revision	str
instrument_manufacturer	str
instrument_model	str
interleaving_offset_correction_enabled	bool
io_resource_descriptor	str
logical_name	str
master enable	bool
max_input_frequency	float
max_real_time_sampling_rate	float
max_ris_rate	float
min_sample_rate	float
onboard_memory_size	int
output_clock_source	str
pll_lock_status	bool
points_done	float
poll_interval	int
probe_attenuation	float
ready_for_advance_event_output_terminal	str
ready_for_ref_event_output_terminal	str
ready_for_start_event_output_terminal	str
records_done	int
record_arm_source	str
ref_clk_rate	float
ref_trigger_detector_location	RefTriggerDetectorLocation
ref_trigger_minimum_quiet_time	float in seconds or datetime.timedelta
ref_trig_tdc_enable	bool
resolution	int
ris_in_auto_setup_enable	bool
ris_method	RISMethod
ris_num_averages	int
sample_clock_timebase_multiplier	int
sample_mode	int
samp_clk_timebase_div	int
samp_clk_timebase_rate	float
samp_clk_timebase_src	str
serial_number	str
simulate	bool
specific_driver_description	str
specific_driver_revision	str
specific_driver_vendor	str
start_to_ref_trigger_holdoff	float in seconds or datetime.timedelta
supported_instrument_models	str
	bool
LIIGGEL AULO LIIGGELEG	
trigger_auto_triggered trigger_coupling	TriggerCoupling
trigger_auto_triggered  trigger_coupling  trigger_delay_time	TriggerCoupling float in seconds or datetime.timedelta

Continued on next page

Table 11 – continued from previous page

Property	Datatype
trigger_hysteresis	float
trigger_impedance	float
trigger_level	float
trigger_modifier	TriggerModifier
trigger_slope	TriggerSlope
trigger_source	str
trigger_type	TriggerType
trigger_window_high_level	float
trigger_window_low_level	float
trigger_window_mode	TriggerWindowMode
tv_trigger_event	VideoTriggerEvent
tv_trigger_line_number	int
tv_trigger_polarity	VideoPolarity
tv_trigger_signal_format	VideoSignalFormat
vertical_coupling	VerticalCoupling
vertical_offset	float
vertical_range	float

# **Public methods**

Method name
abort()
acquisition_status()
auto_setup()
commit()
configure_chan_characteristics()
<pre>configure_equalization_filter_coefficients()</pre>
configure_horizontal_timing()
configure_trigger_digital()
configure_trigger_edge()
configure_trigger_hysteresis()
configure_trigger_immediate()
configure_trigger_software()
configure_trigger_video()
configure_trigger_window()
configure_vertical()
disable()
fetch()
fetch_into()
<pre>get_equalization_filter_coefficients()</pre>
lock()
<pre>probe_compensation_signal_start()</pre>
<pre>probe_compensation_signal_stop()</pre>
read()
reset()
reset_device()
reset_with_defaults()
self_cal()
self_test()
Continued on next page

Continued on next page

Table 12 – continued from previous page

Method name	
send_software_trigger_edge()	
unlock()	

## **Properties**

## absolute sample clock offset

niscope.Session.absolute\_sample\_clock\_offset

Gets or sets the absolute time offset of the sample clock relative to the reference clock in terms of seconds.

**Note:** Configures the sample clock relationship with respect to the reference clock. This parameter is factored into NI-TClk adjustments and is typically used to improve the repeatability of NI-TClk Synchronization. When this parameter is read, the currently programmed value is returned. The range of the absolute sample clock offset is [-.5 sample clock periods, .5 sample clock periods]. The default absolute sample clock offset is 0s.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocking:Advanced:Absolute Sample Clock Offset
- C Attribute: NISCOPE\_ATTR\_ABSOLUTE\_SAMPLE\_CLOCK\_OFFSET

#### accessory gain

 $\verb|niscope.Session.accessory_gain|\\$ 

Returns the calibration gain for the current device configuration. **Related topics:** NI 5122/5124/5142 Calibration

**Note:** This property is only supported by the NI PXI-5900 differential amplifier.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].accessory_gain
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Device:Accessory:Gain
- C Attribute: NISCOPE\_ATTR\_ACCESSORY\_GAIN

## accessory\_offset

niscope.Session.accessory\_offset

Returns the calibration offset for the current device configuration. **Related topics:** NI 5122/5124/5142 Calibration

Note: This property is supported only by the NI PXI-5900 differential amplifier.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].accessory_offset
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Device:Accessory:Offset
- C Attribute: NISCOPE\_ATTR\_ACCESSORY\_OFFSET

## acquisition start time

#### niscope.Session.acquisition\_start\_time

Specifies the length of time from the trigger event to the first point in the waveform record in seconds. If the value is positive, the first point in the waveform record occurs after the trigger event (same as specifying <code>niscope.Session.trigger\_delay\_time</code>). If the value is negative, the first point in the waveform record occurs before the trigger event (same as specifying <code>niscope.Session.horz record ref position</code>).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Horizontal:Advanced:Acquisition Start Time
- C Attribute: NISCOPE\_ATTR\_ACQUISITION\_START\_TIME

## acquisition type

# niscope.Session.acquisition\_type

Specifies how the digitizer acquires data and fills the waveform record.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.AcquisitionType
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Acquisition: Acquisition Type
- C Attribute: NISCOPE\_ATTR\_ACQUISITION\_TYPE

# acq\_arm\_source

# niscope.Session.acq\_arm\_source

Specifies the source the digitizer monitors for a start (acquisition arm) trigger. When the start trigger is received, the digitizer begins acquiring pretrigger samples. Valid Values: NISCOPE\_VAL\_IMMEDIATE ('VAL\_IMMEDIATE') - Triggers immediately NISCOPE\_VAL\_RTSI\_0 ('VAL\_RTSI\_0') - RTSI 0 NISCOPE\_VAL\_RTSI\_1 ('VAL\_RTSI\_1')

- RTSI 1 NISCOPE\_VAL\_RTSI\_2 ('VAL\_RTSI\_2') - RTSI 2 NISCOPE\_VAL\_RTSI\_3 ('VAL\_RTSI\_3') - RTSI 3 NISCOPE\_VAL\_RTSI\_4 ('VAL\_RTSI\_4') - RTSI 4 NISCOPE\_VAL\_RTSI\_5 ('VAL\_RTSI\_5') - RTSI 5 NISCOPE\_VAL\_RTSI\_6 ('VAL\_RTSI\_6') - RTSI 6 NISCOPE\_VAL\_PFI\_0 ('VAL\_PFI\_0') - PFI 0 NISCOPE\_VAL\_PFI\_1 ('VAL\_PFI\_1') - PFI 1 NISCOPE\_VAL\_PFI\_2 ('VAL\_PFI\_2') - PFI 2 NISCOPE\_VAL\_PXI\_STAR ('VAL\_PXI\_STAR') - PXI Star Trigger

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization:Start Trigger (Acq. Arm):Source
- C Attribute: NISCOPE\_ATTR\_ACQ\_ARM\_SOURCE

#### adv trig src

niscope.Session.adv\_trig\_src

Specifies the source the digitizer monitors for an advance trigger. When the advance trigger is received, the digitizer begins acquiring pretrigger samples.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization:Advance Trigger:Source
- C Attribute: NISCOPE\_ATTR\_ADV\_TRIG\_SRC

#### allow more records than memory

```
niscope.Session.allow_more_records_than_memory
```

Indicates whether more records can be configured with niscope. Session. configure\_horizontal\_timing() than fit in the onboard memory. If this property

is set to True, it is necessary to fetch records while the acquisition is in progress. Eventually, some of the records will be overwritten. An error is returned from the fetch method if you attempt to fetch a record that has been overwritten.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Horizontal:Enable Records > Memory
- C Attribute: NISCOPE\_ATTR\_ALLOW\_MORE\_RECORDS\_THAN\_MEMORY

## arm ref trig src

niscope.Session.arm\_ref\_trig\_src

Specifies the source the digitizer monitors for an arm reference trigger. When the arm reference trigger is received, the digitizer begins looking for a reference (stop) trigger from the user-configured trigger source.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization: Arm Reference Trigger: Source
- C Attribute: NISCOPE\_ATTR\_ARM\_REF\_TRIG\_SRC

## backlog

niscope.Session.backlog

Returns the number of samples (niscope.Session.points\_done) that have been acquired but not fetched for the record specified by niscope.Session.fetch\_record\_number.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

LabVIEW Property: Fetch:Fetch Backlog
 C Attribute: NISCOPE\_ATTR\_BACKLOG

# bandpass\_filter\_enabled

#### niscope.Session.bandpass\_filter\_enabled

Enables the bandpass filter on the specificed channel. The default value is FALSE.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].bandpass_filter_enabled = var
var = session.channels[0,1].bandpass_filter_enabled
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Vertical:Advanced:Bandpass Filter Enabled
- C Attribute: NISCOPE\_ATTR\_BANDPASS\_FILTER\_ENABLED

#### binary\_sample\_width

#### niscope.Session.binary\_sample\_width

Indicates the bit width of the binary data in the acquired waveform. Useful for determining which Binary Fetch method to use. Compare to niscope. Session.resolution. To configure the device to store samples with a lower resolution that the native, set this property to the desired binary width. This can be useful for streaming at faster speeds at the cost of resolution. The least significant bits will be lost with this configuration. Valid Values: 8, 16, 32

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Acquisition:Binary Sample Width
- C Attribute: NISCOPE\_ATTR\_BINARY\_SAMPLE\_WIDTH

#### channel count

niscope.Session.channel\_count

Indicates the number of channels that the specific instrument driver supports. For channel-based properties, the IVI engine maintains a separate cache value for each channel.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Capabilities:Channel Count
- C Attribute: NISCOPE\_ATTR\_CHANNEL\_COUNT

# channel\_enabled

niscope.Session.channel\_enabled

Specifies whether the digitizer acquires a waveform for the channel. Valid Values: True (1) - Acquire data on this channel False (0) - Don't acquire data on this channel

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].channel_enabled = var
var = session.channels[0,1].channel_enabled
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Vertical:Channel Enabled

• C Attribute: NISCOPE\_ATTR\_CHANNEL\_ENABLED

# channel\_terminal\_configuration

niscope.Session.channel\_terminal\_configuration

Specifies the terminal configuration for the channel.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].channel_terminal_configuration = var
var = session.channels[0,1].channel_terminal_configuration
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TerminalConfiguration
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Vertical: Channel Terminal Configuration
- C Attribute: NISCOPE\_ATTR\_CHANNEL\_TERMINAL\_CONFIGURATION

# data\_transfer\_block\_size

```
niscope.Session.data_transfer_block_size
```

Specifies the maximum number of samples to transfer at one time from the device to host memory. Increasing this number should result in better fetching performance because the driver does not need

to restart the transfers as often. However, increasing this number may also increase the amount of page-locked memory required from the system.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Fetch:Data Transfer Block Size
- C Attribute: NISCOPE\_ATTR\_DATA\_TRANSFER\_BLOCK\_SIZE

#### data transfer maximum bandwidth

#### niscope.Session.data\_transfer\_maximum\_bandwidth

This property specifies the maximum bandwidth that the device is allowed to consume.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Fetch:Advanced:Maximum Bandwidth
- C Attribute: NISCOPE\_ATTR\_DATA\_TRANSFER\_MAXIMUM\_BANDWIDTH

#### data transfer preferred packet size

# niscope.Session.data\_transfer\_preferred\_packet\_size

This property specifies the size of (read requestlmemory write) data payload. Due to alignment of the data buffers, the hardware may not always generate a packet of this size.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Fetch:Advanced:Preferred Packet Size

• C Attribute: NISCOPE\_ATTR\_DATA\_TRANSFER\_PREFERRED\_PACKET\_SIZE

#### device temperature

## niscope.Session.device\_temperature

Returns the temperature of the device in degrees Celsius from the onboard sensor.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Device:Temperature

• C Attribute: NISCOPE\_ATTR\_DEVICE\_TEMPERATURE

# driver\_setup

#### niscope.Session.driver\_setup

This property indicates the Driver Setup string that the user specified when initializing the driver. Some cases exist where the end-user must specify instrument driver options at initialization. An example of this is specifying a particular instrument model from among a family of instruments that the driver supports. This is useful when using simulation. The end-user can specify driver-specific options through the DriverSetup keyword in the optionsString parameter in niscope.Session. \_\_init\_\_\_(), or through the IVI Configuration Utility. If the user does not specify a Driver Setup string, this property returns an empty string.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	0
Resettable	0

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NISCOPE\_ATTR\_DRIVER\_SETUP

#### enable dc restore

```
niscope.Session.enable_dc_restore
```

Restores the video-triggered data retrieved by the digitizer to the video signal's zero reference point. Valid Values: True - Enable DC restore False - Disable DC restore

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Video:Enable DC Restore
- C Attribute: NISCOPE\_ATTR\_ENABLE\_DC\_RESTORE

## enable\_time\_interleaved\_sampling

```
niscope. Session. enable time interleaved sampling
```

Specifies whether the digitizer acquires the waveform using multiple ADCs for the channel enabling a higher maximum real-time sampling rate. Valid Values: True (1) - Use multiple interleaved ADCs on this channel False (0) - Use only this channel's ADC to acquire data for this channel

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].enable_time_interleaved_sampling = var
var = session.channels[0,1].enable_time_interleaved_sampling
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Horizontal:Enable Time Interleaved Sampling
- C Attribute: NISCOPE\_ATTR\_ENABLE\_TIME\_INTERLEAVED\_SAMPLING

## end of acquisition event output terminal

## niscope.Session.end\_of\_acquisition\_event\_output\_terminal

Specifies the destination for the End of Acquisition Event. When this event is asserted, the digitizer has completed sampling for all records. Consult your device documentation for a specific list of valid destinations.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization:End of Acquisition:Output Terminal
- C Attribute: NISCOPE\_ATTR\_END\_OF\_ACQUISITION\_EVENT\_OUTPUT\_TERMINAL

## end\_of\_record\_event\_output\_terminal

#### niscope.Session.end\_of\_record\_event\_output\_terminal

Specifies the destination for the End of Record Event. When this event is asserted, the digitizer has completed sampling for the current record. Consult your device documentation for a specific list of valid destinations.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

Tip: This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization:End of Record:Output Terminal
- C Attribute: NISCOPE\_ATTR\_END\_OF\_RECORD\_EVENT\_OUTPUT\_TERMINAL

# end\_of\_record\_to\_advance\_trigger\_holdoff

# niscope.Session.end\_of\_record\_to\_advance\_trigger\_holdoff

End of Record to Advance Trigger Holdoff is the length of time (in seconds) that a device waits between the completion of one record and the acquisition of pre-trigger samples for the next record. During this time, the acquisition engine state delays the transition to the Wait for Advance Trigger state, and will not store samples in onboard memory, accept an Advance Trigger, or trigger on the input signal.. **Supported Devices**: NI 5185/5186

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:End of Record to Advance Trigger Holdoff
- C Attribute: NISCOPE\_ATTR\_END\_OF\_RECORD\_TO\_ADVANCE\_TRIGGER\_HOLDOFF

#### equalization filter enabled

#### niscope.Session.equalization\_filter\_enabled

Enables the onboard signal processing FIR block. This block is connected directly to the input signal. This filter is designed to compensate the input signal for artifacts introduced to the signal outside of the digitizer. However, since this is a generic FIR filter any coefficients are valid. Coefficients should be between +1 and -1 in value.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].equalization_filter_enabled = var
var = session.channels[0,1].equalization_filter_enabled
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Onboard Signal Processing: Equalization: Equalization Filter Enabled
- C Attribute: NISCOPE\_ATTR\_EQUALIZATION\_FILTER\_ENABLED

## equalization num coefficients

#### niscope.Session.equalization\_num\_coefficients

Returns the number of coefficients that the FIR filter can accept. This filter is designed to compensate the input signal for artifacts introduced to the signal outside of the digitizer. However, since this is a generic FIR filter any coefficients are valid. Coefficients should be between +1 and -1 in value.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].equalization_num_coefficients
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Onboard Signal Processing: Equalization: Equalization Num Coefficients
- C Attribute: NISCOPE ATTR EQUALIZATION NUM COEFFICIENTS

# exported\_advance\_trigger\_output\_terminal

# niscope.Session.exported\_advance\_trigger\_output\_terminal

Specifies the destination to export the advance trigger. When the advance trigger is received, the digitizer begins acquiring samples for the Nth record. Consult your device documentation for a specific list of valid destinations.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization:Advance Trigger:Output Terminal
- C Attribute: NISCOPE\_ATTR\_EXPORTED\_ADVANCE\_TRIGGER\_OUTPUT\_TERMINAL

## exported ref trigger output terminal

#### niscope.Session.exported\_ref\_trigger\_output\_terminal

Specifies the destination export for the reference (stop) trigger. Consult your device documentation for a specific list of valid destinations.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Output Terminal
- C Attribute: NISCOPE\_ATTR\_EXPORTED\_REF\_TRIGGER\_OUTPUT\_TERMINAL

## exported\_start\_trigger\_output\_terminal

#### niscope. Session. exported start trigger output terminal

Specifies the destination to export the Start trigger. When the start trigger is received, the digitizer begins acquiring samples. Consult your device documentation for a specific list of valid destinations.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization:Start Trigger (Acq. Arm):Output Terminal
- C Attribute: NISCOPE\_ATTR\_EXPORTED\_START\_TRIGGER\_OUTPUT\_TERMINAL

# flex\_fir\_antialias\_filter\_type

## niscope.Session.flex\_fir\_antialias\_filter\_type

The NI 5922 flexible-resolution digitizer uses an onboard FIR lowpass antialias filter. Use this property to select from several types of filters to achieve desired filtering characteristics.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify

a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].flex_fir_antialias_filter_type = var
var = session.channels[0,1].flex_fir_antialias_filter_type
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.FlexFIRAntialiasFilterType
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Vertical:Advanced:Flex FIR Antialias Filter Type
- C Attribute: NISCOPE\_ATTR\_FLEX\_FIR\_ANTIALIAS\_FILTER\_TYPE

## fpga bitfile path

niscope.Session.fpga\_bitfile\_path

Gets the absolute file path to the bitfile loaded on the FPGA.

**Note:** Gets the absolute file path to the bitfile loaded on the FPGA.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Device:FPGA Bitfile Path
- C Attribute: NISCOPE\_ATTR\_FPGA\_BITFILE\_PATH

# high\_pass\_filter\_frequency

niscope.Session.high\_pass\_filter\_frequency

Specifies the frequency for the highpass filter in Hz. The device uses one of the valid values listed below. If an invalid value is specified, no coercion occurs. The default value is 0. (**PXIe-5164**) **Valid Values:** 0 90 450 **Related topics:** Digital Filtering

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Vertical:Advanced:High Pass Filter Frequency
- C Attribute: NISCOPE\_ATTR\_HIGH\_PASS\_FILTER\_FREQUENCY

### horz enforce realtime

# niscope.Session.horz\_enforce\_realtime

Indicates whether the digitizer enforces real-time measurements or allows equivalent-time measurements.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Horizontal:Enforce Realtime
- C Attribute: NISCOPE\_ATTR\_HORZ\_ENFORCE\_REALTIME

# horz\_min\_num\_pts

#### niscope.Session.horz\_min\_num\_pts

Specifies the minimum number of points you require in the waveform record for each channel. NI-SCOPE uses the value you specify to configure the record length that the digitizer uses for waveform acquisition. niscope.Session.horz\_record\_length returns the actual record length. Valid Values: 1 - available onboard memory

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Horizontal:Min Number of Points

• C Attribute: NISCOPE\_ATTR\_HORZ\_MIN\_NUM\_PTS

#### horz num records

# niscope.Session.horz\_num\_records

Specifies the number of records to acquire. Can be used for multi-record acquisition and single-record acquisitions. Setting this to 1 indicates a single-record acquisition.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Horizontal:Number of Records

• C Attribute: NISCOPE\_ATTR\_HORZ\_NUM\_RECORDS

# horz\_record\_length

## niscope.Session.horz\_record\_length

Returns the actual number of points the digitizer acquires for each channel. The value is equal to or greater than the minimum number of points you specify with <code>niscope.Session.horz\_min\_num\_pts</code>. Allocate a ViReal64 array of this size or greater to pass as the WaveformArray parameter of the Read and Fetch methods. This property is only valid after a call to the one of the Configure Horizontal methods.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Horizontal:Actual Record Length

• C Attribute: NISCOPE\_ATTR\_HORZ\_RECORD\_LENGTH

# horz\_record\_ref\_position

#### niscope.Session.horz\_record\_ref\_position

Specifies the position of the Reference Event in the waveform record. When the digitizer detects a trigger, it waits the length of time the <code>niscope.Session.trigger\_delay\_time</code> property specifies. The event that occurs when the delay time elapses is the Reference Event. The Reference Event is relative to the start of the record and is a percentage of the record length. For example, the value 50.0 corresponds to the center of the waveform record and 0.0 corresponds to the first element in the waveform record. Valid Values: 0.0 - 100.0

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Horizontal:Reference Position
- C Attribute: NISCOPE\_ATTR\_HORZ\_RECORD\_REF\_POSITION

#### horz sample rate

## niscope.Session.horz\_sample\_rate

Returns the effective sample rate using the current configuration. The units are samples per second. This property is only valid after a call to the one of the Configure Horizontal methods. Units: Hertz (Samples / Second)

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Horizontal:Actual Sample Rate
- C Attribute: NISCOPE\_ATTR\_HORZ\_SAMPLE\_RATE

#### horz time per record

#### niscope. Session.horz time per record

Specifies the length of time that corresponds to the record length. Units: Seconds

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Horizontal:Advanced:Time Per Record
- C Attribute: NISCOPE\_ATTR\_HORZ\_TIME\_PER\_RECORD

## input\_clock\_source

niscope.Session.input\_clock\_source

Specifies the input source for the PLL reference clock (the 1 MHz to 20 MHz clock on the NI 5122, the 10 MHz clock for the NI 5112/5620/5621/5911) to which the digitizer will be phase-locked; for the NI 5102, this is the source of the board clock.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocking:Reference (Input) Clock Source
- C Attribute: NISCOPE\_ATTR\_INPUT\_CLOCK\_SOURCE

#### input impedance

niscope.Session.input\_impedance

Specifies the input impedance for the channel in Ohms.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].input_impedance = var
var = session.channels[0,1].input_impedance
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Vertical:Input Impedance
- C Attribute: NISCOPE\_ATTR\_INPUT\_IMPEDANCE

# instrument\_firmware\_revision

niscope.Session.instrument\_firmware\_revision

A string that contains the firmware revision information for the instrument you are currently using.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Firmware Revision
- C Attribute: NISCOPE\_ATTR\_INSTRUMENT\_FIRMWARE\_REVISION

# instrument\_manufacturer

niscope.Session.instrument\_manufacturer

A string that contains the name of the instrument manufacturer.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Manufacturer
- C Attribute: NISCOPE\_ATTR\_INSTRUMENT\_MANUFACTURER

#### instrument model

niscope.Session.instrument\_model

A string that contains the model number of the current instrument.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Model
- C Attribute: NISCOPE\_ATTR\_INSTRUMENT\_MODEL

#### interleaving offset correction enabled

niscope.Session.interleaving\_offset\_correction\_enabled

Enables the interleaving offset correction on the specified channel. The default value is TRUE. **Related topics:** Timed Interleaved Sampling

**Note:** If disabled, warranted specifications are not guaranteed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Vertical:Advanced:Interleaving Offset Correction Enabled
- C Attribute: NISCOPE\_ATTR\_INTERLEAVING\_OFFSET\_CORRECTION\_ENABLED

## io resource descriptor

#### niscope.Session.io\_resource\_descriptor

Indicates the resource descriptor the driver uses to identify the physical device. If you initialize the driver with a logical name, this property contains the resource descriptor that corresponds to the entry in the IVI Configuration utility. If you initialize the instrument driver with the resource descriptor, this property contains that value. You can pass a logical name to niscope. Session. Init() or niscope. Session. \_\_init\_\_(). The IVI Configuration utility must contain an entry for the logical name. The logical name entry refers to a virtual instrument section in the IVI Configuration file. The virtual instrument section specifies a physical device and initial user options.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Advanced Session Information:Resource Descriptor
- C Attribute: NISCOPE\_ATTR\_IO\_RESOURCE\_DESCRIPTOR

#### logical name

#### niscope.Session.logical\_name

A string containing the logical name you specified when opening the current IVI session. You can pass a logical name to niscope.Session.Init() or niscope.Session.\_\_init\_\_(). The IVI Configuration utility must contain an entry for the logical name. The logical name entry refers to a virtual instrument section in the IVI Configuration file. The virtual instrument section specifies a physical device and initial user options.

Note: One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Advanced Session Information: Logical Name
- C Attribute: NISCOPE\_ATTR\_LOGICAL\_NAME

#### master enable

```
niscope. Session. master enable
```

Specifies whether you want the device to be a master or a slave. The master typically originates the trigger signal and clock sync pulse. For a standalone device, set this property to False.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization:Master Enable
- C Attribute: NISCOPE\_ATTR\_MASTER\_ENABLE

#### max\_input\_frequency

```
niscope.Session.max_input_frequency
```

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].max_input_frequency = var
var = session.channels[0,1].max_input_frequency
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

- LabVIEW Property: Vertical:Maximum Input Frequency
- C Attribute: NISCOPE\_ATTR\_MAX\_INPUT\_FREQUENCY

### max\_real\_time\_sampling\_rate

niscope.Session.max\_real\_time\_sampling\_rate

Returns the maximum real time sample rate in Hz.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Horizontal:Maximum Real Time Sample Rate
- C Attribute: NISCOPE\_ATTR\_MAX\_REAL\_TIME\_SAMPLING\_RATE

#### max ris rate

niscope.Session.max\_ris\_rate

Returns the maximum sample rate in RIS mode in Hz.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Horizontal:Maximum RIS Rate
- C Attribute: NISCOPE\_ATTR\_MAX\_RIS\_RATE

#### min sample rate

niscope.Session.min\_sample\_rate

Specify the sampling rate for the acquisition in Samples per second. Valid Values: The combination of sampling rate and min record length must allow the digitizer to sample at a valid sampling rate for the acquisition type specified in <code>niscope.Session.ConfigureAcquisition()</code> and not require more memory than the onboard memory module allows.

Note: One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	Yes

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Horizontal:Min Sample Rate

• C Attribute: NISCOPE\_ATTR\_MIN\_SAMPLE\_RATE

## onboard\_memory\_size

niscope.Session.onboard\_memory\_size

Returns the total combined amount of onboard memory for all channels in bytes.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Horizontal:Memory Size

• C Attribute: NISCOPE\_ATTR\_ONBOARD\_MEMORY\_SIZE

### output clock source

niscope.Session.output\_clock\_source

Specifies the output source for the 10 MHz clock to which another digitizer's sample clock can be phased-locked.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocking:Output Clock Source
- C Attribute: NISCOPE\_ATTR\_OUTPUT\_CLOCK\_SOURCE

### pll\_lock\_status

niscope. Session.pll lock status

If TRUE, the PLL has remained locked to the external reference clock since it was last checked. If FALSE, the PLL has become unlocked from the external reference clock since it was last checked.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocking:PLL Lock Status
- C Attribute: NISCOPE\_ATTR\_PLL\_LOCK\_STATUS

## points\_done

niscope.Session.points\_done

Actual number of samples acquired in the record specified by niscope.Session.fetch\_record\_number from the niscope.Session.fetch\_relative\_to and niscope.Session.fetch\_offset properties.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

• LabVIEW Property: Fetch:Points Done

• C Attribute: NISCOPE\_ATTR\_POINTS\_DONE

### poll interval

niscope.Session.poll\_interval

Specifies the poll interval in milliseconds to use during RIS acquisitions to check whether the acquisition is complete.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	0
Resettable	0

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NISCOPE\_ATTR\_POLL\_INTERVAL

## probe\_attenuation

niscope.Session.probe\_attenuation

Specifies the probe attenuation for the input channel. For example, for a 10:1 probe, set this property to 10.0. Valid Values: Any positive real number. Typical values are 1, 10, and 100.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].probe_attenuation = var
var = session.channels[0,1].probe_attenuation
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

- LabVIEW Property: Vertical:Probe Attenuation
- C Attribute: NISCOPE\_ATTR\_PROBE\_ATTENUATION

### ready for advance event output terminal

### niscope.Session.ready\_for\_advance\_event\_output\_terminal

Specifies the destination for the Ready for Advance Event. When this event is asserted, the digitizer is ready to receive an advance trigger. Consult your device documentation for a specific list of valid destinations.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization:Ready for Advance:Output Terminal
- C Attribute: NISCOPE\_ATTR\_READY\_FOR\_ADVANCE\_EVENT\_OUTPUT\_TERMINAL

## ready\_for\_ref\_event\_output\_terminal

# $\verb|niscope.Session.ready_for\_ref_event_output\_terminal|\\$

Specifies the destination for the Ready for Reference Event. When this event is asserted, the digitizer is ready to receive a reference trigger. Consult your device documentation for a specific list of valid destinations.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Synchronization:Ready for Reference:Output Terminal
- C Attribute: NISCOPE\_ATTR\_READY\_FOR\_REF\_EVENT\_OUTPUT\_TERMINAL

## ready\_for\_start\_event\_output\_terminal

## niscope.Session.ready\_for\_start\_event\_output\_terminal

Specifies the destination for the Ready for Start Event. When this event is asserted, the digitizer is ready to receive a start trigger. Consult your device documentation for a specific list of valid destinations.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization:Ready for Start:Output Terminal
- C Attribute: NISCOPE\_ATTR\_READY\_FOR\_START\_EVENT\_OUTPUT\_TERMINAL

# records\_done

niscope.Session.records\_done

Specifies the number of records that have been completely acquired.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Fetch:Records Done
- C Attribute: NISCOPE\_ATTR\_RECORDS\_DONE

#### record arm source

niscope.Session.record\_arm\_source

Specifies the record arm source.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Synchronization:Record Arm Source
- C Attribute: NISCOPE\_ATTR\_RECORD\_ARM\_SOURCE

### ref clk rate

niscope.Session.ref\_clk\_rate

If niscope. Session.input\_clock\_source is an external source, this property specifies the frequency of the input, or reference clock, to which the internal sample clock timebase is synchronized. The frequency is in hertz.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocking:Reference Clock Rate
- C Attribute: NISCOPE\_ATTR\_REF\_CLK\_RATE

## ref\_trigger\_detector\_location

niscope.Session.ref\_trigger\_detector\_location

Indicates which analog compare circuitry to use on the device.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.RefTriggerDetectorLocation
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Triggering:Onboard Signal Processing:Ref Trigger Detection Location
- C Attribute: NISCOPE\_ATTR\_REF\_TRIGGER\_DETECTOR\_LOCATION

## ref\_trigger\_minimum\_quiet\_time

#### niscope.Session.ref\_trigger\_minimum\_quiet\_time

The amount of time the trigger circuit must not detect a signal above the trigger level before the trigger is armed. This property is useful for triggering at the beginning and not in the middle of signal bursts.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Onboard Signal Processing:Ref Trigger Min Quiet Time
- C Attribute: NISCOPE\_ATTR\_REF\_TRIGGER\_MINIMUM\_QUIET\_TIME

## ref\_trig\_tdc\_enable

#### niscope.Session.ref\_trig\_tdc\_enable

This property controls whether the TDC is used to compute an accurate trigger.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Horizontal:Advanced:Enable TDC
- C Attribute: NISCOPE\_ATTR\_REF\_TRIG\_TDC\_ENABLE

#### resolution

niscope.Session.resolution

Indicates the bit width of valid data (as opposed to padding bits) in the acquired waveform. Compare to <code>niscope.Session.binary\_sample\_width</code>.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Acquisition: Resolution
- C Attribute: NISCOPE\_ATTR\_RESOLUTION

#### ris\_in\_auto\_setup\_enable

niscope.Session.ris\_in\_auto\_setup\_enable

Indicates whether the digitizer should use RIS sample rates when searching for a frequency in autosetup. Valid Values: True (1) - Use RIS sample rates in autosetup False (0) - Do not use RIS sample rates in autosetup

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Acquisition: Advanced: Enable RIS in Auto Setup
- C Attribute: NISCOPE\_ATTR\_RIS\_IN\_AUTO\_SETUP\_ENABLE

### ris method

#### niscope.Session.ris\_method

Specifies the algorithm for random-interleaved sampling, which is used if the sample rate exceeds the value of <code>niscope.Session.max\_real\_time\_sampling\_rate</code>.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.RISMethod
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Horizontal:RIS Method

• C Attribute: NISCOPE\_ATTR\_RIS\_METHOD

#### ris num averages

#### niscope. Session. ris num averages

The number of averages for each bin in an RIS acquisition. The number of averages times the oversampling factor is the minimum number of real-time acquisitions necessary to reconstruct the RIS waveform. Averaging is useful in RIS because the trigger times are not evenly spaced, so adjacent points in the reconstructed waveform not be accurately spaced. By averaging, the errors in both time and voltage are smoothed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Horizontal:RIS Num Avg

• C Attribute: NISCOPE\_ATTR\_RIS\_NUM\_AVERAGES

#### sample clock timebase multiplier

# $\verb|niscope.Session.sample_clock_timebase_multiplier|\\$

If Sample Clock Timebase Source <p:py:meth: 'niscope.Session.SampleClockTimebaseSource.html>'\_\_ is an external source, this property specifies the ratio between the Sample Clock Timebase Rate <p:py:meth: 'niscope.Session.SampleClockTimebaseRate.html>'\_\_ and the actual sample rate,

which can be higher. This property can be used in conjunction with the *Sample Clock Timebase Divisor Property <p:py:meth: 'niscope.Session.SampleClockTimebaseDivisor.*html>'\_\_. Some devices use multiple ADCs to sample the same channel at an effective sample rate that is greater than the specified clock rate. When providing an external sample clock use this property to indicate when you want a higher sample rate. Valid values for this property vary by device and current configuration. **Related topics:** Sample Clock

**Note:** One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocking:Sample Clock Timebase Multiplier
- C Attribute: NISCOPE\_ATTR\_SAMPLE\_CLOCK\_TIMEBASE\_MULTIPLIER

### sample mode

niscope.Session.sample\_mode

Indicates the sample mode the digitizer is currently using.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Acquisition:Sample Mode
- C Attribute: NISCOPE\_ATTR\_SAMPLE\_MODE

#### samp clk timebase div

niscope.Session.samp\_clk\_timebase\_div

If niscope. Session. samp\_clk\_timebase\_src is an external source, specifies the ratio between the sample clock timebase rate and the actual sample rate, which can be slower.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Clocking:Sample Clock Timebase Divisor
- C Attribute: NISCOPE\_ATTR\_SAMP\_CLK\_TIMEBASE\_DIV

#### samp clk timebase rate

niscope.Session.samp\_clk\_timebase\_rate

If niscope.Session.samp\_clk\_timebase\_src is an external source, specifies the frequency in hertz of the external clock used as the timebase source.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Clocking:Sample Clock Timebase Rate
- C Attribute: NISCOPE\_ATTR\_SAMP\_CLK\_TIMEBASE\_RATE

# samp\_clk\_timebase\_src

niscope.Session.samp\_clk\_timebase\_src

Specifies the source of the sample clock timebase, which is the timebase used to control waveform sampling. The actual sample rate may be the timebase itself or a divided version of the timebase, depending on the <code>niscope.Session.min\_sample\_rate</code> (for internal sources) or the <code>niscope.Session.samp\_clk\_timebase\_div</code> (for external sources).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

7.4. niscope module

- LabVIEW Property: Clocking:Sample Clock Timebase Source
- C Attribute: NISCOPE\_ATTR\_SAMP\_CLK\_TIMEBASE\_SRC

## serial number

niscope.Session.serial\_number

Returns the serial number of the device.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Device:Serial Number
- C Attribute: NISCOPE\_ATTR\_SERIAL\_NUMBER

# simulate

niscope. Session. simulate

Specifies whether or not to simulate instrument driver I/O operations. If simulation is enabled, instrument driver methods perform range checking and call Ivi\_GetAttribute and Ivi\_SetAttribute methods, but they do not perform instrument I/O. For output parameters that represent instrument data, the instrument driver methods return calculated values. The default value is False. Use the niscope.Session.\_\_init\_\_() method to override this value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:User Options:Simulate
- C Attribute: NISCOPE\_ATTR\_SIMULATE

### specific driver description

niscope.Session.specific\_driver\_description

A string that contains a brief description of the specific driver

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Identification:Description
- C Attribute: NISCOPE\_ATTR\_SPECIFIC\_DRIVER\_DESCRIPTION

#### specific driver revision

niscope.Session.specific\_driver\_revision

A string that contains additional version information about this instrument driver.

The following table lists the characteristics of this property.

Value
str
read only
False
No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Identification:Revision
- C Attribute: NISCOPE\_ATTR\_SPECIFIC\_DRIVER\_REVISION

### specific driver vendor

niscope.Session.specific\_driver\_vendor

A string that contains the name of the vendor that supplies this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

- LabVIEW Property: Inherent IVI Attributes:Driver Identification:Driver Vendor
- C Attribute: NISCOPE ATTR SPECIFIC DRIVER VENDOR

# start\_to\_ref\_trigger\_holdoff

```
niscope.Session.start_to_ref_trigger_holdoff
```

Pass the length of time you want the digitizer to wait after it starts acquiring data until the digitizer enables the trigger system to detect a reference (stop) trigger. Units: Seconds Valid Values: 0.0 - 171.8

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].start_to_ref_trigger_holdoff = var
var = session.channels[0,1].start_to_ref_trigger_holdoff
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Start To Ref Trigger Holdoff
- C Attribute: NISCOPE\_ATTR\_START\_TO\_REF\_TRIGGER\_HOLDOFF

### supported instrument models

niscope.Session.supported\_instrument\_models

A string that contains a comma-separated list of the instrument model numbers supported by this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

- LabVIEW Property: Inherent IVI Attributes:Driver Capabilities:Supported Instrument Models
- C Attribute: NISCOPE ATTR SUPPORTED INSTRUMENT MODELS

### trigger auto triggered

#### niscope.Session.trigger\_auto\_triggered

Specifies if the last acquisition was auto triggered. You can use the Auto Triggered property to find out if the last acquisition was triggered.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Auto Triggered
- C Attribute: NISCOPE\_ATTR\_TRIGGER\_AUTO\_TRIGGERED

## trigger coupling

## niscope.Session.trigger\_coupling

Specifies how the digitizer couples the trigger source. This property affects instrument operation only when <code>niscope.Session.trigger\_type</code> is set to <code>EDGE</code>, <code>HYSTERESIS</code>, or <code>WINDOW</code>.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerCoupling
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Coupling
- C Attribute: NISCOPE\_ATTR\_TRIGGER\_COUPLING

## trigger\_delay\_time

#### niscope.Session.trigger\_delay\_time

Specifies the trigger delay time in seconds. The trigger delay time is the length of time the digitizer waits after it receives the trigger. The event that occurs when the trigger delay elapses is the Reference Event. Valid Values: 0.0 - 171.8

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Delay
- C Attribute: NISCOPE\_ATTR\_TRIGGER\_DELAY\_TIME

## trigger holdoff

## niscope.Session.trigger\_holdoff

Specifies the length of time (in seconds) the digitizer waits after detecting a trigger before enabling the trigger subsystem to detect another trigger. This property affects instrument operation only when the digitizer requires multiple acquisitions to build a complete waveform. The digitizer requires multiple waveform acquisitions when it uses equivalent-time sampling or when the digitizer is configured for a multi-record acquisition through a call to <code>niscope.Session.configure\_horizontal\_timing()</code>. Valid Values: 0.0 - 171.8

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: **Triggering:Trigger Holdoff**
- C Attribute: NISCOPE\_ATTR\_TRIGGER\_HOLDOFF

#### trigger hysteresis

#### niscope. Session. trigger hysteresis

Specifies the size of the hysteresis window on either side of the trigger level. The digitizer triggers

when the trigger signal passes through the threshold you specify with the Trigger Level parameter, has the slope you specify with the Trigger Slope parameter, and passes through the hysteresis window that you specify with this parameter.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Hysteresis
- C Attribute: NISCOPE ATTR TRIGGER HYSTERESIS

### trigger impedance

niscope.Session.trigger\_impedance

Specifies the input impedance for the external analog trigger channel in Ohms. Valid Values: 50 - 50 ohms 1000000 - 1 mega ohm

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Impedance
- C Attribute: NISCOPE\_ATTR\_TRIGGER\_IMPEDANCE

## trigger\_level

niscope.Session.trigger\_level

Specifies the voltage threshold for the trigger subsystem. The units are volts. This property affects instrument behavior only when the <code>niscope.Session.trigger\_type</code> is set to <code>EDGE</code>, <code>HYSTERESIS</code>, or <code>WINDOW</code>. Valid Values: The values of the range and offset parameters in <code>niscope.Session.configure\_vertical()</code> determine the valid range for the trigger level on the channel you use as the Trigger Source. The value you pass for this parameter must meet the following conditions:

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

• LabVIEW Property: Triggering:Trigger Level

• C Attribute: NISCOPE\_ATTR\_TRIGGER\_LEVEL

## trigger\_modifier

#### niscope.Session.trigger\_modifier

Configures the device to automatically complete an acquisition if a trigger has not been received. Valid Values: None (1) - Normal triggering Auto Trigger (2) - Auto trigger acquisition if no trigger arrives

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerModifier
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Triggering:Trigger Modifier

• C Attribute: NISCOPE\_ATTR\_TRIGGER\_MODIFIER

## trigger\_slope

niscope.Session.trigger\_slope

Specifies if a rising or a falling edge triggers the digitizer. This property affects instrument operation only when <code>niscope.Session.trigger\_type</code> is set to <code>EDGE</code>, <code>HYSTERESIS</code>, or <code>WINDOW</code>.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerSlope
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Triggering:Trigger Slope

• C Attribute: NISCOPE\_ATTR\_TRIGGER\_SLOPE

## trigger\_source

niscope.Session.trigger\_source

Specifies the source the digitizer monitors for the trigger event.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Triggering:Trigger Source

• C Attribute: NISCOPE\_ATTR\_TRIGGER\_SOURCE

# trigger\_type

niscope.Session.trigger\_type

Specifies the type of trigger to use.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerType
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Triggering:Trigger Type

• C Attribute: NISCOPE\_ATTR\_TRIGGER\_TYPE

#### trigger window high level

niscope.Session.trigger\_window\_high\_level

Pass the upper voltage threshold you want the digitizer to use for window triggering.

The digitizer triggers when the trigger signal enters or leaves the window you specify with <code>niscope.Session.trigger\_window\_low\_level</code> and <code>niscope.Session.trigger\_window\_low\_level</code> and <code>niscope.Session.trigger\_window\_high\_level</code> Valid Values: The values of the Vertical Range and Vertical Offset parameters in <code>niscope.Session.configure\_vertical()</code> determine the valid range for the High Window Level on the channel you use as the Trigger Source parameter in <code>niscope.Session.ConfigureTriggerSource()</code>. The value you pass for this parameter must meet the following conditions. High Trigger Level <= Vertical Range/2 + Vertical Offset High Trigger Level > Low Trigger Level

Note: One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Window:High Level
- C Attribute: NISCOPE\_ATTR\_TRIGGER\_WINDOW\_HIGH\_LEVEL

## trigger\_window\_low\_level

#### niscope.Session.trigger\_window\_low\_level

Pass the lower voltage threshold you want the digitizer to use for window triggering. The digitizer triggers when the trigger signal enters or leaves the window you specify with <code>niscope.Session.trigger\_window\_low\_level</code> and <code>niscope.Session.trigger\_window\_low\_level</code> and <code>niscope.Session.trigger\_window\_high\_level</code>. Units: Volts Valid Values: The values of the Vertical Range and Vertical Offset parameters in <code>niscope.Session.configure\_vertical()</code> determine the valid range for the Low Window Level on the channel you use as the Trigger Source parameter in <code>niscope.Session.ConfigureTriggerSource()</code>. The value you pass for this parameter must meet the following conditions. Low Trigger Level <= Vertical Range/2 + Vertical Offset Low Trigger Level < High Trigger Level

**Note:** One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	False
Resettable	No

- LabVIEW Property: Triggering:Trigger Window:Low Level
- C Attribute: NISCOPE\_ATTR\_TRIGGER\_WINDOW\_LOW\_LEVEL

# trigger\_window\_mode

## niscope.Session.trigger\_window\_mode

Specifies whether you want a trigger to occur when the signal enters or leaves the window specified by niscope.Session.trigger\_window\_low\_level, or niscope.Session.trigger\_window\_high\_level.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerWindowMode
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Window:Window Mode
- C Attribute: NISCOPE\_ATTR\_TRIGGER\_WINDOW\_MODE

# tv\_trigger\_event

niscope.Session.tv\_trigger\_event

Specifies the condition in the video signal that causes the digitizer to trigger.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.VideoTriggerEvent
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: **Triggering:Trigger Video:Event**
- C Attribute: NISCOPE\_ATTR\_TV\_TRIGGER\_EVENT

### tv trigger line number

#### niscope.Session.tv\_trigger\_line\_number

Specifies the line on which to trigger, if *niscope*. *Session.tv\_trigger\_event* is set to line number. The valid ranges of the property depend on the signal format selected. M-NTSC has a valid range of 1 to 525. B/G-PAL, SECAM, 576i, and 576p have a valid range of 1 to 625. 720p has a valid range of 1 to 750. 1080i and 1080p have a valid range of 1125.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Video:Line Number
- C Attribute: NISCOPE\_ATTR\_TV\_TRIGGER\_LINE\_NUMBER

# tv\_trigger\_polarity

#### niscope.Session.tv\_trigger\_polarity

Specifies whether the video signal sync is positive or negative.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.VideoPolarity
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Video:Polarity
- $\bullet \ \ C \ Attribute: \\ \textbf{NISCOPE\_ATTR\_TV\_TRIGGER\_POLARITY}$

# tv\_trigger\_signal\_format

```
niscope.Session.tv_trigger_signal_format
```

Specifies the type of video signal, such as NTSC, PAL, or SECAM.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify

a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].tv_trigger_signal_format = var
var = session.channels[0,1].tv_trigger_signal_format
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.VideoSignalFormat
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Triggering:Trigger Video:Signal Format
- C Attribute: NISCOPE\_ATTR\_TV\_TRIGGER\_SIGNAL\_FORMAT

### vertical coupling

```
niscope.Session.vertical_coupling
```

Specifies how the digitizer couples the input signal for the channel. When input coupling changes, the input stage takes a finite amount of time to settle.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].vertical_coupling = var
var = session.channels[0,1].vertical_coupling
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.VerticalCoupling
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Vertical: Vertical Coupling
- C Attribute: NISCOPE\_ATTR\_VERTICAL\_COUPLING

### vertical offset

```
niscope.Session.vertical_offset
```

Specifies the location of the center of the range. The value is with respect to ground and is in volts. For example, to acquire a sine wave that spans between 0.0 and 10.0 V, set this property to 5.0 V.

**Note:** This property is not supported by all digitizers.Refer to the NI High-Speed Digitizers Help for a list of vertical offsets supported for each device.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].vertical_offset = var
var = session.channels[0,1].vertical_offset
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Vertical: Vertical Offset

• C Attribute: NISCOPE\_ATTR\_VERTICAL\_OFFSET

# vertical\_range

```
niscope.Session.vertical_range
```

Specifies the absolute value of the input range for a channel in volts. For example, to acquire a sine wave that spans between -5 and +5 V, set this property to 10.0 V. Refer to the NI High-Speed Digitizers Help for a list of supported vertical ranges for each device. If the specified range is not supported by a device, the value is coerced up to the next valid range.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niscope. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].vertical_range = var
var = session.channels[0,1].vertical_range
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Vertical: Vertical Range
- C Attribute: NISCOPE\_ATTR\_VERTICAL\_RANGE

#### **Methods**

#### abort

```
niscope.Session.abort()
```

Aborts an acquisition and returns the digitizer to the Idle state. Call this method if the digitizer times out waiting for a trigger.

## acquisition\_status

```
niscope.Session.acquisition_status()
```

Returns status information about the acquisition to the **status** output parameter.

Return type niscope. Acquisition Status

## Returns

Returns whether the acquisition is complete, in progress, or unknown.

#### **Defined Values**

COMPLETE

IN\_PROGRESS

STATUS\_UNKNOWN

#### auto setup

```
niscope.Session.auto_setup()
```

Automatically configures the instrument. When you call this method, the digitizer senses the input signal and automatically configures many of the instrument settings. If a signal is detected on a channel, the driver chooses the smallest available vertical range that is larger than the signal range. For example, if the signal is a  $1.2~V_{pk-pk}$  sine wave, and the device supports 1~V and 2~V vertical ranges, the driver will choose the 2~V vertical range for that channel.

If no signal is found on any analog input channel, a warning is returned, and all channels are enabled. A channel is considered to have a signal present if the signal is at least 10% of the smallest vertical range available for that channel.

The following settings are changed:

General	
Acquisition mode	Normal
Reference clock	Internal
Vertical	
Vertical coupling	AC (DC for NI 5621)
Vertical bandwidth	Full
Vertical range	Changed by auto setup
Vertical offset	0 V
Probe attenuation	Unchanged by auto setup
Input impedance	Unchanged by auto setup
Horizontal	
Sample rate	Changed by auto setup
Min record length	Changed by auto setup
Enforce realtime	True
Number of Records	Changed to 1
Triggering	
Trigger type	Edge if signal present, otherwise immediate
Trigger channel	Lowest numbered channel with a signal present
Trigger slope	Positive
Trigger coupling	DC
Reference position	50%
Trigger level	50% of signal on trigger channel
Trigger delay	0
Trigger holdoff	0
Trigger output	None

## commit

niscope.Session.commit()

Commits to hardware all the parameter settings associated with the task. Use this method if you want a parameter change to be immediately reflected in the hardware. This method is not supported for Traditional NI-DAQ (Legacy) devices.

## configure\_chan\_characteristics

niscope.Session.configure\_chan\_characteristics(input\_impedance,

max\_input\_frequency)

Configures the properties that control the electrical characteristics of the channel—the input impedance and the bandwidth.

**Tip:** This method requires repeated capabilities (channels). If called directly on the niscope. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].configure_chan_characteristics(input_impedance, \_ \hookrightarrow max_input_frequency)
```

#### **Parameters**

- input\_impedance (float) The input impedance for the channel; NI-SCOPE sets niscope.Session.input\_impedance to this value.
- max\_input\_frequency (float) The bandwidth for the channel; NI-SCOPE sets niscope. Session.max\_input\_frequency to this value. Pass 0 for this value to use the hardware default bandwidth. Pass -1 for this value to achieve full bandwidth.

### configure equalization filter coefficients

```
niscope.Session.configure_equalization_filter_coefficients (coefficients)
```

Configures the custom coefficients for the equalization FIR filter on the device. This filter is designed to compensate the input signal for artifacts introduced to the signal outside of the digitizer. Because this filter is a generic FIR filter, any coefficients are valid. Coefficient values should be between +1 and -1.

**Tip:** This method requires repeated capabilities (channels). If called directly on the niscope. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].configure_equalization_filter_

→coefficients(coefficients)
```

Parameters coefficients (list of float) The custom coefficients for the equalization FIR filter on the device. These coefficients should be between +1 and -1. You can obtain the number of coefficients :py:data: 'niscope.Session.equalization num coefficients <cvi:py:data:niscope.Session.equalization\_num\_coefficients.html>'\_\_ property. The :py:data: 'niscope.Session.equalization filter enabled <cvi:py:data:niscope.Session.equalization\_filter\_enabled.html>'\_\_\_ property must be set to TRUE to enable the filter.

### configure horizontal timing

```
niscope.Session.configure_horizontal_timing (min_sample_rate, min_num_pts, ref_position, num_records, enforce realtime)
```

Configures the common properties of the horizontal subsystem for a multirecord acquisition in terms of minimum sample rate.

#### **Parameters**

• min\_sample\_rate (float) - The sampling rate for the acquisition. Refer to niscope.Session.min\_sample\_rate for more information.

• min\_num\_pts (int) - The minimum number of points you need in the record for each channel; call niscope.Session.ActualRecordLength() to obtain the actual record length used.

Valid Values: Greater than 1; limited by available memory

**Note:** One or more of the referenced methods are not in the Python API for this driver.

- **ref\_position** (*float*) The position of the Reference Event in the waveform record specified as a percentage.
- num\_records (int) The number of records to acquire
- **enforce\_realtime** (bool) Indicates whether the digitizer enforces real-time measurements or allows equivalent-time (RIS) measurements; not all digitizers support RIS—refer to Features Supported by Device for more information.

Default value: True

### **Defined Values**

True—Allow real-time acquisitions only

False—Allow real-time and equivalent-time acquisitions

## configure\_trigger\_digital

```
niscope.Session.configure_trigger_digital(trigger_source,
```

slope=niscope.TriggerSlope.POSITIVE,

hold-

off = date time. time delta (seconds = 0.0),

*delay=datetime.timedelta(seconds=0.0))* 

Configures the common properties of a digital trigger.

When you initiate an acquisition, the digitizer waits for the start trigger, which is configured through the <code>niscope.Session.acq\_arm\_source</code> (Start Trigger Source) property. The default is immediate. Upon receiving the start trigger the digitizer begins sampling pretrigger points. After the digitizer finishes sampling pretrigger points, the digitizer waits for a reference (stop) trigger that you specify with a method such as this one. Upon receiving the reference trigger the digitizer finishes the acquisition after completing posttrigger sampling. With each Configure Trigger method, you specify configuration parameters such as the trigger source and the amount of trigger delay.

**Note:** For multirecord acquisitions, all records after the first record are started by using the Advance Trigger Source. The default is immediate.

You can adjust the amount of pre-trigger and post-trigger samples using the reference position parameter on the <code>niscope.Session.configure\_horizontal\_timing()</code> method. The default is half of the record length.

Some features are not supported by all digitizers. Refer to Features Supported by Device for more information.

Digital triggering is not supported in RIS mode.

#### **Parameters**

- **trigger\_source** (*str*) Specifies the trigger source. Refer to *niscope*. *Session.trigger\_source* for defined values.
- **slope** (niscope.TriggerSlope) Specifies whether you want a rising edge or a falling edge to trigger the digitizer. Refer to niscope.Session. trigger\_slope for more information.
- holdoff (float in seconds or datetime.timedelta) The length of time the digitizer waits after detecting a trigger before enabling NI-SCOPE to detect another trigger. Refer to niscope.Session.trigger\_holdoff for more information.
- **delay** (float in seconds or datetime.timedelta) How long the digitizer waits after receiving the trigger to start acquiring data. Refer to niscope. Session.trigger\_delay\_time for more information.

## configure\_trigger\_edge

niscope.Session.configure\_trigger\_edge (trigger\_source, level, trigger\_coupling, slope=niscope.TriggerSlope.POSITIVE, holdoff=datetime.timedelta(seconds=0.0), delay=datetime.timedelta(seconds=0.0))

Configures common properties for analog edge triggering.

When you initiate an acquisition, the digitizer waits for the start trigger, which is configured through the <code>niscope.Session.acq\_arm\_source</code> (Start Trigger Source) property. The default is immediate. Upon receiving the start trigger the digitizer begins sampling pretrigger points. After the digitizer finishes sampling pretrigger points, the digitizer waits for a reference (stop) trigger that you specify with a method such as this one. Upon receiving the reference trigger the digitizer finishes the acquisition after completing posttrigger sampling. With each Configure Trigger method, you specify configuration parameters such as the trigger source and the amount of trigger delay.

**Note:** Some features are not supported by all digitizers. Refer to Features Supported by Device for more information.

## **Parameters**

- **trigger\_source** (*str*) Specifies the trigger source. Refer to *niscope*. *Session.trigger\_source* for defined values.
- **level** (*float*) The voltage threshold for the trigger. Refer to *niscope*. Session.trigger\_level for more information.
- trigger\_coupling (niscope.TriggerCoupling) Applies coupling and filtering options to the trigger signal. Refer to niscope.Session. trigger\_coupling for more information.
- slope (niscope.TriggerSlope) Specifies whether you want a rising edge or a falling edge to trigger the digitizer. Refer to niscope.Session. trigger\_slope for more information.
- holdoff (float in seconds or datetime.timedelta) The length of time the digitizer waits after detecting a trigger before enabling NI-SCOPE to detect another trigger. Refer to niscope.Session.trigger\_holdoff for more information.

• **delay** (float in seconds or datetime.timedelta) - How long the digitizer waits after receiving the trigger to start acquiring data. Refer to niscope. Session.trigger\_delay\_time for more information.

#### configure\_trigger\_hysteresis

```
niscope. Session. configure_trigger_hysteresis (trigger_source, level, hysteresis, trigger_coupling, slope=niscope. TriggerSlope.POSITIVE, hold-
off=datetime.timedelta(seconds=0.0), de-lay=datetime.timedelta(seconds=0.0))
```

Configures common properties for analog hysteresis triggering. This kind of trigger specifies an additional value, specified in the **hysteresis** parameter, that a signal must pass through before a trigger can occur. This additional value acts as a kind of buffer zone that keeps noise from triggering an acquisition.

When you initiate an acquisition, the digitizer waits for the start trigger, which is configured through the <code>niscope.Session.acq\_arm\_source</code>. The default is immediate. Upon receiving the start trigger the digitizer begins sampling pretrigger points. After the digitizer finishes sampling pretrigger points, the digitizer waits for a reference (stop) trigger that you specify with a method such as this one. Upon receiving the reference trigger the digitizer finishes the acquisition after completing posttrigger sampling. With each Configure Trigger method, you specify configuration parameters such as the trigger source and the amount of trigger delay.

**Note:** Some features are not supported by all digitizers. Refer to Features Supported by Device for more information.

#### **Parameters**

- **trigger\_source** (*str*) Specifies the trigger source. Refer to *niscope*. Session.trigger\_source for defined values.
- **level** (*float*) The voltage threshold for the trigger. Refer to *niscope*. Session.trigger\_level for more information.
- hysteresis (float) The size of the hysteresis window on either side of the level in volts; the digitizer triggers when the trigger signal passes through the hysteresis value you specify with this parameter, has the slope you specify with slope, and passes through the level. Refer to niscope.Session. trigger\_hysteresis for defined values.
- trigger\_coupling (niscope.TriggerCoupling) Applies coupling and filtering options to the trigger signal. Refer to niscope.Session. trigger\_coupling for more information.
- **slope** (niscope.TriggerSlope) Specifies whether you want a rising edge or a falling edge to trigger the digitizer. Refer to niscope.Session. trigger\_slope for more information.
- holdoff (float in seconds or datetime.timedelta) The length of time the digitizer waits after detecting a trigger before enabling NI-SCOPE to detect another trigger. Refer to niscope.Session.trigger\_holdoff for more information.

• **delay** (float in seconds or datetime.timedelta) - How long the digitizer waits after receiving the trigger to start acquiring data. Refer to niscope. Session.trigger\_delay\_time for more information.

### configure\_trigger\_immediate

```
niscope.Session.configure_trigger_immediate()
```

Configures common properties for immediate triggering. Immediate triggering means the digitizer triggers itself.

When you initiate an acquisition, the digitizer waits for a trigger. You specify the type of trigger that the digitizer waits for with a Configure Trigger method, such as niscope. Session. configure\_trigger\_immediate().

## configure\_trigger\_software

```
niscope.Session.configure_trigger_software (holdoff=datetime.timedelta(seconds=0.0), de-lay=datetime.timedelta(seconds=0.0))
```

Configures common properties for software triggering.

When you initiate an acquisition, the digitizer waits for the start trigger, which is configured through the <code>niscope.Session.acq\_arm\_source</code> (Start Trigger Source) property. The default is immediate. Upon receiving the start trigger the digitizer begins sampling pretrigger points. After the digitizer finishes sampling pretrigger points, the digitizer waits for a reference (stop) trigger that you specify with a method such as this one. Upon receiving the reference trigger the digitizer finishes the acquisition after completing posttrigger sampling. With each Configure Trigger method, you specify configuration parameters such as the trigger source and the amount of trigger delay.

To trigger the acquisition, use niscope. Session.send\_software\_trigger\_edge().

**Note:** Some features are not supported by all digitizers. Refer to Features Supported by Device for more information.

### **Parameters**

- holdoff (float in seconds or datetime.timedelta) The length of time the digitizer waits after detecting a trigger before enabling NI-SCOPE to detect another trigger. Refer to niscope.Session.trigger\_holdoff for more information.
- delay (float in seconds or datetime.timedelta) How long the digitizer waits after receiving the trigger to start acquiring data. Refer to niscope. Session.trigger\_delay\_time for more information.

#### configure trigger video

```
niscope. Session. configure_trigger_video (trigger_source, signal_format, event, polarity, trigger_coupling, enable_dc_restore=False, line_number=1, hold-off=datetime.timedelta(seconds=0.0), delay=datetime.timedelta(seconds=0.0))
```

Configures the common properties for video triggering, including the signal format, TV event, line number, polarity, and enable DC restore. A video trigger occurs when the digitizer finds a valid video signal sync.

When you initiate an acquisition, the digitizer waits for the start trigger, which is configured through the <code>niscope.Session.acq\_arm\_source</code> (Start Trigger Source) property. The default is immediate. Upon receiving the start trigger the digitizer begins sampling pretrigger points. After the digitizer finishes sampling pretrigger points, the digitizer waits for a reference (stop) trigger that you specify with a method such as this one. Upon receiving the reference trigger the digitizer finishes the acquisition after completing posttrigger sampling. With each Configure Trigger method, you specify configuration parameters such as the trigger source and the amount of trigger delay.

**Note:** Some features are not supported by all digitizers. Refer to Features Supported by Device for more information.

#### **Parameters**

- **trigger\_source** (*str*) Specifies the trigger source. Refer to *niscope*. Session.trigger\_source for defined values.
- signal\_format (niscope.VideoSignalFormat) Specifies the type of video signal sync the digitizer should look for. Refer to niscope.Session. tv\_trigger\_signal\_format for more information.
- **event** (niscope.VideoTriggerEvent) Specifies the TV event you want to trigger on. You can trigger on a specific or on the next coming line or field of the signal.
- **polarity** (*niscope*. *VideoPolarity*) Specifies the polarity of the video signal sync.
- trigger\_coupling (niscope.TriggerCoupling) Applies coupling and filtering options to the trigger signal. Refer to niscope.Session. trigger\_coupling for more information.
- enable\_dc\_restore (bool) Offsets each video line so the clamping level (the portion of the video line between the end of the color burst and the beginning of the active image) is moved to zero volt. Refer to niscope.Session. enable dc restore for defined values.
- line\_number (int) Selects the line number to trigger on. The line number range covers an entire frame and is referenced as shown on Vertical Blanking and Synchronization Signal. Refer to niscope.Session.

  tv\_trigger\_line\_number for more information.

Default value: 1

 holdoff (float in seconds or datetime.timedelta) – The length of time the digitizer waits after detecting a trigger before enabling NI-SCOPE to

detect another trigger. Refer to niscope.Session.trigger\_holdoff for more information.

• delay (float in seconds or datetime.timedelta) - How long the digitizer waits after receiving the trigger to start acquiring data. Refer to niscope. Session.trigger\_delay\_time for more information.

## configure\_trigger\_window

```
niscope.Session.configure_trigger_window(trigger_source, low_level, high_level, window_mode, trigger_coupling, hold-off=datetime.timedelta(seconds=0.0), delay=datetime.timedelta(seconds=0.0)
```

Configures common properties for analog window triggering. A window trigger occurs when a signal enters or leaves a window you specify with the **high level** or **low level** parameters.

When you initiate an acquisition, the digitizer waits for the start trigger, which is configured through the <code>niscope.Session.acq\_arm\_source</code> (Start Trigger Source) property. The default is immediate. Upon receiving the start trigger the digitizer begins sampling pretrigger points. After the digitizer finishes sampling pretrigger points, the digitizer waits for a reference (stop) trigger that you specify with a method such as this one. Upon receiving the reference trigger the digitizer finishes the acquisition after completing posttrigger sampling. With each Configure Trigger method, you specify configuration parameters such as the trigger source and the amount of trigger delay.

To trigger the acquisition, use niscope. Session. send\_software\_trigger\_edge().

**Note:** Some features are not supported by all digitizers. Refer to Features Supported by Device for more information.

#### **Parameters**

- **trigger\_source** (*str*) Specifies the trigger source. Refer to *niscope*. Session.trigger\_source for defined values.
- **low\_level** (*float*) Passes the voltage threshold you want the digitizer to use for low triggering.
- **high\_level** (float) Passes the voltage threshold you want the digitizer to use for high triggering.
- window\_mode (niscope.TriggerWindowMode) Specifies whether you want the trigger to occur when the signal enters or leaves a window.
- trigger\_coupling (niscope.TriggerCoupling) Applies coupling and filtering options to the trigger signal. Refer to niscope.Session. trigger\_coupling for more information.
- holdoff (float in seconds or datetime.timedelta) The length of time the digitizer waits after detecting a trigger before enabling NI-SCOPE to detect another trigger. Refer to niscope.Session.trigger\_holdoff for more information.
- **delay** (float in seconds or datetime.timedelta) How long the digitizer waits after receiving the trigger to start acquiring data. Refer to niscope. Session.trigger\_delay\_time for more information.

### configure vertical

```
niscope.Session.configure_vertical (range, coupling, offset=0.0, probe_attenuation=1.0, enabled=True)
```

Configures the most commonly configured properties of the digitizer vertical subsystem, such as the range, offset, coupling, probe attenuation, and the channel.

**Tip:** This method requires repeated capabilities (channels). If called directly on the niscope. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].configure_vertical(range, coupling, offset=0.0,\_ \hookrightarrow probe_attenuation=1.0, enabled=True)
```

#### **Parameters**

- range (float) Specifies the vertical range Refer to niscope. Session. vertical range for more information.
- **coupling** (niscope. VerticalCoupling) Specifies how to couple the input signal. Refer to niscope. Session. vertical\_coupling for more information.
- **offset** (float) Specifies the vertical offset. Refer to niscope. Session. vertical\_offset for more information.
- **probe\_attenuation** (*float*) Specifies the probe attenuation. Refer to *niscope.Session.probe\_attenuation* for valid values.
- **enabled** (bool) Specifies whether the channel is enabled for acquisition. Refer to niscope. Session.channel\_enabled for more information.

### disable

```
niscope.Session.disable()
```

Aborts any current operation, opens data channel relays, and releases RTSI and PFI lines.

### fetch

```
niscope. Session. fetch (num\_samples=None, relative\_to=niscope.FetchRelativeTo.PRETRIGGER, offset=0, record\_number=0, num\_records=None, time-out=datetime.timedelta(seconds=5.0))
```

Returns the waveform from a previously initiated acquisition that the digitizer acquires for the specified channel. This method returns scaled voltage waveforms.

This method may return multiple waveforms depending on the number of channels, the acquisition type, and the number of records you specify.

**Note:** Some functionality, such as time stamping, is not supported in all digitizers.

**Tip:** This method requires repeated capabilities (channels). If called directly on the niscope. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].fetch(num_samples=None, relative_to=niscope.

FetchRelativeTo.PRETRIGGER, offset=0, record_number=0, num_
records=None, timeout=datetime.timedelta(seconds=5.0))
```

#### **Parameters**

- num\_samples (int) The maximum number of samples to fetch for each waveform. If the acquisition finishes with fewer points than requested, some devices return partial data if the acquisition finished, was aborted, or a timeout of 0 was used. If it fails to complete within the timeout period, the method raises.
- relative\_to (niscope.FetchRelativeTo) Position to start fetching within one record.
- **offset** (*int*) Offset in samples to start fetching data within each record. The offset can be positive or negative.
- record number (int) Zero-based index of the first record to fetch.
- num\_records (int) Number of records to fetch. Use -1 to fetch all configured records.
- **timeout** (*float or datetime.timedelta*) The time to wait for data to be acquired; using 0 for this parameter tells NI-SCOPE to fetch whatever is currently available. Using -1 seconds for this parameter implies infinite timeout.

### **Return type** list of WaveformInfo

#### Returns

Returns an array of classes with the following timing and scaling information about each waveform:

- relative\_initial\_x (float) the time (in seconds) from the trigger to the first sample in the fetched waveform
- **absolute\_initial\_x** (float) timestamp (in seconds) of the first fetched sample. This timestamp is comparable between records and acquisitions; devices that do not support this parameter use 0 for this output.
- x\_increment (float) the time between points in the acquired waveform in seconds
- channel (str) channel name this waveform was asquire from
- record (int) record number of this waveform
- gain (float) the gain factor of the given channel; useful for scaling binary data with the following formula:

```
voltage = binarydata * gainfactor + offset
```

• **offset** (float) the offset factor of the given channel; useful for scaling binary data with the following formula:

```
voltage = binarydata * qainfactor + offset
```

• samples (array of float) floating point array of samples. Length will be of the actual samples acquired

### fetch into

```
niscope.Session.fetch_into (waveform, relative_to=niscope.FetchRelativeTo.PRETRIGGER, offset=0, record_number=0, num_records=None, time-out=datetime.timedelta(seconds=5.0))
```

Returns the waveform from a previously initiated acquisition that the digitizer acquires for the specified channel. This method returns scaled voltage waveforms.

This method may return multiple waveforms depending on the number of channels, the acquisition type, and the number of records you specify.

**Note:** Some functionality, such as time stamping, is not supported in all digitizers.

**Tip:** This method requires repeated capabilities (channels). If called directly on the niscope. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].fetch(waveform, relative_to=niscope.

→FetchRelativeTo.PRETRIGGER, offset=0, record_number=0, num_

→records=None, timeout=datetime.timedelta(seconds=5.0))
```

#### **Parameters**

• waveform (array.array("d")) - numpy array of the appropriate type and size the should be acquired as a 1D array. Size should be num\_samples times number of waveforms. Call niscope.Session.\_actual\_num\_wfms() to determine the number of waveforms.

Types supported are

- numpy.float64
- numpy.int8
- numpy.in16
- numpy.int32

#### Example:

• relative\_to (niscope.FetchRelativeTo) - Position to start fetching within one record.

- **offset** (*int*) Offset in samples to start fetching data within each record. The offset can be positive or negative.
- record\_number (int) Zero-based index of the first record to fetch.
- num\_records (int) Number of records to fetch. Use -1 to fetch all configured records.
- **timeout** (*float*) The time to wait in seconds for data to be acquired; using 0 for this parameter tells NI-SCOPE to fetch whatever is currently available. Using -1 for this parameter implies infinite timeout.

## Return type list of WaveformInfo

#### Returns

Returns an array of classed with the following timing and scaling information about each waveform:

- relative\_initial\_x (float) the time (in seconds) from the trigger to the first sample in the fetched waveform
- absolute\_initial\_x (float) timestamp (in seconds) of the first fetched sample. This timestamp is comparable between records and acquisitions; devices that do not support this parameter use 0 for this output.
- x\_increment (float) the time between points in the acquired waveform in seconds
- channel (str) channel name this waveform was asquire from
- record (int) record number of this waveform
- gain (float) the gain factor of the given channel; useful for scaling binary data with the following formula:

```
voltage = binarydata * gainfactor + offset
```

• **offset** (float) the offset factor of the given channel; useful for scaling binary data with the following formula:

```
voltage = binarydata * gainfactor + offset
```

• **samples** (array of float) floating point array of samples. Length will be of the actual samples acquired

# get\_equalization\_filter\_coefficients

```
niscope.Session.get_equalization_filter_coefficients()
```

Retrieves the custom coefficients for the equalization FIR filter on the device. This filter is designed to compensate the input signal for artifacts introduced to the signal outside of the digitizer. Because this filter is a generic FIR filter, any coefficients are valid. Coefficient values should be between +1 and -1.

**Tip:** This method requires repeated capabilities (channels). If called directly on the niscope. Session object, then the method will use all repeated capabilities in the session. You can specify a subset

of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].get_equalization_filter_coefficients()
```

#### lock

```
niscope.Session.lock()
```

Obtains a multithread lock on the device session. Before doing so, the software waits until all other execution threads release their locks on the device session.

Other threads may have obtained a lock on this session for the following reasons:

- The application called the niscope. Session.lock() method.
- A call to NI-SCOPE locked the session.
- After a call to the <code>niscope.Session.lock()</code> method returns successfully, no other threads can access the device session until you call the <code>niscope.Session.unlock()</code> method or exit out of the with block when using lock context manager.
- Use the niscope.Session.lock() method and the niscope.Session.unlock() method around a sequence of calls to instrument driver methods if you require that the device retain its settings through the end of the sequence.

You can safely make nested calls to the <code>niscope.Session.lock()</code> method within the same thread. To completely unlock the session, you must balance each call to the <code>niscope.Session.lock()</code> method with a call to the <code>niscope.Session.unlock()</code> method.

One method for ensuring there are the same number of unlock method calls as there is lock calls is to use lock as a context manager

```
with niscope.Session('dev1') as session:
    with session.lock():
        # Calls to session within a single lock context
```

The first with block ensures the session is closed regardless of any exceptions raised

The second with block ensures that unlock is called regardless of any exceptions raised

Return type context manager

**Returns** When used in a *with* statement, *niscope*. Session. lock() acts as a context manager and unlock will be called when the *with* block is exited

# probe\_compensation\_signal\_start

```
niscope.Session.probe_compensation_signal_start()
Starts the 1 kHz square wave output on PFI 1 for probe compensation.
```

### probe\_compensation\_signal\_stop

```
niscope.Session.probe_compensation_signal_stop()
Stops the 1 kHz square wave output on PFI 1 for probe compensation.
```

#### read

niscope.Session.read (num\_samples=None, relative\_to=niscope.FetchRelativeTo.PRETRIGGER, offset=0, record\_number=0, num\_records=None, time-out=datetime.timedelta(seconds=5.0))

Initiates an acquisition, waits for it to complete, and retrieves the data. The process is similar to calling niscope.Session.\_initiate\_acquisition(), niscope.Session.acquisition\_status(), and niscope.Session.fetch(). The only difference is that with niscope.Session.read(), you enable all channels specified with channelList before the acquisition; in the other method, you enable the channels with niscope.Session.configure vertical().

This method may return multiple waveforms depending on the number of channels, the acquisition type, and the number of records you specify.

**Note:** Some functionality, such as time stamping, is not supported in all digitizers.

**Tip:** This method requires repeated capabilities (channels). If called directly on the niscope. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling this method on the result.:

```
session.channels[0,1].read(num_samples=None, relative_to=niscope.

→FetchRelativeTo.PRETRIGGER, offset=0, record_number=0, num_

→records=None, timeout=datetime.timedelta(seconds=5.0))
```

#### **Parameters**

- num\_samples (int) The maximum number of samples to fetch for each waveform. If the acquisition finishes with fewer points than requested, some devices return partial data if the acquisition finished, was aborted, or a timeout of 0 was used. If it fails to complete within the timeout period, the method raises.
- relative\_to (niscope.FetchRelativeTo) Position to start fetching within one record.
- **offset** (*int*) Offset in samples to start fetching data within each record. The offset can be positive or negative.
- **record\_number** (*int*) Zero-based index of the first record to fetch.
- num\_records (int) Number of records to fetch. Use -1 to fetch all configured records
- **timeout** (*float or datetime.timedelta*) The time to wait for data to be acquired; using 0 for this parameter tells NI-SCOPE to fetch whatever is currently available. Using -1 seconds for this parameter implies infinite timeout.

## Return type list of WaveformInfo

### Returns

Returns an array of classes with the following timing and scaling information about each waveform:

- relative\_initial\_x (float) the time (in seconds) from the trigger to the first sample in the fetched waveform
- absolute\_initial\_x (float) timestamp (in seconds) of the first fetched sample. This timestamp is comparable between records and acquisitions; devices that do not support this parameter use 0 for this output.
- x increment (float) the time between points in the acquired waveform in seconds
- channel (str) channel name this waveform was asquire from
- record (int) record number of this waveform
- gain (float) the gain factor of the given channel; useful for scaling binary data with the following formula:

```
voltage = binarydata * gainfactor + offset
```

• **offset** (float) the offset factor of the given channel; useful for scaling binary data with the following formula:

```
voltage = binarydata * gainfactor + offset
```

• samples (array of float) floating point array of samples. Length will be of the actual samples acquired

### reset

```
niscope.Session.reset()
```

Stops the acquisition, releases routes, and all session properties are reset to their default states.

# reset\_device

```
niscope.Session.reset_device()
```

Performs a hard reset of the device. Acquisition stops, all routes are released, RTSI and PFI lines are tristated, hardware is configured to its default state, and all session properties are reset to their default state.

• Thermal Shutdown

### reset with defaults

```
niscope.Session.reset_with_defaults()
```

Performs a software reset of the device, returning it to the default state and applying any initial default settings from the IVI Configuration Store.

#### self cal

niscope. Session. **self\_cal** (*option=niscope.Option.SELF\_CALIBRATE\_ALL\_CHANNELS*) Self-calibrates most NI digitizers, including all SMC-based devices and most Traditional NI-DAQ (Legacy) devices. To verify that your digitizer supports self-calibration, refer to Features Supported by Device.

For SMC-based digitizers, if the self-calibration is performed successfully in a regular session, the calibration constants are immediately stored in the self-calibration area of the EEPROM. If the self-calibration is performed in an external calibration session, the calibration constants take effect immediately for the duration of the session. However, they are not stored in the EEPROM until niscope.Session.CalEnd() is called with action set to NISCOPE\_VAL\_ACTION\_STORE and no errors occur.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

**Tip:** This method requires repeated capabilities (channels). If called directly on the niscope. Session object, then the method will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niscope. Session repeated capabilities container, and calling this method on the result.:

 $\verb|session.channels[0,1].self_cal(option=niscope.Option.SELF_CALIBRATE\_ALL\_ \\ \hookrightarrow CHANNELS)|$ 

Parameters option (niscope.Option) – The calibration option. Use VI\_NULL for a normal self-calibration operation or NISCOPE\_VAL\_CAL\_RESTORE\_EXTERNAL\_CALIBRATION to restore the previous calibration.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

## self\_test

```
niscope.Session.self_test()
```

Runs the instrument self-test routine and returns the test result(s). Refer to the device-specific help topics for an explanation of the message contents.

Raises SelfTestError on self test failure. Properties on exception object:

- code failure code from driver
- message status message from driver

Self-Test Code	Description
0	Passed self-test
1	Self-test failed

## send\_software\_trigger\_edge

```
niscope.Session.send_software_trigger_edge (which_trigger)
```

Sends the selected trigger to the digitizer. Call this method if you called <code>niscope.Session.configure\_trigger\_software()</code> when you want the Reference trigger to occur. You can also call this method to override a misused edge, digital, or hysteresis trigger. If you have configured <code>niscope.Session.acq\_arm\_source</code>, <code>niscope.Session.arm\_ref\_trig\_src</code>, or <code>niscope.Session.adv\_trig\_src</code>, call this method when you want to send the corresponding trigger to the digitizer.

**Parameters which\_trigger** (niscope.WhichTrigger) - Specifies the type of trigger to send to the digitizer.

#### **Defined Values**

```
START (OL)

ARM_REFERENCE (1L)

REFERENCE (2L)

ADVANCE (3L)
```

# unlock

```
niscope.Session.unlock()
```

Releases a lock that you acquired on an device session using niscope.Session.lock(). Refer to niscope.Session.unlock() for additional information on session locks.

## **Properties**

Property	Datatype
niscope.Session.absolute_sample_clock_offset	float in seconds or datetime.timedelta
niscope.Session.accessory_gain	float
niscope.Session.accessory_offset	float
niscope.Session.acquisition_start_time	float in seconds or datetime.timedelta
niscope.Session.acquisition_type	AcquisitionType
niscope.Session.acq_arm_source	str
niscope.Session.adv_trig_src	str
niscope.Session.allow_more_records_than_memory	bool
niscope.Session.arm_ref_trig_src	str
niscope.Session.backlog	float
niscope.Session.bandpass_filter_enabled	bool
niscope.Session.binary_sample_width	int
niscope.Session.channel_count	int
niscope.Session.channel_enabled	bool

Continued on next page

Table 13 – continued from previous page

Property  niscope.Session.channel_terminal_configuration  niscope.Session.data_transfer_block_size  int  niscope.Session.data_transfer_maximum_bandwidth  niscope.Session.data_transfer_preferred_packet_size  int  niscope.Session.device_temperature  niscope.Session.driver_setup  niscope.Session.enable_dc_restore  niscope.Session.enable_time_interleaved_sampling  niscope.Session.end_of_acquisition_event_output_terminal str  niscope.Session.end_of_record_to_advance_trigger_holdoff  niscope.Session.equalization_filter_enabled  niscope.Session.exported_advance_trigger_output_terminal str  niscope.Session.exported_advance_trigger_output_terminal str  niscope.Session.exported_advance_trigger_output_terminal str  niscope.Session.exported_advance_trigger_output_terminal str  niscope.Session.exported_start_trigger_output_terminal str  niscope.Session.exported_start_trigger_output_terminal str  niscope.Session.filex_fir_antialias_filter_type  niscope.Session.filex_fir_antialias_filter_type  niscope.Session.horz_enforce_realtime  niscope.Session.horz_enforce_realtime  niscope.Session.horz_num_records  niscope.Session.horz_num_records  nint  niscope.Session.horz_num_record_length  niscope.Session.horz_record_ref_position  niscope.Session.horz_sample_rate  float	uration
niscope.Session.data_transfer_block_size int niscope.Session.data_transfer_maximum_bandwidth float niscope.Session.data_transfer_preferred_packet_size int niscope.Session.device_temperature float niscope.Session.driver_setup str niscope.Session.enable_dc_restore bool niscope.Session.enable_time_interleaved_sampling bool niscope.Session.enable_time_interleaved_sampling bool niscope.Session.end_of_acquisition_event_output_terminal str niscope.Session.end_of_record_event_output_terminal str niscope.Session.end_of_record_to_advance_trigger_holdoff float in seconds or dat niscope.Session.equalization_filter_enabled bool niscope.Session.equalization_filter_enabled bool niscope.Session.exported_advance_trigger_output_terminal str niscope.Session.exported_ref_trigger_output_terminal str niscope.Session.exported_start_trigger_output_terminal str niscope.Session.flex_fir_antialias_filter_type FlexFIRAntialia niscope.Session.fpga_bitfile_path str niscope.Session.horz_enforce_realtime bool niscope.Session.horz_enforce_realtime bool niscope.Session.horz_min_num_pts int niscope.Session.horz_num_records int niscope.Session.horz_record_length int niscope.Session.horz_record_ref_position float	
niscope.Session.data_transfer_maximum_bandwidth       float         niscope.Session.data_transfer_preferred_packet_size       int         niscope.Session.device_temperature       float         niscope.Session.driver_setup       str         niscope.Session.enable_dc_restore       bool         niscope.Session.enable_time_interleaved_sampling       bool         niscope.Session.end_of_acquisition_event_output_terminal       str         niscope.Session.end_of_record_event_output_terminal       str         niscope.Session.end_of_record_event_output_terminal       str         niscope.Session.equalization_filter_enabled       bool         niscope.Session.equalization_num_coefficients       int         niscope.Session.exported_advance_trigger_output_terminal       str         niscope.Session.exported_ref_trigger_output_terminal       str         niscope.Session.exported_ref_trigger_output_terminal       str         niscope.Session.flex_fir_antialias_filter_type       FlexFIRAntialia         niscope.Session.horg_bitfile_path       str         niscope.Session.horz_enforce_realtime       bool         niscope.Session.horz_min_num_pts       int         niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position <t< td=""><td></td></t<>	
niscope. Session. data_transfer_preferred_packet_size       int         niscope. Session. device_temperature       float         niscope. Session. driver_setup       str         niscope. Session. enable_dc_restore       bool         niscope. Session. enable_time_interleaved_sampling       bool         niscope. Session. end_of_acquisition_event_output_terminal       str         niscope. Session. end_of_record_event_output_terminal       str         niscope. Session. end_of_record_to_advance_trigger_holdoff       float in seconds or date         niscope. Session. equalization_filter_enabled       bool         niscope. Session. equalization_num_coefficients       int         niscope. Session. exported_advance_trigger_output_terminal       str         niscope. Session. exported_ref_trigger_output_terminal       str         niscope. Session. exported_ref_trigger_output_terminal       str         niscope. Session. flex_fir_antialias_filter_type       FlexFIRAntialia         niscope. Session. flex_fir_antialias_filter_type       float         niscope. Session. horz_enforce_realtime       bool         niscope. Session. horz_min_num_pts       int         niscope. Session. horz_num_records       int         niscope. Session. horz_record_length       int         niscope. Session. horz_record_ref_position       float	
niscope.Session.device_temperature       float         niscope.Session.driver_setup       str         niscope.Session.enable_dc_restore       bool         niscope.Session.enable_time_interleaved_sampling       bool         niscope.Session.end_of_acquisition_event_output_terminal       str         niscope.Session.end_of_record_event_output_terminal       str         niscope.Session.end_of_record_to_advance_trigger_holdoff       float in seconds or date         niscope.Session.equalization_filter_enabled       bool         niscope.Session.equalization_num_coefficients       int         niscope.Session.exported_advance_trigger_output_terminal       str         niscope.Session.exported_ref_trigger_output_terminal       str         niscope.Session.exported_start_trigger_output_terminal       str         niscope.Session.flex_fir_antialias_filter_type       FlexFIRAntialia         niscope.Session.high_pass_filter_frequency       float         niscope.Session.horz_enforce_realtime       bool         niscope.Session.horz_min_num_pts       int         niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.driver_setup         str           niscope.Session.enable_dc_restore         bool           niscope.Session.enable_time_interleaved_sampling         bool           niscope.Session.end_of_acquisition_event_output_terminal         str           niscope.Session.end_of_record_event_output_terminal         str           niscope.Session.end_of_record_to_advance_trigger_holdoff         float in seconds or date           niscope.Session.equalization_filter_enabled         bool           niscope.Session.equalization_num_coefficients         int           niscope.Session.exported_advance_trigger_output_terminal         str           niscope.Session.exported_ref_trigger_output_terminal         str           niscope.Session.flex_fir_antialias_filter_type         FlexFIRAntialia           niscope.Session.fpga_bitfile_path         str           niscope.Session.horz_enforce_realtime         bool           niscope.Session.horz_min_num_pts         int           niscope.Session.horz_num_records         int           niscope.Session.horz_record_length         int           niscope.Session.horz_record_ref_position         float	
niscope.Session.enable_dc_restoreboolniscope.Session.enable_time_interleaved_samplingboolniscope.Session.end_of_acquisition_event_output_terminalstrniscope.Session.end_of_record_event_output_terminalstrniscope.Session.end_of_record_to_advance_trigger_holdofffloat in seconds or datniscope.Session.equalization_filter_enabledboolniscope.Session.equalization_num_coefficientsintniscope.Session.exported_advance_trigger_output_terminalstrniscope.Session.exported_ref_trigger_output_terminalstrniscope.Session.exported_start_trigger_output_terminalstrniscope.Session.flex_fir_antialias_filter_typeFlexFIRAntialianiscope.Session.high_pass_filter_frequencyfloatniscope.Session.horz_enforce_realtimeboolniscope.Session.horz_min_num_ptsintniscope.Session.horz_num_recordsintniscope.Session.horz_record_lengthintniscope.Session.horz_record_ref_positionfloat	
niscope.Session.enable_time_interleaved_samplingboolniscope.Session.end_of_acquisition_event_output_terminalstrniscope.Session.end_of_record_event_output_terminalstrniscope.Session.end_of_record_to_advance_trigger_holdofffloat in seconds or datniscope.Session.equalization_filter_enabledboolniscope.Session.equalization_num_coefficientsintniscope.Session.exported_advance_trigger_output_terminalstrniscope.Session.exported_ref_trigger_output_terminalstrniscope.Session.exported_start_trigger_output_terminalstrniscope.Session.flex_fir_antialias_filter_typeFlexFIRAntialianiscope.Session.high_pass_filter_frequencyfloatniscope.Session.horz_enforce_realtimeboolniscope.Session.horz_min_num_ptsintniscope.Session.horz_num_recordsintniscope.Session.horz_record_lengthintniscope.Session.horz_record_ref_positionfloat	
niscope.Session.end_of_acquisition_event_output_terminalstrniscope.Session.end_of_record_event_output_terminalstrniscope.Session.end_of_record_to_advance_trigger_holdofffloat in seconds or datniscope.Session.equalization_filter_enabledboolniscope.Session.equalization_num_coefficientsintniscope.Session.exported_advance_trigger_output_terminalstrniscope.Session.exported_ref_trigger_output_terminalstrniscope.Session.exported_start_trigger_output_terminalstrniscope.Session.flex_fir_antialias_filter_typeFlexFIRAntialianiscope.Session.fpga_bitfile_pathstrniscope.Session.horz_enforce_realtimeboolniscope.Session.horz_enforce_realtimeboolniscope.Session.horz_num_recordsintniscope.Session.horz_record_lengthintniscope.Session.horz_record_ref_positionfloat	
niscope.Session.end_of_record_event_output_terminal       str         niscope.Session.end_of_record_to_advance_trigger_holdoff       float in seconds or dat         niscope.Session.equalization_filter_enabled       bool         niscope.Session.exported_advance_trigger_output_terminal       str         niscope.Session.exported_advance_trigger_output_terminal       str         niscope.Session.exported_ref_trigger_output_terminal       str         niscope.Session.exported_start_trigger_output_terminal       str         niscope.Session.flex_fir_antialias_filter_type       FlexFIRAntialia         niscope.Session.fpga_bitfile_path       str         niscope.Session.high_pass_filter_frequency       float         niscope.Session.horz_enforce_realtime       bool         niscope.Session.horz_min_num_pts       int         niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.end_of_record_to_advance_trigger_holdoff       float in seconds or date to niscope.Session.equalization_filter_enabled       bool         niscope.Session.equalization_num_coefficients       int         niscope.Session.exported_advance_trigger_output_terminal       str         niscope.Session.exported_ref_trigger_output_terminal       str         niscope.Session.exported_start_trigger_output_terminal       str         niscope.Session.flex_fir_antialias_filter_type       FlexFIRAntialias         niscope.Session.fpga_bitfile_path       str         niscope.Session.high_pass_filter_frequency       float         niscope.Session.horz_enforce_realtime       bool         niscope.Session.horz_min_num_pts       int         niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.equalization_filter_enabledboolniscope.Session.equalization_num_coefficientsintniscope.Session.exported_advance_trigger_output_terminalstrniscope.Session.exported_ref_trigger_output_terminalstrniscope.Session.exported_start_trigger_output_terminalstrniscope.Session.flex_fir_antialias_filter_typeFlexFIRAntialianiscope.Session.fpga_bitfile_pathstrniscope.Session.high_pass_filter_frequencyfloatniscope.Session.horz_enforce_realtimeboolniscope.Session.horz_min_num_ptsintniscope.Session.horz_num_recordsintniscope.Session.horz_record_lengthintniscope.Session.horz_record_ref_positionfloat	tatima timadalta
niscope.Session.equalization_num_coefficientsintniscope.Session.exported_advance_trigger_output_terminalstrniscope.Session.exported_ref_trigger_output_terminalstrniscope.Session.exported_start_trigger_output_terminalstrniscope.Session.flex_fir_antialias_filter_typeFlexFIRAntialianiscope.Session.fpga_bitfile_pathstrniscope.Session.high_pass_filter_frequencyfloatniscope.Session.horz_enforce_realtimeboolniscope.Session.horz_min_num_ptsintniscope.Session.horz_num_recordsintniscope.Session.horz_record_lengthintniscope.Session.horz_record_ref_positionfloat	etime.timedena
niscope.Session.exported_advance_trigger_output_terminal       str         niscope.Session.exported_ref_trigger_output_terminal       str         niscope.Session.exported_start_trigger_output_terminal       str         niscope.Session.flex_fir_antialias_filter_type       FlexFIRAntialia         niscope.Session.fpga_bitfile_path       str         niscope.Session.high_pass_filter_frequency       float         niscope.Session.horz_enforce_realtime       bool         niscope.Session.horz_min_num_pts       int         niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.exported_ref_trigger_output_terminal       str         niscope.Session.exported_start_trigger_output_terminal       str         niscope.Session.flex_fir_antialias_filter_type       FlexFIRAntialias         niscope.Session.fpga_bitfile_path       str         niscope.Session.high_pass_filter_frequency       float         niscope.Session.horz_enforce_realtime       bool         niscope.Session.horz_min_num_pts       int         niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.exported_start_trigger_output_terminal       str         niscope.Session.flex_fir_antialias_filter_type       FlexFIRAntialia         niscope.Session.fpga_bitfile_path       str         niscope.Session.high_pass_filter_frequency       float         niscope.Session.horz_enforce_realtime       bool         niscope.Session.horz_min_num_pts       int         niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.flex_fir_antialias_filter_type       FlexFIRAntialias         niscope.Session.fpga_bitfile_path       str         niscope.Session.high_pass_filter_frequency       float         niscope.Session.horz_enforce_realtime       bool         niscope.Session.horz_min_num_pts       int         niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.fpga_bitfile_pathstrniscope.Session.high_pass_filter_frequencyfloatniscope.Session.horz_enforce_realtimeboolniscope.Session.horz_min_num_ptsintniscope.Session.horz_num_recordsintniscope.Session.horz_record_lengthintniscope.Session.horz_record_ref_positionfloat	- m'l m
niscope.Session.high_pass_filter_frequencyfloatniscope.Session.horz_enforce_realtimeboolniscope.Session.horz_min_num_ptsintniscope.Session.horz_num_recordsintniscope.Session.horz_record_lengthintniscope.Session.horz_record_ref_positionfloat	<u>isFilterType</u>
niscope.Session.horz_enforce_realtime       bool         niscope.Session.horz_min_num_pts       int         niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.horz_min_num_pts       int         niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.horz_num_records       int         niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.horz_record_length       int         niscope.Session.horz_record_ref_position       float	
niscope.Session.horz_record_ref_position float	
niscone Session horz sample rate	
niscope.Session.horz_time_per_record float in seconds or dat	etime.timedelta
niscope.Session.input_clock_source str	
niscope.Session.input_impedance float	
niscope.Session.instrument_firmware_revision str	
niscope.Session.instrument_manufacturer str	
niscope.Session.instrument_model str	
niscope.Session.interleaving_offset_correction_enabled bool	
niscope.Session.io_resource_descriptor str	
niscope.Session.logical_name str	
niscope.Session.master_enable bool	
niscope.Session.max_input_frequency float	
niscope.Session.max_real_time_sampling_rate   float	
niscope.Session.max_ris_rate float	
niscope.Session.min_sample_rate float	
niscope.Session.onboard_memory_size int	
niscope.Session.output_clock_source str	
niscope.Session.pll_lock_status bool	
niscope.Session.points_done float	
niscope.Session.poll_interval int	
niscope.Session.probe_attenuation float	
niscope.Session.ready_for_advance_event_output_terminal str	
niscope.Session.ready_for_ref_event_output_terminal str	
niscope.Session.ready_for_start_event_output_terminal str	
niscope.Session.records_done int	

Continued on next page

Table 13 – continued from previous page

Property	Datatype
niscope.Session.record_arm_source	str
niscope.Session.ref_clk_rate	float
niscope.Session.ref_trigger_detector_location	RefTriggerDetectorLocation
niscope.Session.ref_trigger_minimum_quiet_time	float in seconds or datetime.timedelta
niscope.Session.ref_trig_tdc_enable	bool
niscope.Session.resolution	int
niscope.Session.ris_in_auto_setup_enable	bool
niscope.Session.ris_method	RISMethod
niscope.Session.ris_num_averages	int
niscope.Session.sample_clock_timebase_multiplier	int
niscope.Session.sample_mode	int
niscope.Session.samp_clk_timebase_div	int
niscope.Session.samp_clk_timebase_rate	float
niscope.Session.samp_clk_timebase_src	str
niscope.Session.serial_number	str
niscope.Session.simulate	bool
niscope.Session.specific_driver_description	str
niscope.Session.specific_driver_revision	str
niscope.Session.specific_driver_vendor	str
niscope.Session.start_to_ref_trigger_holdoff	float in seconds or datetime.timedelta
niscope.Session.supported_instrument_models	str
niscope.Session.trigger_auto_triggered	bool
niscope.Session.trigger_coupling	TriggerCoupling
niscope.Session.trigger_delay_time	float in seconds or datetime.timedelta
niscope.Session.trigger_holdoff	float in seconds or datetime.timedelta
niscope.Session.trigger_hysteresis	float
niscope.Session.trigger_impedance	float
niscope.Session.trigger_level	float
niscope.Session.trigger_modifier	TriggerModifier
niscope.Session.trigger_slope	TriggerSlope
niscope.Session.trigger_source	str
niscope.Session.trigger_type	TriggerType
niscope.Session.trigger_window_high_level	float
niscope.Session.trigger_window_low_level	float
niscope.Session.trigger_window_mode	TriggerWindowMode
niscope.Session.tv_trigger_event	VideoTriggerEvent
niscope.Session.tv_trigger_line_number	int
niscope.Session.tv_trigger_polarity	VideoPolarity
niscope.Session.tv_trigger_signal_format	VideoSignalFormat
niscope.Session.vertical_coupling	VerticalCoupling
niscope.Session.vertical_offset	float
niscope.Session.vertical_range	float

# Methods

Method name	
niscope.Session.abort()	
niscope.Session.acquisition_status()	

Continued on next page

Table 14 – continued from previous page

Method name
niscope.Session.auto_setup()
niscope.Session.commit()
niscope.Session.configure_chan_characteristics()
niscope.Session.configure_equalization_filter_coefficients()
niscope.Session.configure_horizontal_timing()
niscope.Session.configure_trigger_digital()
niscope.Session.configure_trigger_edge()
niscope.Session.configure_trigger_hysteresis()
niscope.Session.configure_trigger_immediate()
niscope.Session.configure_trigger_software()
niscope.Session.configure_trigger_video()
niscope.Session.configure_trigger_window()
niscope.Session.configure_vertical()
niscope.Session.disable()
niscope.Session.fetch()
niscope.Session.fetch_into()
niscope.Session.get_equalization_filter_coefficients()
niscope.Session.lock()
niscope.Session.probe_compensation_signal_start()
niscope.Session.probe_compensation_signal_stop()
niscope.Session.read()
niscope.Session.reset()
niscope.Session.reset_device()
niscope.Session.reset_with_defaults()
niscope.Session.self_cal()
niscope.Session.self_test()
niscope.Session.send_software_trigger_edge()
niscope.Session.unlock()

# 7.4.3.2 Enums

Enums used in NI-SCOPE

# **AcquisitionStatus**

class niscope.AcquisitionStatus

COMPLETE

IN\_PROGRESS

STATUS\_UNKNOWN

# AcquisitionType

class niscope.AcquisitionType

#### NORMAL

Sets the digitizer to normal resolution mode. The digitizer can use real-time sampling or equivalent-time sampling.

#### **FLEXRES**

Sets the digitizer to flexible resolution mode if supported. The digitizer uses different hardware configurations to change the resolution depending on the sampling rate used.

#### DDC

Sets the digitizer to DDC mode on the NI 5620/5621.

#### **FetchRelativeTo**

#### class niscope.FetchRelativeTo

#### READ POINTER

The read pointer is set to zero when a new acquisition is initiated. After every fetch the read pointer is incremeted to be the sample after the last sample retrieved. Therefore, you can repeatedly fetch relative to the read pointer for a continuous acquisition program.

#### PRETRIGGER

Fetches relative to the first pretrigger point requested with niscope.Session.configure\_horizontal\_timing().

#### NOW

Fetch data at the last sample acquired.

#### START

Fetch data starting at the first point sampled by the digitizer.

#### TRIGGER

Fetch at the first posttrigger sample.

# **FlexFIRAntialiasFilterType**

### class niscope.FlexFIRAntialiasFilterType

## FOURTYEIGHT\_TAP\_STANDARD

This filter is optimized for alias protection and frequency-domain flatness

#### FOURTYEIGHT TAP HANNING

This filter is optimized for the lowest possible bandwidth for a 48 tap filter and maximizes the SNR

### SIXTEEN\_TAP\_HANNING

This filter is optimized for the lowest possible bandwidth for a 16 tap filter and maximizes the SNR

# EIGHT\_TAP\_HANNING

This filter is optimized for the lowest possible bandwidth for a 8 tap filter and maximizes the SNR

# **Option**

# class niscope.Option

#### SELF CALIBRATE ALL CHANNELS

Self Calibrating all Channels

#### RESTORE EXTERNAL CALIBRATION

Restore External Calibration.

### **RISMethod**

#### class niscope.RISMethod

#### **EXACT NUM AVERAGES**

Acquires exactly the specified number of records for each bin in the RIS acquisition. An error is returned from the fetch method if the RIS acquisition does not successfully acquire the specified number of waveforms within the timeout period. You may call the fetch method again to allow more time for the acquisition to finish.

#### MIN\_NUM\_AVERAGES

Each RIS sample is the average of a least a minimum number of randomly distributed points.

#### INCOMPLETE

Returns the RIS waveform after the specified timeout even if it is incomplete. If no waveforms have been acquired in certain bins, these bins will have a NaN (when fetching scaled data) or a zero (when fetching binary data). A warning (positive error code) is returned from the fetch method if the RIS acquisition did not finish. The acquisition aborts when data is returned.

#### LIMITED BIN WIDTH

Limits the waveforms in the various bins to be within 200 ps of the center of the bin.

#### RefTriggerDetectorLocation

#### class niscope.RefTriggerDetectorLocation

#### ANALOG\_DETECTION\_CIRCUIT

use the hardware analog circuitry to implement the reference trigger. This option will trigger before any onboard signal processing.

#### DDC OUTPUT

use the onboard signal processing logic to implement the reference trigger. This option will trigger based on the onboard signal processed data.

# **TerminalConfiguration**

## class niscope.TerminalConfiguration

#### SINGLE ENDED

Channel is single ended

### UNBALANCED\_DIFFERENTIAL

Channel is unbalanced differential

#### DIFFERENTIAL

Channel is differential

# **TriggerCoupling**

```
AC
AC coupling

DC
DC coupling

HF_REJECT
Highpass filter coupling

LF_REJECT
Lowpass filter coupling

AC_PLUS_HF_REJECT
Highpass and lowpass filter coupling
```

# TriggerModifier

```
class niscope.TriggerModifier
```

## NO\_TRIGGER\_MOD

Normal triggering.

AUTO

Software will trigger an acquisition automatically if no trigger arrives after a certain amount of time.

## **TriggerSlope**

```
class niscope.TriggerSlope

NEGATIVE
Falling edge

POSITIVE
```

Rising edge

## **TriggerType**

```
class niscope.TriggerType
```

# EDGE

Configures the digitizer for edge triggering. An edge trigger occurs when the trigger signal crosses the trigger level specified with the set trigger slope. You configure the trigger level and slope with niscope. Session.configure\_trigger\_edge().

TV

Configures the digitizer for video/TV triggering. You configure the video trigger parameters like signal Format, Line to trigger off of, Polarity, and Enable DC Restore with niscope.Session.configure\_trigger\_video().

#### IMMEDIATE

Configures the digitizer for immediate triggering. An immediate trigger occurs as soon as the pretrigger samples are acquired.

# HYSTERESIS

Configures the digitizer for hysteresis triggering. A hysteresis trigger occurs when the trigger signal crosses the trigger level with the specified slope and passes through the hysteresis window you specify. You configure the trigger level, slope, and hysteresis with niscope. Session. configure\_trigger\_hysteresis().

#### DIGITAL

Configures the digitizer for digital triggering. A digital trigger occurs when the trigger signal has the specified slope. You configure the trigger slope with niscope. Session. configure\_trigger\_digital().

#### WINDOW

Configures the digitizer for window triggering. A window trigger occurs when the trigger signal enters or leaves the window defined by the values you specify with the Low Window Level, High Window Level, and Window Mode Parameters. You configure the low window level high window level, and window mode with niscope. Session.configure trigger window().

#### SOFTWARE

Configures the digitizer for software triggering. A software trigger occurs when niscope. Session. SendSoftwareTrigger() is called.

# TriggerWindowMode

```
class niscope.TriggerWindowMode
```

#### ENTERING

Trigger upon entering the window

#### **LEAVING**

Trigger upon leaving the window

#### VerticalCoupling

```
class niscope.VerticalCoupling
```

AC

AC coupling

DC

DC coupling

GND

**GND** coupling

# **VideoPolarity**

class niscope.VideoPolarity

#### POSITIVE

Specifies that the video signal has positive polarity.

#### NEGATIVE

Specifies that the video signal has negative polarity.

# VideoSignalFormat

### class niscope. Video Signal Format

#### NTSC

NTSC signal format supports line numbers from 1 to 525

#### PAL

PAL signal format supports line numbers from 1 to 625

#### SECAM

SECAM signal format supports line numbers from 1 to 625

#### M PAL

M-PAL signal format supports line numbers from 1 to 525

# VIDEO\_480I\_59\_94\_FIELDS\_PER\_SECOND

480 lines, interlaced, 59.94 fields per second

## VIDEO\_480I\_60\_FIELDS\_PER\_SECOND

480 lines, interlaced, 60 fields per second

### VIDEO\_480P\_59\_94\_FRAMES\_PER\_SECOND

480 lines, progressive, 59.94 frames per second

## VIDEO\_480P\_60\_FRAMES\_PER\_SECOND

480 lines, progressive,60 frames per second

# VIDEO\_576I\_50\_FIELDS\_PER\_SECOND

576 lines, interlaced, 50 fields per second

#### VIDEO\_576P\_50\_FRAMES\_PER\_SECOND

576 lines, progressive, 50 frames per second

# VIDEO\_720P\_50\_FRAMES\_PER\_SECOND

720 lines, progressive, 50 frames per second

#### VIDEO\_720P\_59\_94\_FRAMES\_PER\_SECOND

720 lines, progressive, 59.94 frames per second

# VIDEO\_720P\_60\_FRAMES\_PER\_SECOND

720 lines, progressive, 60 frames per second

## VIDEO\_1080I\_50\_FIELDS\_PER\_SECOND

1,080 lines, interlaced, 50 fields per second

# VIDEO\_1080I\_59\_94\_FIELDS\_PER\_SECOND

1,080 lines, interlaced, 59.94 fields per second

## VIDEO\_1080I\_60\_FIELDS\_PER\_SECOND

1,080 lines, interlaced, 60 fields per second

# VIDEO\_1080P\_24\_FRAMES\_PER\_SECOND

1,080 lines, progressive, 24 frames per second

# VideoTriggerEvent

## class niscope.VideoTriggerEvent

#### FIELD1

Trigger on field 1 of the signal

#### FIELD2

Trigger on field 2 of the signal

# ANY\_FIELD

Trigger on the first field acquired

## ANY\_LINE

Trigger on the first line acquired

#### LINE NUMBER

Trigger on a specific line of a video signal. Valid values vary depending on the signal format configured.

# WhichTrigger

```
class niscope.WhichTrigger
```

START

ARM REFERENCE

REFERENCE

ADVANCE

#### 7.4.3.3 Exceptions and Warnings

### **DriverError**

```
exception niscope.DriverError
An error originating from the NI-SCOPE driver
```

# UnsupportedConfigurationError

#### exception niscope.UnsupportedConfigurationError

An error due to using this module in an usupported platform.

#### DriverNotInstalledError

# exception niscope.DriverNotInstalledError

An error due to using this module without the driver runtime installed.

## InvalidRepeatedCapabilityError

```
exception niscope. InvalidRepeatedCapabilityError
An error due to an invalid character in a repeated capability
```

#### SelfTestError

```
exception niscope.SelfTestError
An error due to a failed self-test
```

# **DriverWarning**

```
exception niscope.DriverWarning
A warning originating from the NI-SCOPE driver
```

#### **7.4.3.4 Examples**

niscope\_fetch.py

Listing 8: (niscope\_fetch.py)

```
#!/usr/bin/python
2
   import argparse
   import niscope
   import pprint
   import sys
   pp = pprint.PrettyPrinter(indent=4, width=80)
10
   def example(resource_name, channels, options, length, voltage):
11
       with niscope.Session(resource_name=resource_name, options=options) as session:
12
           session.configure_vertical(range=voltage, coupling=niscope.VerticalCoupling.
13
   →AC)
           session.configure_horizontal_timing(min_sample_rate=50000000, min_num_
14
   →pts=length, ref_position=50.0, num_records=1, enforce_realtime=True)
           with session.initiate():
               waveforms = session.channels[channels].fetch(num_samples=length)
16
           for i in range(len(waveforms)):
17
               print('Waveform {0} information:'.format(i))
18
               print(str(waveforms[i]) + '\n\n')
19
20
21
   def _main(argsv):
22
       parser = argparse.ArgumentParser(description='Acquires one record from the given_
23
   → channels.', formatter_class=argparse.ArgumentDefaultsHelpFormatter)
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource_
24
   →name of a National Instruments Digitizer')
       parser.add_argument('-c', '--channels', default='0', help='Channel(s) to use')
25
       parser.add_argument('-1', '--length', default=1000, type=int, help='Measure_
   →record length')
```

(continues on next page)

```
parser.add_argument('-v', '--voltage', default=1.0, type=float, help='Voltage,
   →range (V)')
       parser.add_argument('-op', '--option-string', default='', type=str, help='Option_
28
   →string')
       args = parser.parse_args(argsv)
       example(args.resource_name, args.channels, args.option_string, args.length, args.
31
32
   def main():
33
       _main(sys.argv[1:])
34
37
   def test_example():
       options = {'simulate': True, 'driver_setup': {'Model': '5164', 'BoardType': 'PXIe
38
   example('PXI1Slot2', '0', options, 1000, 1.0)
39
41
   def test_main():
42
       cmd_line = ['--option-string', 'Simulate=1, DriverSetup=Model:5164; BoardType:PXIe
43
   _main(cmd_line)
44
45
   if __name__ == '__main__':
48
       main()
49
```

## niscope\_fetch\_forever.py

Listing 9: (niscope\_fetch\_forever.py)

```
#!/usr/bin/python
2
   import argparse
   import datetime
   import niscope
   import numpy as np
   import pprint
   import sys
10
   pp = pprint.PrettyPrinter(indent=4, width=80)
11
12
13
   # We use fetch into which allows us to allocate a single buffer per channel and
14
   → "fetch into" it a section at a time without having to
   # reconstruct the waveform once we are done
   def example(resource_name, options, total_acquisition_time_in_seconds, voltage,_
   →sample_rate_in_hz, samples_per_fetch):
       total_samples = int(total_acquisition_time_in_seconds * sample_rate_in_hz)
17
       # 1. Opening session
18
       with niscope.Session(resource_name=resource_name, options=options) as session:
```

(continues on next page)

```
# We will acquire on all channels of the device
20
           channel_list = [c for c in range(session.channel_count)] # Need an actual,
21
   →list and not a range
22
           # 2. Creating numpy arrays
23
           waveforms = [np.ndarray(total_samples, dtype=np.float64) for c in channel_
24
   →list]
25
           # 3. Configuring
26
           session.configure_horizontal_timing(min_sample_rate=sample_rate_in_hz, min_
27
   →num_pts=1, ref_position=0.0, num_records=1, enforce_realtime=True)
           session.channels[channel_list].configure_vertical(voltage, coupling=niscope.
    →VerticalCoupling.DC, enabled=True)
           # Configure software trigger, but never send the trigger.
29
           # This starts an infinite acquisition, until you call session.abort() or.
30
   ⇒session.close()
           session.configure_trigger_software()
31
           current_pos = 0
32
           # 4. initating
33
           with session.initiate():
34
               while current_pos < total_samples:</pre>
35
                    # We fetch each channel at a time so we don't have to de-interleave..
36
   -afterwards
                    # We do not keep the wfm_info returned from fetch_into
37
                    for channel, waveform in zip(channel_list, waveforms):
                        # 5. fetching - we return the slice of the waveform array that we.
   →want to "fetch into"
                        session.channels[channel].fetch_into(waveform[current_pos:current_
40
   →pos + samples_per_fetch], relative_to=niscope.FetchRelativeTo.READ_POINTER,
                                                              offset=0, record_number=0,...
41
   →num_records=1, timeout=datetime.timedelta(seconds=5.0))
42
                   current_pos += samples_per_fetch
43
44
   def _main(argsv):
45
       parser = argparse.ArgumentParser(description='Fetch more samples than will fit in...
46
   →memory.', formatter_class=argparse.ArgumentDefaultsHelpFormatter)
47
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource,
   →name of a National Instruments Digitizer')
       parser.add_argument('-t', '--time', default=10, type=int, help='Time to sample (s)
48
       parser.add_argument('-v', '--voltage', default=1.0, type=float, help='Voltage_
49
   →range (V)')
       parser.add_argument('-op', '--option-string', default='', type=str, help='Option_
50
    →string')
       parser.add_argument('-r', '--sample-rate', default=1000.0, type=float, help=
51
   →'Sample Rate (Hz)')
       parser.add_argument('-s', '--samples-per-fetch', default=100, type=int, help=
52
   →'Samples per fetch')
53
       args = parser.parse_args(argsv)
       example(args.resource_name, args.option_string, args.time, args.voltage, args.
   →sample_rate, args.samples_per_fetch)
55
   def main():
57
58
       _main(sys.argv[1:])
```

(continues on next page)

```
60
   def test example():
61
       options = {'simulate': True, 'driver_setup': {'Model': '5164', 'BoardType': 'PXIe
62
   \hookrightarrow ', }, }
       example('PXI1Slot2', options, 10, 1.0, 1000.0, 100)
63
65
   def test_main():
66
       cmd_line = ['--option-string', 'Simulate=1, DriverSetup=Model:5164; BoardType:PXIe
67
   _main(cmd_line)
68
71
   if __name__ == '__main__':
       main()
72.
```

### niscope read.py

Listing 10: (niscope\_read.py)

```
#!/usr/bin/python
2
   import argparse
3
   import niscope
   import pprint
   import sys
   pp = pprint.PrettyPrinter(indent=4, width=80)
10
   def example(resource_name, channels, options, length, voltage):
11
       with niscope.Session(resource_name=resource_name, options=options) as session:
12
           session.configure_vertical(range=voltage, coupling=niscope.VerticalCoupling.
13
   →AC)
           session.configure_horizontal_timing(min_sample_rate=50000000, min_num_
14
   →pts=length, ref_position=50.0, num_records=1, enforce_realtime=True)
           waveforms = session.channels[channels].read(num_samples=length)
15
           for i in range(len(waveforms)):
16
               print('Waveform {0} information:'.format(i))
17
               print(str(waveforms[i]) + '\n\n')
18
19
20
21
   def _main(argsv):
       parser = argparse.ArgumentParser(description='Acquires one record from the given_
22
   → channels.', formatter_class=argparse.ArgumentDefaultsHelpFormatter)
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource...
23
   ⇒name of a National Instruments Digitizer')
       parser.add_argument('-c', '--channels', default='0', help='Channel(s) to use')
24
       parser.add_argument('-1', '--length', default=1000, type=int, help='Measure_
25
   →record length')
       parser.add_argument('-v', '--voltage', default=1.0, type=float, help='Voltage,
   →range (V)')
       parser.add_argument('-op', '--option-string', default='', type=str, help='Option.
27
    →string')
```

(continues on next page)

```
args = parser.parse_args(argsv)
28
       example(args.resource_name, args.channels, args.option_string, args.length, args.
29
   →voltage)
30
31
   def main():
32
       _main(sys.argv[1:])
33
34
35
   def test_example():
36
       options = {'simulate': True, 'driver_setup': {'Model': '5164', 'BoardType': 'PXIe
37
       example('PXI1Slot2', '0', options, 1000, 1.0)
39
40
   def test_main():
41
       cmd_line = ['--option-string', 'Simulate=1, DriverSetup=Model:5164; BoardType:PXIe
42
       _main(cmd_line)
43
44
45
   if __name__ == '__main___':
46
       main()
47
```

# 7.5 niswitch module

## 7.5.1 Installation

As a prerequisite to using the niswitch module, you must install the NI-SWITCH runtime on your system. Visit ni.com/downloads to download the driver runtime for your devices.

The nimi-python modules (i.e. for **NI-SWITCH**) can be installed with pip:

```
$ python -m pip install niswitch~=1.0.1
```

Or easy install from setuptools:

```
$ python -m easy_install niswitch
```

# 7.5.2 **Usage**

The following is a basic example of using the **niswitch** module to open a session to a Switch and connect channels.

```
import niswitch
with niswitch.Session("Dev1") as session:
    session.connect(channel1='r0', channel2='c0')
```

Additional examples for NI-SWITCH are located in src/niswitch/examples/ directory.

## 7.5.3 API Reference

#### 7.5.3.1 niswitch.Session

**class** niswitch. **Session** (*self, resource\_name, topology="Configured Topology", simulate=False, reset\_device=False*)

Returns a session handle used to identify the switch in all subsequent instrument driver calls and sets the topology of the switch. <code>niswitch.Session.\_\_init\_\_()</code> creates a new IVI instrument driver session for the switch specified in the resourceName parameter. The driver uses the topology specified in the topology parameter and overrides the topology specified in MAX. Note: When initializing an NI SwitchBlock device with topology, you must specify the toplogy created when you configured the device in MAX, using either NISWITCH\_TOPOLOGY\_CONFIGURED\_TOPOLOGY or the toplogy string of the device. Refer to the Initializing with Toplogy for NI SwitchBlock Devices topic in the NI Switches Help for information about determining the topology string of an NI SwitchBlock device. By default, the switch is reset to a known state. Enable simulation by specifying the topology and setting the simulate parameter to True.

#### **Parameters**

- resource\_name (str) Resource name of the switch module to initialize. Default value: None Syntax: Optional fields are shown in square brackets ([]). Configured in MAX Under Valid Syntax Devices and Interfaces DeviceName Traditional NI-DAQ Devices SCXI[chassis ID]::slot number PXI System PXI[bus number]::device number TIP: IVI logical names are also valid for the resource name. Default values for optional fields: chassis ID = 1 bus number = 0 Example resource names: Resource Name Description SC1Mod3 NI-DAQmx module in chassis "SC1" slot 3 MySwitch NI-DAQmx module renamed to "MySwitch" SCXI1::3 Traditional NI-DAQ module in chassis 1, slot 3 SCXI::3 Traditional NI-DAQ module in chassis 1, slot 3 PXI0::16 PXI bus 0, device number 16 PXI::16 PXI bus 0, device number 16
- Pass the topology name for topology (str) you want use switch you specify with Resource Name parameter. You can NISWITCH TOPOLOGY CONFIGURED TOPOLOGY also to use that was configured for the device in MAX. the last topology fault Value: NISWITCH\_TOPOLOGY\_CONFIGURED\_TOPOLOGY NISWITCH TOPOLOGY 1127 1 WIRE 64X1 MUX Valid Values: NISWITCH\_TOPOLOGY\_1127\_2\_WIRE\_32X1\_MUX NISWITCH\_TOPOLOGY\_1127\_2\_WIRE\_4X8\_MATR NISWITCH TOPOLOGY 1127 4 WIRE 16X1 MUX NISWITCH TOPOLOGY 1127 INDEPENDENT NISWITCH\_TOPOLOGY\_1128\_1\_WIRE\_64X1\_MUX NISWITCH\_TOPOLOGY\_1128\_2\_WIRE\_32X1\_MUX NISWITCH TOPOLOGY 1128 2 WIRE 4X8 MATRIX NISWITCH\_TOPOLOGY\_1128\_4\_WIRE\_16X1\_MUX NISWITCH\_TOPOLOGY\_1128\_INDEPENDENT NISWITCH\_TOPOLOGY\_1129\_2\_WIRE\_16X16\_MATRIX NISWITCH\_TOPOLOGY\_1129\_2\_WIRE\_8X32\_MATRIX NISWITCH TOPOLOGY 1129 2 WIRE 4X64 MATRIX NISWITCH\_TOPOLOGY\_1129\_2\_WIRE\_DUAL\_8X16\_MATRIX NISWITCH\_TOPOLOGY\_1129\_2\_WIRE\_DUAL\_4X32\_MATRIX NISWITCH\_TOPOLOGY\_1129\_2\_WIRE\_QUAD\_4X16\_MATRIX NISWITCH\_TOPOLOGY\_1130\_1\_WIRE\_256X1\_MUX NISWITCH\_TOPOLOGY\_1130\_1\_WIRE\_DUAL\_128 NISWITCH\_TOPOLOGY\_1130\_1\_WIRE\_4X64\_MATRIX NISWITCH TOPOLOGY 1130 1 WIRE 8x32 MATRIX NISWITCH TOPOLOGY 1130 1 WIRE OCTAL 32X1 MUX NISWITCH\_TOPOLOGY\_1130\_1\_WIRE\_QUAD\_64X1\_MUX NISWITCH TOPOLOGY 1130 1 WIRE SIXTEEN 16X1 MUX

NISWITCH\_TOPOLOGY\_1130\_2\_WIRE\_128X1\_MUX NISWITCH\_TOPOLOGY\_1130\_2\_WIRE\_OCTAL\_16

7.5. niswitch module 379

NISWITCH\_TOPOLOGY\_1130\_2\_WIRE\_4X32\_MATRIX

NISWITCH\_TOPOLOGY\_1130\_2\_WIRE\_QUAD\_32X1\_MUX

```
NISWITCH TOPOLOGY 1130 4 WIRE 64X1 MUX NISWITCH TOPOLOGY 1130 4 WIRE OUAD 16X
NISWITCH TOPOLOGY 1130 INDEPENDENT NISWITCH TOPOLOGY 1160 16 SPDT
NISWITCH TOPOLOGY 1161 8 SPDT NISWITCH TOPOLOGY 1163R OCTAL 4X1 MUX
NISWITCH_TOPOLOGY_1166_16_DPDT NISWITCH_TOPOLOGY_1166_32_SPDT
NISWITCH TOPOLOGY 1167 INDEPENDENT NISWITCH TOPOLOGY 1169 100 SPST
NISWITCH TOPOLOGY 1169 50 DPST NISWITCH TOPOLOGY 1175 1 WIRE 196X1 MUX
NISWITCH TOPOLOGY 1175 2 WIRE 98X1 MUX NISWITCH TOPOLOGY 1175 2 WIRE 95X1 MUX
NISWITCH TOPOLOGY 1190 QUAD 4X1 MUXNISWITCH TOPOLOGY 1191 QUAD 4X1 MUX
NISWITCH TOPOLOGY 1192 8 SPDT NISWITCH TOPOLOGY 1193 32X1 MUX
NISWITCH_TOPOLOGY_1193_16X1_TERMINATED_MUX
NISWITCH_TOPOLOGY_1193_DUAL_16X1_MUX NISWITCH_TOPOLOGY_1193_DUAL_8X1_TERMINA
NISWITCH_TOPOLOGY_1193_QUAD_8X1_MUX NISWITCH_TOPOLOGY_1193_QUAD_4X1_TERMINAT
NISWITCH TOPOLOGY 1193 INDEPENDENT NISWITCH TOPOLOGY 1194 QUAD 4X1 MUX
NISWITCH_TOPOLOGY_1195_QUAD_4X1_MUX NISWITCH_TOPOLOGY_2501_1_WIRE_48X1_MUX
NISWITCH_TOPOLOGY_2501_1_WIRE_48X1_AMPLIFIED_MUX
NISWITCH_TOPOLOGY_2501_2_WIRE_24X1_MUX NISWITCH_TOPOLOGY_2501_2_WIRE_24X1_AMP
NISWITCH_TOPOLOGY_2501_2_WIRE_DUAL_12X1_MUX
NISWITCH TOPOLOGY 2501 2 WIRE QUAD 6X1 MUX
NISWITCH TOPOLOGY 2501 2 WIRE 4X6 MATRIX
NISWITCH TOPOLOGY 2501 4 WIRE 12X1 MUX NISWITCH TOPOLOGY 2503 1 WIRE 48X1 MUX
NISWITCH_TOPOLOGY_2503_2_WIRE_24X1_MUX NISWITCH_TOPOLOGY_2503_2_WIRE_DUAL_12X
NISWITCH TOPOLOGY 2503 2 WIRE QUAD 6X1 MUX
NISWITCH_TOPOLOGY_2503_2_WIRE_4X6_MATRIX
NISWITCH_TOPOLOGY_2503_4_WIRE_12X1_MUX NISWITCH_TOPOLOGY_2510_INDEPENDENT
NISWITCH TOPOLOGY 2512 INDEPENDENT NISWITCH TOPOLOGY 2514 INDEPENDENT
NISWITCH TOPOLOGY 2515 INDEPENDENT NISWITCH TOPOLOGY 2520 80 SPST
NISWITCH_TOPOLOGY_2523_26_DPDT NISWITCH_TOPOLOGY_2524_1_WIRE_128X1_MUX
NISWITCH_TOPOLOGY_2524_1_WIRE_DUAL_64X1_MUX
NISWITCH TOPOLOGY 2524 1 WIRE QUAD 32X1 MUX
NISWITCH_TOPOLOGY_2524_1_WIRE_OCTAL_16X1_MUX
NISWITCH_TOPOLOGY_2524_1_WIRE_SIXTEEN_8X1_MUX
NISWITCH_TOPOLOGY_2525_2_WIRE_64X1_MUX NISWITCH_TOPOLOGY_2525_2_WIRE_DUAL_32X
NISWITCH_TOPOLOGY_2525_2_WIRE_QUAD_16X1_MUX
NISWITCH TOPOLOGY 2525 2 WIRE OCTAL 8X1 MUX
NISWITCH TOPOLOGY 2525 2 WIRE SIXTEEN 4X1 MUX
NISWITCH TOPOLOGY 2526 1 WIRE 158X1 MUXNISWITCH TOPOLOGY 2526 2 WIRE 79X1 MUX
NISWITCH_TOPOLOGY_2527_1_WIRE_64X1_MUX NISWITCH_TOPOLOGY_2527_1_WIRE_DUAL_32X
NISWITCH TOPOLOGY 2527 2 WIRE 32X1 MUXNISWITCH TOPOLOGY 2527 2 WIRE DUAL 16X
NISWITCH_TOPOLOGY_2527_4_WIRE_16X1_MUX NISWITCH_TOPOLOGY_2527_INDEPENDENT
NISWITCH TOPOLOGY 2529 2 WIRE DUAL 4X16 MATRIX
NISWITCH TOPOLOGY 2529 2 WIRE 8X16 MATRIX
NISWITCH TOPOLOGY 2529 2 WIRE 4X32 MATRIX
NISWITCH_TOPOLOGY_2530_1_WIRE_128X1_MUX NISWITCH_TOPOLOGY_2530_1_WIRE_DUAL_642
NISWITCH_TOPOLOGY_2530_1_WIRE_4x32_MATRIX
NISWITCH_TOPOLOGY_2530_1_WIRE_8x16_MATRIX
NISWITCH TOPOLOGY 2530 1 WIRE OCTAL 16X1 MUX
NISWITCH_TOPOLOGY_2530_1_WIRE_QUAD_32X1_MUX
NISWITCH_TOPOLOGY_2530_2_WIRE_4x16_MATRIX
NISWITCH_TOPOLOGY_2530_2_WIRE_64X1_MUX NISWITCH_TOPOLOGY_2530_2_WIRE_DUAL_32X
```

NISWITCH\_TOPOLOGY\_2530\_4\_WIRE\_32X1\_MUX NISWITCH\_TOPOLOGY\_2530\_4\_WIRE\_DUAL\_16X NISWITCH TOPOLOGY 2530 INDEPENDENT NISWITCH TOPOLOGY 2531 1 WIRE 4X128 MATRIX

380 Chapter 7. License

NISWITCH\_TOPOLOGY\_2530\_2\_WIRE\_QUAD\_16X1\_MUX

NISWITCH TOPOLOGY 2531 1 WIRE 8X64 MATRIX

NISWITCH TOPOLOGY 2531 1 WIRE DUAL 4X64 MATRIX

```
NISWITCH_TOPOLOGY_2531_1_WIRE_DUAL_8X32_MATRIX
NISWITCH TOPOLOGY 2531 2 WIRE 4X64 MATRIX NISWITCH TOPOLOGY 2531 2 WIRE 8X32 M
NISWITCH_TOPOLOGY_2532_1_WIRE_16X32_MATRIX
NISWITCH TOPOLOGY 2532 1 WIRE 4X128 MATRIX NISWITCH TOPOLOGY 2532 1 WIRE 8X64 M
NISWITCH TOPOLOGY 2532 1 WIRE DUAL 16X16 MATRIX NISWITCH TOPOLOGY 2532 1 WIRE
NISWITCH TOPOLOGY 2532 1 WIRE DUAL 8X32 MATRIX NISWITCH TOPOLOGY 2532 1 WIRE S
NISWITCH TOPOLOGY 2532 2 WIRE 16X16 MATRIX NISWITCH TOPOLOGY 2532 2 WIRE 4X64 M
NISWITCH TOPOLOGY 2532 2 WIRE 8X32 MATRIX NISWITCH TOPOLOGY 2532 2 WIRE DUAL 4
NISWITCH_TOPOLOGY_2533_1_WIRE_4X64_MATRIX NISWITCH_TOPOLOGY_2534_1_WIRE_8X32_M
NISWITCH_TOPOLOGY_2535_1_WIRE_4X136_MATRIX NISWITCH_TOPOLOGY_2536_1_WIRE_8X68_N
NISWITCH TOPOLOGY 2540 1 WIRE 8X9 MATRIX NISWITCH TOPOLOGY 2541 1 WIRE 8X12 MATRIX NISWITCH TOPOLOGY 2541 1 WIRE 8
NISWITCH TOPOLOGY 2542 OUAD 2X1 TERMINATED MUX NISWITCH TOPOLOGY 2543 DUAL 4
NISWITCH_TOPOLOGY_2544_8X1_TERMINATED_MUX NISWITCH_TOPOLOGY_2545_4X1_TERMINATED_MUX NISWITCH_TOPOLOGY_2545_4X1_TERMINATED_TOPOLOGY_2545_4X1_TERMINATED_TOPOLOGY_255_5X1_TERMINATED_TOPOLOGY_255_5X1_TERMINATED_TOPOLOGY_255_5X1_TERMINATED_TOPOLOGY_255_5X1_TERMINATED_TOPOLOGY_255_5X1_TERMINATED_TOPOLOGY_255_5X1_TERMINATED_TOPOLOGY_255_5X1_TERMINATED_TOPOLOGY_255_5X1_TERMINATED_TOPOLOGY_255_5X1_TERMINATED_TOPOLOGY_255_5X1_TERMINATED_TOPOLOGY_255_5X1_TE
NISWITCH_TOPOLOGY_2546_DUAL_4X1_MUX NISWITCH_TOPOLOGY_2547_8X1_MUX
NISWITCH_TOPOLOGY_2548_4_SPDT NISWITCH_TOPOLOGY_2549_TERMINATED_2_SPDT
NISWITCH_TOPOLOGY_2554_4X1_MUX NISWITCH_TOPOLOGY_2555_4X1_TERMINATED_MUX
NISWITCH TOPOLOGY 2556 DUAL 4X1 MUX NISWITCH TOPOLOGY 2557 8X1 MUX
NISWITCH TOPOLOGY 2558 4 SPDT NISWITCH TOPOLOGY 2559 TERMINATED 2 SPDT
NISWITCH TOPOLOGY 2564 16 SPST
                                                                      NISWITCH TOPOLOGY 2564 8 DPST
NISWITCH TOPOLOGY 2565 16 SPST NISWITCH TOPOLOGY 2566 16 SPDT
NISWITCH TOPOLOGY 2566 8 DPDT NISWITCH TOPOLOGY 2567 INDEPENDENT
NISWITCH_TOPOLOGY_2569_100_SPST NISWITCH_TOPOLOGY_2569_50_DPST
NISWITCH TOPOLOGY 2570 20 DPDT NISWITCH TOPOLOGY 2570 40 SPDT
NISWITCH TOPOLOGY 2571 66 SPDT NISWITCH TOPOLOGY 2575 1 WIRE 196X1 MUX
NISWITCH_TOPOLOGY_2575_2_WIRE_98X1_MUX NISWITCH_TOPOLOGY_2575_2_WIRE_95X1_MUX
NISWITCH_TOPOLOGY_2576_2_WIRE_64X1_MUX NISWITCH_TOPOLOGY_2576_2_WIRE_DUAL_32X
NISWITCH_TOPOLOGY_2576_2_WIRE_OCTAL_8X1_MUX NISWITCH_TOPOLOGY_2576_2_WIRE_QUA
NISWITCH TOPOLOGY 2576 2 WIRE SIXTEEN 4X1 MUXNISWITCH TOPOLOGY 2576 INDEPEND
NISWITCH_TOPOLOGY_2584_1_WIRE_12X1_MUX NISWITCH_TOPOLOGY_2584_1_WIRE_DUAL_6X1
NISWITCH_TOPOLOGY_2584_2_WIRE_6X1_MUX NISWITCH_TOPOLOGY_2584_INDEPENDENT
NISWITCH_TOPOLOGY_2585_1_WIRE_10X1_MUX NISWITCH_TOPOLOGY_2586_10_SPST
NISWITCH_TOPOLOGY_2586_5_DPST NISWITCH_TOPOLOGY_2590_4X1_MUX
NISWITCH TOPOLOGY 2591 4X1 MUX NISWITCH TOPOLOGY 2593 16X1 MUX
NISWITCH TOPOLOGY 2593 8X1 TERMINATED MUXNISWITCH TOPOLOGY 2593 DUAL 8X1 MU
NISWITCH TOPOLOGY 2593 DUAL 4X1 TERMINATED MUX NISWITCH TOPOLOGY 2593 INDEPE
NISWITCH TOPOLOGY 2594 4X1 MUX NISWITCH TOPOLOGY 2595 4X1 MUX
NISWITCH TOPOLOGY 2596 DUAL 6X1 MUX NISWITCH TOPOLOGY 2597 6X1 TERMINATED MU
NISWITCH_TOPOLOGY_2598_DUAL_TRANSFER NISWITCH_TOPOLOGY_2599_2_SPDT
NISWITCH TOPOLOGY 2720 INDEPENDENT NISWITCH TOPOLOGY 2722 INDEPENDENT
NISWITCH TOPOLOGY 2725 INDEPENDENT NISWITCH TOPOLOGY 2727 INDEPENDENT
NISWITCH TOPOLOGY 2737 2 WIRE_4X64_MATRIX NISWITCH_TOPOLOGY_2738_2_WIRE_8X32_M
NISWITCH_TOPOLOGY_2739_2_WIRE_16X16_MATRIX NISWITCH_TOPOLOGY_2746_QUAD_4X1_MU
NISWITCH_TOPOLOGY_2747_DUAL_8X1_MUX NISWITCH_TOPOLOGY_2748_16X1_MUX
NISWITCH_TOPOLOGY_2790_INDEPENDENT NISWITCH_TOPOLOGY_2796_DUAL_6X1_MUX
NISWITCH TOPOLOGY 2797 6X1 TERMINATED MUX NISWITCH TOPOLOGY 2798 DUAL TRANSI
NISWITCH TOPOLOGY 2799 2 SPDT
```

- **simulate** (bool) Enables simulation of the switch module specified in the resource name parameter. Valid Values: True simulate False Don't simulate (Default Value)
- **reset\_device** (bool) Specifies whether to reset the switch module during the initialization process. Valid Values: True Reset Device (Default Value) False Currently unsupported. The device will not reset.

7.5. niswitch module 381

# **Properties**

Property	Datatype
analog_bus_sharing_enable	bool
bandwidth	float
channel_count	int
characteristic_impedance	float
continuous_scan	bool
digital_filter_enable	bool
driver_setup	str
handshaking_initiation	HandshakingInitiation
instrument_firmware_revision	str
instrument_manufacturer	str
instrument_model	str
io_resource_descriptor	str
is_configuration_channel	bool
is_debounced	bool
is_scanning	bool
is_source_channel	bool
is_waiting_for_trig	bool
logical_name	str
max_ac_voltage	float
max_carry_ac_current	float
max_carry_ac_power	float
max_carry_dc_current	float
max_carry_dc_power	float
max_dc_voltage	float
max_switching_ac_current	float
max_switching_ac_power	float
max_switching_dc_current	float
max_switching_dc_power	float
number_of_relays	int
num_of_columns	int
num_of_rows	int
<pre>power_down_latching_relays_after_debounce</pre>	bool
scan_advanced_output	ScanAdvancedOutput
scan_advanced_polarity	ScanAdvancedPolarity
scan_delay	float in seconds or datetime.timedelta
scan_list	str
scan_mode	ScanMode
serial_number	str
settling_time	float in seconds or datetime.timedelta
simulate	bool
specific_driver_description	str
specific_driver_revision	str
specific_driver_vendor	str
supported_instrument_models	str
temperature	float
trigger_input	TriggerInput
trigger_input_polarity	TriggerInputPolarity
wire_mode	int

#### **Public methods**

Method name
abort()
can_connect()
commit()
connect()
connect_multiple()
disable()
disconnect()
disconnect_all()
disconnect_multiple()
<pre>get_channel_name()</pre>
get_path()
<pre>get_relay_count()</pre>
<pre>get_relay_name()</pre>
<pre>get_relay_position()</pre>
lock()
relay_control()
reset()
reset_with_defaults()
route_scan_advanced_output()
route_trigger_input()
self_test()
send_software_trigger()
set_path()
unlock()
wait_for_debounce()
<pre>wait_for_scan_complete()</pre>

## **Properties**

# analog\_bus\_sharing\_enable

# niswitch.Session.analog\_bus\_sharing\_enable

Enables or disables sharing of an analog bus line so that multiple NI SwitchBlock devices may connect to it simultaneously. To enable multiple NI SwitchBlock devices to share an analog bus line, set this property to True for each device on the channel that corresponds with the shared analog bus line. The default value for all devices is False, which disables sharing of the analog bus. Refer to the Using the Analog Bus on an NI SwitchBlock Carrier topic in the NI Switches Help for more information about sharing the analog bus.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].analog_bus_sharing_enable = var
var = session.channels[0,1].analog_bus_sharing_enable
```

7.5. niswitch module 383

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Channel Configuration: Analog Bus Sharing Enable
- C Attribute: NISWITCH\_ATTR\_ANALOG\_BUS\_SHARING\_ENABLE

#### bandwidth

niswitch.Session.bandwidth

This channel-based property returns the bandwidth for the channel. The units are hertz.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].bandwidth
```

The following table lists the characteristics of this property.

Value
float
read only
True
No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Bandwidth
- C Attribute: NISWITCH\_ATTR\_BANDWIDTH

#### channel count

niswitch.Session.channel\_count

Indicates the number of channels that the specific instrument driver supports.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Capabilities:Channel Count
- C Attribute: NISWITCH\_ATTR\_CHANNEL\_COUNT

### characteristic\_impedance

#### niswitch.Session.characteristic\_impedance

This channel-based property returns the characteristic impedance for the channel. The units are ohms.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].characteristic_impedance
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics: Characteristic Impedance
- C Attribute: NISWITCH ATTR CHARACTERISTIC IMPEDANCE

# continuous\_scan

## niswitch.Session.continuous\_scan

When a switch device is scanning, the swich can either stop scanning when the end of the scan (False) or continue scanning from the top of the scan list again (True). Notice that if you set the scan to continuous (True), the Wait For Scan Complete operation will always time out and you must call Abort to stop the scan.

The following table lists the characteristics of this property.

7.5. niswitch module 385

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Scanning Configuration: Continuous Scan
- C Attribute: NISWITCH\_ATTR\_CONTINUOUS\_SCAN

# digital\_filter\_enable

#### niswitch.Session.digital\_filter\_enable

This property specifies whether to apply the pulse width filter to the Trigger Input. Enabling the Digital Filter (True) prevents the switch module from being triggered by pulses that are less than 150 ns on PXI trigger lines 0–7. When Digital Filter is disabled (False), it is possible for the switch module to be triggered by noise on the PXI trigger lines. If the device triggering the switch is capable of sending pulses greater than 150 ns, you should not disable the Digital Filter.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Scanning Configuration: Digital Filter Enable
- C Attribute: NISWITCH\_ATTR\_DIGITAL\_FILTER\_ENABLE

#### driver setup

#### niswitch.Session.driver setup

This property indicates the Driver Setup string that the user specified when initializing the driver. Some cases exist where the end-user must specify instrument driver options at initialization time. An example of this is specifying a particular instrument model from among a family of instruments that the driver supports. This is useful when using simulation. The end-user can specify driver-specific options through the DriverSetup keyword in the optionsString parameter to the niswitch. Session.InitWithOptions() method, or through the IVI Configuration Utility. If the user does not specify a Driver Setup string, this property returns an empty string.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Characteristic	value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Advanced Session Information: Driver Setup
- C Attribute: NISWITCH\_ATTR\_DRIVER\_SETUP

# handshaking\_initiation

niswitch.Session.handshaking\_initiation

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.HandshakingInitiation
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Scanning Configuration: Handshaking Initiation
- C Attribute: NISWITCH\_ATTR\_HANDSHAKING\_INITIATION

## instrument firmware revision

niswitch.Session.instrument firmware revision

A string that contains the firmware revision information for the instrument you are currently using.

The following table lists the characteristics of this property.

Value
str
read only
False
No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Firmware Revision
- C Attribute: NISWITCH\_ATTR\_INSTRUMENT\_FIRMWARE\_REVISION

7.5. niswitch module 387

## instrument\_manufacturer

#### niswitch.Session.instrument manufacturer

A string that contains the name of the instrument manufacturer you are currently using.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Manufacturer
- C Attribute: NISWITCH\_ATTR\_INSTRUMENT\_MANUFACTURER

## instrument\_model

#### niswitch.Session.instrument\_model

A string that contains the model number or name of the instrument that you are currently using.

The following table lists the characteristics of this property.

Value
str
read only
False
No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Instrument Identification:Model
- C Attribute: NISWITCH\_ATTR\_INSTRUMENT\_MODEL

## io resource descriptor

#### niswitch.Session.io\_resource\_descriptor

Indicates the resource descriptor the driver uses to identify the physical device. If you initialize the driver with a logical name, this property contains the resource descriptor that corresponds to the entry in the IVI Configuration utility. If you initialize the instrument driver with the resource descriptor, this property contains that value.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Advanced Session Information:IO Resource Descriptor
- C Attribute: NISWITCH ATTR IO RESOURCE DESCRIPTOR

### is\_configuration\_channel

```
niswitch.Session.is_configuration_channel
```

This channel-based property specifies whether to reserve the channel for internal path creation. A channel that is available for internal path creation is called a configuration channel. The driver may use configuration channels to create paths between two channels you specify in the <code>niswitch.Session.connect()</code> method. Configuration channels are not available for external connections. Set this property to True to mark the channel as a configuration channel. Set this property to False to mark the channel as available for external connections. After you identify a channel as a configuration channel, you cannot use that channel for external connections. The <code>niswitch.Session.connect()</code> method returns the NISWITCH\_ERROR\_IS\_CONFIGURATION\_CHANNEL error when you attempt to establish a connection between a configuration channel and any other channel.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].is_configuration_channel = var
var = session.channels[0,1].is_configuration_channel
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Channel Configuration: Is Configuration Channel
- C Attribute: NISWITCH ATTR IS CONFIGURATION CHANNEL

7.5. niswitch module 389

### is debounced

#### niswitch.Session.is\_debounced

This property indicates whether the entire switch device has settled since the last switching command. A value of True indicates that all signals going through the switch device are valid.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Is Debounced
- C Attribute: NISWITCH\_ATTR\_IS\_DEBOUNCED

### is scanning

#### niswitch. Session. is scanning

If True, the switch module is currently scanning through the scan list (i.e. it is not in the Idle state). If False, the switch module is not currently scanning through the scan list (i.e. it is in the Idle state).

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Scanning Configuration:Is Scanning
- C Attribute: NISWITCH\_ATTR\_IS\_SCANNING

# is\_source\_channel

### niswitch.Session.is\_source\_channel

This channel-based property specifies whether you want to identify the channel as a source channel. Typically, you set this property to True when you attach the channel to a power supply, a method generator, or an active measurement point on the unit under test, and you do not want to connect the channel to another source. The driver prevents source channels from connecting to each other. The <code>niswitch.Session.connect()</code> method returns the

NISWITCH\_ERROR\_ATTEMPT\_TO\_CONNECT\_SOURCES when you attempt to connect two channels that you identify as source channels.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].is_source_channel = var
var = session.channels[0,1].is_source_channel
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Channel Configuration:Is Source Channel
- C Attribute: NISWITCH ATTR IS SOURCE CHANNEL

# is\_waiting\_for\_trig

## niswitch.Session.is\_waiting\_for\_trig

In a scan list, a semi-colon (;) is used to indicate that at that point in the scan list, the scan engine should pause until a trigger is received from the trigger input. If that trigger is user generated through either a hardware pulse or the Send SW Trigger operation, it is necessary for the user to know when the scan engine has reached such a state.

The following table lists the characteristics of this property.

Value
bool
read only
False
No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Scanning Configuration: Is Waiting for Trigger?
- C Attribute: NISWITCH\_ATTR\_IS\_WAITING\_FOR\_TRIG

7.5. niswitch module 391

### logical name

#### niswitch.Session.logical\_name

A string containing the logical name you specified when opening the current IVI session. You may pass a logical name to the niswitch.Session.init() or niswitch.Session. InitWithOptions() methods. The IVI Configuration utility must contain an entry for the logical name. The logical name entry refers to a virtual instrument section in the IVI Configuration file. The virtual instrument section specifies a physical device and initial user options.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: Advanced Session Information: Logical Name
- C Attribute: NISWITCH\_ATTR\_LOGICAL\_NAME

#### max\_ac\_voltage

#### niswitch.Session.max\_ac\_voltage

This channel-based property returns the maximum AC voltage the channel can switch. The units are volts RMS.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].max_ac_voltage
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Maximum AC Voltage
- C Attribute: NISWITCH\_ATTR\_MAX\_AC\_VOLTAGE

### max\_carry\_ac\_current

#### niswitch. Session.max carry ac current

This channel-based property returns the maximum AC current the channel can carry. The units are amperes RMS.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].max_carry_ac_current
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics: Maximum Carry AC Current
- C Attribute: NISWITCH\_ATTR\_MAX\_CARRY\_AC\_CURRENT

### max\_carry\_ac\_power

### niswitch.Session.max\_carry\_ac\_power

This channel-based property returns the maximum AC power the channel can carry. The units are volt-amperes.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].max_carry_ac_power
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Maximum Carry AC Power
- C Attribute: NISWITCH\_ATTR\_MAX\_CARRY\_AC\_POWER

### max\_carry\_dc\_current

niswitch.Session.max\_carry\_dc\_current

This channel-based property returns the maximum DC current the channel can carry. The units are amperes.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].max_carry_dc_current
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Maximum Carry DC Current
- C Attribute: NISWITCH ATTR MAX CARRY DC CURRENT

## max\_carry\_dc\_power

niswitch.Session.max\_carry\_dc\_power

This channel-based property returns the maximum DC power the channel can carry. The units are watts.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can

specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].max_carry_dc_power
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Maximum Carry DC Power
- C Attribute: NISWITCH\_ATTR\_MAX\_CARRY\_DC\_POWER

### max dc voltage

```
niswitch. Session.max dc voltage
```

This channel-based property returns the maximum DC voltage the channel can switch. The units are volts.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].max_dc_voltage
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Maximum DC Voltage
- C Attribute: NISWITCH\_ATTR\_MAX\_DC\_VOLTAGE

### max switching ac current

### niswitch.Session.max\_switching\_ac\_current

This channel-based property returns the maximum AC current the channel can switch. The units are amperes RMS.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].max_switching_ac_current
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Maximum Switching AC Current
- C Attribute: NISWITCH\_ATTR\_MAX\_SWITCHING\_AC\_CURRENT

### max switching ac power

### niswitch.Session.max\_switching\_ac\_power

This channel-based property returns the maximum AC power the channel can switch. The units are volt-amperes.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].max_switching_ac_power
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics: Maximum Switching AC Power
- C Attribute: NISWITCH\_ATTR\_MAX\_SWITCHING\_AC\_POWER

## max\_switching\_dc\_current

### niswitch.Session.max\_switching\_dc\_current

This channel-based property returns the maximum DC current the channel can switch. The units are amperes.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].max_switching_dc_current
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics: Maximum Switching DC Current
- C Attribute: NISWITCH\_ATTR\_MAX\_SWITCHING\_DC\_CURRENT

## max\_switching\_dc\_power

## niswitch.Session.max\_switching\_dc\_power

This channel-based property returns the maximum DC power the channel can switch. The units are watts.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].max_switching_dc_power
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics: Maximum Switching DC Power
- C Attribute: NISWITCH\_ATTR\_MAX\_SWITCHING\_DC\_POWER

### number\_of\_relays

niswitch.Session.number\_of\_relays

This property returns the number of relays.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics: Number of Relays
- C Attribute: NISWITCH\_ATTR\_NUMBER\_OF\_RELAYS

### num of columns

niswitch.Session.num of columns

This property returns the number of channels on the column of a matrix or scanner. If the switch device is a scanner, this value is the number of input channels. The <code>niswitch.Session.wire\_mode</code> property affects the number of available columns. For example, if your device has 8 input lines and you use the four-wire mode, then the number of columns you have available is 2.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Matrix Configuration: Number of Columns
- C Attribute: NISWITCH\_ATTR\_NUM\_OF\_COLUMNS

#### num of rows

#### niswitch. Session.num of rows

This property returns the number of channels on the row of a matrix or scanner. If the switch device is a scanner, this value is the number of output channels. The <code>niswitch.Session.wire\_mode</code> property affects the number of available rows. For example, if your device has 8 input lines and you use the two-wire mode, then the number of columns you have available is 4.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Matrix Configuration: Number of Rows
- C Attribute: NISWITCH\_ATTR\_NUM\_OF\_ROWS

## power\_down\_latching\_relays\_after\_debounce

### niswitch.Session.power\_down\_latching\_relays\_after\_debounce

This property specifies whether to power down latching relays after calling Wait For Debounce. When Power Down Latching Relays After Debounce is enabled (True), a call to Wait For Debounce ensures that the relays are settled and the latching relays are powered down.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Power Down Latching Relays After Debounce
- C Attribute: NISWITCH\_ATTR\_POWER\_DOWN\_LATCHING\_RELAYS\_AFTER\_DEBOUNCE

### scan advanced output

#### niswitch.Session.scan\_advanced\_output

This property specifies the method you want to use to notify another instrument that all signals going through the switch device have settled following the processing of one entry in the scan list.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ScanAdvancedOutput
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Scanning Configuration:Scan Advanced Output
- C Attribute: NISWITCH\_ATTR\_SCAN\_ADVANCED\_OUTPUT

### scan\_advanced\_polarity

#### niswitch. Session.scan advanced polarity

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ScanAdvancedPolarity
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Scanning Configuration: Scan Advanced Polarity
- C Attribute: NISWITCH\_ATTR\_SCAN\_ADVANCED\_POLARITY

### scan delay

## $\verb|niswitch.Session.scan_delay| \\$

This property specifies the minimum amount of time the switch device waits before it asserts the scan advanced output trigger after opening or closing the switch. The switch device always waits for debounce before asserting the trigger. The units are seconds. the greater value of the settling time and the value you specify as the scan delay.

Note: NI PXI-2501/2503/2565/2590/2591 Users—the actual delay will always be

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• LabVIEW Property: Scanning Configuration:Scan Delay

• C Attribute: NISWITCH\_ATTR\_SCAN\_DELAY

### scan list

#### niswitch. Session.scan list

This property contains a scan list, which is a string that specifies channel connections and trigger conditions. The niswitch.Session.initiate() method makes or breaks connections and waits for triggers according to the instructions in the scan list. The scan list is comprised of channel names that you separate with special characters. These special characters determine the operations the scanner performs on the channels when it executes this scan list. To create a path between two channels, use the following character between the two channel names: -> (a dash followed by a '>' sign) Example: 'CH1->CH2' tells the switch to make a path from channel CH1 to channel CH2. To break or clear a path, use the following character as a prefix before the path: ~ (tilde) Example: '~CH1->CH2' tells the switch to break the path from channel CH1 to channel CH2. To tell the switch device to wait for a trigger event, use the following character as a separator between paths: ; (semi-colon) Example: 'CH1->CH2;CH3->CH4' tells the switch to make the path from channel CH1 to channel CH2, wait for a trigger, and then make the path from CH3 to CH4.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

LabVIEW Property: Scanning Configuration:Scan List

C Attribute: NISWITCH\_ATTR\_SCAN\_LIST

#### scan mode

#### niswitch.Session.scan\_mode

This property specifies what happens to existing connections that conflict with the connections you make in a scan list. For example, if CH1 is already connected to CH2 and the scan list instructs the switch device to connect CH1 to CH3, this property specifies what happens to the connection between CH1 and CH2. If the value of this property is *NONE*, the switch device takes no action on

existing paths. If the value is <code>BREAK\_BEFORE\_MAKE</code>, the switch device breaks conflicting paths before making new ones. If the value is <code>BREAK\_AFTER\_MAKE</code>, the switch device breaks conflicting paths after making new ones. Most switch devices support only one of the possible values. In such cases, this property serves as an indicator of the device's behavior.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.ScanMode
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Scanning Configuration:Scan Mode
- C Attribute: NISWITCH\_ATTR\_SCAN\_MODE

## serial\_number

niswitch.Session.serial\_number

This read-only property returns the serial number for the switch device controlled by this instrument driver. If the device does not return a serial number, the driver returns the IVI\_ERROR\_ATTRIBUTE\_NOT\_SUPPORTED error.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Serial Number
- C Attribute: NISWITCH\_ATTR\_SERIAL\_NUMBER

### settling time

niswitch.Session.settling\_time

This channel-based property returns the maximum length of time from after you make a connection

until the signal flowing through the channel settles. The units are seconds. the greater value of the settling time and the value you specify as the scan delay.

**Note:** NI PXI-2501/2503/2565/2590/2591 Users—the actual delay will always be

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
session.channels[0,1].settling_time = var
var = session.channels[0,1].settling_time
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float in seconds or datetime.timedelta
Permissions	read-write
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Settling Time
- C Attribute: NISWITCH\_ATTR\_SETTLING\_TIME

### simulate

niswitch.Session.simulate

Specifies whether or not to simulate instrument driver I/O operations. If simulation is enabled, instrument driver methods perform range checking and call Ivi\_GetAttribute and Ivi\_SetAttribute methods, but they do not perform instrument I/O. For output parameters that represent instrument data, the instrument driver methods return calculated values. The default value is False. Use the niswitch.Session.InitWithOptions() method to override this value.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	bool
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes: User Options: Simulate
- C Attribute: NISWITCH\_ATTR\_SIMULATE

## specific\_driver\_description

## niswitch.Session.specific\_driver\_description

A string that contains a brief description of the specific driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Identification:Description
- C Attribute: NISWITCH\_ATTR\_SPECIFIC\_DRIVER\_DESCRIPTION

## specific\_driver\_revision

### niswitch.Session.specific\_driver\_revision

A string that contains additional version information about this instrument driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Identification:Revision
- C Attribute: NISWITCH\_ATTR\_SPECIFIC\_DRIVER\_REVISION

## specific\_driver\_vendor

## niswitch.Session.specific\_driver\_vendor

A string that contains the name of the vendor that supplies this driver.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Identification:Driver Vendor
- C Attribute: NISWITCH\_ATTR\_SPECIFIC\_DRIVER\_VENDOR

## supported\_instrument\_models

niswitch.Session.supported\_instrument\_models

Contains a comma-separated list of supported instrument models.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Inherent IVI Attributes:Driver Capabilities:Supported Instrument Models
- $\bullet \ \ C \ Attribute: \ \textbf{NISWITCH\_ATTR\_SUPPORTED\_INSTRUMENT\_MODELS}$

## temperature

niswitch.Session.temperature

This property returns the temperature as read by the Switch module. The units are degrees Celsius.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	float
Permissions	read only
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics:Temperature
- C Attribute: NISWITCH\_ATTR\_TEMPERATURE

## trigger\_input

### niswitch. Session.trigger input

This property specifies the source of the trigger for which the switch device can wait when processing a scan list. The switch device waits for a trigger when it encounters a semi-colon in a scan list. When the trigger occurs, the switch device advances to the next entry in the scan list.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerInput
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Scanning Configuration: Trigger Input
- C Attribute: NISWITCH\_ATTR\_TRIGGER\_INPUT

## trigger\_input\_polarity

niswitch.Session.trigger\_input\_polarity

Determines the behavior of the trigger Input.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	enums.TriggerInputPolarity
Permissions	read-write
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Scanning Configuration: Trigger Input Polarity
- C Attribute: NISWITCH\_ATTR\_TRIGGER\_INPUT\_POLARITY

## wire\_mode

niswitch.Session.wire\_mode

This property returns the wire mode of the switch device. This property affects the values of the

niswitch. Session.num\_of\_rows and niswitch. Session.num\_of\_columns properties. The actual number of input and output lines on the switch device is fixed, but the number of channels depends on how many lines constitute each channel.

**Tip:** This property can use repeated capabilities (channels). If set or get directly on the niswitch. Session object, then the set/get will use all repeated capabilities in the session. You can specify a subset of repeated capabilities using the Python index notation on an niswitch. Session repeated capabilities container, and calling set/get value on the result.:

```
var = session.channels[0,1].wire_mode
```

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read only
Channel Based	True
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

- LabVIEW Property: Module Characteristics: Wire mode
- C Attribute: NISWITCH ATTR WIRE MODE

### **Methods**

## abort

```
niswitch.Session.abort()
```

Aborts the scan in progress. Initiate a scan with niswitch.Session.initiate(). If the switch module is not scanning, NISWITCH\_ERROR\_NO\_SCAN\_IN\_PROGRESS error is returned.

## can connect

```
niswitch.Session.can_connect(channel1, channel2)
```

Verifies that a path between channel 1 and channel 2 can be created. If a path is possible in the switch module, the availability of that path is returned given the existing connections. If the path is possible but in use, a NISWITCH\_WARN\_IMPLICIT\_CONNECTION\_EXISTS warning is returned.

#### **Parameters**

- **channel1** (*stx*) Input one of the channel names of the desired path. Pass the other channel name as the channel 2 parameter. Refer to Devices Overview for valid channel names for the switch module. Examples of valid channel names: ch0, com0, ab0, r1, c2, cjtemp Default value: ""
- **channel2** (str) Input one of the channel names of the desired path. Pass the other channel name as the channel 1 parameter. Refer to Devices Overview for

valid channel names for the switch module. Examples of valid channel names: ch0, com0, ab0, r1, c2, cjtemp Default value: ""

Return type niswitch.PathCapability

#### Returns

Indicates whether path is valid. Possible values include: NISWITCH VAL PATH AVAILABLE NISWITCH VAL PATH EXISTS 2 NISWITCH VAL PATH UNSUPPORTED 3 4 NISWITCH VAL SOURCE CONFLICT NISWITCH\_VAL\_RSRC\_IN\_USE NISWITCH\_VAL\_CHANNEL\_NOT\_AVAILABLE Notes: NISWITCH\_VAL\_PATH\_AVAILABLE indicates that the driver can create the path at this time. (2) NISWITCH\_VAL\_PATH\_EXISTS indicates that the path already exists. (3) NISWITCH\_VAL\_PATH\_UNSUPPORTED indicates that the instrument is not capable of creating a path between the channels you specify. (4) NISWITCH\_VAL\_RSRC\_IN\_USE indicates that although the path is valid, the driver cannot create the path at this moment because the switch device is currently using one or more of the required channels to create another path. You must destroy the other path before creating this one. (5) NISWITCH VAL SOURCE CONFLICT indicates that the instrument cannot create a path because both channels are connected to a different source channel. (6) NISWITCH VAL CHANNEL NOT AVAILABLE indicates that the driver cannot create a path between the two channels because one of the channels is a configuration channel and thus unavailable for external connections.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

#### commit

```
niswitch.Session.commit()
```

Downloads the configured scan list and trigger settings to hardware. Calling <code>niswitch.Session.commit()</code> optional as it is implicitly called during <code>niswitch.Session.initiate()</code>. Use <code>niswitch.Session.commit()</code> to arm triggers in a given order or to control when expensive hardware operations are performed.

### connect

```
niswitch.Session.connect(channel1, channel2)
```

Creates a path between channel 1 and channel 2. The driver calculates and uses the shortest path between the two channels. Refer to Immediate Operations for information about Channel Usage types. If a path is not available, the method returns one of the following errors: - NISWITCH\_ERROR\_EXPLICIT\_CONNECTION\_EXISTS, if the two channels are already explicitly connected by calling either the niswitch. Session.connect() or niswitch.Session.set\_path() method. NISWITCH\_ERROR\_IS\_CONFIGURATION\_CHANNEL, if a channel is a configuration channel. Error elaboration contains information about which of the two channels is a configuration channel. - NISWITCH\_ERROR\_ATTEMPT\_TO\_CONNECT\_SOURCES, if both channels are connected to a different source. Error elaboration contains information about sources channel 1 and 2 connect to. - NISWITCH\_ERROR\_CANNOT\_CONNECT\_TO\_ITSELF, if channels 1 and 2 are one and the same channel. - NISWITCH ERROR PATH NOT FOUND, if the driver cannot

find a path between the two channels. Note: Paths are bidirectional. For example, if a path exists between channels CH1 and CH2, then the path also exists between channels CH2 and CH1.

#### **Parameters**

- **channel1** (*str*) Input one of the channel names of the desired path. Pass the other channel name as the channel 2 parameter. Refer to Devices Overview for valid channel names for the switch module. Examples of valid channel names: ch0, com0, ab0, r1, c2, cjtemp Default value: None
- **channel2** (str) Input one of the channel names of the desired path. Pass the other channel name as the channel 1 parameter. Refer to Devices Overview for valid channel names for the switch module. Examples of valid channel names: ch0, com0, ab0, r1, c2, cjtemp Default value: None

### connect multiple

```
niswitch.Session.connect_multiple(connection_list)
```

Creates the connections between channels specified in Connection List. Specify connections with two endpoints only or the explicit path between two endpoints. NI-SWITCH calculates and uses the shortest path between the channels. Refer to Setting Source and Configuration Channels for information about channel usage types. In the event of an error, connecting stops at the point in the list where the error occurred. If a path is not available, the method returns one of the following errors: - NISWITCH\_ERROR\_EXPLICIT\_CONNECTION\_EXISTS, if the two channels are already explicitly connected. - NISWITCH\_ERROR\_IS\_CONFIGURATION\_CHANNEL, if a channel is a configuration channel. Error elaboration contains information about which of the two channels is a configuration channel. - NISWITCH\_ERROR\_ATTEMPT\_TO\_CONNECT\_SOURCES, if both channels are connected to a different source. Error elaboration contains information about sources channel 1 and 2 to connect. - NISWITCH\_ERROR\_CANNOT\_CONNECT\_TO\_ITSELF, if channels 1 and 2 are one and the same channel. - NISWITCH\_ERROR\_PATH\_NOT\_FOUND, if the driver cannot find a path between the two channels. Note: Paths are bidirectional. For example, if a path exists between channels ch1 and ch2, then the path also exists between channels ch1 and ch2.

**Parameters connection\_list** (str) – Connection List specifies a list of connections between channels to make. NI-SWITCH validates the connection list, and aborts execution of the list if errors are returned. Refer to Connection and Disconnection List Syntax for valid connection list syntax and examples. Refer to Devices Overview for valid channel names for the switch module. Example of a valid connection list: c0 -> r1, [c2 -> r2 -> c3] In this example, r2 is a configuration channel. Default value: None

#### disable

```
niswitch.Session.disable()
```

Places the switch module in a quiescent state where it has minimal or no impact on the system to which it is connected. All channels are disconnected and any scan in progress is aborted.

### disconnect

```
niswitch.Session.disconnect(channel1, channel2)
```

This method destroys the path between two channels that you create with the <code>niswitch.Session.connect()</code> or <code>niswitch.Session.set\_path()</code> method. If a path is not connected or not available, the method returns the IVISWTCH ERROR NO SUCH PATH error.

#### **Parameters**

- **channel1** (*str*) Input one of the channel names of the path to break. Pass the other channel name as the channel 2 parameter. Refer to Devices Overview for valid channel names for the switch module. Examples of valid channel names: ch0, com0, ab0, r1, c2, cjtemp Default value: None
- **channel2** (*str*) Input one of the channel names of the path to break. Pass the other channel name as the channel 1 parameter. Refer to Devices Overview for valid channel names for the switch module. Examples of valid channel names: ch0, com0, ab0, r1, c2, cjtemp Default value: None

## disconnect all

```
niswitch.Session.disconnect_all()
```

Breaks all existing paths. If the switch module cannot break all paths, NISWITCH\_WARN\_PATH\_REMAINS warning is returned.

## disconnect\_multiple

```
niswitch.Session.disconnect multiple (disconnection list)
```

Breaks the connections between channels specified in Disconnection List. If no connections exist between channels, NI-SWITCH returns an error. In the event of an error, the VI stops at the point in the list where the error occurred.

**Parameters disconnection\_list** (str) – Disconnection List specifies a list of connections between channels to break. NI-SWITCH validates the disconnection list, and aborts execution of the list if errors are returned. Refer to Connection and Disconnection List Syntax for valid disconnection list syntax and examples. Refer to Devices Overview for valid channel names for the switch module. Example of a valid disconnection list: c0 -> r1, [c2 -> r2 -> c3] In this example, r2 is a configuration channel. Default value: None

## get\_channel\_name

```
niswitch.Session.get_channel_name (index)
```

Returns the channel string that is in the channel table at the specified index. Use <code>niswitch.Session.get\_channel\_name()</code> in a For Loop to get a complete list of valid channel names for the switch module. Use the Channel Count property to determine the number of channels.

**Parameters index** (*int*) – A 1-based index into the channel table. Default value: 1 Maximum value: Value of Channel Count property.

### get path

```
niswitch.Session.get_path(channel1, channel2)
```

Returns a string that identifies the explicit path created with <code>niswitch.Session.connect()</code>. Pass this string to <code>niswitch.Session.set\_path()</code> to establish the exact same path in future connections. In some cases, multiple paths are available between two channels. When you call <code>niswitch.Session.connect()</code>, the driver selects an available path. With <code>niswitch.Session.connect()</code>, there is no guarantee that the driver selected path will always be the same path through the switch module. <code>niswitch.Session.get\_path()</code> only returns those paths

explicitly created by niSwitch Connect Channels or niswitch. Session.set\_path(). For example, if you connect channels CH1 and CH3, and then channels CH2 and CH3, an explicit path between channels CH1 and CH2 does not exist an error is returned

#### **Parameters**

- **channel1** (*str*) Input one of the channel names of the desired path. Pass the other channel name as the channel 2 parameter. Refer to Devices Overview for valid channel names for the switch module. Examples of valid channel names: ch0, com0, ab0, r1, c2, citemp Default value: ""
- **channel2** (*str*) Input one of the channel names of the desired path. Pass the other channel name as the channel 1 parameter. Refer to Devices Overview for valid channel names for the switch module. Examples of valid channel names: ch0, com0, ab0, r1, c2, cjtemp Default value: ""

## get\_relay\_count

```
niswitch.Session.get_relay_count(relay_name)
```

Returns the number of times the relay has changed from Closed to Open. Relay count is useful for tracking relay lifetime and usage. Call <code>niswitch.Session.wait\_for\_debounce()</code> before <code>niswitch.Session.get\_relay\_count()</code> to ensure an accurate count. Refer to the Relay Count topic in the NI Switches Help to determine if the switch module supports relay counting.

**Parameters relay\_name** (str) – Name of the relay. Default value: None Examples of valid relay names: ch0, ab0, 1wire, hlselect Refer to Devices Overview for a list of valid relay names for the switch module.

Return type int

**Returns** The number of relay cycles.

## get\_relay\_name

```
niswitch.Session.get_relay_name (index)
```

Returns the relay name string that is in the relay list at the specified index. Use niswitch. Session.get\_relay\_name() in a For Loop to get a complete list of valid relay names for the switch module. Use the Number of Relays property to determine the number of relays.

**Parameters index** (*int*) – A 1-based index into the channel table. Default value: 1 Maximum value: Value of Channel Count property.

### get relay position

```
niswitch.Session.get_relay_position(relay_name)
```

Returns the relay position for the relay specified in the Relay Name parameter.

**Parameters relay\_name** (str) – Name of the relay. Default value: None Examples of valid relay names: ch0, ab0, 1wire, hlselect Refer to Devices Overview for a list of valid relay names for the switch module.

Return type niswitch. RelayPosition

## Returns

Indicates whether the relay is open or closed.  $\label{eq:niswitch_val_open} \mbox{10 NI-WITCH\_VAL\_CLOSED 11}$ 

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

#### lock

```
niswitch.Session.lock()
```

Obtains a multithread lock on the device session. Before doing so, the software waits until all other execution threads release their locks on the device session.

Other threads may have obtained a lock on this session for the following reasons:

- The application called the *niswitch*. Session.lock() method.
- A call to NI-SWITCH locked the session.
- After a call to the <code>niswitch.Session.lock()</code> method returns successfully, no other threads can access the device session until you call the <code>niswitch.Session.unlock()</code> method or exit out of the with block when using lock context manager.
- Use the niswitch. Session.lock() method and the niswitch. Session.unlock() method around a sequence of calls to instrument driver methods if you require that the device retain its settings through the end of the sequence.

You can safely make nested calls to the <code>niswitch.Session.lock()</code> method within the same thread. To completely unlock the session, you must balance each call to the <code>niswitch.Session.lock()</code> method with a call to the <code>niswitch.Session.unlock()</code> method.

One method for ensuring there are the same number of unlock method calls as there is lock calls is to use lock as a context manager

```
with niswitch.Session('dev1') as session:
    with session.lock():
        # Calls to session within a single lock context
```

The first with block ensures the session is closed regardless of any exceptions raised

The second with block ensures that unlock is called regardless of any exceptions raised

Return type context manager

**Returns** When used in a *with* statement, *niswitch.Session.lock()* acts as a context manager and unlock will be called when the *with* block is exited

## relay\_control

```
niswitch.Session.relay_control(relay_name, relay_action)
```

Controls individual relays of the switch. When controlling individual relays, the protection offered by setting the usage of source channels and configuration channels, and by enabling or disabling analog bus sharing on the NI SwitchBlock, does not apply. Refer to the device book for your switch in the NI Switches Help to determine if the switch supports individual relay control.

### **Parameters**

• **relay\_name** (*str*) – Name of the relay. Default value: None Examples of valid relay names: ch0, ab0, 1wire, hlselect Refer to Devices Overview for a list of valid relay names for the switch module.

• relay\_action (niswitch.RelayAction) - Specifies whether to open or close a given relay. Default value: Relay Close Defined values: NISWITCH\_VAL\_OPEN\_RELAY NISWITCH\_VAL\_CLOSE\_RELAY (Default Value)

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

#### reset

```
niswitch.Session.reset()
```

Disconnects all created paths and returns the switch module to the state at initialization. Configuration channel and source channel settings remain unchanged.

## reset\_with\_defaults

```
niswitch.Session.reset_with_defaults()
```

Resets the switch module and applies initial user specified settings from the logical name used to initialize the session. If the session was created without a logical name, this method is equivalent to niswitch. Session.reset().

### route scan advanced output

Routes the scan advanced output trigger from a trigger bus line (TTLx) to the front or rear connector.

#### **Parameters**

• scan\_advanced\_output\_connector (niswitch. ScanAdvancedOutput) - The scan advanced trigger destination. Valid locations are the FRONTCONNECTOR and REARCONNECTOR. Default value: FRONTCONNECTOR

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

• scan\_advanced\_output\_bus\_line (niswitch. ScanAdvancedOutput) - The trigger line to route the scan advanced output trigger from the front or rear connector. Select NONE to break an existing route. Default value: None Valid Values: NONE TTLO TTL1 TTL2 TTL3 TTL4 TTL5 TTL6 TTL7

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

• invert (bool) - If True, inverts the input trigger signal from falling to rising or vice versa. Default value: False

#### route trigger input

```
\begin{tabular}{ll} niswitch. Session. {\bf route\_trigger\_input} (trigger\_input\_connector, & trigger\_input\_bus\_line, invert=False) \\ \end{tabular}
```

Routes the input trigger from the front or rear connector to a trigger bus line (TTLx). To disconnect the route, call this method again and specify None for trigger bus line parameter.

#### **Parameters**

• trigger\_input\_connector (niswitch.TriggerInput) - The location of the input trigger source on the switch module. Valid locations are the FRONTCONNECTOR and REARCONNECTOR. Default value: FRONTCONNECTOR

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

• trigger\_input\_bus\_line (niswitch.TriggerInput) - The trigger line to route the input trigger. Select NISWITCH\_VAL\_NONE to break an existing route. Default value: None Valid Values: NISWITCH\_VAL\_NONE TTL0 TTL1 TTL2 TTL3 TTL4 TTL5 TTL6 TTL7

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

invert (bool) – If True, inverts the input trigger signal from falling to rising or vice versa. Default value: False

## self\_test

```
niswitch.Session.self_test()
```

Verifies that the driver can communicate with the switch module.

Raises SelfTestError on self test failure. Properties on exception object:

- code failure code from driver
- · message status message from driver

Self-Test Code	Description
0	Passed self-test
1	Self-test failed

### send software trigger

```
niswitch.Session.send_software_trigger()
```

Sends a software trigger to the switch module specified in the NI-SWITCH session. When the trigger input is set to <code>SOFTWARE\_TRIG</code> through either the <code>niswitch.Session.ConfigureScanTrigger()</code> or the <code>niswitch.Session.trigger\_input</code> property, the scan does not proceed from a semi-colon (wait for trigger) until <code>niswitch.Session.send\_software\_trigger()</code> is called.

**Note:** One or more of the referenced methods are not in the Python API for this driver.

### set\_path

```
niswitch.Session.set_path(path_list)
```

Connects two channels by specifying an explicit path in the path list parameter. niswitch. Session.set\_path() is particularly useful where path repeatability is important, such as in calibrated signal paths. If this is not necessary, use niswitch.Session.connect().

**Parameters** path\_list (str) - A string composed of comma-separated paths between channel 1 and channel 2. The first and last names in the path are the endpoints of the path. Every other channel in the path are configuration channels. Example of a valid path list string: ch0->com0, com0->ab0. In this example, com0 is a configuration channel. Default value: None Obtain the path list for a previously created path with niswitch.Session.get\_path().

### unlock

```
niswitch.Session.unlock()
```

Releases a lock that you acquired on an device session using niswitch. Session.lock(). Refer to niswitch. Session.unlock() for additional information on session locks.

### wait for debounce

niswitch.Session.wait\_for\_debounce (maximum\_time\_ms=datetime.timedelta(milliseconds=5000))

Pauses until all created paths have settled. If the time you specify with the Maximum
Time (ms) parameter elapsed before the switch paths have settled, this method returns the
NISWITCH\_ERROR\_MAX\_TIME\_EXCEEDED error.

Parameters maximum\_time\_ms (float in seconds or datetime. timedelta) - Specifies the maximum length of time to wait for all relays in the switch module to activate or deactivate. If the specified time elapses before all relays active or deactivate, a timeout error is returned. Default Value:5000 ms

### wait\_for\_scan\_complete

niswitch.Session.wait\_for\_scan\_complete (maximum\_time\_ms=datetime.timedelta(milliseconds=5000))

Pauses until the switch module stops scanning or the maximum time has elapsed and returns a timeout error. If the time you specify with the Maximum Time (ms) parameter elapsed before the scanning operation has finished, this method returns the NISWITCH\_ERROR\_MAX\_TIME\_EXCEEDED error.

Parameters maximum\_time\_ms (float in seconds or datetime. timedelta) - Specifies the maximum length of time to wait for the switch module to stop scanning. If the specified time elapses before the scan ends, NISWITCH\_ERROR\_MAX\_TIME\_EXCEEDED error is returned. Default Value:5000 ms

# **Properties**

Property	Datatype
niswitch.Session.analog_bus_sharing_enable	bool
niswitch.Session.bandwidth	float
niswitch.Session.channel_count	int
niswitch.Session.characteristic_impedance	float
niswitch.Session.continuous_scan	bool
niswitch.Session.digital_filter_enable	bool
niswitch.Session.driver_setup	str
niswitch.Session.handshaking_initiation	HandshakingInitiation
niswitch.Session.instrument_firmware_revision	str
niswitch.Session.instrument_manufacturer	str
niswitch.Session.instrument_model	str
niswitch.Session.io_resource_descriptor	str
niswitch.Session.is_configuration_channel	bool
niswitch.Session.is_debounced	bool
niswitch.Session.is_aepounced niswitch.Session.is_scanning	bool
niswitch.Session.is_source_channel	bool
niswitch.Session.is_waiting_for_trig	bool
niswitch.Session.logical_name	str
niswitch.Session.max_ac_voltage	float
niswitch.Session.max_carry_ac_current	float
niswitch.Session.max_carry_ac_power	float
niswitch.Session.max_carry_dc_current	float
niswitch.Session.max_carry_dc_power	float
niswitch.Session.max_dc_voltage	float
niswitch.Session.max_switching_ac_current	float
niswitch.Session.max_switching_ac_power	float
niswitch.Session.max_switching_dc_current	float
niswitch.Session.max_switching_dc_power	float
niswitch.Session.number_of_relays	int
niswitch.Session.num_of_columns	int
niswitch.Session.num of rows	int
niswitch.Session.power_down_latching_relays_after_debounce	bool
niswitch.Session.scan_advanced_output	ScanAdvancedOutput
niswitch.Session.scan_advanced_polarity	ScanAdvancedPolarity
niswitch.Session.scan_delay	float in seconds or datetime.timedelta
niswitch.Session.scan_list	str
niswitch.Session.scan_mode	ScanMode
niswitch.Session.serial_number	str
niswitch.Session.settling_time	float in seconds or datetime.timedelta
niswitch.Session.simulate	bool
niswitch.Session.specific_driver_description	str
niswitch.Session.specific_driver_revision	str
niswitch.Session.specific_driver_vendor	str
niswitch.Session.supported_instrument_models	str
niswitch.Session.temperature	float
niswitch.Session.trigger_input	TriggerInput
niswitch.Session.trigger_input_polarity	TriggerInputPolarity

Continued on next page

Table 16 – continued from previous page

Property	Datatype
niswitch.Session.wire_mode	int

#### Methods

Method name
niswitch.Session.abort()
niswitch.Session.can_connect()
niswitch.Session.commit()
niswitch.Session.connect()
niswitch.Session.connect_multiple()
niswitch.Session.disable()
niswitch.Session.disconnect()
niswitch.Session.disconnect_all()
niswitch.Session.disconnect_multiple()
niswitch.Session.get_channel_name()
niswitch.Session.get_path()
niswitch.Session.get_relay_count()
niswitch.Session.get_relay_name()
niswitch.Session.get_relay_position()
niswitch.Session.lock()
niswitch.Session.relay_control()
niswitch.Session.reset()
niswitch.Session.reset_with_defaults()
<pre>niswitch.Session.route_scan_advanced_output()</pre>
<pre>niswitch.Session.route_trigger_input()</pre>
niswitch.Session.self_test()
<pre>niswitch.Session.send_software_trigger()</pre>
niswitch.Session.set_path()
niswitch.Session.unlock()
niswitch.Session.wait_for_debounce()
niswitch.Session.wait_for_scan_complete()

## 7.5.3.2 Enums

Enums used in NI-SWITCH

# HandshakingInitiation

 ${\bf class}\ {\tt niswitch.} \\ {\bf Handshaking Initiation}$ 

## MEASUREMENT\_DEVICE

The niSwitch Initiate Scan <switchviref.chm::/:py:meth:'niswitch.Session.Initiate\_Scan.html>'\_\_ VI does not return until the switch hardware is waiting for a trigger input. This ensures that if you initiate the measurement device after calling the niSwitch Initiate Scan <switchviref.chm::/:py:meth:'niswitch.Session.Initiate\_Scan.html>'\_\_ VI , the switch is sure to receive the first measurement complete (MC) signal sent by the measurement device. The measurement

device should be configured to first take a measurement, send MC, then wait for scanner advanced output signal. Thus, the first MC of the measurement device initiates handshaking.

#### SWITCH

The niSwitch Initiate Scan <switchviref.chm::/:py:meth:'niswitch.Session.Initiate\_Scan.html>'\_\_ VI returns immediately after beginning scan list execution. It is assumed that the measurement device has already been configured and is waiting for the scanner advanced signal. The measurement should be configured to first wait for a trigger, then take a measurement. Thus, the first scanner advanced output signal of the switch module initiates handshaking.

### **PathCapability**

```
class niswitch.PathCapability
```

```
PATH_AVAILABLE
```

Path Available

PATH EXISTS

Path Exists

### PATH UNSUPPORTED

Path Unsupported

### RESOURCE\_IN\_USE

Resource in use

#### SOURCE CONFLICT

Source conflict

## CHANNEL\_NOT\_AVAILABLE

Channel not available

## RelayAction

```
class niswitch.RelayAction
```

OPEN

Open Relay

CLOSE

Close Relay

## RelayPosition

class niswitch.RelayPosition

OPEN

Open

CLOSED

Closed

### ScanAdvancedOutput

#### class niswitch.ScanAdvancedOutput

#### NONE

The switch device does not produce a Scan Advanced Output trigger.

#### **EXTERNAL**

External Trigger. The switch device produces the Scan Advanced Output trigger on the external trigger output.

#### TTL0

The switch device produces the Scan Advanced Output on the PXI TRIGO line.

#### TTL1

The switch device produces the Scan Advanced Output on the PXI TRIG1 line.

#### TTL2

The switch device produces the Scan Advanced Output on the PXI TRIG2 line.

#### TTL3

The switch device produces the Scan Advanced Output on the PXI TRIG3 line.

#### TTL4

The switch device produces the Scan Advanced Output on the PXI TRIG4 line.

#### TTL5

The switch device produces the Scan Advanced Output on the PXI TRIG5 line.

#### TTL6

The switch device produces the Scan Advanced Output on the PXI TRIG6 line.

#### TTL7

The switch device produces the Scan Advanced Output on the PXI TRIG7 line.

## PXI\_STAR

The switch module produces the Scan Advanced Output Trigger on the PXI Star trigger bus before processing the next entry in the scan list.

### REARCONNECTOR

The switch device produces the Scan Advanced Output trigger on the rear connector.

### FRONTCONNECTOR

The switch device produces the Scan Advanced Output trigger on the front connector.

### REARCONNECTOR MODULE1

The switch module produces the Scan Advanced Output Trigger on the rear connector module 1.

#### REARCONNECTOR MODULE2

The switch module produces the Scan Advanced Output Trigger on the rear connector module 2.

### REARCONNECTOR\_MODULE3

The switch module produces the Scan Advanced Output Trigger on the rear connector module 3.

### REARCONNECTOR MODULE 4

The switch module produces the Scan Advanced Output Trigger on the rear connector module 4.

### REARCONNECTOR\_MODULE5

The switch module produces the Scan Advanced Output Trigger on the rear connector module 5.

## REARCONNECTOR\_MODULE 6

The switch module produces the Scan Advanced Output Trigger on the rear connector module 6.

#### REARCONNECTOR MODULE7

The switch module produces the Scan Advanced Output Trigger on the rear connector module 7.

#### REARCONNECTOR MODULE8

The switch module produces the Scan Advanced Output Trigger on the rear connector module 8.

#### REARCONNECTOR MODULE 9

The switch module produces the Scan Advanced Ouptut Trigger on the rear connector module 9.

#### REARCONNECTOR MODULE10

The switch module produces the Scan Advanced Output Trigger on the rear connector module 10.

#### REARCONNECTOR MODULE11

The switch module produces the Scan Advanced Output Trigger on the rear connector module 11.

#### REARCONNECTOR MODULE12

The switch module produces the Scan Advanced Output Trigger on the rear connector module 12.

#### FRONTCONNECTOR\_MODULE1

The switch module produces the Scan Advanced Output Trigger on the front connector module 1.

#### FRONTCONNECTOR MODULE2

The switch module produces the Scan Advanced Output Trigger on the front connector module 2.

#### FRONTCONNECTOR MODULE3

The switch module produces the Scan Advanced Output Trigger on the front connector module 3.

#### FRONTCONNECTOR MODULE4

The switch module produces the Scan Advanced Output Trigger on the front connector module 4.

#### FRONTCONNECTOR MODULE5

The switch module produces the Scan Advanced Output Trigger on the front connector module 5.

### FRONTCONNECTOR\_MODULE 6

The switch module produces the Scan Advanced Output Trigger on the front connector module 6.

#### FRONTCONNECTOR MODULE7

The switch module produces the Scan Advanced Output Trigger on the front connector module 7.

### FRONTCONNECTOR MODULE8

The switch module produces the Scan Advanced Output Trigger on the front connector module 8.

## FRONTCONNECTOR\_MODULE 9

The switch module produces the Scan Advanced Output Trigger on the front connector module 9.

### FRONTCONNECTOR MODULE10

The switch module produces the Scan Advanced Output Trigger on the front connector module 10.

#### FRONTCONNECTOR MODULE11

The switch module produces the Scan Advanced Output Trigger on the front connector module 11.

### FRONTCONNECTOR MODULE12

The switch module produces the Scan Advanced Output Trigger on the front connector module 12.

### ScanAdvancedPolarity

### class niswitch.ScanAdvancedPolarity

### RISING

The trigger occurs on the rising edge of the signal.

#### **FALLING**

The trigger occurs on the falling edge of the signal.

#### ScanMode

#### class niswitch.ScanMode

#### NONE

No implicit action on connections when scanning.

### BREAK\_BEFORE\_MAKE

When scanning, the switch device breaks existing connections before making new connections.

#### BREAK AFTER MAKE

When scanning, the switch device breaks existing connections after making new connections.

## **TriggerInput**

### class niswitch.TriggerInput

#### IMMEDIATE

Immediate Trigger. The switch device does not wait for a trigger before processing the next entry in the scan list.

#### **EXTERNAL**

External Trigger. The switch device waits until it receives a trigger from an external source through the external trigger input before processing the next entry in the scan list.

### SOFTWARE TRIG

The switch device waits until you call the <code>niswitch.Session.send\_software\_trigger()</code> method before processing the next entry in the scan list.

### TTLO

The switch device waits until it receives a trigger on the PXI TRIGO line before processing the next entry in the scan list.

#### TTL1

The switch device waits until it receives a trigger on the PXI TRIG1 line before processing the next entry in the scan list.

### TTL2

The switch device waits until it receives a trigger on the PXI TRIG2 line before processing the next entry in the scan list.

#### TTL3

The switch device waits until it receives a trigger on the PXI TRIG3 line before processing the next entry in the scan list.

#### TTL4

The switch device waits until it receives a trigger on the PXI TRIG4 line before processing the next entry in the scan list.

## TTL5

The switch device waits until it receives a trigger on the PXI TRIG5 line before processing the next entry in the scan list.

#### TTL6

The switch device waits until it receives a trigger on the PXI TRIG6 line before processing the next entry in the scan list.

#### **TT.7**

The switch device waits until it receives a trigger on the PXI TRIG7 line before processing the next entry in the scan list.

#### PXI STAR

The switch device waits until it receives a trigger on the PXI STAR trigger bus before processing the next entry in the scan list.

#### REARCONNECTOR

The switch device waits until it receives a trigger on the rear connector.

#### FRONTCONNECTOR

The switch device waits until it receives a trigger on the front connector.

#### REARCONNECTOR MODULE1

The switch module waits until it receives a trigger on the rear connector module 1.

#### REARCONNECTOR MODULE2

The switch module waits until it receives a trigger on the rear connector module 2.

#### REARCONNECTOR MODULE3

The switch module waits until it receives a trigger on the rear connector module 3.

#### REARCONNECTOR MODULE 4

The switch module waits until it receives a trigger on the rear connector module 4.

#### REARCONNECTOR MODULE5

The switch module waits until it receives a trigger on the rear connector module 5.

### REARCONNECTOR\_MODULE 6

The switch module waits until it receives a trigger on the rear connector module 6.

#### REARCONNECTOR MODULE7

The switch module waits until it receives a trigger on the rear connector module 7.

#### REARCONNECTOR\_MODULE8

The switch module waits until it receives a trigger on the rear connector module 8.

### REARCONNECTOR MODULE 9

The switch module waits until it receives a trigger on the rear connector module 9.

#### REARCONNECTOR MODULE10

The switch module waits until it receives a trigger on the rear connector module 10.

### REARCONNECTOR MODULE11

The switch module waits until it receives a trigger on the rear connector module 11.

## REARCONNECTOR\_MODULE12

The switch module waits until it receives a trigger on the rear connector module 12.

### FRONTCONNECTOR\_MODULE1

The switch module waits until it receives a trigger on the front connector module 1.

#### FRONTCONNECTOR\_MODULE2

The switch module waits until it receives a trigger on the front connector module 2.

### FRONTCONNECTOR\_MODULE3

The switch module waits until it receives a trigger on the front connector module 3.

#### FRONTCONNECTOR MODULE 4

The switch module waits until it receives a trigger on the front connector module 4.

## FRONTCONNECTOR\_MODULE5

The switch module waits until it receives a trigger on the front connector module 5.

#### FRONTCONNECTOR MODULE 6

The switch module waits until it receives a trigger on the front connector module 6.

#### FRONTCONNECTOR MODULE7

The switch module waits until it receives a trigger on the front connector module 7.

#### FRONTCONNECTOR MODULE8

The switch module waits until it receives a trigger on the front connector module 8.

#### FRONTCONNECTOR MODULE 9

The switch module waits until it receives a trigger on the front connector module 9.

### FRONTCONNECTOR\_MODULE10

The switch module waits until it receives a trigger on the front connector module 10.

#### FRONTCONNECTOR MODULE11

The switch module waits until it receives a trigger on the front connector module 11.

#### FRONTCONNECTOR MODULE12

The switch module waits until it receives a trigger on the front connector module 12.

## **TriggerInputPolarity**

### class niswitch.TriggerInputPolarity

### RISING

The trigger occurs on the rising edge of the signal.

#### **FALLING**

The trigger occurs on the falling edge of the signal.

### 7.5.3.3 Exceptions and Warnings

#### **DriverError**

### exception niswitch.DriverError

An error originating from the NI-SWITCH driver

## UnsupportedConfigurationError

## $\textbf{exception} \ \, \texttt{niswitch}. \textbf{UnsupportedConfigurationError}$

An error due to using this module in an usupported platform.

#### **DriverNotInstalledError**

### exception niswitch.DriverNotInstalledError

An error due to using this module without the driver runtime installed.

## InvalidRepeatedCapabilityError

```
exception niswitch.InvalidRepeatedCapabilityError An error due to an invalid character in a repeated capability
```

#### SelfTestError

```
exception niswitch.SelfTestError
An error due to a failed self-test
```

## **DriverWarning**

```
exception niswitch.DriverWarning
   A warning originating from the NI-SWITCH driver
```

## **7.5.3.4 Examples**

## niswitch\_connect\_channels.py

Listing 11: (niswitch\_connect\_channels.py)

```
#!/usr/bin/python
2
   import argparse
   import niswitch
   import sys
   def example(resource_name, channel1, channel2, topology, simulate):
       # if we are simulating resource name must be blank
9
       resource_name = '' if simulate else resource_name
10
11
       with niswitch.Session(resource_name=resource_name, topology=topology,_
12
   ⇒simulate=simulate) as session:
           session.connect(channel1=channel1, channel2=channel2)
13
           print('Channel', channel1, ' and ', channel2, ' are now connected.')
14
           session.disconnect(channel1=channel1, channel2=channel2)
15
           print('Channel', channel1, ' and ', channel2, ' are now disconnected.')
17
18
   def _main(argsv):
19
       parser = argparse.ArgumentParser(description='Performs a connection with NI-
20
   →SWITCH Channels.', formatter_class=argparse.ArgumentDefaultsHelpFormatter)
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource_
21
   →name of a National Instruments Switch.')
       parser.add_argument('-ch1', '--channel1', default='c0', help='Channel One.')
22
       parser.add_argument('-ch2', '--channel2', default='r0', help='Channel Two.')
23
       parser.add_argument('-t', '--topology', default='Configured Topology', help=
24
   →'Topology.')
       parser.add_argument('-s', '--simulate', default=False, action='store_true', help=
25
   →'Simulate device.')
       args = parser.parse_args(argsv)
```

(continues on next page)

```
example(args.resource_name, args.channel1, args.channel2, args.topology, args.
27
    →simulate)
28
29
   def test_example():
        example('', 'c0', 'r0', '2737/2-Wire 4x64 Matrix', True)
31
32
33
   def test_main():
34
       cmd_line = ['--topology', '2737/2-Wire 4x64 Matrix', '--simulate']
35
        _main(cmd_line)
36
   def main():
39
       _main(sys.argv[1:])
40
41
42
   if __name__ == '__main__':
43
        main()
44
45
```

### niswitch get device info.py

Listing 12: (niswitch\_get\_device\_info.py)

```
#!/usr/bin/python
2
   import argparse
   import niswitch
   import sys
   def example(resource_name, topology, simulate, device, channel, relay):
       # if we are simulating resource name must be blank
9
       resource_name = '' if simulate else resource_name
10
11
       with niswitch.Session(resource_name=resource_name, topology=topology,_
12
   ⇒simulate=simulate) as session:
           if device:
13
               print('Device Info:')
14
               row_format = '{:<18}' * (2)
15
               print(row_format.format('Device Name: ', session.io_resource_descriptor))
16
               print(row_format.format('Device Model: ', session.instrument_model))
17
               print(row_format.format('Driver Revision: ', session.specific_driver_
18
   →revision))
               print(row_format.format('Channel count: ', session.channel_count))
19
               print(row_format.format('Relay count: ', session.number_of_relays))
20
           if channel:
21
               print('Channel Info:')
22
               row_format = '{:6}' + ' ' * 12 + '{:<15}{:<22}{:6}'
23
               print(row_format.format('Number', 'Name', 'Is Configuration', 'Is Source
   ' ) )
               for i in range(1, session.channel_count + 1):
25
```

(continues on next page)

```
channel_name = session.get_channel_name(index=i)
26
                    channel = session.channels[channel_name]
27
                    print(row_format.format(i, channel_name, str(channel.is_configuration_
28
    →channel), str(channel.is_source_channel)))
           if relay:
29
               print('Relay Info:')
30
               row_format = '{:6}' + ' ' * 12 + '{:<15}{:<22}{:6}'
31
               print(row_format.format('Number', 'Name', 'Position', 'Count'))
32
               for i in range(1, session.number_of_relays + 1):
33
                    relay_name = session.get_relay_name(index=i)
3/1
                    print(row_format.format(i, relay_name, session.get_relay_
35
   →position(relay_name=relay_name), session.get_relay_count(relay_name=relay_name)))
37
   def _main(argsv):
38
       parser = argparse.ArgumentParser(description='Prints information for the...
39
   → specified National Instruments Switch module.', formatter_class=argparse.
   → ArgumentDefaultsHelpFormatter)
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource_
   →name of a National Instruments Switch.')
       parser.add_argument('-d', '--device', default=False, action='store_true', help=
41
   → 'Prints information for the device')
       parser.add_argument('-c', '--channel', default=False, action='store_true', help=
42
   →'Prints information for all channels on the device')
       parser.add_argument('-r', '--relay', default=False, action='store_true', help=
   →'Prints information for all relays on the device')
       parser.add_argument('-t', '--topology', default='Configured Topology', help=
44
   → 'Topology.')
       parser.add_argument('-s', '--simulate', default=False, action='store_true', help=
45
   →'Simulate device.')
       args = parser.parse_args(argsv)
46
       if not (args.device or args.channel or args.relay):
48
           print('You must specify at least one of -d, -c, or -r!')
49
           parser.print_help()
50
           sys.exit(1)
51
52
       example(args.resource_name, args.topology, args.simulate, args.device, args.
   ⇔channel, args.relay)
54
55
56
   def test_example():
       example('', '2737/2-Wire 4x64 Matrix', True, True, True, True)
57
58
59
   def test main():
60
       cmd_line = ['--topology', '2737/2-Wire 4x64 Matrix', '--simulate', '--device', '--
61
   ⇔channel', '--relay', ]
       _main(cmd_line)
62.
63
64
   def main():
       _main(sys.argv[1:])
66
67
68
              _ == '___main___':
69
       __name_
       main()
```

(continues on next page)

# niswitch\_relay\_control.py

71 72

Listing 13: (niswitch\_relay\_control.py)

```
#!/usr/bin/python
2
   import argparse
   import niswitch
   import sys
   def example(resource_name, topology, simulate, relay, action):
       # if we are simulating resource name must be blank
       resource_name = '' if simulate else resource_name
11
       with niswitch.Session(resource_name=resource_name, topology=topology,_
12
   ⇒simulate=simulate) as session:
           session.relay_control(relay_name=relay, relay_action=niswitch.
13
   →RelayAction[action])
           print('Relay ', relay, ' has had the action ', action, ' performed.')
14
16
   def _main(argsv):
17
       parser = argparse.ArgumentParser(description='Performs relay control with NI-
18
   →SWITCH relays.', formatter_class=argparse.ArgumentDefaultsHelpFormatter)
       parser.add_argument('-n', '--resource-name', default='PXI1Slot2', help='Resource_
19
   →name of a National Instruments Switch.')
       parser.add_argument('-r', '--relay', default='k0', help='Relay Name.')
       parser.add_argument('-a', '--action', default='OPEN', choices=niswitch.
21
   →RelayAction.__members__.keys(), type=str.upper, help='Relay Action.')
       parser.add_argument('-t', '--topology', default='Configured Topology', help=
22
   →'Topology.')
       parser.add_argument('-s', '--simulate', default=False, action='store_true', help=
23
   →'Simulate device.')
       args = parser.parse_args(argsv)
24
       example (args.resource_name, args.topology, args.simulate, args.relay, args.action)
25
26
27
   def test_example():
28
       example('', '2737/2-Wire 4x64 Matrix', True, 'kr0c0', 'OPEN')
29
31
32
   def test_main():
       cmd_line = ['--topology', '2737/2-Wire 4x64 Matrix', '--simulate', '--relay',
33
   \rightarrow 'kr0c0'1
       _main(cmd_line)
34
35
   def main():
37
       _main(sys.argv[1:])
38
```

(continues on next page)

```
40
41
42
main()

if __name__ == '__main__':
main()
```

# 7.6 nimodinst module

## 7.6.1 Installation

As a prerequisite to using the nimodinst module, you must install the NI-ModInst runtime on your system. Visit ni.com/downloads to download the driver runtime for your devices.

The nimi-python modules (i.e. for **NI-ModInst**) can be installed with pip:

```
$ python -m pip install nimodinst~=1.0.1
```

Or easy\_install from setuptools:

```
$ python -m easy_install nimodinst
```

# 7.6.2 Usage

The following is a basic example of using the **nimodinst** module to retrieve information on all High Speed Digitizers currently in the system.

Additional examples for NI-ModInst are located in src/nimodinst/examples/ directory.

### 7.6.3 API Reference

#### 7.6.3.1 nimodinst.Session

```
class nimodinst.Session(self, driver)
```

Creates a handle to a list of installed devices supported by the specified driver. Call this method and pass in the name of a National Instruments instrument driver, such as "NI-SCOPE". This method searches the system and constructs a list of all the installed devices that are supported by that driver, and then returns both a handle to this list and the number of devices found. The handle is used with other methods to query for properties such as device name and model, and to safely discard the list when finished. Note This handle reflects the system state when the handle is created (that is, when you call this method. If you remove devices from the system or rename them in Measurement & Automation Explorer (MAX), this handle may not refer to an accurate list of devices. You should destroy the handle using nimodinst.Session. \_close\_installed\_devices\_session() and create a new handle using this method.

Parameters driver (str) – A string specifying the driver whose supported devices you want to find. This string is not case-sensitive. Some examples are: NI-SCOPE niScope NI-FGEN niFgen NI-HSDIO niHSDIO NI-DMM niDMM NI-SWITCH niSwitch Note If you use the empty string for this parameter, NI-ModInst creates a list of all Modular Instruments devices installed in the system.

#### **Properties**

Property	Datatype
bus_number	int
chassis_number	int
device_model	str
device_name	str
max_pciexpress_link_width	int
pciexpress_link_width	int
serial_number	str
slot_number	int
socket_number	int

#### **Public methods**

Method name
Method name

#### **Properties**

## bus\_number

nimodinst.Session.bus number

The bus on which the device has been enumerated.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NIMODINST\_ATTR\_BUS\_NUMBER

#### chassis\_number

nimodinst.Session.chassis\_number

The number of the chassis in which the device is installed. This property can only be queried for PXI devices installed in a chassis that has been properly identified in MAX.

The following table lists the characteristics of this property.

7.6. nimodinst module 429

Characteristic	Value
Datatype	int
Permissions	read
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NIMODINST\_ATTR\_CHASSIS\_NUMBER

## device\_model

nimodinst.Session.device model

The model of the device (for example, NI PXI-5122)

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NIMODINST\_ATTR\_DEVICE\_MODEL

# device\_name

nimodinst.Session.device\_name

The name of the device, which can be used to open an instrument driver session for that device

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NIMODINST\_ATTR\_DEVICE\_NAME

# max pciexpress link width

nimodinst.Session.max\_pciexpress\_link\_width MAX\_PCIEXPRESS\_LINK\_WIDTH

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NIMODINST\_ATTR\_MAX\_PCIEXPRESS\_LINK\_WIDTH

# pciexpress\_link\_width

nimodinst.Session.pciexpress\_link\_width PCIEXPRESS LINK WIDTH

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NIMODINST\_ATTR\_PCIEXPRESS\_LINK\_WIDTH

## serial\_number

nimodinst.Session.serial\_number

The serial number of the device

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	str
Permissions	read
Channel Based	False
Resettable	No

7.6. nimodinst module 431

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NIMODINST\_ATTR\_SERIAL\_NUMBER

#### slot number

## nimodinst.Session.slot\_number

The slot (for example, in a PXI chassis) in which the device is installed. This property can only be queried for PXI devices installed in a chassis that has been properly identified in MAX.

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NIMODINST\_ATTR\_SLOT\_NUMBER

## socket\_number

#### nimodinst.Session.socket number

The socket number on which the device has been enumerated

The following table lists the characteristics of this property.

Characteristic	Value
Datatype	int
Permissions	read
Channel Based	False
Resettable	No

**Tip:** This property corresponds to the following LabVIEW Property or C Attribute:

• C Attribute: NIMODINST\_ATTR\_SOCKET\_NUMBER

## Methods

## **Properties**

Property	Datatype
nimodinst.Session.bus_number	int
nimodinst.Session.chassis_number	int
nimodinst.Session.device_model	str
nimodinst.Session.device_name	str
nimodinst.Session.max_pciexpress_link_width	int
nimodinst.Session.pciexpress_link_width	int
nimodinst.Session.serial_number	str
nimodinst.Session.slot_number	int
nimodinst.Session.socket_number	int

#### **Methods**

Method name
Method name

# 7.6.3.2 Exceptions and Warnings

# **DriverError**

**exception** nimodinst.**DriverError**An error originating from the NI-ModInst driver

# UnsupportedConfigurationError

**exception** nimodinst.**UnsupportedConfigurationError**An error due to using this module in an usupported platform.

#### DriverNotInstalledError

**exception** nimodinst.**DriverNotInstalledError**An error due to using this module without the driver runtime installed.

# Invalid Repeated Capability Error

**exception** nimodinst.InvalidRepeatedCapabilityError An error due to an invalid character in a repeated capability

7.6. nimodinst module 433

#### SelfTestError

```
exception nimodinst.SelfTestError
An error due to a failed self-test
```

### **DriverWarning**

```
exception nimodinst.DriverWarning
A warning originating from the NI-ModInst driver
```

### **7.6.3.3 Examples**

## nimodinst\_all\_devices.py

Listing 14: (nimodinst\_all\_devices.py)

```
#!/usr/bin/python
2
   import nimodinst
3
   def example():
6
       with nimodinst.Session('') as session:
7
            if len(session) > 0:
8
                print("%d items" % len(session))
                print("{: >20} {: >15} {: >10}".format('Name', 'Model', 'S/N'))
10
            for d in session:
11
12
                print("{: >20} {: >15} {: >10}".format(d.device_name, d.device_model, d.
    ⇔serial_number))
13
14
   def _main():
15
       example()
16
17
18
   def test_example():
19
       example()
20
21
22
   if __name__ == '__main__':
23
24
       _main()
25
26
```

# 7.7 nise module

## 7.7.1 Installation

As a prerequisite to using the nise module, you must install the NI Switch Executive runtime on your system. Visit ni.com/downloads to download the driver runtime for your devices.

The nimi-python modules (i.e. for **NI Switch Executive**) can be installed with pip:

```
$ python -m pip install nise~=0.1.0
```

#### Or easy\_install from setuptools:

```
$ python -m easy_install nise
```

# **7.7.2 Usage**

The following is a basic example of using the **nise** module to open a session to a Switch Executive Virtual Device and connect a routegroup.

```
import nise
with nise.Session('SwitchExecutiveExample') as session:
    session.connect('DIOTOUUT')
```

Additional examples for NI Switch Executive are located in src/nise/examples/ directory.

# 7.7.3 API Reference

#### 7.7.3.1 nise.Session

```
class nise.Session(self, virtual_device_name, options={})
```

Opens a session to a specified NI Switch Executive virtual device. Opens communications with all of the IVI switches associated with the specified NI Switch Executive virtual device. Returns a session handle that you use to identify the virtual device in all subsequent NI Switch Executive method calls. NI Switch Executive uses a reference counting scheme to manage open session handles to an NI Switch Executive virtual device. Each call to nise.Session.\_\_init\_\_() must be matched with a subsequent call to nise.Session.\_\_init\_\_() with the same virtual device name always returns the same session handle. NI Switch Executive disconnects its communication with the IVI switches after all session handles are closed to a given virtual device. The session handles may be used safely in multiple threads of an application. Sessions may only be opened to a given NI Switch Executive virtual device from a single process at a time.

#### **Parameters**

- virtual\_device\_name (str) The name of the NI Switch Executive virtual device.
- **options** (*str*) Specifies the initial value of certain properties for the session. The syntax for **options** is a dictionary of properties with an assigned value. For example:

```
{ 'simulate': False }
```

You do not have to specify a value for all the properties. If you do not specify a value for a property, the default value is used.

Advanced Example: { 'simulate': True, 'driver\_setup': { 'Model': '<model number>', 'BoardType': '<type>' } }

Property	Default
range_check	True
query_instrument_status	False
cache	True
simulate	False
record_value_coersions	False
driver_setup	{}

7.7. nise module 435

#### **Properties**

Property	Datatype
Property	Datatype

#### **Public methods**

Method name
connect()
connect_and_disconnect()
disconnect()
disconnect_all()
expand_route_spec()
find_route()
<pre>get_all_connections()</pre>
is_connected()
is_debounced()
wait_for_debounce()

## **Properties**

#### **Methods**

#### connect

nise.Session.connect(connect\_spec, multiconnect\_mode=nise.MulticonnectMode.DEFAULT, wait\_for\_debounce=True)

Connects the routes specified by the connection specification. When connecting, it may allow for multiconnection based on the multiconnection mode. In the event of an error, the call to <code>nise.Session.connect()</code> will attempt to undo any connections made so that the system will be left in the same state that it was in before the call was made. Some errors can be caught before manipulating hardware, although it is feasible that a hardware call could fail causing some connections to be momentarily closed and then reopened. If the wait for debounce parameter is set, the method will not return until the switch system has debounced.

#### **Parameters**

- **connect\_spec** (str) String describing the connections to be made. The route specification strings are best summarized as a series of routes delimited by ampersands. The specified routes may be route names, route group names, or fully specified route paths delimited by square brackets. Some examples of route specification strings are: MyRoute MyRouteGroup MyRoute & MyRouteGroup [A->Switch1/r0->B] MyRoute & MyRouteGroup & [A->Switch1/r0->B] Refer to Route Specification Strings in the NI Switch Executive Help for more information.
- multiconnect\_mode (nise.MulticonnectMode) This value sets the connection mode for the method. The mode might be one of the following:
   NISE\_VAL\_USE\_DEFAULT\_MODE (-1) uses the mode selected as the default for the route in the NI Switch Executive virtual device configuration. If a mode has not been selected for the route in the NI Switch Executive virtual device, this parameter defaults to NISE\_VAL\_MULTICONNECT\_ROUTES.
   NISE\_VAL\_NO\_MULTICONNECT (0) routes specified in the connection specification must be disconnected before they can be reconnected. Calling Connect on

a route that was connected using No Multiconnect mode results in an error condition. NISE\_VAL\_MULTICONNECT\_ROUTES (1)- routes specified in the connection specification can be connected multiple times. The first call to Connect performs the physical hardware connection. Successive calls to Connect increase a connection reference count. Similarly, calls to Disconnect decrease the reference count. Once it reaches 0, the hardware is physically disconnected. Multiconnecting routes applies to entire routes and not to route segments.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

• wait\_for\_debounce (bool) – Waits (if true) for switches to debounce between its connect and disconnect operations. If false, it immediately begins the second operation after completing the first. The order of connect and disconnect operation is set by the Operation Order input.

#### connect and disconnect

nise.Session.connect\_and\_disconnect (connect\_spec, disconnect\_spec, multiconnect\_mode=nise.MulticonnectMode.DEFAULT, operation\_order=nise.OperationOrder.AFTER, wait for debounce=True)

Connects routes and disconnects routes in a similar fashion to <code>nise.Session.connect()</code> and <code>nise.Session.disconnect()</code> except that the operations happen in the context of a single method call. This method is useful for switching from one state to another state. <code>nise.Session.connect\_and\_disconnect()</code> manipulates the hardware connections and disconnections only when the routes are different between the connection and disconnection specifications. If any routes are common between the connection and disconnection specifications, NI Switch Executive determines whether or not the relays need to be switched. This functionality has the distinct advantage of increased throughput for shared connections, because hardware does not have to be involved and potentially increases relay lifetime by decreasing the number of times that the relay has to be switched. In the event of an error, the call to <code>nise.Session.connect\_and\_disconnect()</code> attempts to undo any connections made, but does not attempt to reconnect disconnections. Some errors can be caught before manipulating hardware, although it is feasible that a hardware call could fail causing some connections to be momentarily closed and then reopened.

#### **Parameters**

- **connect\_spec** (str) String describing the connections to be made. The route specification strings are best summarized as a series of routes delimited by ampersands. The specified routes may be route names, route group names, or fully specified route paths delimited by square brackets. Some examples of route specification strings are: MyRoute MyRouteGroup MyRoute & MyRouteGroup [A->Switch1/r0->B] MyRoute & MyRouteGroup & [A->Switch1/r0->B] Refer to Route Specification Strings in the NI Switch Executive Help for more information.
- **disconnect\_spec** (str) String describing the disconnections to be made. The route specification strings are best summarized as a series of routes delimited by ampersands. The specified routes may be route names, route group names, or fully specified route paths delimited by square brackets. Some examples of route specification strings are: MyRoute MyRouteGroup MyRoute & MyRouteGroup [A->Switch1/r0->B] MyRoute & MyRouteGroup & [A->Switch1/r0->B] Refer to Route Specification Strings in the NI Switch Executive Help for more information.

7.7. nise module 437

• multiconnect mode (nise.MulticonnectMode) - This value sets the connection mode for the method. The mode might be one of the following: NISE VAL USE DEFAULT MODE (-1) - uses the mode selected as the default for the route in the NI Switch Executive virtual device configuration. If a mode has not been selected for the route in the NI Switch Executive virtual device, this parameter defaults to NISE VAL MULTICONNECT ROUTES. NISE VAL NO MULTICONNECT (0) - routes specified in the connection specification must be disconnected before they can be reconnected. Calling Connect on a route that was connected using No Multiconnect mode results in an error condition. NISE\_VAL\_MULTICONNECT\_ROUTES (1) - routes specified in the connection specification can be connected multiple times. The first call to Connect performs the physical hardware connection. Successive calls to Connect increase a connection reference count. Similarly, calls to Disconnect decrease the reference count. Once it reaches 0, the hardware is physically disconnected. This behavior is slightly different with SPDT relays. For more information, refer to the Exclusions and SPDT Relays topic in the NI Switch Executive Help. Multiconnecting routes applies to entire routes and not to route segments.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

• operation\_order (nise.OperationOrder) - Sets the order of the operation for the method. Defined values are Break Before Make and Break After Make. NISE\_VAL\_BREAK\_BEFORE\_MAKE (1) - The method disconnects the routes specified in the disconnect specification before connecting the routes specified in the connect specification. This is the typical mode of operation. NISE\_VAL\_BREAK\_AFTER\_MAKE (2) - The method connects the routes specified in the connection specification before connecting the routes specified in the disconnection specification. This mode of operation is normally used when you are switching current and want to ensure that a load is always connected to your source. The order of operation is to connect first or disconnect first.

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

• wait\_for\_debounce (bool) – Waits (if true) for switches to debounce between its connect and disconnect operations. If false, it immediately begins the second operation after completing the first. The order of connect and disconnect operation is set by the Operation Order input.

#### disconnect

# nise.Session.disconnect(disconnect\_spec)

Disconnects the routes specified in the Disconnection Specification. If any of the specified routes were originally connected in a multiconnected mode, the call to <code>nise.Session.disconnect()</code> reduces the reference count on the route by 1. If the reference count reaches 0, it is disconnected. If a specified route does not exist, it is an error condition. In the event of an error, the call to <code>nise.Session.disconnect()</code> continues to try to disconnect everything specified by the route specification string but reports the error on completion.

**Parameters disconnect\_spec** (str) – String describing the disconnections to be

made. The route specification strings are best summarized as a series of routes delimited by ampersands. The specified routes may be route names, route group names, or fully specified route paths delimited by square brackets. Some examples of route specification strings are: MyRoute MyRouteGroup MyRoute & MyRouteGroup [A->Switch1/r0->B] MyRoute & MyRouteGroup & [A->Switch1/r0->B] Refer to Route Specification Strings in the NI Switch Executive Help for more information.

#### disconnect\_all

```
nise.Session.disconnect all()
```

Disconnects all connections on every IVI switch device managed by the NISE session reference passed to this method. nise.Session.disconnect\_all() ignores all multiconnect modes. Calling nise.Session.disconnect\_all() resets all of the switch states for the system.

#### expand\_route\_spec

```
nise.Session.expand_route_spec (route_spec, expand_action=nise.ExpandAction.ROUTES, expanded_route_spec_size=[1024])
```

Expands a route spec string to yield more information about the routes and route groups within the spec. The route specification string returned from <code>nise.Session.expand\_route\_spec()</code> can be passed to other Switch Executive API methods (such as <code>nise.Session.connect()</code>, <code>nise.Session.disconnect()</code>, and <code>nise.Session.connect\_and\_disconnect()</code>) that use route specification strings.

#### **Parameters**

- route\_spec (str) String describing the routes and route groups to expand. The route specification strings are best summarized as a series of routes delimited by ampersands. The specified routes may be route names, route group names, or fully specified route paths delimited by square brackets. Some examples of route specification strings are: MyRoute MyRouteGroup MyRoute & MyRouteGroup [A->Switch1/r0->B] MyRoute & MyRouteGroup & [A->Switch1/r0->B] Refer to Route Specification Strings in the NI Switch Executive Help for more information.
- **expand\_action** (nise.ExpandAction) This value sets the expand action for the method. The action might be one of the following: NISE\_VAL\_EXPAND\_TO\_ROUTES (0) expands the route spec to routes. Converts route groups to their constituent routes. NISE\_VAL\_EXPAND\_TO\_PATHS (1) expands the route spec to paths. Converts routes and route groups to their constituent square bracket route spec strings. Example: [Dev1/c0->Dev1/r0->Dev1/c1]

**Note:** One or more of the referenced values are not in the Python API for this driver. Enums that only define values, or represent True/False, have been removed.

• **expanded\_route\_spec\_size** (list of int) – The routeSpecSize is an ViInt32 that is passed by reference into the method. As an input, it is the size of the route spec string buffer being passed. If the route spec string is larger than the string buffer being passed, only the portion of the route spec string that can fit in the string buffer is copied into it. On return from the method, routeSpecSize holds the size required to hold the entire route spec string. Note that this size may be larger than the buffer size as the method always returns the size needed to hold the entire buffer. You may pass NULL for this parameter if you are not interested in the return value for routeSpecSize and routeSpec.

7.7. nise module 439

#### Return type str

**Returns** The expanded route spec. Route specification strings can be directly passed to nise. Session.connect(), nise. Session.disconnect(), or nise. Session.connect\_and\_disconnect() Refer to Route Specification Strings in the NI Switch Executive Help for more information. You may pass NULL for this parameter if you are not interested in the return value. To obtain the route specification string, you should pass a buffer to this parameter. The size of the buffer required may be obtained by calling the method with NULL for this parameter and a valid ViInt32 to routeSpecSize. The routeSpecSize will contain the size needed to hold the entire route specification (including the NULL termination character). Common operation is to call the method twice. The first time you call the method you can determine the size needed to hold the route specification string. Allocate a buffer of the appropriate size and then re-call the method to obtain the entire buffer.

## find\_route

```
nise. Session. find route (channel1, channel2, route spec size=[1024])
```

Finds an existing or potential route between channel 1 and channel 2. The returned route specification contains the route specification and the route capability determines whether or not the route existed, is possible, or is not possible for various reasons. The route specification string returned from <code>nise.Session.find\_route()</code> can be passed to other Switch Executive API methods (such as <code>nise.Session.connect()</code>, <code>nise.Session.disconnect()</code>, and <code>nise.Session.connect()</code> that use route specification strings.

#### **Parameters**

- **channel1** (*str*) Channel name of one of the endpoints of the route to find. The channel name must either be a channel alias name or a name in the device/ivichannel syntax. Examples: MyChannel Switch1/R0
- **channel2** (*str*) Channel name of one of the endpoints of the route to find. The channel name must either be a channel alias name or a name in the device/ivichannel syntax. Examples: MyChannel Switch1/R0
- route\_spec\_size (list of int) The routeSpecSize is an ViInt32 that is passed by reference into the method. As an input, it is the size of the route string buffer being passed. If the route string is larger than the string buffer being passed, only the portion of the route string that can fit in the string buffer is copied into it. On return from the method, routeSpecSize holds the size required to hold the entire route string. Note that this size may be larger than the buffer size as the method always returns the size needed to hold the entire buffer. You may pass NULL for this parameter if you are not interested in the return value for routeSpecSize and routeSpec.

# Return type

```
tuple (route_spec, path_capability)
WHERE
route_spec (str):
```

The fully specified route path complete with delimiting square brackets if the route exists or is possible. An example of a fully specified route string is: [A->Switch1/r0->B] Route specification strings can be directly passed to nise.Session.connect(), nise.Session.disconnect(), or

nise.Session.connect\_and\_disconnect() Refer to Route Specification Strings in the NI Switch Executive Help for more information. You may pass NULL for this parameter if you are not interested in the return value. To obtain the route specification string, you should pass a buffer to this parameter. The size of the buffer required may be obtained by calling the method with NULL for this parameter and a valid ViInt32 to routeSpecSize. The routeSpecSize will contain the size needed to hold the entire route specification (including the NULL termination character). Common operation is to call the method twice. The first time you call the method you can determine the size needed to hold the route specification string. Allocate a buffer of the appropriate size and then re-call the method to obtain the entire buffer.

path\_capability (nise.PathCapability):

The return value which expresses the capability of finding a valid route between Channel 1 and Channel 2. Refer to the table below for value descriptions. You may pass NULL for this parameter if you are not interested in the return value. Route capability might be one of the following: Path Available (1) A path between channel 1 and channel 2 is available. The route specification parameter returns a string describing the available path. Path Exists (2) A path between channel 1 and channel 2 already exists. The route specification parameter returns a string describing the existing path. Path Unsupported (3) There is no potential path between channel 1 and channel 2 given the current configuration. Resource In Use (4) There is a potential path between channel 1 and channel 2, although a resource needed to complete the path is already in use. Source Conflict (5) Channel 1 and channel 2 cannot be connected because their connection would result in an exclusion violation. Channel Not Available (6) One of the channels is not useable as an endpoint channel. Make sure that it is not marked as a reserved for routing. Channels Hardwired (7) The two channels reside on the same hardwire. An implicit path already exists.

#### get\_all\_connections

```
nise.Session.get_all_connections(route_spec_size=[1024])
```

Returns the top-level connected routes and route groups. The route specification string returned from nise. Session.get\_all\_connections() can be passed to other Switch Executive API methods (such as nise. Session.connect(), nise. Session. disconnect(), nise. Session.connect\_and\_disconnect(), and nise. Session. expand route spec()) that use route specification strings.

Parameters route\_spec\_size (list of int) - The routeSpecSize is an ViInt32 that is passed by reference into the method. As an input, it is the size of the route spec string buffer being passed. If the route spec string is larger than the string buffer being passed, only the portion of the route spec string that can fit in the string buffer is copied into it. On return from the method, routeSpecSize holds the size required to hold the entire route spec string. Note that this size may be larger than the buffer size as the method always returns the size needed to hold the entire buffer. You may pass NULL for this parameter if you are not interested in the return value for routeSpecSize and routeSpec.

#### Return type str

Returns The route spec of all currently connected routes and route groups. Route specification strings can be directly passed to nise.

Session.connect(), nise.Session.disconnect(), nise.

7.7. nise module 441

Session.connect\_and\_disconnect(), or nise.Session.expand\_route\_spec() Refer to Route Specification Strings in the NI Switch Executive Help for more information. You may pass NULL for this parameter if you are not interested in the return value. To obtain the route specification string, you should pass a buffer to this parameter. The size of the buffer required may be obtained by calling the method with NULL for this parameter and a valid ViInt32 to routeSpecSize. The routeSpecSize will contain the size needed to hold the entire route specification (including the NULL termination character). Common operation is to call the method twice. The first time you call the method you can determine the size needed to hold the route specification string. Allocate a buffer of the appropriate size and then re-call the method to obtain the entire buffer.

# is\_connected

```
nise.Session.is_connected(route_spec)
```

Checks whether the specified routes and routes groups are connected. It returns true if connected.

**Parameters route\_spec** (str) – String describing the connections to check. The route specification strings are best summarized as a series of routes delimited by ampersands. The specified routes may be route names, route group names, or fully specified route paths delimited by square brackets. Some examples of route specification strings are: MyRoute MyRouteGroup MyRoute & MyRouteGroup [A->Switch1/r0->B] MyRoute & MyRouteGroup & [A->Switch1/r0->B] Refer to Route Specification Strings in the NI Switch Executive Help for more information.

#### Return type bool

**Returns** Returns TRUE if the routes and routes groups are connected or FALSE if they are not.

#### is debounced

```
nise.Session.is_debounced()
```

Checks to see if the switching system is debounced or not. This method does not wait for debouncing to occur. It returns true if the system is fully debounced. This method is similar to the IviSwtch specific method.

# Return type bool

**Returns** Returns TRUE if the system is fully debounced or FALSE if it is still settling.

#### wait for debounce

```
nise.Session.wait_for_debounce(maximum_time_ms=datetime.timedelta(milliseconds=-
```

Waits for all of the switches in the NI Switch Executive virtual device to debounce. This method does not return until either the switching system is completely debounced and settled or the maximum time has elapsed and the system is not yet debounced. In the event that the maximum time elapses, the method returns an error indicating that a timeout has occurred. To ensure that all of the switches have settled, NI recommends calling <code>nise.Session.wait\_for\_debounce()</code> after a series of connection or disconnection operations and before taking any measurements of the signals connected to the switching system.

Parameters maximum\_time\_ms

(float in seconds or datetime.

timedelta) – The amount of time to wait (in milliseconds) for the debounce to complete. A value of 0 checks for debouncing once and returns an error if the system is not debounced at that time. A value of -1 means to block for an infinite period of time until the system is debounced.

# **Properties**

Property	Datatype
Property	Datatype

# Methods

Method name
nise.Session.connect()
nise.Session.connect_and_disconnect()
nise.Session.disconnect()
nise.Session.disconnect_all()
nise.Session.expand_route_spec()
nise.Session.find_route()
nise.Session.get_all_connections()
nise.Session.is_connected()
nise.Session.is_debounced()
nise.Session.wait_for_debounce()

# 7.7.3.2 Enums

Enums used in NI Switch Executive

# **ExpandAction**

class nise.ExpandAction

ROUTES

Expand to routes

PATHS

Expand to paths

#### MulticonnectMode

 ${\tt class} \ {\tt nise}. {\tt MulticonnectMode}$ 

DEFAULT

Default

7.7. nise module 443

#### NO MULTICONNECT

No multiconnect

#### MULTICONNECT

Multiconnect

# **OperationOrder**

# class nise.OperationOrder

#### BEFORE

Break before make

#### AFTER

Break after make

# **PathCapability**

## class nise.PathCapability

# PATH\_NEEDS\_HARDWIRE

Path needs hardwire

#### PATH NEEDS CONFIG CHANNEL

Path needs config channel

# PATH\_AVAILABLE

Path available

#### PATH EXISTS

Path exists

## PATH\_UNSUPPORTED

Path Unsupported

# RESOURCE\_IN\_USE

Resource in use

# EXCLUSION\_CONFLICT

**Exclusion conflict** 

#### CHANNEL NOT AVAILABLE

Channel not available

# CHANNELS\_HARDWIRED

Channels hardwired

## 7.7.3.3 Exceptions and Warnings

# DriverError

## exception nise.DriverError

An error originating from the NI Switch Executive driver

## UnsupportedConfigurationError

# exception nise.UnsupportedConfigurationError

An error due to using this module in an usupported platform.

#### **DriverNotInstalledError**

#### exception nise.DriverNotInstalledError

An error due to using this module without the driver runtime installed.

## InvalidRepeatedCapabilityError

```
exception nise.InvalidRepeatedCapabilityError
```

An error due to an invalid character in a repeated capability

#### SelfTestError

```
exception nise.SelfTestError
```

An error due to a failed self-test

# **DriverWarning**

#### exception nise.DriverWarning

A warning originating from the NI Switch Executive driver

#### **7.7.3.4 Examples**

#### nise basic example.py

Listing 15: (nise\_basic\_example.py)

```
#!/usr/bin/python
   import argparse
   import nise
   import sys
   def example(virtual_device_name, connection):
       with nise.Session(virtual_device_name=virtual_device_name) as session:
           session.connect(connection)
           print(connection, ' is now connected.')
10
12
   def _main(argsv):
13
       parser = argparse.ArgumentParser(description='Connects the specified connection,
14
   →specification', formatter_class=argparse.ArgumentDefaultsHelpFormatter)
       parser.add_argument('-n', '--virtual-device', default='SwitchExecutiveExample',_
15
   →help='NI Switch Executive Virtual Device name')
       parser.add_argument('-c', '--connection', default='DIOTOUUT', help='Connection...
   →Specification')
```

7.7. nise module 445

(continues on next page)

(continued from previous page)

```
args = parser.parse_args(argsv)
17
        example(args.virtual_device, args.connection)
18
19
20
   def main():
21
        _main(sys.argv[1:])
22
23
24
   def test_example():
25
        example('SwitchExecutiveExample', 'DIOToUUT')
26
27
29
   def test_main():
30
        cmd_line = []
        _main(cmd_line)
31
32
33
   if __name__ == '__main__':
34
35
        main()
36
37
```

# CHAPTER 8

# Indices and tables

- genindex
- modindex
- search

NI Modular Instruments Python API Documentation, Release 1.0.1		

# Python Module Index

# n

nidcpower, 14 nidmm, 123 nifgen, 191 nimodinst, 428 niscope, 293 nise, 435 niswitch, 379

450 Python Module Index

Α	allocate_waveform() (in module nifgen.Session), 251
	allow_more_records_than_memory (in module nis-
abort() (in module nidcpower.Session), 98 abort() (in module nidmm.Session), 159	cope.Session), 301
	analog_bus_sharing_enable (in module
abort() (in module nifgen.Session), 251	niswitch.Session), 383
abort() (in module niscope.Session), 347	analog_data_mask (in module nifgen.Session), 197
abort() (in module niswitch.Session), 407	ANALOG_DETECTION_CIRCUIT (nis-
absolute_sample_clock_offset (in module nis-	cope.RefTriggerDetectorLocation attribute),
cope.Session), 298	369
AC (nidmm.WaveformCoupling attribute), 186	analog_filter_enabled (in module nifgen.Session), 198
AC (niscope.TriggerCoupling attribute), 370	analog_path (in module nifgen.Session), 198
AC (NISCOPE. Vertical Coupling attribute), 371	analog_static_value (in module nifgen.Session), 198
AC_CURRENT (nidmm.Function attribute), 181	AnalogPath (class in nifgen), 280
ac_max_freq (in module nidmm.Session), 127	ANY_FIELD (niscope.VideoTriggerEvent attribute), 373
ac_min_freq (in module nidmm.Session), 128	ANY_LINE (niscope. Video TriggerEvent attribute), 373
AC_PLUS_HF_REJECT (niscope.TriggerCoupling at-	aperture_time (in module nidcpower.Session), 18
tribute), 370	aperture_time (in module nidmm.Session), 128
AC_VOLTS (nidmm.Function attribute), 181	aperture_time_units (in module nidcpower.Session), 19
AC_VOLTS_DC_COUPLED (nidmm.Function at-	aperture_time_units (in module nidmm.Session), 129
tribute), 181	ApertureTimeUnits (class in nidcpower), 115
accessory_gain (in module niscope.Session), 298	ApertureTimeUnits (class in nidmm), 180
accessory_offset (in module niscope.Session), 299	ARB (nifgen.OutputMode attribute), 282
acq_arm_source (in module niscope.Session), 300	arb_gain (in module nifgen.Session), 199
acquisition_start_time (in module niscope.Session), 300	arb_marker_position (in module nifgen.Session), 199
acquisition_status() (in module niscope.Session), 347	arb_offset (in module nifgen.Session), 200
acquisition_type (in module niscope.Session), 300	arb_repeat_count (in module nifgen.Session), 201
AcquisitionStatus (class in nidmm), 179	arb_sample_rate (in module nifgen.Session), 201
AcquisitionStatus (class in niscope), 367	arb_sequence_handle (in module nifgen.Session), 201
AcquisitionType (class in niscope), 367	arb_waveform_handle (in module nifgen.Session), 202
adc_calibration (in module nidmm.Session), 128	arm_ref_trig_src (in module niscope.Session), 302
ADCCalibration (class in nidmm), 179	ARM_REFERENCE (niscope.WhichTrigger attribute),
adv_trig_src (in module niscope.Session), 301	373
ADVANCE (niscope. Which Trigger attribute), 373	ASYMMETRIC (nidcpower.ComplianceLimitSymmetry
AFTER (nise.OperationOrder attribute), 444 all marker events latched status (in module nif-	attribute), 115
`	AT (nifgen.BusType attribute), 281
gen.Session), 196 all marker events live status (in module nif-	AUTO (nidmm.ADCCalibration attribute), 179
`	AUTO (nidmm.AutoZero attribute), 180
gen.Session), 197	AUTO (nidmm.DCNoiseRejection attribute), 181
allocate_named_waveform() (in module nifgen.Session), 251	AUTO (nidmm.LCCalculationModel attribute), 182
231	AUTO (niscope.TriggerModifier attribute), 370
	11010 (motopolitionalitic autitority), 570

auto_range_value (in module nidmm.Session), 130	CHANNEL_NOT_AVAILABLE
auto_setup() (in module niscope.Session), 347	(niswitch.PathCapability attribute), 418
auto_zero (in module nidcpower.Session), 20	channel_terminal_configuration (in module nis-
auto_zero (in module nidmm.Session), 130	cope.Session), 305
AUTOMATIC (nidcpower.PowerSource attribute), 117	CHANNELS_HARDWIRED (nise.PathCapability
AUTOMATIC (nifgen.ClockMode attribute), 281	attribute), 444
AUTOMATICALLY_AFTER_SOURCE_COMPLETE	characteristic_impedance (in module niswitch.Session),
(nidcpower.MeasureWhen attribute), 116	385
AutoZero (class in nidcpower), 115	chassis_number (in module nimodinst.Session), 429
AutoZero (class in nidmm), 180	clear_arb_memory() (in module nifgen.Session), 252
aux_power_enabled (in module nifgen.Session), 203	clear_arb_sequence() (in module nifgen.Session), 252
AUX_TRIG1 (nidmm.SampleTrigger attribute), 184	clear_freq_list() (in module nifgen.Session), 252
AUX_TRIG1 (nidmm.TriggerSource attribute), 186	clear_user_standard_waveform() (in module nif-
AUXILIARY (nidcpower.PowerSource attribute), 117	gen.Session), 253
AUXILIARY (nidcpower.PowerSourceInUse attribute),	CLOCK_IN (nifgen.ReferenceClockSource attribute),
117	282
auxiliary_power_source_available (in module nid-	CLOCK_IN (nifgen.SampleClockSource attribute), 283
cpower.Session), 20	CLOCK_IN (nifgen.SampleClockTimebaseSource
epower.session), 20	attribute), 284
В	clock_mode (in module nifgen.Session), 204
	ClockMode (class in nifgen), 281
B (nidmm.ThermocoupleType attribute), 184	CLOSE (niswitch.RelayAction attribute), 418
backlog (in module niscope.Session), 302	CLOSED (niswitch.RelayPosition attribute), 418
bandpass_filter_enabled (in module niscope.Session),	commit() (in module nidcpower.Session), 99
303	commit() (in module indepower.session), 39
bandwidth (in module niswitch.Session), 384	
BEFORE (nise.OperationOrder attribute), 444	commit() (in module niscope.Session), 348
BIG (nifgen.ByteOrder attribute), 281	commit() (in module niswitch.Session), 408
binary_sample_width (in module niscope.Session), 303	common_mode_offset (in module nifgen.Session), 205
BREAK_AFTER_MAKE (niswitch.ScanMode at-	COMPLETE (niscope. Acquisition Status attribute), 367
tribute), 421	compliance_limit_symmetry (in module nid-
BREAK_BEFORE_MAKE (niswitch.ScanMode at-	cpower.Session), 21
tribute), 421	ComplianceLimitSymmetry (class in nidcpower), 115
buffer_size (in module nidmm.Session), 130	configure_aperture_time() (in module nid-
BURST (nifgen.TriggerMode attribute), 285	cpower.Session), 99
bus_number (in module nimodinst.Session), 429	configure_arb_sequence() (in module nifgen.Session),
bus_type (in module nifgen.Session), 203	254
BusType (class in nifgen), 280	configure_arb_waveform() (in module nifgen.Session),
ByteOrder (class in nifgen), 281	255
•	configure_chan_characteristics() (in module nis-
C	cope.Session), 348
cable_comp_type (in module nidmm.Session), 131	configure_custom_fir_filter_coefficients() (in module nif-
CableCompensationType (class in nidmm), 180	gen.Session), 256
can_connect() (in module niswitch.Session), 407	configure_equalization_filter_coefficients() (in module
CAPACITANCE (nidmm.Function attribute), 182	niscope.Session), 349
channel_count (in module nidcpower.Session), 21	configure_freq_list() (in module nifgen.Session), 256
channel_count (in module nidmm.Session), 131	configure_horizontal_timing() (in module nis-
channel_count (in module nifgen.Session), 203	cope.Session), 349
channel_count (in module niscope.Session), 203	configure_measurement_absolute() (in module
channel_count (in module niswitch.Session), 384	nidmm.Session), 159
channel_delay (in module nifgen.Session), 204	configure_measurement_digits() (in module
channel_enabled (in module niscope.Session), 304	nidmm.Session), 160
CHANNEL_NOT_AVAILABLE (nise.PathCapability at-	configure_multi_point() (in module nidmm.Session), 161
tribute), 444	configure_rtd_custom() (in module nidmm.Session), 162
uiouc), <del>111</del>	configure_rtd_type() (in module nidmm.Session), 162

configure_standard_waveform() (in module nif-	26
gen.Session), 258	current_limit_high (in module nidcpower.Session), 27
configure_thermistor_custom() (in module	current_limit_low (in module nidcpower.Session), 28
nidmm.Session), 164	current_limit_range (in module nidcpower.Session), 29
configure_thermocouple() (in module nidmm.Session),	current_pole_zero_ratio (in module nidcpower.Session),
164	30
configure_trigger() (in module nidmm.Session), 165	current_source (in module nidmm.Session), 132
configure_trigger_digital() (in module niscope.Session),	CUSTOM (nidcpower.TransientResponse attribute), 119
350	CUSTOM (nidmm.RTDType attribute), 183
configure_trigger_edge() (in module niscope.Session),	CUSTOM (nidmm.ThermistorType attribute), 184
351	COSTONI (mamm. Thermistor Type attribute), 101
configure_trigger_hysteresis() (in module nis-	D
cope.Session), 352	_
configure_trigger_immediate() (in module nis-	data_marker_event_data_bit_number (in module nif-
cope.Session), 353	gen.Session), 206
	data_marker_event_level_polarity (in module nif-
configure_trigger_software() (in module nis-	gen.Session), 206
cope.Session), 353	data_marker_event_output_terminal (in module nif-
configure_trigger_video() (in module niscope.Session),	gen.Session), 207
354	data_marker_events_count (in module nifgen.Session),
configure_trigger_window() (in module niscope.Session),	205
355	data_transfer_block_size (in module nifgen.Session), 207
configure_vertical() (in module niscope.Session), 356	data_transfer_block_size (in module niscope.Session),
configure_waveform_acquisition() (in module	305
nidmm.Session), 165	data_transfer_maximum_bandwidth (in module nif-
connect() (in module nise.Session), 436	gen.Session), 208
connect() (in module niswitch.Session), 408	data_transfer_maximum_bandwidth (in module nis-
connect_and_disconnect() (in module nise.Session), 437	cope.Session), 306
connect_multiple() (in module niswitch.Session), 409	data_transfer_maximum_in_flight_reads (in module nif-
CONTINUOUS (nifgen.TriggerMode attribute), 285	gen.Session), 208
continuous_scan (in module niswitch.Session), 385	data_transfer_preferred_packet_size (in module nif-
create_advanced_arb_sequence() (in module nif-	gen.Session), 209
gen.Session), 260	data_transfer_preferred_packet_size (in module nis-
create_arb_sequence() (in module nifgen.Session), 262	cope.Session), 306
create_freq_list() (in module nifgen.Session), 263	DataMarkerEventLevelPolarity (class in nifgen), 281
create_waveform_from_file_f64() (in module nif-	DC (nidmm.WaveformCoupling attribute), 186
gen.Session), 264	DC (nifgen.Waveform attribute), 286
create_waveform_from_file_i16() (in module nif-	DC (niscope.TriggerCoupling attribute), 370
gen.Session), 265	DC (niscope. Vertical Coupling attribute), 371
create_waveform_numpy() (in module nifgen.Session),	dc_bias (in module nidmm.Session), 132
266	DC_CURRENT (nidcpower.OutputFunction attribute),
CURRENT (nidcpower.MeasurementTypes attribute),	116
116	DC_CURRENT (nidmm.Function attribute), 181
CURRENT (nidcpower.OutputStates attribute), 117	dc_noise_rejection (in module nidcpower.Session), 31
CURRENT (nifgen.RelativeTo attribute), 283	dc_noise_rejection (in module nidmm.Session), 33
current_compensation_frequency (in module nid-	DC_VOLTAGE (nidcpower.OutputFunction attribute),
cpower.Session), 22	116
current_gain_bandwidth (in module nidcpower.Session),	DC_VOLTS (nidmm.Function attribute), 181
23	DCNoiseRejection (class in nidepower), 115
current_level (in module nidcpower.Session), 23	DCNoiseRejection (class in nidmm), 181
current_level_autorange (in module nidcpower.Session),	DDC (niscope.AcquisitionType attribute), 368
24	DDC_CLOCK_IN (nifgen.SampleClockSource at-
current_level_range (in module nidcpower.Session), 25	tribute), 283
current_limit (in module nidcpower.Session), 26	DDC_OUTPUT (niscope.RefTriggerDetectorLocation
current limit autorange (in module nidcpower.Session),	attribute) 369

DEFAULT (nise.MulticonnectMode attribute), 443	disconnect() (in module niswitch.Session), 409
define_user_standard_waveform() (in module nif-	disconnect_all() (in module nise.Session), 439
gen.Session), 266	disconnect_all() (in module niswitch.Session), 410
delete_script() (in module nifgen.Session), 267	disconnect_multiple() (in module niswitch.Session), 410
delete_waveform() (in module nifgen.Session), 267	DIVIDE_DOWN (nifgen.ClockMode attribute), 281
device_model (in module nimodinst.Session), 430	DONE (nifgen.HardwareState attribute), 282
device_name (in module nimodinst.Session), 430	done_event_output_terminal (in module nifgen.Session),
device_temperature (in module niscope.Session), 307	214
DIFFERENTIAL (nifgen.TerminalConfiguration at-	driver_setup (in module nidcpower.Session), 34
tribute), 285	driver_setup (in module nidmm.Session), 133
DIFFERENTIAL (niscope.TerminalConfiguration	driver_setup (in module nifgen.Session), 214
attribute), 369	driver_setup (in module niscope.Session), 307
DIGITAL (niscope.TriggerType attribute), 371	driver_setup (in module niswitch.Session), 386
digital_data_mask (in module nifgen.Session), 209	DriverError, 119, 186, 286, 373, 423, 433, 444
DIGITAL_EDGE (nidcpower.TriggerType attribute), 119	DriverNotInstalledError, 119, 186, 287, 373, 423, 433,
DIGITAL_EDGE (nifgen.ScriptTriggerType attribute),	445
284	DriverWarning, 120, 187, 287, 374, 424, 434, 445
DIGITAL_EDGE (nifgen.StartTriggerType attribute), 285	E
digital_edge_measure_trigger_input_terminal (in module	E (nidmm.ThermocoupleType attribute), 185
nidcpower.Session), 31	EDGE (niscope.TriggerType attribute), 370
digital_edge_pulse_trigger_input_terminal (in module	EIGHT_TAP_HANNING (nis-
nidcpower.Session), 32	cope.FlexFIRAntialiasFilterType attribute),
digital_edge_script_trigger_edge (in module nif- gen.Session), 210	368 enable_dc_restore (in module niscope.Session), 308
digital_edge_script_trigger_source (in module nif-	enable_time_interleaved_sampling (in module nis-
gen.Session), 210	cope.Session), 308
digital_edge_sequence_advance_trigger_input_terminal (in module nidcpower.Session), 32	end_of_acquisition_event_output_terminal (in module niscope.Session), 309
digital_edge_source_trigger_input_terminal (in module	end_of_record_event_output_terminal (in module nis-
nidcpower.Session), 33	cope.Session), 309
digital_edge_start_trigger_edge (in module nif-	end_of_record_to_advance_trigger_holdoff (in module
gen.Session), 211	niscope.Session), 309
digital_edge_start_trigger_input_terminal (in module	ENTERING (niscope.TriggerWindowMode attribute),
nidcpower.Session), 33	371
digital_edge_start_trigger_source (in module nif-	equalization_filter_enabled (in module niscope.Session),
gen.Session), 211	310
digital_filter_enable (in module niswitch.Session), 386	equalization_num_coefficients (in module nis-
digital_filter_enabled (in module nifgen.Session), 212	cope.Session), 311
digital_filter_interpolation_factor (in module nif-	Event (class in nidcpower), 115
gen.Session), 212	EXACT_NUM_AVERAGES (niscope.RISMethod
digital_gain (in module nifgen.Session), 213	attribute), 369
DIGITAL_LEVEL (nifgen.ScriptTriggerType attribute),	EXCLUSION_CONFLICT (nise.PathCapability at-
284	tribute), 444
digital_pattern_enabled (in module nifgen.Session), 213	expand_route_spec() (in module nise.Session), 439
digital_static_value (in module nifgen.Session), 214	ExpandAction (class in nise), 443
DIODE (nidmm.Function attribute), 181	exported_advance_trigger_output_terminal (in module
DIRECT (nifgen.AnalogPath attribute), 280	niscope.Session), 311
disable() (in module nidcpower.Session), 100	exported_measure_trigger_output_terminal (in module
disable() (in module nidmm.Session), 166	nidcpower.Session), 35
disable() (in module nifgen.Session), 268	exported_onboard_reference_clock_output_terminal (in
disable() (in module niscope.Session), 356	module nifgen.Session), 215
disable() (in module niswitch.Session), 409	exported_pulse_trigger_output_terminal (in module nid-
disconnect() (in module nise.Session), 438	cpower.Session). 35

exported_ref_trigger_output_terminal (in module nis-	FIELD1 (niscope. Video Trigger Event attribute), 373
cope.Session), 312 exported_reference_clock_output_terminal (in module	FIELD2 (niscope.VideoTriggerEvent attribute), 373 file_transfer_block_size (in module nifgen.Session), 220
nifgen.Session), 215	filter_correction_frequency (in module nifgen.Session), 220
exported_sample_clock_divisor (in module nif-	220
gen.Session), 216	find_route() (in module nise.Session), 440
exported_sample_clock_output_terminal (in module nif-	FINISHED_WITH_BACKLOG
gen.Session), 216	(nidmm.AcquisitionStatus attribute), 179
exported_sample_clock_timebase_divisor (in module nif-	FINISHED_WITH_NO_BACKLOG
gen.Session), 217	(nidmm.AcquisitionStatus attribute), 180
exported_sample_clock_timebase_output_terminal (in	FIXED (nidmm.ThermocoupleReferenceJunctionType
module nifgen.Session), 217	attribute), 184
exported_script_trigger_output_terminal (in module nif-	FIXED_HIGH_GAIN (nifgen.AnalogPath attribute), 280
gen.Session), 218	FIXED_LOW_GAIN (nifgen.AnalogPath attribute), 280
exported_sequence_advance_trigger_output_terminal (in	flatness_correction_enabled (in module nifgen.Session),
module nidcpower.Session), 36	220
exported_source_trigger_output_terminal (in module nid-	flex_fir_antialias_filter_type (in module niscope.Session),
cpower.Session), 36	312
exported_start_trigger_output_terminal (in module nid-	FlexFIRAntialiasFilterType (class in niscope), 368
cpower.Session), 37	FLEXRES (niscope.AcquisitionType attribute), 368
exported_start_trigger_output_terminal (in module nif-	FOUR_WIRE_RES (nidmm.Function attribute), 181
gen.Session), 218	FOUR_WIRE_RTD (nidmm.TransducerType attribute),
exported_start_trigger_output_terminal (in module nis-	185
cope.Session), 312	FOURTYEIGHT_TAP_HANNING (nis-
EXTERNAL (nidmm.MeasurementCompleteDest	cope.FlexFIRAntialiasFilterType attribute),
attribute), 182 EXTERNAL (nidmm.SampleTrigger attribute), 183	368 FOURTYEIGHT_TAP_STANDARD (nis-
EXTERNAL (nidmm.TriggerSource attribute), 185	cope.FlexFIRAntialiasFilterType attribute),
EXTERNAL (niswitch.ScanAdvancedOutput attribute),	368
419	fpga_bitfile_path (in module nifgen.Session), 221
EXTERNAL (niswitch.TriggerInput attribute), 421	fpga_bitfile_path (in module niscope.Session), 313
external_clock_delay_binary_value (in module nif-	FREQ (nidmm.Function attribute), 181
gen.Session), 219	FREQ_LIST (nifgen.OutputMode attribute), 282
external_sample_clock_multiplier (in module nif-	freq_list_duration_quantum (in module nifgen.Session),
gen.Session), 219	221
	freq_list_handle (in module nifgen.Session), 222
F	freq_voltage_auto_range (in module nidmm.Session),
FALLING (nifgen.ScriptTriggerDigitalEdgeEdge at-	134
tribute), 284	freq_voltage_range (in module nidmm.Session), 134
FALLING (nifgen.StartTriggerDigitalEdgeEdge at-	FRONTCONNECTOR (niswitch.ScanAdvancedOutput
tribute), 284	attribute), 419
FALLING (niswitch.ScanAdvancedPolarity attribute),	FRONTCONNECTOR (niswitch.TriggerInput attribute),
420	422
FALLING (niswitch.TriggerInputPolarity attribute), 423	FRONTCONNECTOR_MODULE1
FAST (nidcpower.TransientResponse attribute), 119	(niswitch.ScanAdvancedOutput attribute),
fetch() (in module nidmm.Session), 166	420 FRONTCONNECTOR_MODULE1
fetch() (in module niscope.Session), 356	(niswitch.TriggerInput attribute), 422
fetch_backlog (in module nidcpower.Session), 37	FRONTCONNECTOR_MODULE10
fetch_into() (in module niscope.Session), 358	(niswitch.ScanAdvancedOutput attribute),
fetch_multi_point() (in module nidmm.Session), 167	420
fetch_multiple() (in module nidcpower.Session), 100 fetch_waveform() (in module nidmm.Session), 168	FRONTCONNECTOR_MODULE10
fetch_waveform_into() (in module nidmin.Session), 168	(niswitch.TriggerInput attribute), 423
FetchRelativeTo (class in niscope), 368	FRONTCONNECTOR_MODULE11

(niswitch.ScanAdvancedOutput 420	attribute),	func_frequency (in module nifgen.Session), 224 func_max_buffer_size (in module nifgen.Session), 225
FRONTCONNECTOR_MODULE11		func_start_phase (in module nifgen.Session), 225
(niswitch.TriggerInput attribute), 423		func_waveform (in module nifgen.Session), 226
FRONTCONNECTOR_MODULE12		Function (class in nidmm), 181
(niswitch.ScanAdvancedOutput 420	attribute),	function (in module nidmm.Session), 134
FRONTCONNECTOR_MODULE12		G
(niswitch.TriggerInput attribute), 423		get_all_connections() (in module nise.Session), 441
FRONTCONNECTOR_MODULE2		get_cal_date_and_time() (in module nidmm.Session),
(niswitch.ScanAdvancedOutput	attribute),	169
420		get_channel_name() (in module nidcpower.Session), 101
FRONTCONNECTOR_MODULE2		get_channel_name() (in module niswitch.Session), 410
(niswitch.TriggerInput attribute), 422		get_dev_temp() (in module nidmm.Session), 170
FRONTCONNECTOR_MODULE3		get_equalization_filter_coefficients() (in module nis-
(niswitch.ScanAdvancedOutput	attribute),	cope.Session), 359
420		get_ext_cal_last_date_and_time() (in module nid-
FRONTCONNECTOR_MODULE3		cpower.Session), 101
(niswitch.TriggerInput attribute), 422		get_ext_cal_last_date_and_time() (in module nif-
FRONTCONNECTOR_MODULE4		gen.Session), 268
(niswitch.ScanAdvancedOutput 420	attribute),	get_ext_cal_last_temp() (in module nidcpower.Session),
FRONTCONNECTOR_MODULE4		101
(niswitch.TriggerInput attribute), 422		get_ext_cal_last_temp() (in module nifgen.Session), 268
FRONTCONNECTOR_MODULE5		get_ext_cal_recommended_interval() (in module nid-
(niswitch.ScanAdvancedOutput	attribute),	cpower.Session), 102 get ext cal recommended interval() (in module
420	attiioute),	get_ext_cal_recommended_interval() (in module nidmm.Session), 170
FRONTCONNECTOR_MODULE5		get_ext_cal_recommended_interval() (in module nif-
(niswitch.TriggerInput attribute), 423		gen.Session), 268
FRONTCONNECTOR_MODULE6		get_hardware_state() (in module nifgen.Session), 268
(niswitch.ScanAdvancedOutput	attribute),	get_last_cal_temp() (in module nidmm.Session), 208
420	,,,	get_path() (in module niswitch.Session), 410
FRONTCONNECTOR_MODULE6		get_path() (in module hiswitch.Session), 410 get_relay_count() (in module niswitch.Session), 411
(niswitch.TriggerInput attribute), 423		get_relay_count() (in module hiswitch.Session), 411 get_relay_name() (in module niswitch.Session), 411
FRONTCONNECTOR_MODULE7		get_relay_position() (in module niswitch.Session), 411
(niswitch.ScanAdvancedOutput	attribute),	get_self_cal_last_date_and_time() (in module nid-
420	,,,	cpower.Session), 102
FRONTCONNECTOR_MODULE7		get_self_cal_last_date_and_time() (in module nif-
(niswitch.TriggerInput attribute), 423		gen.Session), 269
FRONTCONNECTOR_MODULE8		get_self_cal_last_temp() (in module nidcpower.Session),
(niswitch.ScanAdvancedOutput	attribute),	get_sen_ear_tast_temp() (in module indepower.session),
420	,,	get_self_cal_last_temp() (in module nifgen.Session), 269
FRONTCONNECTOR_MODULE8		get_self_cal_supported() (in module nidmm.Session),
(niswitch.TriggerInput attribute), 423		get_sen_ear_supported() (in module mamm.session),
FRONTCONNECTOR_MODULE9		get_self_cal_supported() (in module nifgen.Session), 269
(niswitch.ScanAdvancedOutput	attribute),	GND (niscope. Vertical Coupling attribute), 371
420	,,,	GIVD (inscope, vertical coupling authority), 371
FRONTCONNECTOR_MODULE9		Н
(niswitch.TriggerInput attribute), 423		
FUNC (nifgen.OutputMode attribute), 282		handshaking_initiation (in module niswitch.Session), 387 HandshakingInitiation (class in niswitch), 417
func_amplitude (in module nifgen.Session), 222	2	HARDWARE_ERROR (nifgen.HardwareState attribute),
func_buffer_size (in module nifgen.Session), 22		1282
func_dc_offset (in module nifgen.Session), 223		HardwareState (class in nifgen), 281
func_duty_cycle_high (in module nifgen.Session	on), 224	HF REJECT (niscope.TriggerCoupling attribute), 370

HIGH (nidcpower.OutputCapacitance attribute), 116 HIGH (nidcpower.Polarity attribute), 117	instrument_manufacturer (in module niscope.Session), 318
HIGH (nifgen.DataMarkerEventLevelPolarity attribute),	instrument_manufacturer (in module niswitch.Session), 388
HIGH_ORDER (nidmm.DCNoiseRejection attribute),	instrument_model (in module nidcpower.Session), 38
181	instrument_model (in module nidmm.Session), 136
high_pass_filter_frequency (in module niscope.Session),	instrument_model (in module nifgen.Session), 228
313	instrument_model (in module niscope.Session), 319
HIGH_RESOLUTION (nifgen.ClockMode attribute),	instrument_model (in module niswitch.Session), 388
281	instrument_product_id (in module nidmm.Session), 137
HOLD_LAST (nifgen.IdleBehavior attribute), 282 HOLD_LAST (nifgen.WaitBehavior attribute), 286	interleaving_offset_correction_enabled (in module niscope.Session), 319
horz_enforce_realtime (in module niscope.Session), 314	interlock_input_open (in module nidcpower.Session), 39
horz_min_num_pts (in module niscope.Session), 314	INTERNAL (nidcpower.PowerSource attribute), 117
horz_num_records (in module niscope.Session), 315	INTERNAL (nidcpower.PowerSourceInUse attribute),
horz_record_length (in module niscope.Session), 315	117
horz_record_ref_position (in module niscope.Session),	INTERVAL (nidmm.SampleTrigger attribute), 183
316	INVALID (nifgen.BusType attribute), 280
horz_sample_rate (in module niscope.Session), 316	InvalidRepeatedCapabilityError, 119, 187, 287, 374, 424,
horz_time_per_record (in module niscope.Session), 316	433, 445
HYSTERESIS (niscope.TriggerType attribute), 371	io_resource_descriptor (in module nidcpower.Session),
	io_resource_descriptor (in module nidmm.Session), 137
IDI E (nifaan HandwanaStata attnihuta) 201	io_resource_descriptor (in module nifgen.Session), 228
IDLE (nifgen.HardwareState attribute), 281 idle_behavior (in module nifgen.Session), 226	io_resource_descriptor (in module niscope.Session), 320
idle_value (in module nifgen.Session), 226	io_resource_descriptor (in module niswitch.Session), 388
IdleBehavior (class in nifgen), 282	is_configuration_channel (in module niswitch.Session),
IMMEDIATE (nidmm.SampleTrigger attribute), 183	389
IMMEDIATE (indmin.sampleTrigger attribute), 185	is_connected() (in module nise.Session), 442
IMMEDIATE (inscope.TriggerType attribute), 370	is_debounced (in module niswitch.Session), 390
IMMEDIATE (niswitch.TriggerInput attribute), 421	is_debounced() (in module nise.Session), 442
IN_PROGRESS (niscope.AcquisitionStatus attribute),	is_done() (in module nifgen.Session), 270
367	is_scanning (in module niswitch.Session), 390
INCOMPLETE (niscope.RISMethod attribute), 369	is_source_channel (in module niswitch.Session), 390
INDUCTANCE (nidmm.Function attribute), 182	is_waiting_for_trig (in module niswitch.Session), 391
input_clock_source (in module niscope.Session), 317	IVIDMM (nidmm.OperationMode attribute), 183
input_impedance (in module niscope.Session), 317	
input_resistance (in module nidmm.Session), 135	J
instrument_firmware_revision (in module nid-	J (nidmm.ThermocoupleType attribute), 185
cpower.Session), 38	JUMP_TO (nifgen.IdleBehavior attribute), 282
instrument_firmware_revision (in module	JUMP_TO (nifgen.WaitBehavior attribute), 286
nidmm.Session), 135	
instrument_firmware_revision (in module nif-	K
gen.Session), 227	K (nidmm.ThermocoupleType attribute), 185
instrument_firmware_revision (in module nis-	KEEP_IN_MEMORY (nid-
cope.Session), 318	cpower.SelfCalibrationPersistence attribute),
instrument_firmware_revision (in module	118
niswitch.Session), 387	I
instrument_manufacturer (in module nidcpower.Session),	L
38	LBR_TRIG0 (nidmm.MeasurementCompleteDest
instrument_manufacturer (in module nidmm.Session),	attribute), 182
136 instrument manufacturar (in modula nifaan Sassian) 227	LBR_TRIG1 (nidmm.SampleTrigger attribute), 184
instrument_manufacturer (in module nifgen.Session), 227	LBR_TRIG1 (nidmm.TriggerSource attribute), 186
	lc calculation model (in module nidmm Session), 137

lc_number_meas_to_average nidmm.Session), 138	(in n	nodule	max_switching_ac_current (in module niswitch.Session), 396
LCCalculationModel (class in nidmm)	182		max_switching_ac_power (in module niswitch.Session).
LEAVING (niscope.TriggerWindowM		371	396
LF_REJECT (niscope.TriggerCoupling			max_switching_dc_current (in module niswitch.Session),
LIMITED_BIN_WIDTH (niscope.RIS			397
369	oniemou am	ibute),	
	romErromt often	:huta)	max_switching_dc_power (in module niswitch.Session),
LINE_NUMBER (niscope.VideoTrigg 373	gerevent atti	ibute),	
	201		max_waveform_size (in module nifgen.Session), 234
LITTLE (nifgen.ByteOrder attribute),			meas_complete_dest (in module nidmm.Session), 139
load_impedance (in module nifgen.Ses LOCAL (nidcpower.Sense attribute), 1			MEASURE (nidcpower.SendSoftwareEdgeTriggerType
			attribute), 118
lock() (in module nidepower.Session),			measure() (in module nidepower.Session), 103
lock() (in module nidmm.Session), 171			measure_buffer_size (in module nidcpower.Session), 40
lock() (in module nifgen.Session), 270			MEASURE_COMPLETE (nidcpower.Event attribute),
lock() (in module niscope.Session), 36			115
lock() (in module niswitch.Session), 4			measure_complete_event_delay (in module nid-
logical_name (in module nidcpower.Se			cpower.Session), 41
logical_name (in module nidmm.Sessi			measure_complete_event_output_terminal (in module
logical_name (in module nifgen.Session			nidcpower.Session), 41
logical_name (in module niscope.Sessi			measure_complete_event_pulse_polarity (in module nid-
logical_name (in module niswitch.Sess	* *		cpower.Session), 42
LOW (nidcpower.OutputCapacitance a		5	measure_complete_event_pulse_width (in module nid-
LOW (nidcpower.Polarity attribute), 11			cpower.Session), 42
LOW (nifgen.DataMarkerEventLevell	Polarity attr	ibute),	measure_multiple() (in module nidcpower.Session), 104
281			measure_record_delta_time (in module nid- cpower.Session), 43
M			measure_record_length (in module nidcpower.Session),
M_PAL (niscope.VideoSignalFormat a	ttribute), 372	2	43
MAIN (nifgen.AnalogPath attribute), 2			measure_record_length_is_finite (in module nid-
major_version (in module nifgen.Sessi			cpower.Session), 44
marker_event_output_terminal (in	module	nif-	measure_trigger_type (in module nidcpower.Session), 45
gen.Session), 230			measure_when (in module nidcpower.Session), 45
marker_events_count (in module nifge	n.Session), 2	230	MEASUREMENT_DEVICE
master_enable (in module niscope.Sess			(niswitch.HandshakingInitiation attribute),
max_ac_voltage (in module niswitch.S			417
max_carry_ac_current (in module nisw			MeasurementCompleteDest (class in nidmm), 182
max_carry_ac_power (in module nisw			MeasurementTypes (class in nidcpower), 116
max_carry_dc_current (in module nisv			MeasureWhen (class in nidcpower), 116
max_carry_dc_power (in module nisw			memory_size (in module nifgen.Session), 234
max_dc_voltage (in module niswitch.S			min_freq_list_duration (in module nifgen.Session), 235
max_freq_list_duration (in module nif			min_freq_list_length (in module nifgen.Session), 236
max_freq_list_length (in module nifge			MIN_NUM_AVERAGES (niscope.RISMethod at-
max_input_frequency (in module nisco			tribute), 369
max_loop_count (in module nifgen.Se		,	min_sample_rate (in module niscope.Session), 323
max_num_freq_lists (in module nifgen		32	min_sequence_length (in module nifgen.Session), 236
max_num_sequences (in module nifge			min_waveform_size (in module nifgen.Session), 236
max_num_waveforms (in module nifge			minor_version (in module nifgen.Session), 235
max_pciexpress_link_width (in	module	ni-	module_revision (in module nifgen.Session), 237
modinst.Session), 431			MULTICONNECT (nise.MulticonnectMode attribute),
max_real_time_sampling_rate (in	module	nis-	444
cope.Session), 322			MulticonnectMode (class in nise), 443
max_ris_rate (in module niscope.Sessi	on), 322		· //
max_sequence_length (in module nifge		234	
=			

N	ONBOARD_REFERENCE_CLOCK (nif-
N (nidmm.ThermocoupleType attribute), 185	gen.ReferenceClockSource attribute), 282
NEGATIVE (niscope. TriggerSlope attribute), 370	ONCE (nidcpower.AutoZero attribute), 115
NEGATIVE (niscope. VideoPolarity attribute), 372	ONCE (nidmm.AutoZero attribute), 180
nidcpower (module), 14	OPEN (nidmm.CableCompensationType attribute), 180
nidmm (module), 123	OPEN (niswitch.RelayAction attribute), 418
nifgen (module), 191	OPEN (niswitch.RelayPosition attribute), 418
nimodinst (module), 428	OPEN_AND_SHORT (nidmm.CableCompensationType
niscope (module), 293	attribute), 180
nise (module), 435 niswitch (module), 379	open_cable_comp_conductance (in module nidmm.Session), 140
NO_ACQUISITION_IN_PROGRESS (nidmm.AcquisitionStatus attribute), 180	open_cable_comp_susceptance (in module nidmm.Session), 140
NO_MULTICONNECT (nise.MulticonnectMode at-	operation_mode (in module nidmm.Session), 141 OperationMode (class in nidmm), 183
tribute), 443	OperationOrder (class in nise), 444
NO_TRIGGER_MOD (niscope.TriggerModifier at-	Option (class in niscope), 368
tribute), 370	output_capacitance (in module nidcpower.Session), 46
NOISE (nifgen.Waveform attribute), 286 NONE (nidcpower.TriggerType attribute), 119	output_clock_source (in module niscope.Session), 324
NONE (nidepower. Higger Type attribute), 119 NONE (nidmm.CableCompensationType attribute), 180	output_connected (in module nidcpower.Session), 46
NONE (nidmm.MeasurementCompleteDest attribute),	output_enabled (in module nidcpower.Session), 47 output_enabled (in module nifgen.Session), 237
NONE (niferan Defendance Clock Source attribute) 282	output_function (in module nidcpower.Session), 48
NONE (nifgen.ReferenceClockSource attribute), 282 NONE (niswitch.ScanAdvancedOutput attribute), 419	output_impedance (in module nifgen.Session), 238
NONE (niswitch.ScanMode attribute), 421	output_mode (in module nifgen.Session), 238
NORMAL (nidcpower.DCNoiseRejection attribute), 115	output_resistance (in module nidcpower.Session), 49
NORMAL (nidepower.TransientResponse attribute), 119	OutputCapacitance (class in nidcpower), 116
NORMAL (nidmm.DCNoiseRejection attribute), 181	OutputFunction (class in nidcpower), 116
NORMAL (niscope.AcquisitionType attribute), 367	OutputMode (class in nifgen), 282
NOW (niscope.FetchRelativeTo attribute), 368	OutputStates (class in nidcpower), 117
NTSC (niscope. Video Signal Format attribute), 372	overranging_enabled (in module nidcpower.Session), 49
num_of_columns (in module niswitch.Session), 398	ovp_enabled (in module nidcpower.Session), 50
num_of_rows (in module niswitch.Session), 399	ovp_limit (in module nidcpower.Session), 50
number_of_averages (in module nidmm.Session), 139	D.
number_of_relays (in module niswitch.Session), 398	P
0	P2P_ENDPOINT_FULLNESS (nifgen.StartTriggerType attribute), 285
OFF (nidcpower.AutoZero attribute), 115	PAL (niscope. Video Signal Format attribute), 372
OFF (nidmm.ADCCalibration attribute), 179	PARALLEL (nidmm.LCCalculationModel attribute), 182
OFF (nidmm.AutoZero attribute), 180	PATH_AVAILABLE (nise.PathCapability attribute), 444
offset_comp_ohms (in module nidmm.Session), 140	PATH_AVAILABLE (niswitch.PathCapability attribute),
ON (nidcpower.AutoZero attribute), 115	418
ON (nidmm.ADCCalibration attribute), 179	PATH_EXISTS (nise.PathCapability attribute), 444
ON (nidmm.AutoZero attribute), 180	PATH_EXISTS (niswitch.PathCapability attribute), 418
ON_DEMAND (nidcpower.MeasureWhen attribute), 116	PATH_NEEDS_CONFIG_CHANNEL
ON_MEASURE_TRIGGER (nidcpower.MeasureWhen	(nise.PathCapability attribute), 444
attribute), 116	PATH_NEEDS_HARDWIRE (nise.PathCapability
ONBOARD_CLOCK (nifgen.SampleClockSource	attribute), 444
attribute), 283	PATH_UNSUPPORTED (nise.PathCapability attribute),
ONBOARD_CLOCK (nif-	444
gen.SampleClockTimebaseSource attribute),	PATH_UNSUPPORTED (niswitch.PathCapability
284	attribute), 418
onboard_memory_size (in module niscope.Session), 323	PathCapability (class in nise), 444 PathCapability (class in niswitch), 418

PATHS (nise.ExpandAction attribute), 443 PAUSED (nidmm.AcquisitionStatus attribute), 180	pulse_bias_voltage_limit (in module nidcpower.Session), 57
PCI (nifgen.BusType attribute), 281 pciexpress_link_width (in module nimodinst.Session),	pulse_bias_voltage_limit_high (in module nid- cpower.Session), 58
431	pulse_bias_voltage_limit_low (in module nid-
PCMCIA (nifgen.BusType attribute), 281	cpower.Session), 59
perform_open_cable_comp() (in module	PULSE_COMPLETE (nidcpower.Event attribute), 116
nidmm.Session), 172	pulse_complete_event_output_terminal (in module nid-
perform_short_cable_comp() (in module	cpower.Session), 60
nidmm.Session), 172	pulse_complete_event_pulse_polarity (in module nid-
PERIOD (nidmm.Function attribute), 181	cpower.Session), 61
pll_lock_status (in module niscope.Session), 324	pulse_complete_event_pulse_width (in module nid-
points_done (in module niscope.Session), 324	cpower.Session), 61
Polarity (class in nidcpower), 117	PULSE_CURRENT (nidcpower.OutputFunction at-
poll_interval (in module niscope.Session), 325	tribute), 117
POSITIVE (niscope.TriggerSlope attribute), 370	pulse_current_level (in module nidcpower.Session), 62
POSITIVE (niscope. Video Polarity attribute), 371	pulse_current_level_range (in module nid-
power_down_latching_relays_after_debounce (in module	cpower.Session), 62
niswitch.Session), 399	pulse_current_limit (in module nidcpower.Session), 63
POWER_LINE_CYCLES (nid-	pulse_current_limit_high (in module nidcpower.Session),
cpower.ApertureTimeUnits attribute), 115	64
POWER_LINE_CYCLES (nidmm.ApertureTimeUnits	pulse_current_limit_low (in module nidcpower.Session),
attribute), 180	65
power_line_frequency (in module nidcpower.Session), 51	pulse_current_limit_range (in module nid-
power_source (in module nidcpower.Session), 52	cpower.Session), 66
power_source_in_use (in module nidcpower.Session), 52	pulse_off_time (in module nidcpower.Session), 67
powerline_freq (in module nidmm.Session), 141	pulse_on_time (in module nidcpower.Session), 67
PowerSource (class in nidcpower), 117	pulse_trigger_type (in module nidcpower.Session), 68
PowerSourceInUse (class in nidcpower), 117	PULSE_VOLTAGE (nidcpower.OutputFunction at-
PRETRIGGER (niscope.FetchRelativeTo attribute), 368	tribute), 117
probe_attenuation (in module niscope.Session), 325	pulse_voltage_level (in module nidcpower.Session), 69
probe_compensation_signal_start() (in module nis-	pulse_voltage_level_range (in module nid-
cope.Session), 360	cpower.Session), 69
probe_compensation_signal_stop() (in module nis-	pulse_voltage_limit (in module nidcpower.Session), 70
cope.Session), 360	pulse_voltage_limit_high (in module nidcpower.Session),
PT3750 (nidmm.RTDType attribute), 183	71
PT3851 (nidmm.RTDType attribute), 183	pulse_voltage_limit_low (in module nidcpower.Session),
PT3911 (nidmm.RTDType attribute), 183	72
PT3916 (nidmm.RTDType attribute), 183	pulse_voltage_limit_range (in module nid-
PT3920 (nidmm.RTDType attribute), 183	cpower.Session), 73
PT3928 (nidmm.RTDType attribute), 183	PXI (nifgen.BusType attribute), 281
PULSE (nidcpower.SendSoftwareEdgeTriggerType at-	PXI_CLOCK (nifgen.ReferenceClockSource attribute),
tribute), 118	282
pulse_bias_current_level (in module nidcpower.Session),	PXI_STAR (nidmm.SampleTrigger attribute), 184
53	PXI_STAR (nidmm.TriggerSource attribute), 186
pulse_bias_current_limit (in module nidcpower.Session), 53	PXI_STAR (niswitch.ScanAdvancedOutput attribute), 419
	PXI_STAR (niswitch.TriggerInput attribute), 422
cpower.Session), 54	PXI_STAR_LINE (nifgen.SampleClockSource attribute),
pulse_bias_current_limit_low (in module nid-	283
cpower.Session), 55	PXI_TRIGO (nidmm.MeasurementCompleteDest at-
pulse_bias_delay (in module nidcpower.Session), 56	tribute), 182
pulse_bias_voltage_level (in module nidcpower.Session),	PXI_TRIG0 (nidmm.SampleTrigger attribute), 183
57	PXI_TRIG0 (nidmm.TriggerSource attribute), 185

PXI_TRIG1 (nidmm.MeasurementCompleteDest tribute), 182	t at-	query_in_compliance() (in module nidcpower.Session),
PXI_TRIG1 (nidmm.SampleTrigger attribute), 183		query_instrument_status (in module nidcpower.Session),
PXI_TRIG1 (nidmm.TriggerSource attribute), 185		query_mstrument_status (in module indepower.session),
PXI_TRIG2 (nidmm.MeasurementCompleteDest	t at-	query_max_current_limit() (in module nid-
tribute), 182	ı aı-	cpower.Session), 105
PXI_TRIG2 (nidmm.SampleTrigger attribute), 183		query_max_voltage_level() (in module nid-
PXI_TRIG2 (nidmm.TriggerSource attribute), 185		cpower.Session), 106
PXI_TRIG2 (nidmin. 111gger source attribute), 183 PXI_TRIG3 (nidmm. Measurement Complete Dest	t at-	query_min_current_limit() (in module nid-
tribute), 182	ı aı-	cpower.Session), 106
PXI_TRIG3 (nidmm.SampleTrigger attribute), 184		query_output_state() (in module nidcpower.Session), 106
PXI_TRIG3 (nidmm.TriggerSource attribute), 186		query_output_state() (in module indepower.session), 100
PXI_TRIG4 (nidmm.MeasurementCompleteDest	t at-	R
tribute), 182	ı aı-	
PXI_TRIG4 (nidmm.SampleTrigger attribute), 184		R (nidmm.ThermocoupleType attribute), 185
PXI_TRIG4 (nidmm.TriggerSource attribute), 186		RAMP_DOWN (nifgen.Waveform attribute), 286
PXI_TRIG5 (nidmm.MeasurementCompleteDest	t at-	RAMP_UP (nifgen.Waveform attribute), 286
tribute), 182	ı aı-	range (in module nidmm.Session), 142
PXI_TRIG5 (nidmm.SampleTrigger attribute), 184		read() (in module nidmm.Session), 173
PXI_TRIG5 (nidmm.TriggerSource attribute), 186		read() (in module niscope.Session), 361
PXI_TRIG6 (nidmm.MeasurementCompleteDest	t at-	read_current_temperature() (in module nid-
tribute), 182	ı aı-	cpower.Session), 107
PXI_TRIG6 (nidmm.SampleTrigger attribute), 184		read_current_temperature() (in module nifgen.Session),
PXI_TRIG6 (nidmm.TriggerSource attribute), 186		273
PXI_TRIGO (indimin. Higgersource attribute), 160 PXI_TRIG7 (nidmm. Measurement Complete Dest	t at-	read_multi_point() (in module nidmm.Session), 173
tribute), 182	ı aı-	READ_POINTER (niscope.FetchRelativeTo attribute),
PXI_TRIG7 (nidmm.SampleTrigger attribute), 184		368
PXI_TRIG7 (nidmm.TriggerSource attribute), 186		read_status() (in module nidmm.Session), 174
PXI_TRIGGER_LINE_0RTSI_0	(nif-	read_waveform() (in module nidmm.Session), 175
gen.SampleClockSource attribute), 283	(1111	ready_for_advance_event_output_terminal (in module
PXI_TRIGGER_LINE_1RTSI_1	(nif-	niscope.Session), 326
gen.SampleClockSource attribute), 283	(1111	READY_FOR_PULSE_TRIGGER (nidcpower.Event at-
PXI_TRIGGER_LINE_2RTSI_2	(nif-	tribute), 116
gen.SampleClockSource attribute), 283	(1111	ready_for_pulse_trigger_event_output_terminal (in mod-
PXI TRIGGER LINE 3RTSI 3	(nif-	ule nidepower.Session), 74
gen.SampleClockSource attribute), 283	(1111	ready_for_pulse_trigger_event_pulse_polarity (in module nidcpower.Session), 75
PXI_TRIGGER_LINE_4RTSI_4	(nif-	ready_for_pulse_trigger_event_pulse_width (in module
gen.SampleClockSource attribute), 283	(	nidcpower.Session), 75
PXI_TRIGGER_LINE_5RTSI_5	(nif-	ready_for_ref_event_output_terminal (in module nis-
gen.SampleClockSource attribute), 283	(	cope.Session), 326
PXI_TRIGGER_LINE_6RTSI_6	(nif-	ready_for_start_event_output_terminal (in module nif-
gen.SampleClockSource attribute), 283		gen.Session), 239
PXI_TRIGGER_LINE_7RTSI_7	(nif-	ready_for_start_event_output_terminal (in module nis-
gen.SampleClockSource attribute), 283		cope.Session), 327
PXIE (nifgen.BusType attribute), 281		REARCONNECTOR (niswitch.ScanAdvancedOutput at-
		tribute), 419
Q		REARCONNECTOR (niswitch.TriggerInput attribute),
query_arb_seq_capabilities() (in module nifgen.Ses	(roise	422
271	,51011),	REARCONNECTOR_MODULE1
query_arb_wfm_capabilities() (in module	nif-	(niswitch.ScanAdvancedOutput attribute),
gen.Session), 271	1111	419
query_freq_list_capabilities() (in module nifgen.Ses	ssion).	REARCONNECTOR_MODULE1
272	/1	(niswitch.TriggerInput attribute), 422

REARCONNECTOR_MODULE10		(niswitch.TriggerInput attribute), 422
(niswitch.ScanAdvancedOutput	attribute),	record_arm_source (in module niscope.Session), 328
420		records_done (in module niscope.Session), 327
REARCONNECTOR_MODULE10		ref_clk_rate (in module niscope.Session), 328
(niswitch.TriggerInput attribute), 422		ref_clock_frequency (in module nifgen.Session), 240
REARCONNECTOR_MODULE11		ref_trig_tdc_enable (in module niscope.Session), 329
(niswitch.ScanAdvancedOutput	attribute),	ref_trigger_detector_location (in module nis-
420		cope.Session), 328
REARCONNECTOR_MODULE11		ref_trigger_minimum_quiet_time (in module nis-
(niswitch.TriggerInput attribute), 422	2	cope.Session), 329
REARCONNECTOR_MODULE12		REFERENCE (niscope.WhichTrigger attribute), 373
(niswitch.ScanAdvancedOutput	attribute),	reference_clock_source (in module nifgen.Session), 239
420		ReferenceClockSource (class in nifgen), 282
REARCONNECTOR_MODULE12	1	RefTriggerDetectorLocation (class in niscope), 369
(niswitch.TriggerInput attribute), 422 REARCONNECTOR_MODULE2		Relative To (class in nifgen), 283
(niswitch.ScanAdvancedOutput	attribute),	relay_control() (in module niswitch.Session), 412 RelayAction (class in niswitch), 418
419	attiribute),	RelayPosition (class in niswitch), 418
REARCONNECTOR_MODULE2		REMOTE (nidcpower.Sense attribute), 118
(niswitch.TriggerInput attribute), 422	)	reset() (in module nidcpower.Session), 107
REARCONNECTOR_MODULE3	•	reset() (in module nidmm.Session), 107
(niswitch.ScanAdvancedOutput	attribute),	reset() (in module nifgen.Session), 273
419	utti10 utt),	reset() (in module niscope.Session), 362
REARCONNECTOR_MODULE3		reset() (in module niswitch.Session), 413
(niswitch.TriggerInput attribute), 422	2	reset_average_before_measurement (in module nid-
REARCONNECTOR_MODULE4		cpower.Session), 76
(niswitch.ScanAdvancedOutput	attribute),	reset_device() (in module nidcpower.Session), 107
419		reset_device() (in module nifgen.Session), 273
REARCONNECTOR_MODULE4		reset_device() (in module niscope.Session), 362
(niswitch.TriggerInput attribute), 422	2	reset_with_defaults() (in module nidcpower.Session), 108
REARCONNECTOR_MODULE5		reset_with_defaults() (in module nidmm.Session), 176
(niswitch.ScanAdvancedOutput	attribute),	reset_with_defaults() (in module nifgen.Session), 273
419		reset_with_defaults() (in module niscope.Session), 362
REARCONNECTOR_MODULE5		reset_with_defaults() (in module niswitch.Session), 413
(niswitch.TriggerInput attribute), 422		resolution (in module niscope.Session), 330
REARCONNECTOR_MODULE6	. 44	resolution_absolute (in module nidmm.Session), 143
(niswitch.ScanAdvancedOutput	attribute),	resolution_digits (in module nidmm.Session), 143
419 REARCONNECTOR_MODULE6		RESOURCE_IN_USE (nise.PathCapability attribute), 444
(niswitch.TriggerInput attribute), 422	)	RESOURCE_IN_USE (niswitch.PathCapability at-
REARCONNECTOR_MODULE7	•	tribute), 418
(niswitch.ScanAdvancedOutput	attribute),	RESTORE_EXTERNAL_CALIBRATION (nis-
419	utilioute),	cope.Option attribute), 369
REARCONNECTOR_MODULE7		ris_in_auto_setup_enable (in module niscope.Session).
(niswitch.TriggerInput attribute), 422	)	330
REARCONNECTOR_MODULE8		ris_method (in module niscope.Session), 331
(niswitch.ScanAdvancedOutput	attribute),	ris_num_averages (in module niscope.Session), 331
420		RISING (nifgen.ScriptTriggerDigitalEdgeEdge attribute).
REARCONNECTOR_MODULE8		284
(niswitch.TriggerInput attribute), 422	2	RISING (nifgen.StartTriggerDigitalEdgeEdge attribute).
REARCONNECTOR_MODULE9		284
(niswitch.ScanAdvancedOutput	attribute),	RISING (niswitch.ScanAdvancedPolarity attribute), 420
420		RISING (niswitch.TriggerInputPolarity attribute), 423
REARCONNECTOR_MODULE9		RISMethod (class in niscope), 369

route_scan_advanced_output() (in module niswitch.Session), 413	self_cal() (in module nidmm.Session), 176 self_cal() (in module nifgen.Session), 274
route_trigger_input() (in module niswitch.Session), 414	self_cal() (in module niscope.Session), 363
ROUTES (nise.ExpandAction attribute), 443	SELF_CALIBRATE_ALL_CHANNELS (nis-
RTDType (class in nidmm), 183	cope.Option attribute), 368
RTSI_7 (nifgen.ReferenceClockSource attribute), 283	self_calibration_persistence (in module nid-
RUNNING (nidmm.AcquisitionStatus attribute), 179	cpower.Session), 77
RUNNING (nifgen.HardwareState attribute), 282	self_test() (in module nidcpower.Session), 108
	self_test() (in module nidmm.Session), 176
S	self_test() (in module nifgen.Session), 274
S (nidmm.ThermocoupleType attribute), 185	self_test() (in module niscope.Session), 363
samp_clk_timebase_div (in module niscope.Session), 332	self_test() (in module niswitch.Session), 414
samp_clk_timebase_rate (in module niscope.Session),	SelfCalibrationPersistence (class in nidcpower), 118
333	SelfTestError, 120, 187, 287, 374, 424, 434, 445
samp_clk_timebase_src (in module niscope.Session), 333	send_software_edge_trigger() (in module nid-
sample_clock_source (in module nifgen.Session), 240	cpower.Session), 109
sample_clock_timebase_multiplier (in module nis-	send_software_edge_trigger() (in module nif-
cope.Session), 331	gen.Session), 274
sample_clock_timebase_rate (in module nifgen.Session),	send_software_trigger() (in module nidmm.Session), 176
241	send_software_trigger() (in module niswitch.Session),
sample_clock_timebase_source (in module nif-	414
gen.Session), 241	send_software_trigger_edge() (in module nis-
· · · · · · · · · · · · · · · · · · ·	cope.Session), 364
sample_count (in module nidmm.Session), 144 sample_interval (in module nidmm.Session), 144	SendSoftwareEdgeTriggerType (class in nidcpower), 118
· · · · · · · · · · · · · · · · · · ·	Sense (class in nidcpower), 118
sample_mode (in module niscope.Session), 332	sense (in module nidcpower.Session), 78
sample_trigger (in module nidmm.Session), 144	SEQ (nifgen.OutputMode attribute), 282
SampleClockSource (class in nifgen), 283	SEQUENCE (nidcpower.SourceMode attribute), 118
SampleClockTimebaseSource (class in nifgen), 284	SEQUENCE_ADVANCE (nid-
samples_to_average (in module nidcpower.Session), 76	cpower.SendSoftwareEdgeTriggerType at-
SampleTrigger (class in nidmm), 183	tribute), 118
scan_advanced_output (in module niswitch.Session), 400	sequence_advance_trigger_type (in module nid-
scan_advanced_polarity (in module niswitch.Session),	cpower.Session), 78
400	SEQUENCE_ENGINE_DONE (nidcpower.Event at-
scan_delay (in module niswitch.Session), 400	tribute), 116
scan_list (in module niswitch.Session), 401	sequence_engine_done_event_output_terminal (in mod-
scan_mode (in module niswitch.Session), 401	ule nidcpower.Session), 79
ScanAdvancedOutput (class in niswitch), 419	sequence_engine_done_event_pulse_polarity (in module
ScanAdvancedPolarity (class in niswitch), 420	nidcpower.Session), 79
ScanMode (class in niswitch), 421	sequence_engine_done_event_pulse_width (in module
SCRIPT (nifgen.OutputMode attribute), 282	nidcpower.Session), 80
script_to_generate (in module nifgen.Session), 242	SEQUENCE_ITERATION_COMPLETE (nid-
script_trigger_type (in module nifgen.Session), 243	cpower.Event attribute), 116
script_triggers_count (in module nifgen.Session), 242	sequence_iteration_complete_event_output_terminal (in
ScriptTriggerDigitalEdgeEdge (class in nifgen), 284	module nidepower.Session), 80
ScriptTriggerType (class in nifgen), 284	sequence_iteration_complete_event_pulse_polarity (in
SECAM (niscope. Video Signal Format attribute), 372	module nidepower.Session), 81
SECOND_ORDER (nidcpower.DCNoiseRejection at-	sequence_iteration_complete_event_pulse_width (in
tribute), 115	module nidepower.Session), 81
SECOND_ORDER (nidmm.DCNoiseRejection at-	sequence_loop_count (in module nidcpower.Session), 82
tribute), 181	sequence_loop_count_is_finite (in module nid-
SECONDS (nidcpower.ApertureTimeUnits attribute),	cpower.Session), 83
115	serial_number (in module nidmm.Session), 145
SECONDS (nidmm.ApertureTimeUnits attribute), 180	serial_number (in module nifgen.Session), 243
self_cal() (in module nidcpower.Session), 108	seriar_number (in module inigen.session), 243

serial_number (in module nimodinst.Session), 431 serial_number (in module niscope.Session), 334	SOURCE (nidcpower.SendSoftwareEdgeTriggerType attribute), 118
serial_number (in module niswitch.Session), 402	SOURCE_COMPLETE (nidcpower.Event attribute), 115
SERIES (nidmm.LCCalculationModel attribute), 182	source_complete_event_output_terminal (in module nid-
Session (class in nidepower), 14	cpower.Session), 83
Session (class in nidmm), 123	source_complete_event_pulse_polarity (in module nid-
Session (class in nifgen), 191	cpower.Session), 84
Session (class in nimodinst), 428	source_complete_event_pulse_width (in module nid-
Session (class in niscope), 293	cpower.Session), 84
Session (class in nise), 435	SOURCE_CONFLICT (niswitch.PathCapability at-
Session (class in niswitch), 379	tribute), 418
set_next_write_position() (in module nifgen.Session),	source_delay (in module nidcpower.Session), 85
275	source_mode (in module nidepower.Session), 86
set_path() (in module niswitch.Session), 415	source_trigger_type (in module nidcpower.Session), 86
set_sequence() (in module nidcpower.Session), 419	SourceMode (class in nidcpower), 118
settle_time (in module nidmm.Session), 145	specific_driver_description (in module nid-
settling_time (in module niswitch.Session), 402	cpower.Session), 87
SHORT (nidmm.CableCompensationType attribute), 180	specific_driver_description (in module nidmm.Session),
short_cable_comp_reactance (in module	147
nidmm.Session), 146	specific_driver_description (in module nifgen.Session),
short_cable_comp_resistance (in module	244
nidmm.Session), 146	specific_driver_description (in module niscope.Session),
simulate (in module nidcpower.Session), 83	335
simulate (in module nidmm.Session), 147	specific_driver_description (in module niswitch.Session),
simulate (in module nifgen.Session), 244	404
simulate (in module niscope.Session), 334	specific_driver_major_version (in module
simulate (in module niswitch.Session), 403	nidmm.Session), 148
SINE (nifgen.Waveform attribute), 286	specific_driver_minor_version (in module
SINGLE (nifgen.TriggerMode attribute), 285	nidmm.Session), 148
SINGLE_ENDED (nifgen.TerminalConfiguration at-	specific_driver_prefix (in module nidcpower.Session), 87
tribute), 285	specific_driver_revision (in module nidcpower.Session),
SINGLE_ENDED (niscope.TerminalConfiguration at-	87
tribute), 369	specific_driver_revision (in module nidmm.Session), 149
SINGLE_POINT (nidcpower.SourceMode attribute), 118	specific_driver_revision (in module nifgen.Session), 244
SIXTEEN_TAP_HANNING (nis-	specific_driver_revision (in module niscope.Session), 335
cope.FlexFIRAntialiasFilterType attribute),	specific_driver_revision (in module niswitch.Session),
368	404
slot_number (in module nimodinst.Session), 432	specific_driver_vendor (in module nidcpower.Session),
SLOW (nidcpower.TransientResponse attribute), 119	88
socket_number (in module nimodinst.Session), 432	specific_driver_vendor (in module nidmm.Session), 149
SOFTWARE (niscope.TriggerType attribute), 371	specific_driver_vendor (in module nifgen.Session), 245
SOFTWARE_EDGE (nidcpower.TriggerType attribute),	specific_driver_vendor (in module niscope.Session), 335
119	specific_driver_vendor (in module niswitch.Session), 404
SOFTWARE_EDGE (nifgen.ScriptTriggerType at-	SQUARE (nifgen.Waveform attribute), 286
tribute), 284	START (nidcpower.SendSoftwareEdgeTriggerType at-
SOFTWARE_EDGE (nifgen.StartTriggerType attribute),	tribute), 118
285	START (nifgen.RelativeTo attribute), 283
SOFTWARE_TRIG (nidmm.SampleTrigger attribute),	START (niscope.FetchRelativeTo attribute), 368
183	START (niscope.WhichTrigger attribute), 373
SOFTWARE_TRIG (nidmm.TriggerSource attribute),	start_to_ref_trigger_holdoff (in module niscope.Session),
185	336
SOFTWARE_TRIG (niswitch.TriggerInput attribute),	start_trigger_type (in module nidcpower.Session), 88
421	start trigger type (in module nifgen. Session), 246

THERMISTOR_44007 (nidmm.ThermistorType at-
tribute), 184
ThermistorType (class in nidmm), 184 THERMOCOUPLE (nidmm.TransducerType attribute),
185
ThermocoupleReferenceJunctionType (class in nidmm),
184
ThermocoupleType (class in nidmm), 184
TransducerType (class in nidmm), 185
transient_response (in module nidcpower.Session), 89
TransientResponse (class in nidcpower), 119
TRIANGLE (nifgen. Waveform attribute), 286
TRIG_NONE (nifgen.ScriptTriggerType attribute), 284
TRIG_NONE (nifgen.StartTriggerType attribute), 284
TRIGGER (niscope.FetchRelativeTo attribute), 368
trigger_auto_triggered (in module niscope.Session), 337
trigger_count (in module nidmm.Session), 156
trigger_coupling (in module niscope.Session), 337
trigger_delay (in module nidmm.Session), 156
trigger_delay_time (in module niscope.Session), 338
trigger_holdoff (in module niscope.Session), 338
trigger_hysteresis (in module niscope.Session), 338
trigger_impedance (in module niscope.Session), 339
trigger_input (in module niswitch.Session), 406
trigger_input_polarity (in module niswitch.Session), 406
trigger_level (in module niscope.Session), 339 trigger_mode (in module nifgen.Session), 249
trigger_modifier (in module niscope.Session), 340
trigger_slope (in module niscope.Session), 340
trigger_source (in module nidmm.Session), 157
trigger_source (in module niscope.Session), 341
trigger_type (in module niscope.Session), 341
trigger_window_high_level (in module niscope.Session),
341
trigger_window_low_level (in module niscope.Session),
342
trigger_window_mode (in module niscope.Session), 343
TriggerCoupling (class in niscope), 370
TriggerInput (class in niswitch), 421
TriggerInputPolarity (class in niswitch), 423
TriggerMode (class in nifgen), 285
TriggerModifier (class in niscope), 370
TriggerSlope (class in niscope), 370
TriggerSource (class in nidmm), 185
TriggerType (class in nidcpower), 119 TriggerType (class in niscope), 370
TriggerWindowMode (class in niscope), 370  TriggerWindowMode (class in niscope), 371
TTL0 (niswitch.ScanAdvancedOutput attribute), 419
TTL0 (niswitch.TriggerInput attribute), 421
TTL1 (niswitch.ScanAdvancedOutput attribute), 419
TTL1 (niswitch.TriggerInput attribute), 421
TTL2 (niswitch.ScanAdvancedOutput attribute), 419
TTL2 (niswitch.TriggerInput attribute), 421
TTL3 (niswitch.ScanAdvancedOutput attribute), 419

TTT 2 ( ' ' ' 1 TT '		AMBEO SECOND CO ED ANGER DED RECOND CO	
TTL3 (niswitch.TriggerInput attribute), 421		VIDEO_576P_50_FRAMES_PER_SECOND (nis	s-
TTL4 (niswitch.ScanAdvancedOutput attribute), 419	)	cope.VideoSignalFormat attribute), 372	
TTL4 (niswitch.TriggerInput attribute), 421		VIDEO_720P_50_FRAMES_PER_SECOND (nis	S-
TTL5 (niswitch.ScanAdvancedOutput attribute), 419	)	cope. Video Signal Format attribute), 372	
TTL5 (niswitch.TriggerInput attribute), 421		VIDEO_720P_59_94_FRAMES_PER_SECOND (nis	s-
TTL6 (niswitch.ScanAdvancedOutput attribute), 419	)	cope. Video Signal Format attribute), 372	
TTL6 (niswitch.TriggerInput attribute), 421		VIDEO_720P_60_FRAMES_PER_SECOND (nis	s-
TTL7 (niswitch.ScanAdvancedOutput attribute), 419	)	cope.VideoSignalFormat attribute), 372	
TTL7 (niswitch.TriggerInput attribute), 422		VideoPolarity (class in niscope), 371	
TV (niscope.TriggerType attribute), 370		VideoSignalFormat (class in niscope), 372	
tv_trigger_event (in module niscope.Session), 343		VideoTriggerEvent (class in niscope), 372	
	244		6
tv_trigger_line_number (in module niscope.Session)		VOLTAGE (nidcpower.MeasurementTypes attribute), 11	U
tv_trigger_polarity (in module niscope.Session), 344		VOLTAGE (nidcpower.OutputStates attribute), 117	1
tv_trigger_signal_format (in module niscope.Sess	sion),	voltage_compensation_frequency (in module nic	1-
344		cpower.Session), 90	
TWO_WIRE_RES (nidmm.Function attribute), 181		voltage_gain_bandwidth (in module nidcpower.Session	),
TWO_WIRE_RTD (nidmm.TransducerType attrib	oute),	91	
185		voltage_level (in module nidcpower.Session), 91	
		voltage_level_autorange (in module nidcpower.Session	),
U		92	
UNBALANCED_DIFFERENTIAL	(nis-	voltage_level_range (in module nidcpower.Session), 93	
cope.TerminalConfiguration attribute), 369	`	voltage_limit (in module nidcpower.Session), 93	
unlock() (in module nidcpower.Session), 110		voltage_limit_autorange (in module nidcpower.Session	)
• •		94	,
unlock() (in module nidmm.Session), 177		- ·	
unlock() (in module nifgen.Session), 275		voltage_limit_high (in module nidepower.Session), 95	
unlock() (in module niscope.Session), 364		voltage_limit_low (in module nidcpower.Session), 96	
unlock() (in module niswitch.Session), 415		voltage_limit_range (in module nidcpower.Session), 97	,
UnsupportedConfigurationError, 119, 186, 287, 373,	423,	voltage_pole_zero_ratio (in module nidcpower.Session	),
433, 445		98	
USER (nifgen.Waveform attribute), 286		VXI (nifgen.BusType attribute), 281	
V		W	
vertical_coupling (in module niscope.Session), 345		wait_behavior (in module nifgen.Session), 249	
vertical_offset (in module niscope.Session), 346		wait_for_debounce() (in module nise.Session), 442	
vertical_range (in module niscope.Session), 346			
VerticalCoupling (class in niscope), 371		wait_for_debounce() (in module niswitch.Session), 415	
VIDEO_1080I_50_FIELDS_PER_SECOND		wait_for_event() (in module nidcpower.Session), 110	`
	(nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session	),
cope.VideoSignalFormat attribute), 372	`	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415	),
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND	(nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275	),
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372	(nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250	),
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND	`	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286	),
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372	(nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250	
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372	(nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286	
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND	(nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nit	
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080P_24_FRAMES_PER_SECOND cope.VideoSignalFormat attribute), 372	(nis- (nis- (nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nifgen.HardwareState attribute), 282 Waveform (class in nifgen), 286	
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080P_24_FRAMES_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480I_59_94_FIELDS_PER_SECOND	(nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nifgen.HardwareState attribute), 282 Waveform (class in nifgen), 286 WAVEFORM (nidmm.OperationMode attribute), 183	
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080P_24_FRAMES_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372	(nis- (nis- (nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nit gen.HardwareState attribute), 282 Waveform (class in nifgen), 286 WAVEFORM (nidmm.OperationMode attribute), 183 waveform_coupling (in module nidmm.Session), 157	f-
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080P_24_FRAMES_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480I_60_FIELDS_PER_SECOND	(nis- (nis- (nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nifter gen.HardwareState attribute), 282 Waveform (class in nifgen), 286 WAVEFORM (nidmm.OperationMode attribute), 183 waveform_coupling (in module nidmm.Session), 157 WAVEFORM_CURRENT (nidmm.Function attribute)	f-
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080P_24_FRAMES_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372	(nis- (nis- (nis- (nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nifgen.HardwareState attribute), 282 Waveform (class in nifgen), 286 WAVEFORM (nidmm.OperationMode attribute), 183 waveform_coupling (in module nidmm.Session), 157 WAVEFORM_CURRENT (nidmm.Function attribute) 181	f-
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080P_24_FRAMES_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480P_59_94_FRAMES_PER_SECOND	(nis- (nis- (nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nifter gen.HardwareState attribute), 282 Waveform (class in nifgen), 286 WAVEFORM (nidmm.OperationMode attribute), 183 waveform_coupling (in module nidmm.Session), 157 WAVEFORM_CURRENT (nidmm.Function attribute 181 waveform_points (in module nidmm.Session), 158	f-
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_1080P_24_FRAMES_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480I_59_94_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480I_60_FIELDS_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480P_59_94_FRAMES_PER_SECOND cope.VideoSignalFormat attribute), 372 VIDEO_480P_59_94_FRAMES_PER_SECOND cope.VideoSignalFormat attribute), 372	(nis- (nis- (nis- (nis- (nis- (nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nifter gen.HardwareState attribute), 282 Waveform (class in nifgen), 286 WAVEFORM (nidmm.OperationMode attribute), 183 waveform_coupling (in module nidmm.Session), 157 WAVEFORM_CURRENT (nidmm.Function attribute) 181 waveform_points (in module nidmm.Session), 158 waveform_quantum (in module nifgen.Session), 250	f-
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND	(nis- (nis- (nis- (nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nit gen.HardwareState attribute), 282 Waveform (class in nifgen), 286 WAVEFORM (nidmm.OperationMode attribute), 183 waveform_coupling (in module nidmm.Session), 157 WAVEFORM_CURRENT (nidmm.Function attribute 181 waveform_points (in module nidmm.Session), 158 waveform_quantum (in module nifgen.Session), 250 waveform_rate (in module nidmm.Session), 158	f-
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND	(nis- (nis- (nis- (nis- (nis- (nis- (nis- (nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nift gen.HardwareState attribute), 282 Waveform (class in nifgen), 286 WAVEFORM (nidmm.OperationMode attribute), 183 waveform_coupling (in module nidmm.Session), 157 WAVEFORM_CURRENT (nidmm.Function attribute 181 waveform_points (in module nidmm.Session), 158 waveform_quantum (in module nidmm.Session), 250 waveform_rate (in module nidmm.Session), 158 WAVEFORM_VOLTAGE (nidmm.Function attribute	f-
cope.VideoSignalFormat attribute), 372 VIDEO_1080I_59_94_FIELDS_PER_SECOND	(nis- (nis- (nis- (nis- (nis- (nis-	wait_for_event() (in module nidcpower.Session), 110 wait_for_scan_complete() (in module niswitch.Session 415 wait_until_done() (in module nifgen.Session), 275 wait_value (in module nifgen.Session), 250 WaitBehavior (class in nifgen), 286 WAITING_FOR_START_TRIGGER (nit gen.HardwareState attribute), 282 Waveform (class in nifgen), 286 WAVEFORM (nidmm.OperationMode attribute), 183 waveform_coupling (in module nidmm.Session), 157 WAVEFORM_CURRENT (nidmm.Function attribute 181 waveform_points (in module nidmm.Session), 158 waveform_quantum (in module nifgen.Session), 250 waveform_rate (in module nidmm.Session), 158	f-

```
WhichTrigger (class in niscope), 373
WINDOW (niscope.TriggerType attribute), 371
wire_mode (in module niswitch.Session), 406
write_script() (in module nifgen.Session), 276
WRITE_TO_EEPROM (nid-
cpower.SelfCalibrationPersistence attribute),
118
write_waveform() (in module nifgen.Session), 276
```